

Data Path : Z:\svoasrv\HPCHEM1\BNA_G\Data\BG030822\
 Data File : BG052602.D
 Acq On : 8 Mar 2022 12:41
 Operator : CG/JU
 Sample : SSTDCCC040
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_G
 LabSampleId :
 SSTDCCC040

Quant Time: Mar 09 04:17:45 2022
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_G\Methods\8270-BG030722.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Tue Mar 08 01:09:10 2022
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	20.000	20.000	0.0	108	0.00
2	1,4-Dioxane	40.000	39.839	0.4	105	0.00
3	Pyridine	40.000	41.790	-4.5	106	0.00
4	n-Nitrosodimethylamine	40.000	42.169	-5.4	115	0.00
5 S	2-Fluorophenol	80.000	80.996	-1.2	108	0.00
6	Aniline	40.000	42.017	-5.0	110	0.00
7 S	Phenol-d6	80.000	85.003	-6.3	113	0.00
8	2-Chlorophenol	40.000	40.400	-1.0	107	0.00
9	Benzaldehyde	40.000	36.915	7.7	107	0.00
10 C	Phenol	40.000	41.360	-3.4	109	0.00
11	bis(2-Chloroethyl)ether	40.000	40.117	-0.3	107	0.00
12	1,3-Dichlorobenzene	40.000	40.005	-0.0	108	0.00
13 C	1,4-Dichlorobenzene	40.000	40.765	-1.9	108	0.00
14	1,2-Dichlorobenzene	40.000	40.048	-0.1	107	0.00
15	Benzyl Alcohol	40.000	42.697	-6.7	112	0.00
16	2,2'-oxybis(1-Chloropropane	40.000	41.042	-2.6	110	0.00
17	2-Methylphenol	40.000	40.725	-1.8	110	0.00
18	Hexachloroethane	40.000	39.747	0.6	109	0.00
19 P	n-Nitroso-di-n-propylamine	40.000	41.967	-4.9	112	0.00
20	3+4-Methylphenols	40.000	41.668	-4.2	111	0.00
21 I	Naphthalene-d8	20.000	20.000	0.0	110	0.00
22	Acetophenone	40.000	40.414	-1.0	112	0.00
23 S	Nitrobenzene-d5	80.000	82.672	-3.3	112	0.00
24	Nitrobenzene	40.000	40.776	-1.9	111	0.00
25	Isophorone	40.000	40.535	-1.3	111	0.00
26 C	2-Nitrophenol	40.000	41.340	-3.4	107	0.00
27	2,4-Dimethylphenol	40.000	42.084	-5.2	114	0.00
28	bis(2-Chloroethoxy)methane	40.000	39.350	1.6	110	0.00
29 C	2,4-Dichlorophenol	40.000	41.512	-3.8	108	0.00
30	1,2,4-Trichlorobenzene	40.000	40.708	-1.8	112	0.00
31	Naphthalene	40.000	40.382	-1.0	111	0.00
32	Benzoic acid	40.000	39.291	1.8	111	0.00
33	4-Chloroaniline	40.000	40.888	-2.2	109	0.00
34 C	Hexachlorobutadiene	40.000	39.850	0.4	111	0.00
35	Caprolactam	40.000	39.991	0.0	102	0.00
36 C	4-Chloro-3-methylphenol	40.000	40.814	-2.0	110	0.00
37	2-Methylnaphthalene	40.000	39.902	0.2	109	0.00
38	1-Methylnaphthalene	40.000	39.768	0.6	110	0.00
39 I	Acenaphthene-d10	20.000	20.000	0.0	108	0.00
40	1,2,4,5-Tetrachlorobenzene	40.000	40.474	-1.2	111	0.00
41 P	Hexachlorocyclopentadiene	40.000	40.543	-1.4	115	0.00
42 S	2,4,6-Tribromophenol	80.000	81.562	-2.0	105	0.00
43 C	2,4,6-Trichlorophenol	40.000	42.604	-6.5	111	0.00
44	2,4,5-Trichlorophenol	40.000	41.706	-4.3	109	0.00
45 S	2-Fluorobiphenyl	80.000	80.691	-0.9	110	0.00
46	1,1'-Biphenyl	40.000	40.790	-2.0	110	0.00
47	2-Chloronaphthalene	40.000	40.470	-1.2	109	0.00

Data Path : Z:\svoasrv\HPCHEM1\BNA_G\Data\BG030822\
 Data File : BG052602.D
 Acq On : 8 Mar 2022 12:41
 Operator : CG/JU
 Sample : SSTDCCC040
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_G
 LabSampleId :
 SSTDCCC040

Quant Time: Mar 09 04:17:45 2022
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_G\Methods\8270-BG030722.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Tue Mar 08 01:09:10 2022
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
48	2-Nitroaniline	40.000	41.588	-4.0	108	0.00
49	Acenaphthylene	40.000	40.450	-1.1	109	0.00
50	Dimethylphthalate	40.000	39.681	0.8	107	0.00
51	2,6-Dinitrotoluene	40.000	40.235	-0.6	109	0.00
52 C	Acenaphthene	40.000	40.416	-1.0	108	0.00
53	3-Nitroaniline	40.000	38.700	3.2	99	0.00
54 P	2,4-Dinitrophenol	40.000	39.182	2.0	107	0.00
55	Dibenzofuran	40.000	39.421	1.4	107	0.00
56 P	4-Nitrophenol	40.000	40.851	-2.1	100	0.00
57	2,4-Dinitrotoluene	40.000	40.378	-0.9	106	0.00
58	Fluorene	40.000	40.310	-0.8	108	0.00
59	2,3,4,6-Tetrachlorophenol	40.000	39.951	0.1	102	0.00
60	Diethylphthalate	40.000	39.012	2.5	104	0.00
61	4-Chlorophenyl-phenylether	40.000	40.538	-1.3	111	0.00
62	4-Nitroaniline	40.000	39.801	0.5	102	0.00
63	Azobenzene	40.000	39.671	0.8	108	0.00
64 I	Phenanthrene-d10	20.000	20.000	0.0	104	0.00
65	4,6-Dinitro-2-methylphenol	40.000	43.236	-8.1	105	0.00
66 c	n-Nitrosodiphenylamine	40.000	40.984	-2.5	106	0.00
67	4-Bromophenyl-phenylether	40.000	40.709	-1.8	106	0.00
68	Hexachlorobenzene	40.000	40.518	-1.3	105	0.00
69	Atrazine	40.000	41.710	-4.3	104	0.00
70 C	Pentachlorophenol	40.000	40.570	-1.4	105	0.00
71	Phenanthrene	40.000	40.795	-2.0	106	0.00
72	Anthracene	40.000	40.247	-0.6	104	0.00
73	Carbazole	40.000	40.260	-0.6	103	0.00
74	Di-n-butylphthalate	40.000	40.108	-0.3	102	0.00
75 C	Fluoranthene	40.000	39.958	0.1	103	0.00
76 I	Chrysene-d12	20.000	20.000	0.0	99	0.00
77	Benzydine	40.000	43.127	-7.8	96	0.00
78	Pyrene	40.000	41.743	-4.4	103	0.00
79 S	Terphenyl-d14	80.000	82.334	-2.9	102	0.00
80	Butylbenzylphthalate	40.000	40.791	-2.0	99	0.00
81	Benzo(a)anthracene	40.000	40.228	-0.6	99	0.00
82	3,3'-Dichlorobenzidine	40.000	40.833	-2.1	98	0.00
83	Chrysene	40.000	40.196	-0.5	100	0.00
84	Bis(2-ethylhexyl)phthalate	40.000	40.433	-1.1	97	0.00
85 c	Di-n-octyl phthalate	40.000	40.769	-1.9	99	0.00
86 I	Perylene-d12	20.000	20.000	0.0	100	0.00
87	Indeno(1,2,3-cd)pyrene	40.000	40.525	-1.3	100	0.00
88	Benzo(b)fluoranthene	40.000	40.941	-2.4	101	0.00
89	Benzo(k)fluoranthene	40.000	39.396	1.5	97	0.00
90 C	Benzo(a)pyrene	40.000	40.597	-1.5	99	0.00
91	Dibenzo(a,h)anthracene	40.000	40.897	-2.2	99	0.00
92	Benzo(g,h,i)perylene	40.000	40.179	-0.4	98	0.00

Data Path : Z:\svoasrv\HPCHEM1\BNA_G\Data\BG030822\
Data File : BG052602.D
Acq On : 8 Mar 2022 12:41
Operator : CG/JU
Sample : SSTDCCC040
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_G
LabSampleId :
SSTDCCC040

Quant Time: Mar 09 04:17:45 2022
Quant Method : Z:\svoasrv\HPCHEM1\BNA_G\Methods\8270-BG030722.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Tue Mar 08 01:09:10 2022
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

Compound	Amount	Calc.	%Dev	Area%	Dev(min)
----------	--------	-------	------	-------	----------

(#) = Out of Range

SPCC's out = 0 CCC's out = 0