

Data Path : Z:\svoasrv\HPCHEM1\BNA_G\Data\BG030823\
 Data File : BGCCC886.D
 Acq On : 8 Mar 2023 18:30
 Operator : CG/JU
 Sample : SSTDCCC020
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_G
 ClientSampleId :
 SSTD020462

Quant Time: Mar 09 00:01:45 2023
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_G\Methods\SFAM-EPA-BG030723.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Wed Mar 08 23:58:51 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
Internal Standards						
1) 1,4-Dichlorobenzene-d4	8.420	152	13614	20.000	ng/u1	0.00
20) Naphthalene-d8	11.263	136	57352	20.000	ng/u1	0.00
38) Acenaphthene-d10	15.035	164	49359	20.000	ng/u1	0.00
64) Phenanthrene-d10	17.773	188	135400	20.000	ng/u1	0.00
79) Chrysene-d12	22.098	240	130802	20.000	ng/u1	0.00
88) Perylene-d12	25.647	264	141202	20.000	ng/u1	0.00
System Monitoring Compounds						
3) 1,4-Dioxane-d8	3.690	96	2851	8.121	ng/uL	0.00
4) Pyridine-d5	4.131	84	23381	20.473	ng/u1	0.00
7) Phenol-d5	7.538	99	28719	20.582	ng/u1	0.00
9) Bis-(2-Chloroethyl)eth...	7.721	67	17945	20.475	ng/u1	0.00
11) 2-Chlorophenol-d4	7.938	132	18601	21.300	ng/u1	0.00
15) 4-Methylphenol-d8	9.107	113	22210	21.095	ng/u1	0.00
21) Nitrobenzene-d5	9.589	128	10096	21.224	ng/u1	0.00
24) 2-Nitrophenol-d4	10.323	143	12219	21.713	ng/u1	0.00
28) 2,4-Dichlorophenol-d3	10.864	165	23790	21.825	ng/u1	0.00
31) 4-Chloroaniline-d4	11.381	131	30607	21.310	ng/u1	0.00
46) Dimethylphthalate-d6	14.419	166	88633	20.993	ng/u1	0.00
49) Acenaphthylene-d8	14.730	160	95264	20.550	ng/u1	0.00
54) 4-Nitrophenol-d4	15.194	143	14842	20.746	ng/u1	0.00
60) Fluorene-d10	16.017	176	82030	20.953	ng/u1	0.00
65) 4,6-Dinitro-2-methylph...	16.122	200	18756	19.167	ng/u1	0.00
73) Anthracene-d10	17.873	188	134295	20.522	ng/u1	0.00
81) Pyrene-d10	20.135	212	166199	20.860	ng/u1	0.00
92) Benzo(a)pyrene-d12	25.400	264	160295	20.732	ng/u1	0.00
Target Compounds						
					Qvalue	
2) 1,4-Dioxane	3.731	88	3593	9.184	ng/uL#	77
5) Pyridine	4.148	79	25559	20.889	ng/u1	94
6) Benzaldehyde	7.538	77	18136	24.693	ng/u1#	88
8) Phenol	7.568	94	28872	20.296	ng/u1	91
10) Bis(2-Chloroethyl)ether	7.815	93	21047	21.060	ng/u1	98
12) 2-Chlorophenol	7.973	128	18581	20.709	ng/u1	91
13) 2-Methylphenol	8.843	108	21622	20.674	ng/u1	84
14) 2,2'-oxybis(1-Chloropr...	8.937	45	27999	20.413	ng/u1	99
16) Acetophenone	9.242	105	38186	21.193	ng/u1	95
17) N-Nitroso-di-n-propyla...	9.213	70	21283	20.309	ng/u1#	97
18) 4-Methylphenol	9.172	108	23677	20.857	ng/u1	85
19) Hexachloroethane	9.524	117	7809	21.074	ng/u1#	83
22) Nitrobenzene	9.630	77	31479	21.207	ng/u1	91
23) Isophorone	10.153	82	61129	21.096	ng/u1	97
25) 2-Nitrophenol	10.353	139	12783	21.727	ng/u1	95
26) 2,4-Dimethylphenol	10.394	107	27776	21.535	ng/u1	99
27) Bis(2-Chloroethoxy)met...	10.635	93	30899	21.906	ng/u1	93
29) 2,4-Dichlorophenol	10.887	162	22906	21.797	ng/u1	92
30) Naphthalene	11.310	128	68077	21.489	ng/u1	98
32) 4-Chloroaniline	11.404	127	31044	22.061	ng/u1	100
33) Hexachlorobutadiene	11.587	225	18062	21.840	ng/u1	95
34) Caprolactam	12.145	113	8602	22.340	ng/u1	85
35) 4-Chloro-3-methylphenol	12.491	107	27564	22.592	ng/u1	97
36) 2-Methylnaphthalene	12.885	142	49697	21.808	ng/u1	98


Data Path : Z:\svoasrv\HPCHEM1\BNA_G\Data\BG030823\
 Data File : BGCCC886.D
 Acq On : 8 Mar 2023 18:30
 Operator : CG/JU
 Sample : SSTDCCC020
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_G
 ClientSampleId :
 SSTD020462

Quant Time: Mar 09 00:01:45 2023
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_G\Methods\SFAM-EPA-BG030723.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Wed Mar 08 23:58:51 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
37) 1-Methylnaphthalene	13.102	142	48895	21.474	ng/ul	99
39) 1,2,4,5-Tetrachloroben...	13.243	216	36256	19.878	ng/ul	98
40) Hexachlorocyclopentadiene	13.220	237	23271	18.764	ng/ul	91
41) 2,4,6-Trichlorophenol	13.473	196	25488	21.459	ng/ul	92
42) 2,4,5-Trichlorophenol	13.543	196	27549	20.837	ng/ul	95
43) 1,1'-Biphenyl	13.872	154	74522	20.159	ng/ul	98
44) 2-Chloronaphthalene	13.919	162	61617	20.215	ng/ul	97
45) 2-Nitroaniline	14.107	65	23428	20.844	ng/ul	91
47) Dimethylphthalate	14.466	163	90652	21.240	ng/ul	98
48) 2,6-Dinitrotoluene	14.589	165	18379	21.415	ng/ul	89
50) Acenaphthylene	14.759	152	97331	20.776	ng/ul	99
51) 3-Nitroaniline	14.924	138	17394	22.462	ng/ul	95
52) Acenaphthene	15.094	153	68368	20.866	ng/ul	96
53) 2,4-Dinitrophenol	15.124	184	11868	18.864	ng/ul	94
55) 4-Nitrophenol	15.206	109	18583	20.152	ng/ul	88
56) Dibenzofuran	15.429	168	104195	20.802	ng/ul	97
57) 2,4-Dinitrotoluene	15.376	165	27305	20.637	ng/ul	91
58) 2,3,4,6-Tetrachlorophenol	15.647	232	26316	20.655	ng/ul#	97
59) Diethylphthalate	15.817	149	89042	20.641	ng/ul	98
61) Fluorene	16.070	166	83064	20.826	ng/ul	98
62) 4-Chlorophenyl-phenyle...	16.052	204	50706	21.190	ng/ul	95
63) 4-Nitroaniline	16.075	138	17983	23.410	ng/ul	91
66) 4,6-Dinitro-2-methylph...	16.134	198	19001	20.289	ng/ul	100
67) N-Nitrosodiphenylamine	16.263	169	75742	20.830	ng/ul	95
68) 4-Bromophenyl-phenylether	16.951	248	34853	20.753	ng/ul	99
69) Hexachlorobenzene	17.074	284	37219	20.430	ng/ul	92
70) Atrazine	17.198	200	33639	20.576	ng/ul	98
71) Pentachlorophenol	17.415	266	20592	19.176	ng/ul	96
72) Phenanthrene	17.815	178	153211	21.026	ng/ul	99
74) Anthracene	17.909	178	151962	20.472	ng/ul	98
75) 1,2,3,4-Tetrachloroben...	13.843	216	39083	20.140	ng/uL	96
76) Pentachlorobenzene	15.347	250	41691	20.578	ng/uL	94
77) Carbazole	18.167	167	132763	21.076	ng/ul	97
78) Di-n-butylphthalate	18.696	149	149002	20.449	ng/ul	98
80) Fluoranthene	19.800	202	200024	21.062	ng/ul	99
82) Pyrene	20.165	202	200720	21.107	ng/ul	99
83) Butylbenzylphthalate	21.023	149	66030	20.936	ng/ul	99
84) 3,3'-Dichlorobenzidine	21.969	252	70991	22.404	ng/ul	99
85) Benzo(a)anthracene	22.074	228	197154	20.618	ng/ul	98
86) Bis(2-ethylhexyl)phtha...	21.933	149	92966	20.626	ng/ul	98
87) Chrysene	22.145	228	184255	20.794	ng/ul	99
89) Di-n-octyl phthalate	23.255	149	159423	20.731	ng/ul	100
90) Benzo(b)fluoranthene	24.507	252	198837	20.600	ng/ul	100
91) Benzo(k)fluoranthene	24.583	252	197893	21.341	ng/ul	98
93) Benzo(a)pyrene	25.482	252	171917	20.846	ng/ul#	99
94) Indeno(1,2,3-cd)pyrene	29.742	276	236786	21.011	ng/ul	95
95) Dibenzo(a,h)anthracene	29.812	278	194569	20.921	ng/ul#	96
96) Benzo(g,h,i)perylene	31.017	276	98958	10.951	ng/ul	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed



#2

1,4-Dioxane

Concen: 9.184 ng/uL

RT: 3.731 min Scan# 31

Delta R.T. -0.000 min

Lab File: BGCCC886.D

Acq: 8 Mar 2023 18:30

Instrument :

BNA_G

ClientSampleId :

SSTD020462





Tgt Ion: 88 Resp: 3593

Ion Ratio Lower Upper

88 100

43 28.9 40.2 60.4#

58 68.9 69.0 103.6#



#5

Pyridine

Concen: 20.889 ng/uL

RT: 4.148 min Scan# 104

Delta R.T. -0.000 min

Lab File: BGCCC886.D

Acq: 8 Mar 2023 18:30




Tgt Ion: 79 Resp: 25559


Ion Ratio Lower Upper

79 100

52 68.4 60.3 90.5

51 37.5 31.0 46.4









#6
Benzaldehyde
Concen: 24.693 ng/ul
RT: 7.538 min Scan# 681
Delta R.T. -0.000 min
Lab File: BGCCC886.D
Acq: 8 Mar 2023 18:30




Instrument :
BNA_G
ClientSampleId :
SSTD020462


Tgt Ion: 77 Resp: 18136
Ion Ratio Lower Upper
77 100
105 89.1 70.3 105.5
106 95.5 59.0 88.6#



#8
Phenol
Concen: 20.296 ng/ul
RT: 7.568 min Scan# 686
Delta R.T. -0.000 min
Lab File: BGCCC886.D
Acq: 8 Mar 2023 18:30

Tgt Ion: 94 Resp: 28872
Ion Ratio Lower Upper
94 100
65 37.5 27.2 40.8
66 48.3 44.6 66.8





#10

Bis(2-Chloroethyl)ether

Concen: 21.060 ng/ul

RT: 7.815 min Scan# 728

Delta R.T. -0.000 min

Lab File: BGCCC886.D

Acq: 8 Mar 2023 18:30

Instrument :

BNA_G

ClientSampleId :

SSTD020462





Tgt Ion: 93 Resp: 21047

Ion Ratio Lower Upper

93 100

63 87.1 69.0 103.4

95 28.4 25.6 38.4



#12

2-Chlorophenol

Concen: 20.709 ng/ul

RT: 7.973 min Scan# 755

Delta R.T. -0.000 min

Lab File: BGCCC886.D

Acq: 8 Mar 2023 18:30




Tgt Ion:128 Resp: 18581


Ion Ratio Lower Upper

128 100

64 64.8 45.0 67.4

130 31.3 23.0 34.4









#13
2-Methylphenol
Concen: 20.674 ng/ul
RT: 8.843 min Scan# 903
Delta R.T. -0.000 min
Lab File: BGCCC886.D
Acq: 8 Mar 2023 18:30




Instrument :
BNA_G
ClientSampleId :
SSTD020462


Tgt Ion:108 Resp: 21622
Ion Ratio Lower Upper
108 100
107 93.0 63.4 95.0



#14
2,2'-oxybis(1-Chloropropane)
Concen: 20.413 ng/ul
RT: 8.937 min Scan# 919
Delta R.T. 0.006 min
Lab File: BGCCC886.D
Acq: 8 Mar 2023 18:30

Tgt Ion: 45 Resp: 27999
Ion Ratio Lower Upper
45 100
77 18.8 15.0 22.4
79 11.8 8.3 12.5









#16
Acetophenone
Concen: 21.193 ng/ul
RT: 9.242 min Scan# 911
Delta R.T. -0.000 min
Lab File: BGCCC886.D
Acq: 8 Mar 2023 18:30




Instrument :
BNA_G
ClientSampleId :
SSTD020462


Tgt Ion:105 Resp: 38186
Ion Ratio Lower Upper
105 100
77 86.9 65.9 98.9
51 31.0 23.3 34.9



#17
N-Nitroso-di-n-propylamine
Concen: 20.309 ng/ul
RT: 9.213 min Scan# 966
Delta R.T. -0.006 min
Lab File: BGCCC886.D
Acq: 8 Mar 2023 18:30

Tgt Ion: 70 Resp: 21283
Ion Ratio Lower Upper
70 100
42 62.1 47.8 71.6
101 8.2 5.4 8.2#
130 15.8 13.8 20.8









#18
4-Methylphenol
Concen: 20.857 ng/ul
RT: 9.172 min Scan# 911
Delta R.T. -0.000 min
Lab File: BGCCC886.D
Acq: 8 Mar 2023 18:30




Instrument :
BNA_G
ClientSampleId :
SSTD020462


Tgt Ion:108 Resp: 23677
Ion Ratio Lower Upper
108 100
107 115.2 80.2 120.2



#19
Hexachloroethane
Concen: 21.074 ng/ul
RT: 9.524 min Scan# 1019
Delta R.T. -0.000 min
Lab File: BGCCC886.D
Acq: 8 Mar 2023 18:30

Tgt Ion:117 Resp: 7809
Ion Ratio Lower Upper
117 100
201 136.0 87.5 131.3#
199 74.8 63.8 95.6





#22

Nitrobenzene

Concen: 21.207 ng/ul

RT: 9.630 min Scan# 1103

Delta R.T. -0.006 min

Lab File: BGCCC886.D

Acq: 8 Mar 2023 18:30

Instrument :

BNA_G

ClientSampleId :

SSTD020462





Tgt Ion: 77 Resp: 31479

Ion Ratio Lower Upper

77 100

123 31.3 31.0 46.4

65 14.4 11.1 16.7



#23

Isophorone

Concen: 21.096 ng/ul

RT: 10.153 min Scan# 1126

Delta R.T. -0.000 min

Lab File: BGCCC886.D

Acq: 8 Mar 2023 18:30




Tgt Ion: 82 Resp: 61129


Ion Ratio Lower Upper

82 100

95 7.9 6.1 9.1

138 14.0 12.6 18.8









#25
2-Nitrophenol
Concen: 21.727 ng/ul
RT: 10.353 min Scan# 1160
Delta R.T. -0.000 min
Lab File: BGCCC886.D
Acq: 8 Mar 2023 18:30




Instrument :
BNA_G
ClientSampleId :
SSTD020462


Tgt Ion:139 Resp: 12783
Ion Ratio Lower Upper
139 100
65 68.1 59.4 89.2
109 45.6 35.3 52.9



#26
2,4-Dimethylphenol
Concen: 21.535 ng/ul
RT: 10.394 min Scan# 1167
Delta R.T. -0.000 min
Lab File: BGCCC886.D
Acq: 8 Mar 2023 18:30

Tgt Ion:107 Resp: 27776
Ion Ratio Lower Upper
107 100
121 46.2 36.6 55.0
122 78.2 63.2 94.8









#27
 Bis(2-Chloroethoxy)methane
 Concen: 21.906 ng/ul
 RT: 10.635 min Scan# 1208
 Delta R.T. -0.000 min
 Lab File: BGCCC886.D
 Acq: 8 Mar 2023 18:30




Instrument :
 BNA_G
 ClientSampleId :
 SSTD020462


Tgt Ion:	93	Resp:	30899
Ion	Ratio	Lower	Upper
93	100		
95	28.9	27.3	40.9
123	8.9	7.1	10.7



#29
 2,4-Dichlorophenol
 Concen: 21.797 ng/ul
 RT: 10.887 min Scan# 1251
 Delta R.T. -0.000 min
 Lab File: BGCCC886.D
 Acq: 8 Mar 2023 18:30

Tgt Ion:	162	Resp:	22906
Ion	Ratio	Lower	Upper
162	100		
164	60.9	53.7	80.5
98	40.1	28.7	43.1









#30
Naphthalene
Concen: 21.489 ng/ul
RT: 11.310 min Scan# 11
Delta R.T. -0.000 min
Lab File: BGCCC886.D
Acq: 8 Mar 2023 18:30




Instrument :
BNA_G
ClientSampleId :
SSTD020462


Tgt Ion:128 Resp: 68077
Ion Ratio Lower Upper
128 100
129 11.9 9.4 14.0
127 13.7 11.8 17.6



#32
4-Chloroaniline
Concen: 22.061 ng/ul
RT: 11.404 min Scan# 1339
Delta R.T. -0.000 min
Lab File: BGCCC886.D
Acq: 8 Mar 2023 18:30

Tgt Ion:127 Resp: 31044
Ion Ratio Lower Upper
127 100
129 31.4 25.1 37.7









#33
Hexachlorobutadiene
Concen: 21.840 ng/ul
RT: 11.587 min Scan# 11
Delta R.T. -0.006 min
Lab File: BGCCC886.D
Acq: 8 Mar 2023 18:30




Instrument :
BNA_G
ClientSampleId :
SSTD020462


Tgt Ion:225 Resp: 18062
Ion Ratio Lower Upper
225 100
223 57.0 49.8 74.6
227 59.0 49.0 73.4



#34
Caprolactam
Concen: 22.340 ng/ul
RT: 12.145 min Scan# 1465
Delta R.T. 0.006 min
Lab File: BGCCC886.D
Acq: 8 Mar 2023 18:30

Tgt Ion:113 Resp: 8602
Ion Ratio Lower Upper
113 100
55 166.2 150.8 226.2
56 154.8 110.0 165.0









#35
4-Chloro-3-methylphenol
Concen: 22.592 ng/ul
RT: 12.491 min Scan# 11
Delta R.T. -0.000 min
Lab File: BGCCC886.D
Acq: 8 Mar 2023 18:30




Instrument :
BNA_G
ClientSampleId :
SSTD020462


Tgt Ion:107 Resp: 27564
Ion Ratio Lower Upper
107 100
144 24.2 20.6 30.8
142 71.1 59.0 88.4



#36
2-Methylnaphthalene
Concen: 21.808 ng/ul
RT: 12.885 min Scan# 1591
Delta R.T. -0.000 min
Lab File: BGCCC886.D
Acq: 8 Mar 2023 18:30

Tgt Ion:142 Resp: 49697
Ion Ratio Lower Upper
142 100
141 91.4 74.6 111.8









#37
1-Methylnaphthalene
Concen: 21.474 ng/ul
RT: 13.102 min Scan# 1
Delta R.T. -0.000 min
Lab File: BGCCC886.D
Acq: 8 Mar 2023 18:30




Instrument :
BNA_G
ClientSampleId :
SSTD020462


Tgt Ion:142 Resp: 48895
Ion Ratio Lower Upper
142 100
141 91.8 73.0 109.4
116 5.9 4.6 7.0



#39
1,2,4,5-Tetrachlorobenzene
Concen: 19.878 ng/ul
RT: 13.243 min Scan# 1652
Delta R.T. -0.000 min
Lab File: BGCCC886.D
Acq: 8 Mar 2023 18:30


Tgt Ion:216 Resp: 36256
Ion Ratio Lower Upper
216 100
214 78.4 64.1 96.1
179 19.5 16.0 24.0
108 13.5 9.5 14.3








#40
Hexachlorocyclopentadiene
Concen: 18.764 ng/ul
RT: 13.220 min Scan# 1648
Delta R.T. -0.000 min
Lab File: BGCCC886.D
Acq: 8 Mar 2023 18:30


Instrument :
BNA_G
ClientSampleId :
SSTD020462





Tgt Ion:237 Resp: 23271
Ion Ratio Lower Upper
237 100
235 61.0 55.2 82.8
272 12.5 10.6 16.0




#41
2,4,6-Trichlorophenol
Concen: 21.459 ng/ul
RT: 13.473 min Scan# 1691
Delta R.T. -0.000 min
Lab File: BGCCC886.D
Acq: 8 Mar 2023 18:30



Tgt Ion:196 Resp: 25488
Ion Ratio Lower Upper
196 100
198 92.8 80.9 121.3
200 30.1 27.9 41.9









#42
2,4,5-Trichlorophenol
Concen: 20.837 ng/ul
RT: 13.543 min Scan# 1703
Delta R.T. -0.000 min
Lab File: BGCCC886.D
Acq: 8 Mar 2023 18:30




Instrument :
BNA_G
ClientSampleId :
SSTD020462


Tgt Ion:196 Resp: 27549
Ion Ratio Lower Upper
196 100
198 94.9 72.7 109.1
200 33.5 24.4 36.6



#43
1,1'-Biphenyl
Concen: 20.159 ng/ul
RT: 13.872 min Scan# 1759
Delta R.T. -0.000 min
Lab File: BGCCC886.D
Acq: 8 Mar 2023 18:30

Tgt Ion:154 Resp: 74522
Ion Ratio Lower Upper
154 100
153 41.5 34.2 51.4
76 11.3 9.2 13.8









#44
2-Chloronaphthalene
Concen: 20.215 ng/ul
RT: 13.919 min Scan# 1767
Delta R.T. -0.000 min
Lab File: BGCCC886.D
Acq: 8 Mar 2023 18:30




Instrument :
BNA_G
ClientSampleId :
SSTD020462


Tgt Ion:162 Resp: 61617
Ion Ratio Lower Upper
162 100
164 33.9 28.3 42.5
127 35.9 26.8 40.2



#45
2-Nitroaniline
Concen: 20.844 ng/ul
RT: 14.107 min Scan# 1799
Delta R.T. -0.000 min
Lab File: BGCCC886.D
Acq: 8 Mar 2023 18:30

Tgt Ion: 65 Resp: 23428
Ion Ratio Lower Upper
65 100
92 56.7 54.7 82.1
138 83.5 63.4 95.0









#47
Dimethylphthalate
Concen: 21.240 ng/ul
RT: 14.466 min Scan# 1860
Delta R.T. -0.000 min
Lab File: BGCCC886.D
Acq: 8 Mar 2023 18:30




Instrument :
BNA_G
ClientSampleId :
SSTD020462


Tgt Ion:163 Resp: 90652
Ion Ratio Lower Upper
163 100
194 4.2 3.1 4.7
164 9.7 8.4 12.6



#48
2,6-Dinitrotoluene
Concen: 21.415 ng/ul
RT: 14.589 min Scan# 1881
Delta R.T. -0.000 min
Lab File: BGCCC886.D
Acq: 8 Mar 2023 18:30

Tgt Ion:165 Resp: 18379
Ion Ratio Lower Upper
165 100
89 78.8 53.7 80.5
121 23.6 19.8 29.8









#50
Acenaphthylene
Concen: 20.776 ng/ul
RT: 14.759 min Scan# 1910
Delta R.T. -0.000 min
Lab File: BGCCC886.D
Acq: 8 Mar 2023 18:30




Instrument :
BNA_G
ClientSampleId :
SSTD020462


Tgt Ion:152 Resp: 97331
Ion Ratio Lower Upper
152 100
151 19.0 15.2 22.8
153 12.6 10.6 16.0



#51
3-Nitroaniline
Concen: 22.462 ng/ul
RT: 14.924 min Scan# 1938
Delta R.T. -0.000 min
Lab File: BGCCC886.D
Acq: 8 Mar 2023 18:30

Tgt Ion:138 Resp: 17394
Ion Ratio Lower Upper
138 100
108 10.5 9.0 13.6
92 132.0 110.3 165.5









#52
Acenaphthene
Concen: 20.866 ng/ul
RT: 15.094 min Scan# 1967
Delta R.T. -0.000 min
Lab File: BGCCC886.D
Acq: 8 Mar 2023 18:30




Instrument :
BNA_G
ClientSampleId :
SSTD020462


Tgt Ion:153 Resp: 68368
Ion Ratio Lower Upper
153 100
152 44.4 40.5 60.7
154 83.9 67.8 101.8



#53
2,4-Dinitrophenol
Concen: 18.864 ng/ul
RT: 15.124 min Scan# 1972
Delta R.T. -0.000 min
Lab File: BGCCC886.D
Acq: 8 Mar 2023 18:30

Tgt Ion:184 Resp: 11868
Ion Ratio Lower Upper
184 100
63 68.6 52.9 79.3
154 74.5 54.1 81.1









#55
4-Nitrophenol
Concen: 20.152 ng/ul
RT: 15.206 min Scan# 1986
Delta R.T. -0.000 min
Lab File: BGCCC886.D
Acq: 8 Mar 2023 18:30




Instrument :
BNA_G
ClientSampleId :
SSTD020462


Tgt Ion:109 Resp: 18583
Ion Ratio Lower Upper
109 100
139 86.7 63.3 94.9
65 117.2 82.6 123.8



#56
Dibenzofuran
Concen: 20.802 ng/ul
RT: 15.429 min Scan# 2024
Delta R.T. 0.006 min
Lab File: BGCCC886.D
Acq: 8 Mar 2023 18:30

Tgt Ion:168 Resp: 104195
Ion Ratio Lower Upper
168 100
139 35.4 29.5 44.3









#57
2,4-Dinitrotoluene
Concen: 20.637 ng/ul
RT: 15.376 min Scan# 2014
Delta R.T. 0.006 min
Lab File: BGCCC886.D
Acq: 8 Mar 2023 18:30




Instrument :
BNA_G
ClientSampleId :
SSTD020462


Tgt Ion:165 Resp: 27305
Ion Ratio Lower Upper
165 100
63 49.0 44.5 66.7
182 2.3 1.8 2.8



#58
2,3,4,6-Tetrachlorophenol
Concen: 20.655 ng/ul
RT: 15.647 min Scan# 2061
Delta R.T. -0.000 min
Lab File: BGCCC886.D
Acq: 8 Mar 2023 18:30

Tgt Ion:232 Resp: 26316
Ion Ratio Lower Upper
232 100
131 33.2 27.2 40.8
130 0.0 2.4 3.6#
166 26.5 19.1 28.7









#59
Diethylphthalate
Concen: 20.641 ng/ul
RT: 15.817 min Scan# 2090
Delta R.T. -0.000 min
Lab File: BGCCC886.D
Acq: 8 Mar 2023 18:30




Instrument :
BNA_G
ClientSampleId :
SSTD020462


Tgt Ion:149 Resp: 89042
Ion Ratio Lower Upper
149 100
177 24.1 19.0 28.4
150 11.6 10.6 15.8



#61
Fluorene
Concen: 20.826 ng/ul
RT: 16.070 min Scan# 2133
Delta R.T. -0.006 min
Lab File: BGCCC886.D
Acq: 8 Mar 2023 18:30

Tgt Ion:166 Resp: 83064
Ion Ratio Lower Upper
166 100
165 98.8 80.5 120.7
167 13.9 11.3 16.9









#62
4-Chlorophenyl-phenylether
Concen: 21.190 ng/ul
RT: 16.052 min Scan# 2130
Delta R.T. -0.000 min
Lab File: BGCCC886.D
Acq: 8 Mar 2023 18:30




Instrument :
BNA_G
ClientSampleId :
SSTD020462


Tgt Ion:204 Resp: 50706
Ion Ratio Lower Upper
204 100
206 30.5 26.6 39.8
141 49.0 36.1 54.1



#63
4-Nitroaniline
Concen: 23.410 ng/ul
RT: 16.075 min Scan# 2134
Delta R.T. -0.006 min
Lab File: BGCCC886.D
Acq: 8 Mar 2023 18:30

Tgt Ion:138 Resp: 17983
Ion Ratio Lower Upper
138 100
92 57.6 47.3 70.9
108 116.3 104.6 157.0









#66
 4,6-Dinitro-2-methylphenol
 Concen: 20.289 ng/ul
 RT: 16.134 min Scan# 2144
 Delta R.T. -0.000 min
 Lab File: BGCCC886.D
 Acq: 8 Mar 2023 18:30




Instrument :
 BNA_G
 ClientSampleId :
 SST020462


Tgt Ion:198 Resp: 19001
 Ion Ratio Lower Upper
 198 100
 182 3.6 3.1 4.7
 77 25.7 20.5 30.7



#67
 N-Nitrosodiphenylamine
 Concen: 20.830 ng/ul
 RT: 16.263 min Scan# 2166
 Delta R.T. -0.000 min
 Lab File: BGCCC886.D
 Acq: 8 Mar 2023 18:30

Tgt Ion:169 Resp: 75742
 Ion Ratio Lower Upper
 169 100
 168 71.5 53.9 80.9
 167 38.2 28.2 42.4









#68
4-Bromophenyl-phenylether
Concen: 20.753 ng/ul
RT: 16.951 min Scan# 211
Delta R.T. -0.000 min
Lab File: BGCCC886.D
Acq: 8 Mar 2023 18:30




Instrument :
BNA_G
ClientSampleId :
SSTD020462


Tgt Ion:248 Resp: 34853
Ion Ratio Lower Upper
248 100
250 97.9 78.6 117.8
141 60.6 47.9 71.9



#69
Hexachlorobenzene
Concen: 20.430 ng/ul
RT: 17.074 min Scan# 2304
Delta R.T. -0.000 min
Lab File: BGCCC886.D
Acq: 8 Mar 2023 18:30

Tgt Ion:284 Resp: 37219
Ion Ratio Lower Upper
284 100
142 30.1 20.2 30.4
249 34.0 24.2 36.4





#70

Atrazine

Concen: 20.576 ng/ul

RT: 17.198 min Scan# 21

Delta R.T. -0.000 min

Lab File: BGCCC886.D

Acq: 8 Mar 2023 18:30

Instrument :

BNA_G

ClientSampleId :

SSTD020462





Tgt Ion:200 Resp: 33639

Ion Ratio Lower Upper

200 100

173 20.4 17.5 26.3

215 50.9 39.6 59.4



#71

Pentachlorophenol

Concen: 19.176 ng/ul

RT: 17.415 min Scan# 2362

Delta R.T. -0.000 min

Lab File: BGCCC886.D

Acq: 8 Mar 2023 18:30




Tgt Ion:266 Resp: 20592


Ion Ratio Lower Upper

266 100

264 61.2 50.6 76.0

268 64.7 48.6 73.0









#72
Phenanthrene
Concen: 21.026 ng/ul
RT: 17.815 min Scan# 2430
Delta R.T. -0.000 min
Lab File: BGCCC886.D
Acq: 8 Mar 2023 18:30




Instrument :
BNA_G
ClientSampleId :
SSTD020462


Tgt Ion:178 Resp: 153211
Ion Ratio Lower Upper
178 100
179 15.2 11.9 17.9
176 18.0 14.6 22.0



#74
Anthracene
Concen: 20.472 ng/ul
RT: 17.909 min Scan# 2446
Delta R.T. -0.000 min
Lab File: BGCCC886.D
Acq: 8 Mar 2023 18:30

Tgt Ion:178 Resp: 151962
Ion Ratio Lower Upper
178 100
179 15.7 12.4 18.6
176 18.8 14.2 21.2









#75
1,2,3,4-Tetrachlorobenzene
Concen: 20.140 ng/uL
RT: 13.843 min Scan# 11
Delta R.T. -0.000 min
Lab File: BGCCC886.D
Acq: 8 Mar 2023 18:30




Instrument :
BNA_G
ClientSampleId :
SSTD020462


Tgt Ion:216 Resp: 39083
Ion Ratio Lower Upper
216 100
214 80.1 65.4 98.0
218 45.0 39.4 59.2



#76
Pentachlorobenzene
Concen: 20.578 ng/uL
RT: 15.347 min Scan# 2010
Delta R.T. -0.000 min
Lab File: BGCCC886.D
Acq: 8 Mar 2023 18:30

Tgt Ion:250 Resp: 41691
Ion Ratio Lower Upper
250 100
248 65.5 46.8 70.2
252 60.5 50.6 75.8









#77
 Carbazole
 Concen: 21.076 ng/ul
 RT: 18.167 min Scan# 2490
 Delta R.T. -0.000 min
 Lab File: BGCCC886.D
 Acq: 8 Mar 2023 18:30




Instrument :
 BNA_G
 ClientSampleId :
 SST020462


Tgt Ion:167 Resp: 132763
 Ion Ratio Lower Upper
 167 100
 166 20.9 18.1 27.1
 139 11.5 10.2 15.2



#78
 Di-n-butylphthalate
 Concen: 20.449 ng/ul
 RT: 18.696 min Scan# 2580
 Delta R.T. -0.000 min
 Lab File: BGCCC886.D
 Acq: 8 Mar 2023 18:30

Tgt Ion:149 Resp: 149002
 Ion Ratio Lower Upper
 149 100
 150 8.5 7.4 11.2
 104 7.2 5.4 8.2









#80
Fluoranthene
Concen: 21.062 ng/ul
RT: 19.800 min Scan# 21
Delta R.T. -0.006 min
Lab File: BGCCC886.D
Acq: 8 Mar 2023 18:30




Instrument :
BNA_G
ClientSampleId :
SSTD020462


Tgt Ion:202 Resp: 200024
Ion Ratio Lower Upper
202 100
101 6.0 5.0 7.6
100 4.7 3.6 5.4



#82
Pyrene
Concen: 21.107 ng/ul
RT: 20.165 min Scan# 2830
Delta R.T. -0.000 min
Lab File: BGCCC886.D
Acq: 8 Mar 2023 18:30

Tgt Ion:202 Resp: 200720
Ion Ratio Lower Upper
202 100
101 6.8 5.2 7.8
100 6.2 4.5 6.7





#83

Butylbenzylphthalate

Concen: 20.936 ng/ul

RT: 21.023 min Scan# 2976

Delta R.T. -0.006 min

Lab File: BGCCC886.D

Acq: 8 Mar 2023 18:30

Instrument :

BNA_G

ClientSampleId :

SSTD020462





Tgt Ion:149 Resp: 66030

Ion Ratio Lower Upper

149 100

91 82.6 66.2 99.2

206 24.5 18.8 28.2



#84

3,3'-Dichlorobenzidine

Concen: 22.404 ng/ul

RT: 21.969 min Scan# 3137

Delta R.T. -0.006 min

Lab File: BGCCC886.D

Acq: 8 Mar 2023 18:30




Tgt Ion:252 Resp: 70991


Ion Ratio Lower Upper

252 100

254 66.5 52.6 79.0

126 5.8 4.6 7.0









#85
Benzo(a)anthracene
Concen: 20.618 ng/ul
RT: 22.074 min Scan# 3155
Delta R.T. -0.000 min
Lab File: BGCCC886.D
Acq: 8 Mar 2023 18:30




Instrument :
BNA_G
ClientSampleId :
SSTD020462


Tgt Ion:228 Resp: 197154
Ion Ratio Lower Upper
228 100
229 19.1 14.9 22.3
226 25.3 21.0 31.6



#86
Bis(2-ethylhexyl)phthalate
Concen: 20.626 ng/ul
RT: 21.933 min Scan# 3131
Delta R.T. -0.006 min
Lab File: BGCCC886.D
Acq: 8 Mar 2023 18:30

Tgt Ion:149 Resp: 92966
Ion Ratio Lower Upper
149 100
167 26.5 20.3 30.5
279 4.6 4.0 6.0









#87
 Chrysene
 Concen: 20.794 ng/ul
 RT: 22.145 min Scan# 3167
 Delta R.T. -0.006 min
 Lab File: BGCCC886.D
 Acq: 8 Mar 2023 18:30




Instrument :
 BNA_G
 ClientSampleId :
 SSTD020462


Tgt Ion:	228	Resp:	184255
Ion Ratio	Lower	Upper	
228	100		
226	29.5	23.3	34.9
229	20.5	15.8	23.6



#89
 Di-n-octyl phthalate
 Concen: 20.731 ng/ul
 RT: 23.255 min Scan# 3356
 Delta R.T. -0.006 min
 Lab File: BGCCC886.D
 Acq: 8 Mar 2023 18:30

Tgt Ion:149 Resp: 159423





#90

Benzo(b)fluoranthene

Concen: 20.600 ng/ul

RT: 24.507 min Scan# 3569

Delta R.T. -0.006 min

Lab File: BGCCC886.D

Acq: 8 Mar 2023 18:30

Instrument :

BNA_G

ClientSampleId :

SSTD020462





Tgt Ion:252 Resp: 198837

Ion Ratio Lower Upper

252 100

253 21.4 17.3 25.9

125 4.3 3.6 5.4



#91

Benzo(k)fluoranthene

Concen: 21.341 ng/ul

RT: 24.583 min Scan# 3582

Delta R.T. -0.006 min

Lab File: BGCCC886.D

Acq: 8 Mar 2023 18:30




Tgt Ion:252 Resp: 197893


Ion Ratio Lower Upper

252 100

253 21.7 16.6 24.8

125 4.6 3.5 5.3





#93

Benzo(a)pyrene

Concen: 20.846 ng/ul

RT: 25.482 min Scan# 31

Delta R.T. -0.000 min

Lab File: BGCCC886.D

Acq: 8 Mar 2023 18:30

Instrument :

BNA_G

ClientSampleId :

SSTD020462





Tgt Ion:252 Resp: 171917

Ion Ratio Lower Upper

252 100

253 21.3 17.2 25.8

125 5.1 3.4 5.0#



#94

Indeno(1,2,3-cd)pyrene

Concen: 21.011 ng/ul

RT: 29.742 min Scan# 4460

Delta R.T. -0.012 min

Lab File: BGCCC886.D

Acq: 8 Mar 2023 18:30




Tgt Ion:276 Resp: 236786


Ion Ratio Lower Upper

276 100

138 9.4 6.4 9.6

277 22.4 19.8 29.8









#95
Dibenzo(a,h)anthracene
Concen: 20.921 ng/ul
RT: 29.812 min Scan# 4472
Delta R.T. -0.012 min
Lab File: BGCCC886.D
Acq: 8 Mar 2023 18:30

Instrument :
BNA_G
ClientSampleId :
SSTD020462

Tgt Ion:278 Resp: 194569
Ion Ratio Lower Upper
278 100
139 7.2 3.8 5.8#
279 23.4 20.0 30.0



#96
Benzo(g,h,i)perylene
Concen: 10.951 ng/ul
RT: 31.017 min Scan# 4677
Delta R.T. -0.018 min
Lab File: BGCCC886.D
Acq: 8 Mar 2023 18:30

Tgt Ion:276 Resp: 98958
Ion Ratio Lower Upper
276 100
138 8.1 6.1 9.1
277 21.0 16.5 24.7

