

Data Path : Z:\svoasrv\HPCHEM1\BNA_G\Data\BG030823\
 Data File : BG056912.D
 Acq On : 9 Mar 2023 13:05
 Operator : CG/JU
 Sample : 01713-21
 Misc :
 ALS Vial : 39 Sample Multiplier: 1

Instrument :
 BNA_G
 ClientSampleId :

Quant Time: Mar 09 14:59:16 2023
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_G\Methods\SFAM-EPA-BG030723.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Wed Mar 08 23:58:51 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	8.416	152	16415	20.000	ng/ul	0.00
20) Naphthalene-d8	11.260	136	69378	20.000	ng/ul	0.00
38) Acenaphthene-d10	15.032	164	59349	20.000	ng/ul	0.00
64) Phenanthrene-d10	17.770	188	160770	20.000	ng/ul	0.00
79) Chrysene-d12	22.094	240	153636	20.000	ng/ul	0.00
88) Perylene-d12	25.643	264	166805	20.000	ng/ul	0.00
System Monitoring Compounds						
3) 1,4-Dioxane-d8	3.698	96	1540	3.638	ng/uL	0.00
4) Pyridine-d5	4.133	84	27630	20.065	ng/ul	0.00
7) Phenol-d5	7.541	99	35104	20.865	ng/ul	0.00
9) Bis-(2-Chloroethyl)eth...	7.717	67	22403	21.200	ng/ul	0.00
11) 2-Chlorophenol-d4	7.940	132	22907	21.755	ng/ul	0.00
15) 4-Methylphenol-d8	9.110	113	27599	21.740	ng/ul	0.00
21) Nitrobenzene-d5	9.591	128	12671	22.020	ng/ul	0.00
24) 2-Nitrophenol-d4	10.326	143	15768	23.163	ng/ul	0.00
28) 2,4-Dichlorophenol-d3	10.860	165	28765	21.815	ng/ul	0.00
31) 4-Chloroaniline-d4	11.378	131	35027	20.160	ng/ul	0.00
46) Dimethylphthalate-d6	14.415	166	115099	22.673	ng/ul	0.00
49) Acenaphthylene-d8	14.727	160	124402	22.319	ng/ul	0.00
54) 4-Nitrophenol-d4	15.191	143	18255	21.222	ng/ul	0.00
60) Fluorene-d10	16.013	176	104030	22.099	ng/ul	0.00
65) 4,6-Dinitro-2-methylph...	16.113	200	21754	18.723	ng/ul	0.00
73) Anthracene-d10	17.870	188	181592	23.370	ng/ul	0.00
81) Pyrene-d10	20.132	212	223971	23.933	ng/ul	0.00
92) Benzo(a)pyrene-d12	25.390	264	215155	23.557	ng/ul	-0.02

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_G\Data\BG030823\
Data File : BG056912.D
Acq On : 9 Mar 2023 13:05
Operator : CG/JU
Sample : 01713-21
Misc :
ALS Vial : 39 Sample Multiplier: 1

Instrument :
BNA_G
ClientSampleId :

Quant Time: Mar 09 14:59:16 2023
Quant Method : Z:\svoasrv\HPCHEM1\BNA_G\Methods\SFAM-EPA-BG030723.M
Quant Title : SVOA CALIBRATION
QLast Update : Wed Mar 08 23:58:51 2023
Response via : Initial Calibration

