

Data Path : Z:\svoasrv\HPCHEM1\BNA_G\Data\BG030823\
 Data File : BG056907.D
 Acq On : 9 Mar 2023 9:36
 Operator : CG/JU
 Sample : 01713-04
 Misc :
 ALS Vial : 34 Sample Multiplier: 1

Instrument :
 BNA_G
 ClientSampleId :
 DBZY4

Quant Time: Mar 10 01:18:39 2023
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_G\Methods\SFAM-EPA-BG030723.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Wed Mar 08 23:58:51 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	8.414	152	11800	20.000	ng/u1	0.00
20) Naphthalene-d8	11.258	136	50587	20.000	ng/u1	0.00
38) Acenaphthene-d10	15.030	164	42140	20.000	ng/u1	0.00
64) Phenanthrene-d10	17.768	188	115104	20.000	ng/u1	0.00
79) Chrysene-d12	22.092	240	117098	20.000	ng/u1	0.00
88) Perylene-d12	25.641	264	126159	20.000	ng/u1	-0.01
System Monitoring Compounds						
3) 1,4-Dioxane-d8	3.690	96	1391	4.572	ng/uL	0.00
4) Pyridine-d5	4.131	84	20891	21.105	ng/u1	0.00
7) Phenol-d5	7.539	99	31192	25.791	ng/u1	0.00
9) Bis-(2-Chloroethyl)eth...	7.715	67	19947	26.258	ng/u1	0.00
11) 2-Chlorophenol-d4	7.938	132	20574	27.181	ng/u1	0.00
15) 4-Methylphenol-d8	9.102	113	23255	25.483	ng/u1	0.00
21) Nitrobenzene-d5	9.595	128	10961	26.124	ng/u1	0.00
24) 2-Nitrophenol-d4	10.318	143	14150	28.508	ng/u1	0.00
28) 2,4-Dichlorophenol-d3	10.864	165	25529	26.553	ng/u1	0.00
31) 4-Chloroaniline-d4	11.381	131	27081	21.376	ng/u1	0.00
46) Dimethylphthalate-d6	14.413	166	92562	25.680	ng/u1	0.00
49) Acenaphthylene-d8	14.730	160	103865	26.244	ng/u1	0.00
54) 4-Nitrophenol-d4	15.189	143	14929	24.443	ng/u1	0.00
60) Fluorene-d10	16.011	176	87441	26.161	ng/u1	0.00
65) 4,6-Dinitro-2-methylph...	16.117	200	19316	23.220	ng/u1	0.00
73) Anthracene-d10	17.868	188	144166	25.915	ng/u1	0.00
81) Pyrene-d10	20.130	212	188335	26.405	ng/u1	0.00
92) Benzo(a)pyrene-d12	25.394	264	177716	25.726	ng/u1	-0.01

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_G\Data\BG030823\
Data File : BG056907.D
Acq On : 9 Mar 2023 9:36
Operator : CG/JU
Sample : 01713-04
Misc :
ALS Vial : 34 Sample Multiplier: 1

Instrument :
BNA_G
ClientSampleId :
DBZY4

Quant Time: Mar 10 01:18:39 2023
Quant Method : Z:\svoasrv\HPCHEM1\BNA_G\Methods\SFAM-EPA-BG030723.M
Quant Title : SVOA CALIBRATION
QLast Update : Wed Mar 08 23:58:51 2023
Response via : Initial Calibration

