

Data Path : Z:\SVOASRV\HPCHEM1\BNA G\DATA\BG031020\
 Data File : BG044716.D
 Acq On : 10 Mar 2020 16:33
 Operator : CG/JU
 Sample : SSTDICV040
 Misc :
 ALS Vial : 10 Sample Multiplier: 1

Instrument :
 BNA_G
 ClientSampleId :
 ICVBG031020

Quant Time: Mar 10 18:15:12 2020
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA G\METHODS\8270-BG031020.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Tue Mar 10 16:27:39 2020
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	20.000	20.000	0.0	100	0.01
2	1,4-Dioxane	40.000	38.206	4.5	96	0.00
3	Pyridine	40.000	39.393	1.5	98	0.00
4	n-Nitrosodimethylamine	40.000	39.515	1.2	97	0.00
5 S	2-Fluorophenol	80.000	77.850	2.7	99	0.00
6	Aniline	40.000	38.567	3.6	97	0.00
7 S	Phenol-d6	80.000	77.274	3.4	99	0.00
8	2-Chlorophenol	40.000	38.438	3.9	100	0.01
9	Benzaldehyde	40.000	36.732	8.2	96	0.00
10 C	Phenol	40.000	38.620	3.5	100	0.00
11	bis(2-Chloroethyl)ether	40.000	38.153	4.6	98	0.01
12	1,3-Dichlorobenzene	40.000	38.049	4.9	98	0.01
13 C	1,4-Dichlorobenzene	40.000	38.458	3.9	99	0.00
14	1,2-Dichlorobenzene	40.000	38.560	3.6	100	0.00
15	Benzyl Alcohol	40.000	39.165	2.1	99	0.01
16	2,2'-oxybis(1-Chloropropane)	40.000	37.669	5.8	96	0.00
17	2-Methylphenol	40.000	38.733	3.2	99	0.00
18	Hexachloroethane	40.000	38.367	4.1	98	0.00
19 P	n-Nitroso-di-n-propylamine	40.000	38.550	3.6	96	0.01
20	3+4-Methylphenols	40.000	38.186	4.5	95	0.00
21 I	Naphthalene-d8	20.000	20.000	0.0	100	0.00
22	Acetophenone	40.000	37.470	6.3	98	0.01
23 S	Nitrobenzene-d5	80.000	77.164	3.5	100	0.00
24	Nitrobenzene	40.000	38.395	4.0	100	0.00
25	Isophorone	40.000	37.715	5.7	98	0.01
26 C	2-Nitrophenol	40.000	39.093	2.3	105	0.01
27	2,4-Dimethylphenol	40.000	38.424	3.9	101	0.00
28	bis(2-Chloroethoxy)methane	40.000	38.996	2.5	100	0.01
29 C	2,4-Dichlorophenol	40.000	38.709	3.2	99	0.00
30	1,2,4-Trichlorobenzene	40.000	38.583	3.5	102	0.00
31	Naphthalene	40.000	37.842	5.4	99	0.00
32	Benzoic acid	40.000	40.596	-1.5	119	0.00
33	4-Chloroaniline	40.000	37.566	6.1	99	0.00
34 C	Hexachlorobutadiene	40.000	38.788	3.0	101	0.00
35	Caprolactam	40.000	40.181	-0.5	101	0.00
36 C	4-Chloro-3-methylphenol	40.000	38.518	3.7	99	0.00
37	2-Methylnaphthalene	40.000	38.210	4.5	98	0.00
38	1-Methylnaphthalene	40.000	38.288	4.3	99	0.00
39 I	Acenaphthene-d10	20.000	20.000	0.0	97	0.00
40	1,2,4,5-Tetrachlorobenzene	40.000	38.253	4.4	99	0.00
41 P	Hexachlorocyclopentadiene	40.000	37.889	5.3	99	0.01
42 S	2,4,6-Tribromophenol	80.000	77.370	3.3	99	0.00
43 C	2,4,6-Trichlorophenol	40.000	39.425	1.4	101	0.00
44	2,4,5-Trichlorophenol	40.000	39.585	1.0	103	0.01

Data Path : Z:\SVOASRV\HPCHEM1\BNA G\DATA\BG031020\
 Data File : BG044716.D
 Acq On : 10 Mar 2020 16:33
 Operator : CG/JU
 Sample : SSTDICV040
 Misc :
 ALS Vial : 10 Sample Multiplier: 1

Instrument :
 BNA_G
 ClientSampleId :
 ICVBG031020

Quant Time: Mar 10 18:15:12 2020
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA G\METHODS\8270-BG031020.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Tue Mar 10 16:27:39 2020
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
45 S	2-Fluorobiphenyl	80.000	77.889	2.6	98	0.00
46	1,1'-Biphenyl	40.000	38.449	3.9	98	0.00
47	2-Chloronaphthalene	40.000	38.122	4.7	98	0.00
48	2-Nitroaniline	40.000	39.093	2.3	100	0.00
49	Acenaphthylene	40.000	38.137	4.7	98	0.00
50	Dimethylphthalate	40.000	38.576	3.6	99	0.00
51	2,6-Dinitrotoluene	40.000	39.598	1.0	100	0.00
52 C	Acenaphthene	40.000	38.421	3.9	100	0.00
53	3-Nitroaniline	40.000	38.814	3.0	96	0.00
54 P	2,4-Dinitrophenol	40.000	37.789	5.5	104	0.00
55	Dibenzofuran	40.000	38.700	3.2	99	0.00
56 P	4-Nitrophenol	40.000	40.097	-0.2	102	0.00
57	2,4-Dinitrotoluene	40.000	39.635	0.9	100	0.00
58	Fluorene	40.000	38.009	5.0	98	0.00
59	2,3,4,6-Tetrachlorophenol	40.000	39.210	2.0	99	0.00
60	Diethylphthalate	40.000	38.281	4.3	97	0.00
61	4-Chlorophenyl-phenylether	40.000	37.598	6.0	98	0.00
62	4-Nitroaniline	40.000	38.154	4.6	97	0.00
63	Azobenzene	40.000	37.825	5.4	97	0.00
64 I	Phenanthrene-d10	20.000	20.000	0.0	97	0.00
65	4,6-Dinitro-2-methylphenol	40.000	38.509	3.7	106	0.00
66 c	n-Nitrosodiphenylamine	40.000	38.031	4.9	97	0.00
67	4-Bromophenyl-phenylether	40.000	38.111	4.7	99	0.00
68	Hexachlorobenzene	40.000	37.613	6.0	98	0.00
69	Atrazine	40.000	39.060	2.3	95	0.00
70 C	Pentachlorophenol	40.000	39.953	0.1	99	0.00
71	Phenanthrene	40.000	38.167	4.6	97	0.00
72	Anthracene	40.000	38.289	4.3	97	0.00
73	Carbazole	40.000	37.876	5.3	96	0.00
74	Di-n-butylphthalate	40.000	38.645	3.4	96	0.00
75 C	Fluoranthene	40.000	38.308	4.2	97	0.00
76 I	Chrysene-d12	20.000	20.000	0.0	96	0.00
77	Benzidine	40.000	38.335	4.2	86	0.00
78	Pyrene	40.000	39.008	2.5	97	0.00
79 S	Terphenyl-d14	80.000	75.720	5.4	96	0.00
80	Butylbenzylphthalate	40.000	38.994	2.5	96	0.00
81	Benzo(a)anthracene	40.000	37.860	5.4	94	0.00
82	3,3'-Dichlorobenzidine	40.000	37.619	6.0	91	0.00
83	Chrysene	40.000	38.341	4.1	96	0.00
84	Bis(2-ethylhexyl)phthalate	40.000	38.732	3.2	96	0.00
85 c	Di-n-octyl phthalate	40.000	39.307	1.7	96	0.00
86	Indeno(1,2,3-cd)pyrene	40.000	37.002	7.5	93	0.01
87 I	Perylene-d12	20.000	20.000	0.0	93	0.00

Data Path : Z:\SVOASRV\HPCHEM1\BNA G\DATA\BG031020\
Data File : BG044716.D
Acq On : 10 Mar 2020 16:33
Operator : CG/JU
Sample : SSTDICV040
Misc :
ALS Vial : 10 Sample Multiplier: 1

Instrument :
BNA_G
ClientSampleId :
ICVBG031020

Quant Time: Mar 10 18:15:12 2020
Quant Method : Z:\SVOASRV\HPCHEM1\BNA G\METHODS\8270-BG031020.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Tue Mar 10 16:27:39 2020
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
88	Benzo(b)fluoranthene	40.000	37.712	5.7	92	0.00
89	Benzo(k)fluoranthene	40.000	39.050	2.4	98	0.00
90 C	Benzo(a)pyrene	40.000	38.533	3.7	95	0.00
91	Dibenzo(a,h)anthracene	40.000	37.772	5.6	93	0.01
92	Benzo(q,h,i)perylene	40.000	37.396	6.5	93	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 0