

Data Path : Z:\svoasrv\HPCHEM1\BNA_G\Data\BG031122\
 Data File : BG052646.D
 Acq On : 11 Mar 2022 14:16
 Operator : CG/JU
 Sample : SSTDCCC040
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_G
 LabSampleId :
 SSTDCCC040

Quant Time: Mar 12 06:16:06 2022
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_G\Methods\8270-BG030922.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Thu Mar 10 06:37:54 2022
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	20.000	20.000	0.0	113	0.00
2	1,4-Dioxane	40.000	32.520	18.7	95	0.00
3	Pyridine	40.000	34.897	12.8	97	0.00
4	n-Nitrosodimethylamine	40.000	35.386	11.5	99	0.00
5 S	2-Fluorophenol	80.000	77.476	3.2	110	0.00
6	Aniline	40.000	41.304	-3.3	114	0.00
7 S	Phenol-d6	80.000	81.734	-2.2	116	0.00
8	2-Chlorophenol	40.000	40.927	-2.3	115	0.00
9	Benzaldehyde	40.000	37.994	5.0	110	0.00
10 C	Phenol	40.000	39.918	0.2	114	0.00
11	bis(2-Chloroethyl)ether	40.000	39.142	2.1	111	0.00
12	1,3-Dichlorobenzene	40.000	39.628	0.9	114	0.00
13 C	1,4-Dichlorobenzene	40.000	39.515	1.2	113	0.00
14	1,2-Dichlorobenzene	40.000	39.932	0.2	114	0.00
15	Benzyl Alcohol	40.000	40.925	-2.3	115	0.00
16	2,2'-oxybis(1-Chloropropane	40.000	38.889	2.8	111	0.00
17	2-Methylphenol	40.000	41.904	-4.8	118	0.00
18	Hexachloroethane	40.000	39.328	1.7	113	0.00
19 P	n-Nitroso-di-n-propylamine	40.000	40.305	-0.8	115	0.00
20	3+4-Methylphenols	40.000	41.659	-4.1	116	0.01
21 I	Naphthalene-d8	20.000	20.000	0.0	116	0.00
22	Acetophenone	40.000	39.387	1.5	115	0.00
23 S	Nitrobenzene-d5	80.000	78.754	1.6	112	0.00
24	Nitrobenzene	40.000	39.613	1.0	112	0.00
25	Isophorone	40.000	40.627	-1.6	114	0.00
26 C	2-Nitrophenol	40.000	42.693	-6.7	122	0.01
27	2,4-Dimethylphenol	40.000	42.096	-5.2	117	0.00
28	bis(2-Chloroethoxy)methane	40.000	40.360	-0.9	115	0.00
29 C	2,4-Dichlorophenol	40.000	41.914	-4.8	116	0.01
30	1,2,4-Trichlorobenzene	40.000	39.281	1.8	115	0.00
31	Naphthalene	40.000	40.438	-1.1	116	0.00
32	Benzoic acid	40.000	31.163	22.1	138	0.01
33	4-Chloroaniline	40.000	40.606	-1.5	115	0.00
34 C	Hexachlorobutadiene	40.000	39.120	2.2	113	0.00
35	Caprolactam	40.000	44.199	-10.5	124	0.01
36 C	4-Chloro-3-methylphenol	40.000	42.088	-5.2	119	0.00
37	2-Methylnaphthalene	40.000	41.152	-2.9	117	0.00
38	1-Methylnaphthalene	40.000	40.276	-0.7	115	0.00
39 I	Acenaphthene-d10	20.000	20.000	0.0	115	0.00
40	1,2,4,5-Tetrachlorobenzene	40.000	38.769	3.1	113	0.00
41 P	Hexachlorocyclopentadiene	40.000	32.907	17.7	90	0.00
42 S	2,4,6-Tribromophenol	80.000	83.190	-4.0	114	0.00
43 C	2,4,6-Trichlorophenol	40.000	43.154	-7.9	119	0.00
44	2,4,5-Trichlorophenol	40.000	41.972	-4.9	117	0.00
45 S	2-Fluorobiphenyl	80.000	79.119	1.1	114	0.00
46	1,1'-Biphenyl	40.000	39.938	0.2	113	0.00
47	2-Chloronaphthalene	40.000	39.763	0.6	114	0.00

Data Path : Z:\svoasrv\HPCHEM1\BNA_G\Data\BG031122\
 Data File : BG052646.D
 Acq On : 11 Mar 2022 14:16
 Operator : CG/JU
 Sample : SSTDCCC040
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_G
 LabSampleId :
 SSTDCCC040

Quant Time: Mar 12 06:16:06 2022
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_G\Methods\8270-BG030922.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Thu Mar 10 06:37:54 2022
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
48	2-Nitroaniline	40.000	39.692	0.8	110	0.00
49	Acenaphthylene	40.000	40.436	-1.1	115	0.00
50	Dimethylphthalate	40.000	39.872	0.3	114	0.00
51	2,6-Dinitrotoluene	40.000	40.769	-1.9	119	0.00
52 C	Acenaphthene	40.000	40.169	-0.4	115	0.00
53	3-Nitroaniline	40.000	41.053	-2.6	113	0.00
54 P	2,4-Dinitrophenol	40.000	41.917	-4.8	126	0.00
55	Dibenzofuran	40.000	39.515	1.2	114	0.00
56 P	4-Nitrophenol	40.000	42.378	-5.9	112	0.00
57	2,4-Dinitrotoluene	40.000	40.767	-1.9	116	0.00
58	Fluorene	40.000	39.072	2.3	112	0.00
59	2,3,4,6-Tetrachlorophenol	40.000	42.992	-7.5	117	0.00
60	Diethylphthalate	40.000	39.685	0.8	113	0.00
61	4-Chlorophenyl-phenylether	40.000	38.940	2.7	112	0.00
62	4-Nitroaniline	40.000	38.746	3.1	109	0.00
63	Azobenzene	40.000	38.563	3.6	110	0.00
64 I	Phenanthrene-d10	20.000	20.000	0.0	112	0.00
65	4,6-Dinitro-2-methylphenol	40.000	44.322	-10.8	117	0.01
66 c	n-Nitrosodiphenylamine	40.000	41.454	-3.6	113	0.00
67	4-Bromophenyl-phenylether	40.000	40.627	-1.6	114	0.00
68	Hexachlorobenzene	40.000	39.940	0.2	112	0.00
69	Atrazine	40.000	43.379	-8.4	116	0.00
70 C	Pentachlorophenol	40.000	41.376	-3.4	111	0.00
71	Phenanthrene	40.000	40.134	-0.3	111	0.00
72	Anthracene	40.000	40.333	-0.8	110	0.00
73	Carbazole	40.000	40.245	-0.6	112	0.00
74	Di-n-butylphthalate	40.000	41.077	-2.7	112	0.00
75 C	Fluoranthene	40.000	38.396	4.0	105	0.00
76 I	Chrysene-d12	20.000	20.000	0.0	93	0.00
77	Benzydine	40.000	41.995	-5.0	89	0.00
78	Pyrene	40.000	45.328	-13.3	105	0.00
79 S	Terphenyl-d14	80.000	87.890	-9.9	102	0.00
80	Butylbenzylphthalate	40.000	43.659	-9.1	98	0.00
81	Benzo(a)anthracene	40.000	40.425	-1.1	92	0.00
82	3,3'-Dichlorobenzidine	40.000	41.393	-3.5	91	0.00
83	Chrysene	40.000	40.848	-2.1	94	0.00
84	Bis(2-ethylhexyl)phthalate	40.000	41.613	-4.0	93	0.00
85 c	Di-n-octyl phthalate	40.000	42.238	-5.6	94	0.00
86 I	Perylene-d12	20.000	20.000	0.0	89	0.01
87	Indeno(1,2,3-cd)pyrene	40.000	39.753	0.6	88	0.03
88	Benzo(b)fluoranthene	40.000	40.415	-1.0	89	0.01
89	Benzo(k)fluoranthene	40.000	39.764	0.6	89	0.00
90 C	Benzo(a)pyrene	40.000	39.884	0.3	88	0.01
91	Dibenzo(a,h)anthracene	40.000	39.531	1.2	88	0.04
92	Benzo(g,h,i)perylene	40.000	41.487	-3.7	91	0.03

Data Path : Z:\svoasrv\HPCHEM1\BNA_G\Data\BG031122\
Data File : BG052646.D
Acq On : 11 Mar 2022 14:16
Operator : CG/JU
Sample : SSTDCCC040
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_G
LabSampleId :
SSTDCCC040

Quant Time: Mar 12 06:16:06 2022
Quant Method : Z:\svoasrv\HPCHEM1\BNA_G\Methods\8270-BG030922.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Thu Mar 10 06:37:54 2022
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

Compound	Amount	Calc.	%Dev	Area%	Dev(min)
----------	--------	-------	------	-------	----------

(#) = Out of Range

SPCC's out = 0 CCC's out = 0