

Data Path : Z:\SVOASRV\HPCHEM1\BNA G\DATA\BG031519\
 Data File : BG039965.D
 Acq On : 16 Mar 2019 7:11
 Operator : JU/SJ
 Sample : K1892-01
 Misc :
 ALS Vial : 28 Sample Multiplier: 1

Instrument :
 BNA_G
 ClientSampleId :
 GARAGE-FLOOR

Manual Integrations
 APPROVED

Sohil
 3/18/2019 3:02:50 PM

Quant Time: Mar 18 01:15:59 2019
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA G\METHODS\8270-BG030419.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Mon Mar 04 14:38:15 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	8.14	152	44923	20.00	ng	-0.03
21) Naphthalene-d8	10.96	136	187994	20.00	ng	-0.03
39) Acenaphthene-d10	14.77	164	125173	20.00	ng	-0.02
64) Phenanthrene-d10	17.52	188	288704	20.00	ng	-0.01
76) Chrysene-d12	21.80	240	290310	20.00	ng	-0.03
87) Perylene-d12	25.17	264	297984	20.00	ng	-0.03
System Monitoring Compounds						
5) 2-Fluorophenol	5.70	112	336184	127.67	ng	-0.01
7) Phenol-d6	7.30	99	446163	113.63	ng	-0.01
23) Nitrobenzene-d5	9.33	82	352216	101.33	ng	-0.02
42) 2,4,6-Tribromophenol	16.26	330	223681	153.31	ng	-0.01
45) 2-Fluorobiphenyl	13.40	172	807579	91.86	ng	-0.02
79) Terphenyl-d14	20.11	244	1185402	89.90	ng	-0.01
Target Compounds						
20) 3+4-Methylphenols	8.91	107	10141	2.692	ng	92
32) Benzoic acid	10.34	122	17930m	14.824	ng	
70) Pentachlorophenol	17.16	266	99425	46.998	ng	97

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\SVOASRV\HPCHEM1\BNA G\DATA\BG031519\
Data File : BG039965.D
Acq On : 16 Mar 2019 7:11
Operator : JU/SJ
Sample : K1892-01
Misc :
ALS Vial : 28 Sample Multiplier: 1

Instrument :
BNA_G
ClientSampleId :
GARAGE-FLOOR

Manual Integrations
APPROVED

Sohil
3/18/2019 3:02:50 PM

Quant Time: Mar 18 01:15:59 2019
Quant Method : Z:\SVOASRV\HPCHEM1\BNA G\METHODS\8270-BG030419.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Mon Mar 04 14:38:15 2019
Response via : Initial Calibration

