

Data Path : Z:\svoasrv\HPCHEM1\BNA_G\Data\BG031621\
 Data File : BG048390.D
 Acq On : 16 Mar 2021 14:09
 Operator : CG/JU
 Sample : SSTDICC060
 Misc :
 ALS Vial : 7 Sample Multiplier: 1

Instrument :
 BNA_G
 ClientSampleId :
 SSTDICC060

Manual Integrations
 APPROVED

mohammad
 3/17/2021 10:51:21 AM

Quant Time: Mar 16 14:53:32 2021
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_G\METHODS\8270-BG031621.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Tue Mar 16 14:17:30 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	8.116	152	59929	20.00	ng	# 0.00
21) Naphthalene-d8	10.936	136	209099	20.00	ng	# 0.00
39) Acenaphthene-d10	14.755	164	148921	20.00	ng	0.00
64) Phenanthrene-d10	17.505	188	368354	20.00	ng	# 0.00
76) Chrysene-d12	21.800	240	374683	20.00	ng	# 0.00
86) Perylene-d12	25.137	264	385090	20.00	ng	0.00
System Monitoring Compounds						
5) 2-Fluorophenol	5.660	112	395260	112.27	ng	0.00
7) Phenol-d6	7.258	99	583582	109.10	ng	0.00
23) Nitrobenzene-d5	9.285	82	582731	117.89	ng	0.00
42) 2,4,6-Tribromophenol	16.242	330	237950	99.90	ng	0.00
45) 2-Fluorobiphenyl	13.381	172	1188913	111.92	ng	0.00
79) Terphenyl-d14	20.114	244	2094346	102.14	ng	0.00
Target Compounds						Qvalue
2) 1,4-Dioxane	3.539	88	90862	56.32	ng	# 94
3) Pyridine	3.945	79	257494	57.73	ng	91
4) n-Nitrosodimethylamine	3.856	42	147306	62.21	ng	# 63
6) Aniline	7.435	93	345555	54.07	ng	# 90
8) 2-Chlorophenol	7.676	128	200142	51.76	ng	84
9) Benzaldehyde	7.247	77	158909	41.94	ng	# 79
10) Phenol	7.288	94	293865	54.33	ng	# 68
11) bis(2-Chloroethyl)ether	7.529	93	213790	50.74	ng	93
12) 1,3-Dichlorobenzene	8.005	146	243239	54.45	ng	# 88
13) 1,4-Dichlorobenzene	8.151	146	246666	54.75	ng	# 93
14) 1,2-Dichlorobenzene	8.475	146	231570m	54.05	ng	
15) Benzyl Alcohol	8.351	79	260367	55.23	ng	# 81
16) 2,2'-oxybis(1-Chloropr...	8.639	45	253889	46.47	ng	89
17) 2-Methylphenol	8.545	107	185971	52.73	ng	# 82
18) Hexachloroethane	9.209	117	92288	53.73	ng	96
19) n-Nitroso-di-n-propyla...	8.927	70	221252	53.59	ng	# 95
20) 3+4-Methylphenols	8.880	107	259835	52.87	ng	84
22) Acetophenone	8.939	105	339137	56.50	ng	# 90
24) Nitrobenzene	9.327	77	312475	59.24	ng	# 83
25) Isophorone	9.849	82	582407	59.41	ng	# 89
26) 2-Nitrophenol	10.037	139	103305	47.35	ng	# 61
27) 2,4-Dimethylphenol	10.090	122	187666	59.76	ng	# 85
28) bis(2-Chloroethoxy)met...	10.331	93	256475	53.52	ng	96
29) 2,4-Dichlorophenol	10.572	162	213633	53.19	ng	92
30) 1,2,4-Trichlorobenzene	10.795	180	255424	55.94	ng	93
31) Naphthalene	10.989	128	645182	55.89	ng	98
32) Benzoic acid	10.225	122	134986	70.17	ng	# 87
33) 4-Chloroaniline	11.089	127	280415	56.00	ng	# 88
34) Hexachlorobutadiene	11.271	225	204130	59.77	ng	97
35) Caprolactam	11.871	113	72361	54.39	ng	91
36) 4-Chloro-3-methylphenol	12.200	107	250072	56.32	ng	85
37) 2-Methylnaphthalene	12.587	142	490165	54.62	ng	# 91
38) 1-Methylnaphthalene	12.805	142	463235	54.56	ng	# 99
40) 1,2,4,5-Tetrachloroben...	12.952	216	318234	59.83	ng	# 98
41) Hexachlorocyclopentadiene	12.934	237	192448	59.33	ng	97
43) 2,4,6-Trichlorophenol	13.181	196	209923	54.92	ng	95

Data Path : Z:\svoasrv\HPCHEM1\BNA_G\Data\BG031621\
 Data File : BG048390.D
 Acq On : 16 Mar 2021 14:09
 Operator : CG/JU
 Sample : SSTDICC060
 Misc :
 ALS Vial : 7 Sample Multiplier: 1

Instrument :
 BNA_G
 ClientSampleId :
 SSTDICC060


Manual Integrations
 APPROVED

mohammad
 3/17/2021 10:51:21 AM

Quant Time: Mar 16 14:53:32 2021
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_G\METHODS\8270-BG031621.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Tue Mar 16 14:17:30 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)	
44) 2,4,5-Trichlorophenol	13.251	196	223646	54.25	ng	#	91
46) 1,1'-Biphenyl	13.592	154	612370	57.40	ng		99
47) 2-Chloronaphthalene	13.633	162	495800	54.61	ng	#	93
48) 2-Nitroaniline	13.827	65	177626	54.51	ng	#	70
49) Acenaphthylene	14.479	152	787903	56.10	ng		98
50) Dimethylphthalate	14.203	163	673481	54.09	ng		99
51) 2,6-Dinitrotoluene	14.321	165	143206	51.18	ng	#	50
52) Acenaphthene	14.820	154	529419	55.68	ng		100
53) 3-Nitroaniline	14.656	138	140409	51.09	ng	#	77
54) 2,4-Dinitrophenol	14.849	184	78991	43.36	ng	#	72
55) Dibenzofuran	15.155	168	788819	53.05	ng		92
56) 4-Nitrophenol	14.943	139	119535	53.06	ng	#	38
57) 2,4-Dinitrotoluene	15.108	165	198599	49.13	ng	#	65
58) Fluorene	15.801	166	657716	55.76	ng		96
59) 2,3,4,6-Tetrachlorophenol	15.372	232	215025	52.08	ng	#	94
60) Diethylphthalate	15.566	149	660720	51.90	ng		95
61) 4-Chlorophenyl-phenyle...	15.795	204	386641	54.44	ng		96
62) 4-Nitroaniline	15.819	138	149096	52.16	ng	#	41
63) Azobenzene	16.089	77	744613	58.61	ng		86
65) 4,6-Dinitro-2-methylph...	15.872	198	115181	43.08	ng		91
66) n-Nitrosodiphenylamine	16.007	169	610756	57.33	ng		98
67) 4-Bromophenyl-phenylether	16.689	248	242311	51.31	ng		94
68) Hexachlorobenzene	16.806	284	271344	55.70	ng	#	95
69) Atrazine	16.953	200	228683	51.62	ng		99
70) Pentachlorophenol	17.147	266	180886	56.46	ng		98
71) Phenanthrene	17.546	178	1072449	53.61	ng		99
72) Anthracene	17.640	178	1057990	53.27	ng		100
73) Carbazole	17.905	167	1000499	53.26	ng		97
74) Di-n-butylphthalate	18.463	149	1101875	52.93	ng	#	97
75) Fluoranthene	19.556	202	1398908	53.72	ng		97
77) Benzidine	19.732	184	528654	43.18	ng		98
78) Pyrene	19.920	202	1405096	53.69	ng		99
80) Butylbenzylphthalate	20.801	149	492476	53.74	ng	#	83
81) Benzo(a)anthracene	21.777	228	1434865	54.90	ng		99
82) 3,3'-Dichlorobenzidine	21.689	252	509178	54.39	ng	#	98
83) Chrysene	21.847	228	1367340	54.76	ng		99
84) Bis(2-ethylhexyl)phtha...	21.683	149	720775	55.99	ng		96
85) Di-n-octyl phthalate	22.946	149	1224685	55.87	ng	#	94
87) Indeno(1,2,3-cd)pyrene	28.974	276	1726964	59.18	ng	#	84
88) Benzo(b)fluoranthene	24.080	252	1526268	57.32	ng	#	96
89) Benzo(k)fluoranthene	24.150	252	1460256	57.49	ng	#	97
90) Benzo(a)pyrene	24.985	252	1412089	57.18	ng	#	96
91) Dibenzo(a,h)anthracene	29.050	278	1458797	58.61	ng	#	93
92) Benzo(g,h,i)perylene	30.173	276	1417960	58.62	ng	#	93

(#) = qualifier out of range (m) = manual integration (+) = signals summed







#1
1,4-Dichlorobenzene-d4
Concen: 20.00 ng
RT: 8.116 min Scan# 813
Delta R.T. 0.001 min
Lab File: BG048390.D
Acq: 16 Mar 2021 14:09

Instrument :
BNA_G
ClientSampleId :
SSTDIC060

Tgt Ion:152 Resp: 59929
Ion Ratio Lower Upper
152 100
150 152.8 125.0 187.6
115 63.0 41.8 62.6#




Manual Integrations
APPROVED


mohammad
3/17/2021 10:51:21 AM



#2
1,4-Dioxane
Concen: 56.32 ng
RT: 3.539 min Scan# 34
Delta R.T. 0.001 min
Lab File: BG048390.D
Acq: 16 Mar 2021 14:09

Tgt Ion: 88 Resp: 90862
Ion Ratio Lower Upper
88 100
58 75.4 61.6 92.4
43 34.9 21.3 31.9#









#3
Pyridine
Concen: 57.73 ng
RT: 3.945 min Scan# 103
Delta R.T. 0.001 min
Lab File: BG048390.D
Acq: 16 Mar 2021 14:09

Tgt Ion: 79 Resp: 257494
Ion Ratio Lower Upper
79 100
52 68.7 61.3 91.9
51 40.7 28.4 42.6

Instrument :
BNA_G
ClientSampleId :
SSTDIC060




Manual Integrations
APPROVED


mohammad
3/17/2021 10:51:21 AM



#4
n-Nitrosodimethylamine
Concen: 62.21 ng
RT: 3.856 min Scan# 88
Delta R.T. 0.001 min
Lab File: BG048390.D
Acq: 16 Mar 2021 14:09

Tgt Ion: 42 Resp: 147306
Ion Ratio Lower Upper
42 100
74 102.1 120.4 180.6#
44 4.4 6.2 9.2#









#5
2-Fluorophenol
Concen: 112.27 ng
RT: 5.660 min Scan# 395
Delta R.T. 0.001 min
Lab File: BG048390.D
Acq: 16 Mar 2021 14:09

Tgt Ion: 112 Resp: 395260
Ion Ratio Lower Upper
112 100
64 71.8 46.7 70.1#
63 43.4 22.8 34.2#

Instrument :
BNA_G
ClientSampleId :
SSTDIC060




Manual Integrations
APPROVED


mohammad
3/17/2021 10:51:21 AM



#6
Aniline
Concen: 54.07 ng
RT: 7.435 min Scan# 697
Delta R.T. 0.001 min
Lab File: BG048390.D
Acq: 16 Mar 2021 14:09


Tgt Ion: 93 Resp: 345555
Ion Ratio Lower Upper
93 100
66 42.4 29.5 44.3
65 24.5 15.5 23.3#





#7
Phenol-d6
Concen: 109.10 ng
RT: 7.258 min Scan# 667
Delta R.T. 0.001 min
Lab File: BG048390.D
Acq: 16 Mar 2021 14:09




Instrument :
BNA_G
ClientSampleId :
SSTDIC060




Tgt Ion: 99 Resp: 583582
Ion Ratio Lower Upper
99 100
42 23.9 12.8 19.2#
71 50.3 24.6 36.8#

Manual Integrations
APPROVED



mohammad
3/17/2021 10:51:21 AM




#8
2-Chlorophenol
Concen: 51.76 ng
RT: 7.676 min Scan# 738
Delta R.T. 0.001 min
Lab File: BG048390.D
Acq: 16 Mar 2021 14:09



Tgt Ion: 128 Resp: 200142
Ion Ratio Lower Upper
128 100
130 34.1 12.6 52.6
64 61.9 24.8 64.8









#9
Benzaldehyde
Concen: 41.94 ng
RT: 7.247 min Scan# 665
Delta R.T. 0.001 min
Lab File: BG048390.D
Acq: 16 Mar 2021 14:09

Tgt Ion: 77 Resp: 158909
Ion Ratio Lower Upper
77 100
105 83.5 85.9 125.9#
106 79.6 80.9 120.9#

Instrument :
BNA_G
ClientSampleId :
SSTDIC060




Manual Integrations
APPROVED


mohammad
3/17/2021 10:51:21 AM



#10
Phenol
Concen: 54.33 ng
RT: 7.288 min Scan# 672
Delta R.T. 0.007 min
Lab File: BG048390.D
Acq: 16 Mar 2021 14:09


Tgt Ion: 94 Resp: 293865
Ion Ratio Lower Upper
94 100
65 37.9 6.6 46.6
66 56.7 13.9 53.9#





#11
bis(2-Chloroethyl)ether
Concen: 50.74 ng
RT: 7.529 min Scan# 713
Delta R.T. 0.001 min
Lab File: BG048390.D
Acq: 16 Mar 2021 14:09




Instrument :
BNA_G
ClientSampleId :
SSTDIC060




Tgt Ion: 93 Resp: 213790
Ion Ratio Lower Upper
93 100
63 84.6 56.4 96.4
95 33.6 13.8 53.8

Manual Integrations
APPROVED



mohammad
3/17/2021 10:51:21 AM




#12
1,3-Dichlorobenzene
Concen: 54.45 ng
RT: 8.005 min Scan# 794
Delta R.T. 0.001 min
Lab File: BG048390.D
Acq: 16 Mar 2021 14:09




Tgt Ion: 146 Resp: 243239
Ion Ratio Lower Upper
146 100
148 64.5 50.1 75.1
75 42.9 20.1 30.1#





#13
1,4-Dichlorobenzene
Concen: 54.75 ng
RT: 8.151 min Scan# 819
Delta R.T. 0.001 min
Lab File: BG048390.D
Acq: 16 Mar 2021 14:09




Instrument :
BNA_G
ClientSampleId :
SSTDIC060




Tgt Ion:146 Resp: 246666
Ion Ratio Lower Upper
146 100
148 65.4 51.1 76.7
111 46.2 29.7 44.5#

Manual Integrations
APPROVED



mohammad
3/17/2021 10:51:21 AM




#14
1,2-Dichlorobenzene
Concen: 54.05 ng m
RT: 8.475 min Scan# 874
Delta R.T. 0.001 min
Lab File: BG048390.D
Acq: 16 Mar 2021 14:09




Tgt Ion:146 Resp: 231570
Ion Ratio Lower Upper
146 100
148 65.3 53.0 79.6
111 46.5 30.7 46.1#





#15
Benzyl Alcohol
Concen: 55.23 ng
RT: 8.351 min Scan# 853
Delta R.T. 0.007 min
Lab File: BG048390.D
Acq: 16 Mar 2021 14:09




Instrument :
BNA_G
ClientSampleId :
SSTDIC060




Tgt Ion: 79 Resp: 260367
Ion Ratio Lower Upper
79 100
108 60.8 65.5 98.3#
77 72.4 50.1 75.1

Manual Integrations
APPROVED



mohammad
3/17/2021 10:51:21 AM




#16
2,2'-oxybis(1-Chloropropane)
Concen: 46.47 ng
RT: 8.639 min Scan# 902
Delta R.T. -0.005 min
Lab File: BG048390.D
Acq: 16 Mar 2021 14:09




Tgt Ion: 45 Resp: 253889
Ion Ratio Lower Upper
45 100
77 19.3 0.0 33.4
79 13.3 0.0 30.8





#17
2-Methylphenol
Concen: 52.73 ng
RT: 8.545 min Scan# 886
Delta R.T. 0.001 min
Lab File: BG048390.D
Acq: 16 Mar 2021 14:09




Instrument :
BNA_G
ClientSampleId :
SSTDIC060



Tgt Ion	Ratio	Lower	Upper
107	100		
108	111.3	87.2	130.8
77	66.7	33.8	50.8#
79	65.3	32.7	49.1#




Manual Integrations
APPROVED


mohammad
3/17/2021 10:51:21 AM



#18
Hexachloroethane
Concen: 53.73 ng
RT: 9.209 min Scan# 999
Delta R.T. 0.001 min
Lab File: BG048390.D
Acq: 16 Mar 2021 14:09

Tgt Ion	Ratio	Lower	Upper
117	100		
119	92.3	77.5	116.3
201	127.8	104.9	157.3









#19
n-Nitroso-di-n-propylamine
Concen: 53.59 ng
RT: 8.927 min Scan# 951
Delta R.T. 0.007 min
Lab File: BG048390.D
Acq: 16 Mar 2021 14:09

Tgt Ion: 70 Resp: 221252
Ion Ratio Lower Upper
70 100
42 56.1 45.6 68.4
101 8.1 8.4 12.6#
130 15.4 17.8 26.8#

Instrument :
BNA_G
ClientSampleId :
SSTDICC060




Manual Integrations
APPROVED


mohammad
3/17/2021 10:51:21 AM



#20
3+4-Methylphenols
Concen: 52.87 ng
RT: 8.880 min Scan# 943
Delta R.T. 0.007 min
Lab File: BG048390.D
Acq: 16 Mar 2021 14:09


Tgt Ion: 107 Resp: 259835
Ion Ratio Lower Upper
107 100
108 82.1 70.6 110.6
77 45.7 10.2 50.2
79 41.1 6.8 46.8





#21
Naphthalene-d8
Concen: 20.00 ng
RT: 10.936 min Scan# 1293
Delta R.T. 0.001 min
Lab File: BG048390.D
Acq: 16 Mar 2021 14:09




Instrument :
BNA_G
ClientSampleId :
SSTDIC060




Tgt Ion:136 Resp: 209099
Ion Ratio Lower Upper
136 100
137 10.3 9.1 13.7
54 8.7 6.2 9.4
68 4.7 5.0 7.4#

Manual Integrations
APPROVED



mohammad
3/17/2021 10:51:21 AM




#22
Acetophenone
Concen: 56.50 ng
RT: 8.939 min Scan# 953
Delta R.T. 0.001 min
Lab File: BG048390.D
Acq: 16 Mar 2021 14:09



Tgt Ion:105 Resp: 339137
Ion Ratio Lower Upper
105 100
71 5.7 2.7 4.1#
51 34.3 21.0 31.6#
120 22.1 19.3 28.9





#23

Nitrobenzene-d5

Concen: 117.89 ng

RT: 9.285 min Scan# 1012

Delta R.T. 0.001 min

Lab File: BG048390.D

Acq: 16 Mar 2021 14:09

Tgt Ion:	82	Resp:	582731
Ion	Ratio	Lower	Upper
82	100		
128	32.5	37.0	55.6#
54	52.4	39.0	58.4

Instrument :





BNA_G

ClientSampleId :

SSTDIC060

Manual Integrations
APPROVED

mohammad
3/17/2021 10:51:21 AM



#24

Nitrobenzene

Concen: 59.24 ng




RT: 9.327 min Scan# 1019


Delta R.T. 0.001 min

Lab File: BG048390.D

Acq: 16 Mar 2021 14:09

Tgt Ion:	77	Resp:	312475
Ion	Ratio	Lower	Upper
77	100		
123	34.9	39.7	59.5#
65	14.0	10.4	15.6









#25
Isophorone
Concen: 59.41 ng
RT: 9.849 min Scan# 1108
Delta R.T. 0.001 min
Lab File: BG048390.D
Acq: 16 Mar 2021 14:09

Tgt Ion: 82 Resp: 582407
Ion Ratio Lower Upper
82 100
95 8.6 5.6 8.4#
138 13.8 15.8 23.6#

Instrument :
BNA_G
ClientSampleId :
SSTDICC060




Manual Integrations
APPROVED


mohammad
3/17/2021 10:51:21 AM



#26
2-Nitrophenol
Concen: 47.35 ng
RT: 10.037 min Scan# 1140
Delta R.T. 0.001 min
Lab File: BG048390.D
Acq: 16 Mar 2021 14:09


Tgt Ion:139 Resp: 103305
Ion Ratio Lower Upper
139 100
109 42.6 20.7 31.1#
65 72.5 36.1 54.1#





#27
2,4-Dimethylphenol
Concen: 59.76 ng
RT: 10.090 min Scan# 1149
Delta R.T. 0.007 min
Lab File: BG048390.D
Acq: 16 Mar 2021 14:09




Instrument :
BNA_G
ClientSampleId :
SSTDICC060



Tgt Ion:122 Resp: 187666
Ion Ratio Lower Upper
122 100
107 135.0 89.2 133.8#
121 60.8 46.9 70.3




Manual Integrations
APPROVED


mohammad
3/17/2021 10:51:21 AM



#28
bis(2-Chloroethoxy)methane
Concen: 53.52 ng
RT: 10.331 min Scan# 1190
Delta R.T. 0.001 min
Lab File: BG048390.D
Acq: 16 Mar 2021 14:09

Tgt Ion: 93 Resp: 256475
Ion Ratio Lower Upper
93 100
95 35.1 26.5 39.7
123 9.4 9.4 14.0









#29
2,4-Dichlorophenol
Concen: 53.19 ng
RT: 10.572 min Scan# 1231
Delta R.T. 0.001 min
Lab File: BG048390.D
Acq: 16 Mar 2021 14:09

Tgt Ion:162 Resp: 213633
Ion Ratio Lower Upper
162 100
164 67.7 45.6 85.6
98 41.0 11.0 51.0

Instrument :
BNA_G
ClientSampleId :
SSTDIC060




Manual Integrations
APPROVED


mohammad
3/17/2021 10:51:21 AM



#30
1,2,4-Trichlorobenzene
Concen: 55.94 ng
RT: 10.795 min Scan# 1269
Delta R.T. 0.001 min
Lab File: BG048390.D
Acq: 16 Mar 2021 14:09

Tgt Ion:180 Resp: 255424
Ion Ratio Lower Upper
180 100
182 104.8 76.9 115.3
145 27.6 21.4 32.0









#31
Naphthalene
Concen: 55.89 ng
RT: 10.989 min Scan# 1302
Delta R.T. 0.001 min
Lab File: BG048390.D
Acq: 16 Mar 2021 14:09

Tgt Ion:128 Resp: 645182
Ion Ratio Lower Upper
128 100
129 11.8 9.0 13.6
127 14.9 11.0 16.4

Instrument :
BNA_G
ClientSampleId :
SSTDIC060




Manual Integrations
APPROVED


mohammad
3/17/2021 10:51:21 AM



#32
Benzoic acid
Concen: 70.17 ng
RT: 10.225 min Scan# 1172
Delta R.T. 0.013 min
Lab File: BG048390.D
Acq: 16 Mar 2021 14:09


Tgt Ion:122 Resp: 134986
Ion Ratio Lower Upper
122 100
105 124.2 103.7 143.7
77 110.6 63.0 103.0#





#33
4-Chloroaniline
Concen: 56.00 ng
RT: 11.089 min Scan# 1319
Delta R.T. 0.007 min
Lab File: BG048390.D
Acq: 16 Mar 2021 14:09




Instrument :
BNA_G
ClientSampleId :
SSTDIC060




Tgt Ion:127 Resp: 280415
Ion Ratio Lower Upper
127 100
129 32.2 26.8 40.2
65 41.6 23.8 35.6#
92 23.2 14.5 21.7#

Manual Integrations
APPROVED



mohammad
3/17/2021 10:51:21 AM




#34
Hexachlorobutadiene
Concen: 59.77 ng
RT: 11.271 min Scan# 1350
Delta R.T. 0.001 min
Lab File: BG048390.D
Acq: 16 Mar 2021 14:09




Tgt Ion:225 Resp: 204130
Ion Ratio Lower Upper
225 100
223 64.5 49.5 74.3
227 66.9 52.0 78.0









#35
Caprolactam
Concen: 54.39 ng
RT: 11.871 min Scan# 1452
Delta R.T. 0.013 min
Lab File: BG048390.D
Acq: 16 Mar 2021 14:09





Tgt Ion:113 Resp: 72361
Ion Ratio Lower Upper
113 100
55 175.8 149.2 189.2
56 144.3 107.2 147.2



#36
4-Chloro-3-methylphenol
Concen: 56.32 ng
RT: 12.200 min Scan# 1508
Delta R.T. 0.007 min
Lab File: BG048390.D
Acq: 16 Mar 2021 14:09




Tgt Ion:107 Resp: 250072
Ion Ratio Lower Upper
107 100
144 23.6 22.5 33.7
142 72.6 70.8 106.2



Instrument :
BNA_G
ClientSampleId :
SSTDIC060

Manual Integrations
APPROVED

mohammad
3/17/2021 10:51:21 AM







#37
2-Methylnaphthalene
Concen: 54.62 ng
RT: 12.587 min Scan# 1574
Delta R.T. 0.001 min
Lab File: BG048390.D
Acq: 16 Mar 2021 14:09

Tgt Ion:142 Resp: 490165
Ion Ratio Lower Upper
142 100
141 91.0 71.6 107.4
115 48.0 25.4 38.2#

Instrument :
BNA_G
ClientSampleId :
SSTDICC060




Manual Integrations
APPROVED


mohammad
3/17/2021 10:51:21 AM



#38
1-Methylnaphthalene
Concen: 54.56 ng
RT: 12.805 min Scan# 1611
Delta R.T. 0.001 min
Lab File: BG048390.D
Acq: 16 Mar 2021 14:09


Tgt Ion:142 Resp: 463235
Ion Ratio Lower Upper
142 100
141 93.8 75.4 113.2
116 5.3 2.9 4.3#





#39
Acenaphthene-d10
Concen: 20.00 ng
RT: 14.755 min Scan# 1943
Delta R.T. 0.001 min
Lab File: BG048390.D
Acq: 16 Mar 2021 14:09




Instrument :
BNA_G
ClientSampleId :
SSTDICC060




Tgt Ion:164 Resp: 148921
Ion Ratio Lower Upper
164 100
162 102.9 82.3 123.5
160 44.6 35.3 52.9

Manual Integrations
APPROVED



mohammad
3/17/2021 10:51:21 AM




#40
1,2,4,5-Tetrachlorobenzene
Concen: 59.83 ng
RT: 12.952 min Scan# 1636
Delta R.T. 0.001 min
Lab File: BG048390.D
Acq: 16 Mar 2021 14:09




Tgt Ion:216 Resp: 318234
Ion Ratio Lower Upper
216 100
214 78.4 61.9 92.9
179 19.2 14.8 22.2
108 17.0 10.2 15.4#





#41
Hexachlorocyclopentadiene
Concen: 59.33 ng
RT: 12.934 min Scan# 1633
Delta R.T. 0.001 min
Lab File: BG048390.D
Acq: 16 Mar 2021 14:09




Instrument :
BNA_G
ClientSampleId :
SSTDIC060




Tgt Ion:237 Resp: 192448
Ion Ratio Lower Upper
237 100
235 65.8 43.1 83.1
272 12.8 0.0 33.1

Manual Integrations
APPROVED



mohammad
3/17/2021 10:51:21 AM




#42
2,4,6-Tribromophenol
Concen: 99.90 ng
RT: 16.242 min Scan# 2196
Delta R.T. 0.001 min
Lab File: BG048390.D
Acq: 16 Mar 2021 14:09



Tgt Ion:330 Resp: 237950
Ion Ratio Lower Upper
330 100
332 97.3 76.5 114.7
141 36.8 26.0 39.0









#43
2,4,6-Trichlorophenol
Concen: 54.92 ng
RT: 13.181 min Scan# 1675
Delta R.T. 0.001 min
Lab File: BG048390.D
Acq: 16 Mar 2021 14:09

Tgt Ion:196 Resp: 209923
Ion Ratio Lower Upper
196 100
198 98.3 74.9 112.3
200 29.2 25.3 37.9

Instrument :
BNA_G
ClientSampleId :
SSTDIC060




Manual Integrations
APPROVED


mohammad
3/17/2021 10:51:21 AM



#44
2,4,5-Trichlorophenol
Concen: 54.25 ng
RT: 13.251 min Scan# 1687
Delta R.T. 0.001 min
Lab File: BG048390.D
Acq: 16 Mar 2021 14:09


Tgt Ion:196 Resp: 223646
Ion Ratio Lower Upper
196 100
198 93.0 77.6 116.4
97 52.0 28.6 43.0#
132 22.0 17.8 26.8





#45
2-Fluorobiphenyl
Concen: 111.92 ng
RT: 13.381 min Scan# 1709
Delta R.T. 0.001 min
Lab File: BG048390.D
Acq: 16 Mar 2021 14:09




Instrument :
BNA_G
ClientSampleId :
SSTDIC060




Tgt Ion:172 Resp: 1188913
Ion Ratio Lower Upper
172 100
171 38.4 31.3 46.9
170 24.9 20.8 31.2

Manual Integrations
APPROVED



mohammad
3/17/2021 10:51:21 AM




#46
1,1'-Biphenyl
Concen: 57.40 ng
RT: 13.592 min Scan# 1745
Delta R.T. 0.001 min
Lab File: BG048390.D
Acq: 16 Mar 2021 14:09



Tgt Ion:154 Resp: 612370
Ion Ratio Lower Upper
154 100
153 43.4 22.6 62.6
76 11.9 0.0 32.6









#47
2-Chloronaphthalene
Concen: 54.61 ng
RT: 13.633 min Scan# 1752
Delta R.T. 0.001 min
Lab File: BG048390.D
Acq: 16 Mar 2021 14:09

Tgt Ion:162 Resp: 495800
Ion Ratio Lower Upper
162 100
127 41.9 27.8 41.6#
164 34.2 27.0 40.4

Instrument :
BNA_G
ClientSampleId :
SSTDIC060




Manual Integrations
APPROVED


mohammad
3/17/2021 10:51:21 AM



#48
2-Nitroaniline
Concen: 54.51 ng
RT: 13.827 min Scan# 1785
Delta R.T. 0.001 min
Lab File: BG048390.D
Acq: 16 Mar 2021 14:09


Tgt Ion: 65 Resp: 177626
Ion Ratio Lower Upper
65 100
92 56.8 58.6 88.0#
138 81.3 98.2 147.4#





#49
Acenaphthylene
Concen: 56.10 ng
RT: 14.479 min Scan# 1896
Delta R.T. 0.001 min
Lab File: BG048390.D
Acq: 16 Mar 2021 14:09




Instrument :
BNA_G
ClientSampleId :
SSTDIC060




Tgt Ion:152 Resp: 787903
Ion Ratio Lower Upper
152 100
151 20.4 17.4 26.0
153 14.2 11.5 17.3

Manual Integrations
APPROVED



mohammad
3/17/2021 10:51:21 AM




#50
Dimethylphthalate
Concen: 54.09 ng
RT: 14.203 min Scan# 1849
Delta R.T. 0.001 min
Lab File: BG048390.D
Acq: 16 Mar 2021 14:09




Tgt Ion:163 Resp: 673481
Ion Ratio Lower Upper
163 100
194 3.9 3.8 5.8
164 10.9 8.8 13.2





#51
2,6-Dinitrotoluene
Concen: 51.18 ng
RT: 14.321 min Scan# 1869
Delta R.T. 0.001 min
Lab File: BG048390.D
Acq: 16 Mar 2021 14:09




Instrument :
BNA_G
ClientSampleId :
SSTDIC060




Tgt Ion:165 Resp: 143206
Ion Ratio Lower Upper
165 100
63 77.2 33.7 50.5#
89 75.8 36.6 54.8#

Manual Integrations
APPROVED



mohammad
3/17/2021 10:51:21 AM




#52
Acenaphthene
Concen: 55.68 ng
RT: 14.820 min Scan# 1954
Delta R.T. 0.001 min
Lab File: BG048390.D
Acq: 16 Mar 2021 14:09




Tgt Ion:154 Resp: 529419
Ion Ratio Lower Upper
154 100
153 112.9 90.3 135.5
152 54.4 43.8 65.8





#53
3-Nitroaniline
Concen: 51.09 ng
RT: 14.656 min Scan# 1926
Delta R.T. 0.007 min
Lab File: BG048390.D
Acq: 16 Mar 2021 14:09




Instrument :
BNA_G
ClientSampleId :
SSTDICC060




Tgt Ion:138 Resp: 140409
Ion Ratio Lower Upper
138 100
108 10.5 7.4 11.0
92 131.9 85.2 127.8#

Manual Integrations
APPROVED



mohammad
3/17/2021 10:51:21 AM




#54
2,4-Dinitrophenol
Concen: 43.36 ng
RT: 14.849 min Scan# 1959
Delta R.T. 0.001 min
Lab File: BG048390.D
Acq: 16 Mar 2021 14:09




Tgt Ion:184 Resp: 78991
Ion Ratio Lower Upper
184 100
63 63.0 32.3 48.5#
154 64.9 40.4 60.6#







#55
Dibenzofuran
Concen: 53.05 ng
RT: 15.155 min Scan# 2011
Delta R.T. 0.001 min
Lab File: BG048390.D
Acq: 16 Mar 2021 14:09




Tgt Ion:168 Resp: 788819
Ion Ratio Lower Upper
168 100
139 41.7 28.6 43.0
169 14.5 11.0 16.6




Instrument :
BNA_G
ClientSampleId :
SSTDIC060

Manual Integrations
APPROVED



mohammad
3/17/2021 10:51:21 AM




#56
4-Nitrophenol
Concen: 53.06 ng
RT: 14.943 min Scan# 1975
Delta R.T. 0.007 min
Lab File: BG048390.D
Acq: 16 Mar 2021 14:09




Tgt Ion:139 Resp: 119535
Ion Ratio Lower Upper
139 100
109 120.3 45.9 85.9#
65 137.8 65.4 105.4#





#57
2,4-Dinitrotoluene
Concen: 49.13 ng
RT: 15.108 min Scan# 2003
Delta R.T. 0.007 min
Lab File: BG048390.D
Acq: 16 Mar 2021 14:09




Instrument :
BNA_G
ClientSampleId :
SSTDIC060



Tgt Ion:165 Resp: 198599
Ion Ratio Lower Upper
165 100
63 56.2 26.4 39.6#
89 84.9 47.8 71.8#
182 3.1 2.6 4.0




Manual Integrations
APPROVED


mohammad
3/17/2021 10:51:21 AM



#58
Fluorene
Concen: 55.76 ng
RT: 15.801 min Scan# 2121
Delta R.T. 0.001 min
Lab File: BG048390.D
Acq: 16 Mar 2021 14:09


Tgt Ion:166 Resp: 657716
Ion Ratio Lower Upper
166 100
165 101.0 77.2 115.8
167 13.9 11.4 17.2





#59
2,3,4,6-Tetrachlorophenol
Concen: 52.08 ng
RT: 15.372 min Scan# 2048
Delta R.T. 0.001 min
Lab File: BG048390.D
Acq: 16 Mar 2021 14:09




Instrument :
BNA_G
ClientSampleId :
SSTDIC060




Tgt Ion:232 Resp: 215025
Ion Ratio Lower Upper
232 100
131 39.0 26.6 40.0
130 2.8 1.4 2.0#
166 27.3 21.0 31.6

Manual Integrations
APPROVED



mohammad
3/17/2021 10:51:21 AM




#60
Diethylphthalate
Concen: 51.90 ng
RT: 15.566 min Scan# 2081
Delta R.T. 0.001 min
Lab File: BG048390.D
Acq: 16 Mar 2021 14:09



Tgt Ion:149 Resp: 660720
Ion Ratio Lower Upper
149 100
177 23.8 21.4 32.2
150 14.2 10.7 16.1









#61
4-Chlorophenyl-phenylether
Concen: 54.44 ng
RT: 15.795 min Scan# 2120
Delta R.T. 0.001 min
Lab File: BG048390.D
Acq: 16 Mar 2021 14:09

Instrument :
BNA_G
ClientSampleId :
SSTDICC060

Tgt Ion	Ratio	Lower	Upper
204	100		
206	31.9	27.8	41.8
141	54.5	41.5	62.3




Manual Integrations
APPROVED


mohammad
3/17/2021 10:51:21 AM



#62
4-Nitroaniline
Concen: 52.16 ng
RT: 15.819 min Scan# 2124
Delta R.T. 0.007 min
Lab File: BG048390.D
Acq: 16 Mar 2021 14:09


Tgt Ion	Ratio	Lower	Upper
138	100		
92	65.5	26.9	66.9
108	134.1	49.4	89.4





#63
Azobenzene
Concen: 58.61 ng
RT: 16.089 min Scan# 2170
Delta R.T. 0.007 min
Lab File: BG048390.D
Acq: 16 Mar 2021 14:09




Instrument :
BNA_G
ClientSampleId :
SSTDIC060




Tgt Ion: 77 Resp: 744613
Ion Ratio Lower Upper
77 100
182 25.8 14.0 54.0
105 16.6 3.0 43.0
51 35.1 8.1 48.1

Manual Integrations
APPROVED



mohammad
3/17/2021 10:51:21 AM




#64
Phenanthrene-d10
Concen: 20.00 ng
RT: 17.505 min Scan# 2411
Delta R.T. 0.001 min
Lab File: BG048390.D
Acq: 16 Mar 2021 14:09



Tgt Ion: 188 Resp: 368354
Ion Ratio Lower Upper
188 100
94 4.4 6.3 9.5#
80 6.8 7.0 10.6#









#65
4,6-Dinitro-2-methylphenol
Concen: 43.08 ng
RT: 15.872 min Scan# 2133
Delta R.T. 0.001 min
Lab File: BG048390.D
Acq: 16 Mar 2021 14:09

Tgt Ion	Ratio	Lower	Upper
198	100		
51	37.0	10.4	50.4
105	36.6	12.9	52.9

Instrument :
BNA_G
ClientSampleId :
SSTDIC060




Manual Integrations
APPROVED


mohammad
3/17/2021 10:51:21 AM



#66
n-Nitrosodiphenylamine
Concen: 57.33 ng
RT: 16.007 min Scan# 2156
Delta R.T. 0.001 min
Lab File: BG048390.D
Acq: 16 Mar 2021 14:09

Tgt Ion	Ratio	Lower	Upper
169	100		
168	66.9	55.5	83.3
167	38.0	30.5	45.7









#67
4-Bromophenyl-phenylether
Concen: 51.31 ng
RT: 16.689 min Scan# 2272
Delta R.T. 0.001 min
Lab File: BG048390.D
Acq: 16 Mar 2021 14:09

Instrument :
BNA_G
ClientSampleId :
SSTDIC060

Tgt Ion	Ratio	Lower	Upper
248	100		
250	94.8	78.9	118.3
141	65.8	46.2	69.4




Manual Integrations
APPROVED


mohammad
3/17/2021 10:51:21 AM



#68
Hexachlorobenzene
Concen: 55.70 ng
RT: 16.806 min Scan# 2292
Delta R.T. 0.001 min
Lab File: BG048390.D
Acq: 16 Mar 2021 14:09

Tgt Ion	Ratio	Lower	Upper
284	100		
142	31.7	21.1	31.7#
249	32.9	25.9	38.9









#69
Atrazine
Concen: 51.62 ng
RT: 16.953 min Scan# 2317
Delta R.T. 0.001 min
Lab File: BG048390.D
Acq: 16 Mar 2021 14:09

Tgt Ion:200 Resp: 228683
Ion Ratio Lower Upper
200 100
173 24.0 4.2 44.2
215 48.6 30.0 70.0

Instrument :
BNA_G
ClientSampleId :
SSTDIC060




Manual Integrations
APPROVED


mohammad
3/17/2021 10:51:21 AM



#70
Pentachlorophenol
Concen: 56.46 ng
RT: 17.147 min Scan# 2350
Delta R.T. 0.001 min
Lab File: BG048390.D
Acq: 16 Mar 2021 14:09

Tgt Ion:266 Resp: 180886
Ion Ratio Lower Upper
266 100
268 61.0 49.8 74.6
264 59.6 49.3 73.9









#71
Phenanthrene
Concen: 53.61 ng
RT: 17.546 min Scan# 2418
Delta R.T. 0.001 min
Lab File: BG048390.D
Acq: 16 Mar 2021 14:09

Tgt Ion:178 Resp: 1072449
Ion Ratio Lower Upper
178 100
176 20.9 16.7 25.1
179 16.9 14.0 21.0

Instrument :
BNA_G
ClientSampleId :
SSTDICC060




Manual Integrations
APPROVED


mohammad
3/17/2021 10:51:21 AM



#72
Anthracene
Concen: 53.27 ng
RT: 17.640 min Scan# 2434
Delta R.T. 0.001 min
Lab File: BG048390.D
Acq: 16 Mar 2021 14:09


Tgt Ion:178 Resp: 1057990
Ion Ratio Lower Upper
178 100
176 20.5 16.4 24.6
179 16.9 13.8 20.6





#73
 Carbazole
 Concen: 53.26 ng
 RT: 17.905 min Scan# 2479
 Delta R.T. 0.001 min
 Lab File: BG048390.D
 Acq: 16 Mar 2021 14:09

Instrument :
 BNA_G
 ClientSampleId :
 SSTDICC060






Tgt Ion:167 Resp: 1000499


Ion	Ratio	Lower	Upper
167	100		
166	24.5	18.7	28.1
139	13.6	9.6	14.4

Manual Integrations
 APPROVED

mohammad
 3/17/2021 10:51:21 AM






#74
 Di-n-butylphthalate
 Concen: 52.93 ng
 RT: 18.463 min Scan# 2574
 Delta R.T. 0.001 min
 Lab File: BG048390.D
 Acq: 16 Mar 2021 14:09



Tgt Ion:149 Resp: 1101875

Ion	Ratio	Lower	Upper
149	100		
150	10.5	8.2	12.4
104	7.5	3.8	5.8









#75
Fluoranthene
Concen: 53.72 ng
RT: 19.556 min Scan# 2760
Delta R.T. 0.001 min
Lab File: BG048390.D
Acq: 16 Mar 2021 14:09

Tgt Ion:202 Resp: 1398908
Ion Ratio Lower Upper
202 100
101 6.3 0.0 29.7
203 19.4 0.0 39.2

Instrument :
BNA_G
ClientSampleId :
SSTDIC060




Manual Integrations
APPROVED


mohammad
3/17/2021 10:51:21 AM



#76
Chrysene-d12
Concen: 20.00 ng
RT: 21.800 min Scan# 3142
Delta R.T. 0.001 min
Lab File: BG048390.D
Acq: 16 Mar 2021 14:09

Tgt Ion:240 Resp: 374683
Ion Ratio Lower Upper
240 100
120 4.3 6.2 9.4#
236 24.6 20.5 30.7









#77
Benzidine
Concen: 43.18 ng
RT: 19.732 min Scan# 2790
Delta R.T. 0.001 min
Lab File: BG048390.D
Acq: 16 Mar 2021 14:09

Tgt Ion:	184	Resp:	528654
Ion Ratio	Lower	Upper	
184	100		
185	14.2	12.4	18.6
183	13.5	10.6	15.8

Instrument :
BNA_G
ClientSampleId :
SSTDIC060




Manual Integrations
APPROVED


mohammad
3/17/2021 10:51:21 AM



#78
Pyrene
Concen: 53.69 ng
RT: 19.920 min Scan# 2822
Delta R.T. 0.007 min
Lab File: BG048390.D
Acq: 16 Mar 2021 14:09


Tgt Ion:	202	Resp:	1405096
Ion Ratio	Lower	Upper	
202	100		
200	22.3	18.6	27.8
203	19.0	15.3	22.9





#79
 Terphenyl-d14
 Concen: 102.14 ng
 RT: 20.114 min Scan# 2855
 Delta R.T. 0.001 min
 Lab File: BG048390.D
 Acq: 16 Mar 2021 14:09




Instrument :
 BNA_G
 ClientSampleId :
 SSTDICC060




Tgt Ion:244 Resp: 2094346
 Ion Ratio Lower Upper
 244 100
 212 9.6 6.7 10.1
 122 7.5 7.8 11.6#

Manual Integrations
 APPROVED



mohammad
 3/17/2021 10:51:21 AM




#80
 Butylbenzylphthalate
 Concen: 53.74 ng
 RT: 20.801 min Scan# 2972
 Delta R.T. 0.001 min
 Lab File: BG048390.D
 Acq: 16 Mar 2021 14:09




Tgt Ion:149 Resp: 492476
 Ion Ratio Lower Upper
 149 100
 91 82.7 51.5 77.3#
 206 26.8 21.2 31.8





#81
Benzo(a)anthracene
Concen: 54.90 ng
RT: 21.777 min Scan# 3138
Delta R.T. 0.001 min
Lab File: BG048390.D
Acq: 16 Mar 2021 14:09




Instrument :
BNA_G
ClientSampleId :
SSTDIC060




Tgt Ion:228 Resp: 1434865
Ion Ratio Lower Upper
228 100
226 29.3 23.3 34.9
229 22.3 17.4 26.0

Manual Integrations
APPROVED



mohammad
3/17/2021 10:51:21 AM




#82
3,3'-Dichlorobenzidine
Concen: 54.39 ng
RT: 21.689 min Scan# 3123
Delta R.T. 0.001 min
Lab File: BG048390.D
Acq: 16 Mar 2021 14:09




Tgt Ion:252 Resp: 509178
Ion Ratio Lower Upper
252 100
254 65.4 53.2 79.8
126 6.7 6.8 10.2#





#83
Chrysene
Concen: 54.76 ng
RT: 21.847 min Scan# 3150
Delta R.T. 0.001 min
Lab File: BG048390.D
Acq: 16 Mar 2021 14:09




Instrument :
BNA_G
ClientSampleId :
SSTDIC060




Tgt Ion:228 Resp: 1367340
Ion Ratio Lower Upper
228 100
226 31.8 25.0 37.4
229 21.4 17.3 25.9

Manual Integrations
APPROVED



mohammad
3/17/2021 10:51:21 AM




#84
Bis(2-ethylhexyl)phthalate
Concen: 55.99 ng
RT: 21.683 min Scan# 3122
Delta R.T. 0.001 min
Lab File: BG048390.D
Acq: 16 Mar 2021 14:09




Tgt Ion:149 Resp: 720775
Ion Ratio Lower Upper
149 100
167 28.1 24.8 37.2
279 5.7 4.6 7.0





#85
Di-n-octyl phthalate
Concen: 55.87 ng
RT: 22.946 min Scan# 3337
Delta R.T. 0.001 min
Lab File: BG048390.D
Acq: 16 Mar 2021 14:09




Instrument :
BNA_G
ClientSampleId :
SSTDIC060




Tgt Ion:149 Resp: 1224685
Ion Ratio Lower Upper
149 100
167 1.8 1.4 2.2
43 11.1 6.8 10.2#

Manual Integrations
APPROVED



mohammad
3/17/2021 10:51:21 AM




#86
Perylene-d12
Concen: 20.00 ng
RT: 25.137 min Scan# 3710
Delta R.T. 0.007 min
Lab File: BG048390.D
Acq: 16 Mar 2021 14:09



Tgt Ion:264 Resp: 385090
Ion Ratio Lower Upper
264 100
260 22.5 18.8 28.2
265 21.6 17.8 26.8









#87
 Indeno(1,2,3-cd)pyrene
 Concen: 59.18 ng
 RT: 28.974 min Scan# 4363
 Delta R.T. 0.007 min
 Lab File: BG048390.D
 Acq: 16 Mar 2021 14:09

Tgt Ion: 276 Resp: 1726964
 Ion Ratio Lower Upper
 276 100
 138 4.0 16.2 24.2#
 277 25.9 20.3 30.5

Instrument :
 BNA_G
 ClientSampleId :
 SSTDICC060




Manual Integrations
 APPROVED


mohammad
 3/17/2021 10:51:21 AM



#88
 Benzo(b)fluoranthene
 Concen: 57.32 ng
 RT: 24.080 min Scan# 3530
 Delta R.T. 0.007 min
 Lab File: BG048390.D
 Acq: 16 Mar 2021 14:09

Tgt Ion: 252 Resp: 1526268
 Ion Ratio Lower Upper
 252 100
 253 22.7 18.7 28.1
 125 4.5 6.5 9.7#









#89
Benzo(k)fluoranthene
Concen: 57.49 ng
RT: 24.150 min Scan# 3542
Delta R.T. 0.001 min
Lab File: BG048390.D
Acq: 16 Mar 2021 14:09

Instrument :
BNA_G
ClientSampleId :
SSTDIC060

Tgt Ion:252 Resp: 1460256
Ion Ratio Lower Upper
252 100
253 22.8 18.6 27.8
125 5.1 6.8 10.2#




Manual Integrations
APPROVED


mohammad
3/17/2021 10:51:21 AM



#90
Benzo(a)pyrene
Concen: 57.18 ng
RT: 24.985 min Scan# 3684
Delta R.T. 0.001 min
Lab File: BG048390.D
Acq: 16 Mar 2021 14:09

Tgt Ion:252 Resp: 1412089
Ion Ratio Lower Upper
252 100
253 23.0 17.9 26.9
125 5.1 7.0 10.4#









#91
Dibenzo(a,h)anthracene
Concen: 58.61 ng
RT: 29.050 min Scan# 4376
Delta R.T. 0.007 min
Lab File: BG048390.D
Acq: 16 Mar 2021 14:09

Tgt Ion:278 Resp: 1458797
Ion Ratio Lower Upper
278 100
139 7.9 10.5 15.7#
279 25.8 19.4 29.0

Instrument :
BNA_G
ClientSampleId :
SSTDIC060

Manual Integrations
APPROVED

mohammad
3/17/2021 10:51:21 AM



#92
Benzo(g,h,i)perylene
Concen: 58.62 ng
RT: 30.173 min Scan# 4567
Delta R.T. 0.013 min
Lab File: BG048390.D
Acq: 16 Mar 2021 14:09

Tgt Ion:276 Resp: 1417960
Ion Ratio Lower Upper
276 100
277 24.0 19.6 29.4
138 9.8 13.4 20.2#

