

Data Path : Z:\svoasrv\HPCHEM1\BNA_G\Data\BG031622\
 Data File : BG052670.D
 Acq On : 16 Mar 2022 8:27
 Operator : CG/JU
 Sample : SSTDCCC040
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_G
 LabSampleId :
 SSTDCCC040

Quant Time: Mar 16 11:42:18 2022
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_G\Methods\8270-BG031522.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Tue Mar 15 15:56:10 2022
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	20.000	20.000	0.0	106	0.00
2	1,4-Dioxane	40.000	39.170	2.1	110	0.00
3	Pyridine	40.000	42.844	-7.1	115	0.00
4	n-Nitrosodimethylamine	40.000	40.891	-2.2	111	0.00
5 S	2-Fluorophenol	80.000	78.125	2.3	108	0.00
6	Aniline	40.000	38.850	2.9	106	0.00
7 S	Phenol-d6	80.000	77.239	3.5	105	0.00
8	2-Chlorophenol	40.000	37.711	5.7	102	0.00
9	Benzaldehyde	40.000	38.733	3.2	110	0.00
10 C	Phenol	40.000	38.116	4.7	106	0.00
11	bis(2-Chloroethyl)ether	40.000	40.125	-0.3	112	0.00
12	1,3-Dichlorobenzene	40.000	38.592	3.5	107	0.00
13 C	1,4-Dichlorobenzene	40.000	39.086	2.3	108	0.00
14	1,2-Dichlorobenzene	40.000	38.317	4.2	107	0.00
15	Benzyl Alcohol	40.000	36.955	7.6	102	0.00
16	2,2'-oxybis(1-Chloropropane	40.000	38.438	3.9	106	0.00
17	2-Methylphenol	40.000	37.655	5.9	102	0.00
18	Hexachloroethane	40.000	38.446	3.9	106	0.00
19 P	n-Nitroso-di-n-propylamine	40.000	37.727	5.7	104	0.00
20	3+4-Methylphenols	40.000	38.343	4.1	105	0.00
21 I	Naphthalene-d8	20.000	20.000	0.0	106	0.00
22	Acetophenone	40.000	38.506	3.7	105	0.00
23 S	Nitrobenzene-d5	80.000	80.309	-0.4	106	0.00
24	Nitrobenzene	40.000	39.266	1.8	104	0.00
25	Isophorone	40.000	38.525	3.7	104	0.00
26 C	2-Nitrophenol	40.000	36.002	10.0	96	0.00
27	2,4-Dimethylphenol	40.000	38.195	4.5	107	0.00
28	bis(2-Chloroethoxy)methane	40.000	38.160	4.6	103	0.00
29 C	2,4-Dichlorophenol	40.000	38.507	3.7	103	0.00
30	1,2,4-Trichlorobenzene	40.000	39.663	0.8	105	0.00
31	Naphthalene	40.000	38.779	3.1	105	0.00
32	Benzoic acid	40.000	33.626	15.9	95	0.00
33	4-Chloroaniline	40.000	38.198	4.5	102	0.00
34 C	Hexachlorobutadiene	40.000	38.485	3.8	104	0.00
35	Caprolactam	40.000	38.844	2.9	99	0.00
36 C	4-Chloro-3-methylphenol	40.000	38.958	2.6	102	0.00
37	2-Methylnaphthalene	40.000	38.470	3.8	105	0.00
38	1-Methylnaphthalene	40.000	38.158	4.6	104	0.00
39 I	Acenaphthene-d10	20.000	20.000	0.0	101	0.00
40	1,2,4,5-Tetrachlorobenzene	40.000	38.382	4.0	102	0.00
41 P	Hexachlorocyclopentadiene	40.000	35.741	10.6	95	0.00
42 S	2,4,6-Tribromophenol	80.000	70.689	11.6	91	0.00
43 C	2,4,6-Trichlorophenol	40.000	37.609	6.0	100	0.00
44	2,4,5-Trichlorophenol	40.000	39.789	0.5	105	0.00
45 S	2-Fluorobiphenyl	80.000	77.185	3.5	103	0.00
46	1,1'-Biphenyl	40.000	39.077	2.3	105	0.00
47	2-Chloronaphthalene	40.000	38.311	4.2	103	0.00

Data Path : Z:\svoasrv\HPCHEM1\BNA_G\Data\BG031622\
 Data File : BG052670.D
 Acq On : 16 Mar 2022 8:27
 Operator : CG/JU
 Sample : SSTDCCC040
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_G
 LabSampleId :
 SSTDCCC040

Quant Time: Mar 16 11:42:18 2022
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_G\Methods\8270-BG031522.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Tue Mar 15 15:56:10 2022
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
48	2-Nitroaniline	40.000	39.514	1.2	99	0.00
49	Acenaphthylene	40.000	39.252	1.9	105	0.00
50	Dimethylphthalate	40.000	38.367	4.1	100	0.00
51	2,6-Dinitrotoluene	40.000	40.572	-1.4	99	0.00
52 C	Acenaphthene	40.000	38.855	2.9	104	0.00
53	3-Nitroaniline	40.000	41.323	-3.3	103	0.00
54 P	2,4-Dinitrophenol	40.000	35.952	10.1	98	0.00
55	Dibenzofuran	40.000	39.342	1.6	104	0.00
56 P	4-Nitrophenol	40.000	40.237	-0.6	98	0.00
57	2,4-Dinitrotoluene	40.000	40.420	-1.1	99	0.00
58	Fluorene	40.000	38.181	4.5	103	0.00
59	2,3,4,6-Tetrachlorophenol	40.000	38.464	3.8	97	0.00
60	Diethylphthalate	40.000	39.325	1.7	102	0.00
61	4-Chlorophenyl-phenylether	40.000	37.967	5.1	102	0.00
62	4-Nitroaniline	40.000	41.117	-2.8	105	0.00
63	Azobenzene	40.000	39.433	1.4	104	0.00
64 I	Phenanthrene-d10	20.000	20.000	0.0	104	0.00
65	4,6-Dinitro-2-methylphenol	40.000	37.383	6.5	99	0.00
66 c	n-Nitrosodiphenylamine	40.000	37.323	6.7	102	0.00
67	4-Bromophenyl-phenylether	40.000	33.989	15.0	93	0.00
68	Hexachlorobenzene	40.000	37.474	6.3	101	0.00
69	Atrazine	40.000	39.101	2.2	102	0.00
70 C	Pentachlorophenol	40.000	36.342	9.1	94	0.00
71	Phenanthrene	40.000	38.231	4.4	104	0.00
72	Anthracene	40.000	38.130	4.7	104	0.00
73	Carbazole	40.000	38.186	4.5	104	0.00
74	Di-n-butylphthalate	40.000	38.338	4.2	101	0.00
75 C	Fluoranthene	40.000	38.584	3.5	105	0.00
76 I	Chrysene-d12	20.000	20.000	0.0	107	0.00
77	Benzydine	40.000	41.611	-4.0	117	0.00
78	Pyrene	40.000	38.472	3.8	106	0.00
79 S	Terphenyl-d14	80.000	76.303	4.6	104	0.00
80	Butylbenzylphthalate	40.000	36.607	8.5	98	0.00
81	Benzo(a)anthracene	40.000	38.723	3.2	108	0.00
82	3,3'-Dichlorobenzidine	40.000	38.974	2.6	108	0.00
83	Chrysene	40.000	38.864	2.8	108	0.00
84	Bis(2-ethylhexyl)phthalate	40.000	37.816	5.5	103	0.00
85 c	Di-n-octyl phthalate	40.000	38.006	5.0	103	0.00
86 I	Perylene-d12	20.000	20.000	0.0	114	0.00
87	Indeno(1,2,3-cd)pyrene	40.000	39.217	2.0	117	0.00
88	Benzo(b)fluoranthene	40.000	37.857	5.4	112	0.00
89	Benzo(k)fluoranthene	40.000	37.920	5.2	113	0.00
90 C	Benzo(a)pyrene	40.000	38.548	3.6	114	0.00
91	Dibenzo(a,h)anthracene	40.000	39.175	2.1	117	0.00
92	Benzo(g,h,i)perylene	40.000	39.540	1.2	117	0.00

Data Path : Z:\svoasrv\HPCHEM1\BNA_G\Data\BG031622\
Data File : BG052670.D
Acq On : 16 Mar 2022 8:27
Operator : CG/JU
Sample : SSTDCCC040
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_G
LabSampleId :
SSTDCCC040

Quant Time: Mar 16 11:42:18 2022
Quant Method : Z:\svoasrv\HPCHEM1\BNA_G\Methods\8270-BG031522.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Tue Mar 15 15:56:10 2022
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

Compound	Amount	Calc.	%Dev	Area%	Dev(min)
----------	--------	-------	------	-------	----------

(#) = Out of Range SPCC's out = 0 CCC's out = 0