

Data Path : Z:\svoasrv\HPCHEM1\BNA_G\Data\BG031821\
 Data File : BG048419.D
 Acq On : 18 Mar 2021 16:32
 Operator : CG/JU
 Sample : SSTDICV040
 Misc :
 ALS Vial : 10 Sample Multiplier: 1

Instrument :
 BNA_G
 ClientSampleId :
 ICVBG031821

Quant Time: Mar 18 17:10:01 2021
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_G\METHODS\8270-BG031821.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Thu Mar 18 16:38:55 2021
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	88	0.00
2	1,4-Dioxane	0.510	0.481	5.7	88	0.00
3	Pyridine	1.422	1.440	-1.3	93	0.00
4	n-Nitrosodimethylamine	0.827	0.833	-0.7	94	0.00
5 S	2-Fluorophenol	1.209	1.206	0.2	94	0.00
6	Aniline	2.081	2.097	-0.8	93	0.00
7 S	Phenol-d6	1.760	1.766	-0.3	92	0.00
8	2-Chlorophenol	1.261	1.229	2.5	90	0.00
9	Benzaldehyde	1.109	1.039	6.3	94	0.00
10 C	Phenol	1.803	1.816	-0.7	94	0.00
11	bis(2-Chloroethyl)ether	1.264	1.253	0.9	91	0.00
12	1,3-Dichlorobenzene	1.500	1.459	2.7	90	0.00
13 C	1,4-Dichlorobenzene	1.489	1.528	-2.6	97	0.00
14	1,2-Dichlorobenzene	1.429	1.437	-0.6	93	0.00
15	Benzyl Alcohol	1.562	1.556	0.4	92	0.00
16	2,2'-oxybis(1-Chloropropane	1.445	1.435	0.7	92	0.02
17	2-Methylphenol	1.143	1.156	-1.1	96	0.00
18	Hexachloroethane	0.573	0.568	0.9	95	0.00
19 P	n-Nitroso-di-n-propylamine	1.306	1.297	0.7	90	0.00
20	3+4-Methylphenols	1.566	1.596	-1.9	95	0.00
21 I	Naphthalene-d8	1.000	1.000	0.0	93	0.00
22	Acetophenone	0.562	0.556	1.1	94	0.00
23 S	Nitrobenzene-d5	0.500	0.506	-1.2	96	0.00
24	Nitrobenzene	0.536	0.518	3.4	91	0.00
25	Isophorone	0.957	0.961	-0.4	95	0.00
26 C	2-Nitrophenol	0.204	0.201	1.5	93	0.00
27	2,4-Dimethylphenol	0.318	0.321	-0.9	96	0.00
28	bis(2-Chloroethoxy)methane	0.434	0.426	1.8	96	0.00
29 C	2,4-Dichlorophenol	0.369	0.381	-3.3	95	0.00
30	1,2,4-Trichlorobenzene	0.446	0.448	-0.4	95	0.00
31	Naphthalene	1.109	1.091	1.6	92	0.00
32	Benzoic acid	0.247	0.257	-4.0	96	0.00
33	4-Chloroaniline	0.474	0.473	0.2	94	0.00
34 C	Hexachlorobutadiene	0.353	0.350	0.8	92	0.00
35	Caprolactam	0.132	0.139	-5.3	101	0.00
36 C	4-Chloro-3-methylphenol	0.428	0.436	-1.9	97	0.00
37	2-Methylnaphthalene	0.846	0.853	-0.8	96	0.00
38	1-Methylnaphthalene	0.796	0.800	-0.5	95	0.00
39 I	Acenaphthene-d10	1.000	1.000	0.0	100	0.00
40	1,2,4,5-Tetrachlorobenzene	0.730	0.691	5.3	95	0.00
41 P	Hexachlorocyclopentadiene	0.472	0.450	4.7	94	0.00
42 S	2,4,6-Tribromophenol	0.325	0.323	0.6	101	0.00
43 C	2,4,6-Trichlorophenol	0.495	0.501	-1.2	99	0.00
44	2,4,5-Trichlorophenol	0.545	0.517	5.1	94	0.00
45 S	2-Fluorobiphenyl	1.390	1.337	3.8	97	0.00
46	1,1'-Biphenyl	1.435	1.346	6.2	95	0.00
47	2-Chloronaphthalene	1.157	1.096	5.3	95	0.00

Data Path : Z:\svoasrv\HPCHEM1\BNA_G\Data\BG031821\
 Data File : BG048419.D
 Acq On : 18 Mar 2021 16:32
 Operator : CG/JU
 Sample : SSTDICV040
 Misc :
 ALS Vial : 10 Sample Multiplier: 1

Instrument :
 BNA_G
 ClientSampleId :
 ICBVG031821

Quant Time: Mar 18 17:10:01 2021
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_G\METHODS\8270-BG031821.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Thu Mar 18 16:38:55 2021
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
48	2-Nitroaniline	0.428	0.424	0.9	97	0.00
49	Acenaphthylene	1.874	1.769	5.6	95	0.00
50	Dimethylphthalate	1.628	1.559	4.2	97	0.00
51	2,6-Dinitrotoluene	0.350	0.348	0.6	98	0.00
52 C	Acenaphthene	1.259	1.245	1.1	98	0.00
53	3-Nitroaniline	0.335	0.327	2.4	99	0.00
54 P	2,4-Dinitrophenol	0.214	0.218	-1.9	99	0.00
55	Dibenzofuran	1.876	1.797	4.2	97	0.00
56 P	4-Nitrophenol	0.300	0.300	0.0	102	0.00
57	2,4-Dinitrotoluene	0.497	0.516	-3.8	101	0.00
58	Fluorene	1.574	1.513	3.9	97	0.00
59	2,3,4,6-Tetrachlorophenol	0.538	0.528	1.9	98	0.00
60	Diethylphthalate	1.591	1.553	2.4	99	0.00
61	4-Chlorophenyl-phenylether	0.906	0.891	1.7	99	0.00
62	4-Nitroaniline	0.367	0.368	-0.3	101	0.00
63	Azobenzene	1.732	1.666	3.8	98	0.00
64 I	Phenanthrene-d10	1.000	1.000	0.0	99	0.00
65	4,6-Dinitro-2-methylphenol	0.126	0.134	-6.3	103	0.00
66 c	n-Nitrosodiphenylamine	0.579	0.575	0.7	99	0.00
67	4-Bromophenyl-phenylether	0.252	0.243	3.6	97	0.00
68	Hexachlorobenzene	0.270	0.264	2.2	99	0.00
69	Atrazine	0.233	0.231	0.9	99	0.00
70 C	Pentachlorophenol	0.179	0.193	-7.8	104	0.00
71	Phenanthrene	1.056	1.046	0.9	99	0.00
72	Anthracene	1.048	1.024	2.3	98	0.00
73	Carbazole	1.000	0.975	2.5	99	0.00
74	Di-n-butylphthalate	1.106	1.102	0.4	100	0.00
75 C	Fluoranthene	1.407	1.407	0.0	101	0.00
76 I	Chrysene-d12	1.000	1.000	0.0	99	0.00
77	Benzydine	0.599	0.578	3.5	96	0.00
78	Pyrene	1.374	1.393	-1.4	101	0.00
79 S	Terphenyl-d14	1.109	1.083	2.3	101	0.00
80	Butylbenzylphthalate	0.488	0.500	-2.5	102	0.00
81	Benzo(a)anthracene	1.387	1.411	-1.7	103	0.00
82	3,3'-Dichlorobenzidine	0.501	0.491	2.0	100	0.00
83	Chrysene	1.330	1.340	-0.8	102	0.00
84	Bis(2-ethylhexyl)phthalate	0.675	0.740	-9.6	109	0.00
85 c	Di-n-octyl phthalate	1.130	1.233	-9.1	108	0.00
86 I	Perylene-d12	1.000	1.000	0.0	103	0.00
87	Indeno(1,2,3-cd)pyrene	1.525	1.550	-1.6	105	0.00
88	Benzo(b)fluoranthene	1.409	1.376	2.3	101	0.00
89	Benzo(k)fluoranthene	1.330	1.337	-0.5	104	0.00
90 C	Benzo(a)pyrene	1.285	1.293	-0.6	103	0.00
91	Dibenzo(a,h)anthracene	1.301	1.328	-2.1	105	0.02
92	Benzo(g,h,i)perylene	1.262	1.281	-1.5	105	0.00

Data Path : Z:\svoasrv\HPCHEM1\BNA_G\Data\BG031821\
Data File : BG048419.D
Acq On : 18 Mar 2021 16:32
Operator : CG/JU
Sample : SSTDICV040
Misc :
ALS Vial : 10 Sample Multiplier: 1

Instrument :
BNA_G
ClientSampleId :
ICVBG031821

Quant Time: Mar 18 17:10:01 2021
Quant Method : Z:\SVOASRV\HPCHEM1\BNA_G\METHODS\8270-BG031821.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Thu Mar 18 16:38:55 2021
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

Compound	AvgRF	CCRF	%Dev Area	% Dev(min)
----------	-------	------	-----------	------------

(#) = Out of Range

SPCC's out = 0 CCC's out = 0