

Data Path : Z:\svoasrv\HPCHEM1\BNA_G\Data\BG032322\
 Data File : BG052805.D
 Acq On : 24 Mar 2022 11:36
 Operator : CG/JU
 Sample : SP5850
 Misc : SFAM SURROGATE
 ALS Vial : 16 Sample Multiplier: 1

Instrument :
 BNA_G
 ClientSampleId :
 SP5850

Quant Time: Mar 24 12:26:06 2022
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_G\Methods\SFAM-EPA-BG031822.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Wed Mar 23 13:16:32 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	8.145	152	31105	20.000	ng/u1	0.00
20) Naphthalene-d8	10.953	136	127048	20.000	ng/u1	0.00
38) Acenaphthene-d10	14.766	164	101380	20.000	ng/u1	0.00
64) Phenanthrene-d10	17.509	188	254815	20.000	ng/u1	0.00
79) Chrysene-d12	21.792	240	275032	20.000	ng/u1	0.00
88) Perylene-d12	25.111	264	265897	20.000	ng/u1	0.00
System Monitoring Compounds						
3) 1,4-Dioxane-d8	3.552	96	11652	13.232	ng/uL	0.00
4) Pyridine-d5	3.969	84	164176	74.680	ng/u1	0.00
7) Phenol-d5	7.288	99	199269	71.527	ng/u1	0.00
9) Bis-(2-Chloroethyl)eth...	7.458	67	129721	76.238	ng/u1	0.00
11) 2-Chlorophenol-d4	7.670	132	137926	72.912	ng/u1	0.00
15) 4-Methylphenol-d8	8.839	113	149412	74.036	ng/u1	0.00
21) Nitrobenzene-d5	9.303	128	74222	81.157	ng/u1	0.00
24) 2-Nitrophenol-d4	10.025	143	85248	88.376	ng/u1	0.00
28) 2,4-Dichlorophenol-d3	10.566	165	148854	70.933	ng/u1	0.00
31) 4-Chloroaniline-d4	11.077	131	213054	76.817	ng/u1	0.00
46) Dimethylphthalate-d6	14.161	166	575927	73.946	ng/u1	0.00
49) Acenaphthylene-d8	14.460	160	616510	65.142	ng/u1	0.00
54) 4-Nitrophenol-d4	14.936	143	100979	77.113	ng/u1	0.00
60) Fluorene-d10	15.753	176	488812	71.478	ng/u1	0.00
65) 4,6-Dinitro-2-methylph...	15.858	200	95156	73.380	ng/u1	0.00
73) Anthracene-d10	17.609	188	812472	68.472	ng/u1	0.00
81) Pyrene-d10	19.888	212	999323	64.951	ng/u1	0.00
92) Benzo(a)pyrene-d12	24.893	264	965713	70.290	ng/u1	0.02

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_G\Data\BG032322\
Data File : BG052805.D
Acq On : 24 Mar 2022 11:36
Operator : CG/JU
Sample : SP5850
Misc : SFAM SURROGATE
ALS Vial : 16 Sample Multiplier: 1

Instrument :
BNA_G
ClientSampleId :
SP5850

Quant Time: Mar 24 12:26:06 2022
Quant Method : Z:\svoasrv\HPCHEM1\BNA_G\Methods\SFAM-EPA-BG031822.M
Quant Title : SVOA CALIBRATION
QLast Update : Wed Mar 23 13:16:32 2022
Response via : Initial Calibration

