

Data Path : Z:\svoasrv\HPCHEM1\BNA_G\Data\BG032421\
 Data File : BG048485.D
 Acq On : 24 Mar 2021 10:40
 Operator : CG/JU
 Sample : SSTDCCC040
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_G
 LabSampleId :
 SSTDCCC040

Quant Time: Mar 24 12:36:59 2021
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_G\METHODS\8270-BG031821.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed Mar 24 12:36:47 2021
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	77	0.00
2	1,4-Dioxane	0.510	0.484	5.1	78	0.00
3	Pyridine	1.422	1.418	0.3	80	0.00
4	n-Nitrosodimethylamine	0.827	0.787	4.8	78	0.00
5 S	2-Fluorophenol	1.209	1.242	-2.7	84	0.00
6	Aniline	2.081	2.136	-2.6	83	0.00
7 S	Phenol-d6	1.760	1.828	-3.9	83	0.00
8	2-Chlorophenol	1.261	1.349	-7.0	86	0.00
9	Benzaldehyde	1.109	0.976	12.0	77	0.00
10 C	Phenol	1.803	1.839	-2.0	83	0.00
11	bis(2-Chloroethyl)ether	1.264	1.298	-2.7	83	0.00
12	1,3-Dichlorobenzene	1.500	1.465	2.3	79	0.00
13 C	1,4-Dichlorobenzene	1.489	1.515	-1.7	84	0.00
14	1,2-Dichlorobenzene	1.429	1.425	0.3	80	0.00
15	Benzyl Alcohol	1.562	1.587	-1.6	82	0.00
16	2,2'-oxybis(1-Chloropropane	1.445	1.401	3.0	78	0.00
17	2-Methylphenol	1.143	1.196	-4.6	87	0.00
18	Hexachloroethane	0.573	0.606	-5.8	88	0.00
19 P	n-Nitroso-di-n-propylamine	1.306	1.369	-4.8	83	0.00
20	3+4-Methylphenols	1.566	1.737	-10.9	90	0.00
21 I	Naphthalene-d8	1.000	1.000	0.0	85	0.00
22	Acetophenone	0.562	0.566	-0.7	88	0.00
23 S	Nitrobenzene-d5	0.500	0.499	0.2	87	0.00
24	Nitrobenzene	0.536	0.525	2.1	85	0.00
25	Isophorone	0.957	0.959	-0.2	87	0.00
26 C	2-Nitrophenol	0.204	0.201	1.5	86	0.00
27	2,4-Dimethylphenol	0.318	0.332	-4.4	91	0.00
28	bis(2-Chloroethoxy)methane	0.434	0.445	-2.5	92	0.00
29 C	2,4-Dichlorophenol	0.369	0.376	-1.9	87	0.00
30	1,2,4-Trichlorobenzene	0.446	0.410	8.1	80	0.00
31	Naphthalene	1.109	1.106	0.3	85	0.00
32	Benzoic acid	0.247	0.266	-7.7	92	0.01
33	4-Chloroaniline	0.474	0.491	-3.6	90	0.00
34 C	Hexachlorobutadiene	0.353	0.295	16.4	71	0.00
35	Caprolactam	0.132	0.136	-3.0	91	0.00
36 C	4-Chloro-3-methylphenol	0.428	0.456	-6.5	93	0.00
37	2-Methylnaphthalene	0.846	0.874	-3.3	90	0.00
38	1-Methylnaphthalene	0.796	0.824	-3.5	90	0.00
39 I	Acenaphthene-d10	1.000	1.000	0.0	81	0.00
40	1,2,4,5-Tetrachlorobenzene	0.730	0.677	7.3	76	0.00
41 P	Hexachlorocyclopentadiene	0.472	0.442	6.4	75	0.00
42 S	2,4,6-Tribromophenol	0.325	0.332	-2.2	84	0.00
43 C	2,4,6-Trichlorophenol	0.495	0.498	-0.6	80	0.00
44	2,4,5-Trichlorophenol	0.545	0.546	-0.2	81	0.00
45 S	2-Fluorobiphenyl	1.390	1.423	-2.4	84	0.00
46	1,1'-Biphenyl	1.435	1.500	-4.5	86	0.00
47	2-Chloronaphthalene	1.157	1.252	-8.2	88	0.00

Data Path : Z:\svoasrv\HPCHEM1\BNA_G\Data\BG032421\
 Data File : BG048485.D
 Acq On : 24 Mar 2021 10:40
 Operator : CG/JU
 Sample : SSTDCCC040
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_G
 LabSampleId :
 SSTDCCC040

Quant Time: Mar 24 12:36:59 2021
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_G\METHODS\8270-BG031821.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed Mar 24 12:36:47 2021
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
48	2-Nitroaniline	0.428	0.484	-13.1	90	0.00
49	Acenaphthylene	1.874	1.993	-6.4	87	0.00
50	Dimethylphthalate	1.628	1.643	-0.9	83	0.00
51	2,6-Dinitrotoluene	0.350	0.369	-5.4	84	0.00
52 C	Acenaphthene	1.259	1.358	-7.9	87	0.00
53	3-Nitroaniline	0.335	0.363	-8.4	89	0.00
54 P	2,4-Dinitrophenol	0.214	0.235	-9.8	87	0.00
55	Dibenzofuran	1.876	1.990	-6.1	88	0.00
56 P	4-Nitrophenol	0.300	0.317	-5.7	88	0.00
57	2,4-Dinitrotoluene	0.497	0.526	-5.8	84	0.00
58	Fluorene	1.574	1.576	-0.1	82	0.00
59	2,3,4,6-Tetrachlorophenol	0.538	0.511	5.0	77	0.00
60	Diethylphthalate	1.591	1.759	-10.6	91	0.00
61	4-Chlorophenyl-phenylether	0.906	0.881	2.8	80	0.00
62	4-Nitroaniline	0.367	0.385	-4.9	86	0.00
63	Azobenzene	1.732	1.794	-3.6	86	0.00
64 I	Phenanthrene-d10	1.000	1.000	0.0	76	0.00
65	4,6-Dinitro-2-methylphenol	0.126	0.148	-17.5	87	0.00
66 c	n-Nitrosodiphenylamine	0.579	0.632	-9.2	84	0.00
67	4-Bromophenyl-phenylether	0.252	0.243	3.6	75	0.00
68	Hexachlorobenzene	0.270	0.269	0.4	78	0.00
69	Atrazine	0.233	0.241	-3.4	80	0.00
70 C	Pentachlorophenol	0.179	0.182	-1.7	75	0.00
71	Phenanthrene	1.056	1.126	-6.6	82	0.00
72	Anthracene	1.048	1.111	-6.0	82	0.00
73	Carbazole	1.000	1.092	-9.2	85	0.00
74	Di-n-butylphthalate	1.106	1.249	-12.9	87	0.00
75 C	Fluoranthene	1.407	1.377	2.1	76	0.00
76 I	Chrysene-d12	1.000	1.000	0.0	66	0.00
77	Benzydine	0.599	0.729	-21.7	80	0.00
78	Pyrene	1.374	1.564	-13.8	76	0.00
79 S	Terphenyl-d14	1.109	1.180	-6.4	73	0.00
80	Butylbenzylphthalate	0.488	0.612	-25.4#	83	0.00
81	Benzo(a)anthracene	1.387	1.468	-5.8	71	0.00
82	3,3'-Dichlorobenzidine	0.501	0.569	-13.6	77	0.00
83	Chrysene	1.330	1.406	-5.7	71	0.00
84	Bis(2-ethylhexyl)phthalate	0.675	0.867	-28.4#	85	0.00
85 c	Di-n-octyl phthalate	1.130	1.468	-29.9#	86	0.00
86 I	Perylene-d12	1.000	1.000	0.0	73	0.00
87	Indeno(1,2,3-cd)pyrene	1.525	1.539	-0.9	75	0.00
88	Benzo(b)fluoranthene	1.409	1.427	-1.3	75	0.00
89	Benzo(k)fluoranthene	1.330	1.337	-0.5	74	0.00
90 C	Benzo(a)pyrene	1.285	1.319	-2.6	75	0.00
91	Dibenzo(a,h)anthracene	1.301	1.330	-2.2	75	0.00
92	Benzo(g,h,i)perylene	1.262	1.286	-1.9	75	0.00

Data Path : Z:\svoasrv\HPCHEM1\BNA_G\Data\BG032421\
Data File : BG048485.D
Acq On : 24 Mar 2021 10:40
Operator : CG/JU
Sample : SSTDCCC040
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_G
LabSampleId :
SSTDCCC040

Quant Time: Mar 24 12:36:59 2021
Quant Method : Z:\SVOASRV\HPCHEM1\BNA_G\METHODS\8270-BG031821.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Wed Mar 24 12:36:47 2021
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
----------	-------	------	------	-------	----------

(#) = Out of Range

SPCC's out = 0 CCC's out = 1