

Data Path : Z:\svoasrv\HPCHEM1\BNA_G\Data\BG032522\
 Data File : BG052866.D
 Acq On : 26 Mar 2022 16:03
 Operator : CG/JU
 Sample : PB143594BS
 Misc :
 ALS Vial : 44 Sample Multiplier: 1

Instrument :
 BNA_G
 ClientSampleId :
 SLCS594

Quant Time: Mar 28 01:01:45 2022
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_G\Methods\SFAM-EPA-BG031822.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Wed Mar 23 13:16:32 2022
 Response via : Initial Calibration

Manual Integrations APPROVED

Reviewed By :Jagrut Upadhyay 03/28/2022
 Supervised By :mohammad ahmed 03/29/2022

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	8.138	152	30582	20.000	ng/u1	0.00
20) Naphthalene-d8	10.957	136	143792	20.000	ng/u1	0.00
38) Acenaphthene-d10	14.764	164	114791	20.000	ng/u1	0.00
64) Phenanthrene-d10	17.507	188	277727	20.000	ng/u1	0.00
79) Chrysene-d12	21.796	240	277329	20.000	ng/u1	# 0.00
88) Perylene-d12	25.109	264	279782	20.000	ng/u1	0.00
System Monitoring Compounds						
3) 1,4-Dioxane-d8	3.550	96	4801	5.545	ng/uL	0.00
4) Pyridine-d5	3.967	84	69779	32.284	ng/u1	0.00
7) Phenol-d5	7.286	99	109076	39.822	ng/u1	0.00
9) Bis-(2-Chloroethyl)eth...	7.456	67	64711	38.682	ng/u1	0.00
11) 2-Chlorophenol-d4	7.673	132	69933	37.601	ng/u1	0.00
15) 4-Methylphenol-d8	8.837	113	86541	43.616	ng/u1	0.00
21) Nitrobenzene-d5	9.301	128	40057	38.700	ng/u1	0.00
24) 2-Nitrophenol-d4	10.023	143	47534	43.540	ng/u1	0.00
28) 2,4-Dichlorophenol-d3	10.564	165	92025	38.746	ng/u1	0.00
31) 4-Chloroaniline-d4	11.081	131	114914	36.608	ng/u1	0.00
46) Dimethylphthalate-d6	14.159	166	333248	37.789	ng/u1	0.00
49) Acenaphthylene-d8	14.458	160	377442	35.222	ng/u1	0.00
54) 4-Nitrophenol-d4	14.934	143	59825	40.348	ng/u1	0.00
60) Fluorene-d10	15.751	176	282157	36.439	ng/u1	0.00
65) 4,6-Dinitro-2-methylph...	15.862	200	59972	42.432	ng/u1	0.00
73) Anthracene-d10	17.607	188	453051	35.032	ng/u1	0.00
81) Pyrene-d10	19.886	212	565245	36.434	ng/u1	0.00
92) Benzo(a)pyrene-d12	24.880	264	561059	38.810	ng/u1	0.00
Target Compounds						
					Qvalue	
2) 1,4-Dioxane	3.585	88	10540m	9.936	ng/uL	
5) Pyridine	3.990	79	79678	34.116	ng/u1	97
6) Benzaldehyde	7.274	77	74761m	41.056	ng/u1	
8) Phenol	7.315	94	111828	42.361	ng/u1	99
10) Bis(2-Chloroethyl)ether	7.550	93	75993	38.158	ng/u1	95
12) 2-Chlorophenol	7.703	128	73736	38.803	ng/u1	94
13) 2-Methylphenol	8.572	108	85160	41.507	ng/u1	100
14) 2,2'-oxybis(1-Chloropr...	8.666	45	127245	39.308	ng/u1	97
16) Acetophenone	8.960	105	133467	43.118	ng/u1	95
17) N-Nitroso-di-n-propyla...	8.948	70	80191	44.439	ng/u1	98
18) 4-Methylphenol	8.901	108	94052	43.999	ng/u1	95
19) Hexachloroethane	9.230	117	31386	37.226	ng/u1	95
22) Nitrobenzene	9.342	77	117504	38.852	ng/u1	91
23) Isophorone	9.870	82	232738	40.259	ng/u1	97
25) 2-Nitrophenol	10.058	139	47375	41.846	ng/u1#	82
26) 2,4-Dimethylphenol	10.105	107	101798	36.288	ng/u1	96
27) Bis(2-Chloroethoxy)met...	10.346	93	121107	37.672	ng/u1	97
29) 2,4-Dichlorophenol	10.593	162	91753	39.733	ng/u1	99
30) Naphthalene	11.004	128	263383	35.395	ng/u1	97
32) 4-Chloroaniline	11.104	127	109525	35.508	ng/u1	98
33) Hexachlorobutadiene	11.292	225	66241	35.124	ng/u1	98
34) Caprolactam	11.856	113	35284m	50.733	ng/u1	
35) 4-Chloro-3-methylphenol	12.214	107	112003	44.713	ng/u1	91
36) 2-Methylnaphthalene	12.602	142	198380	37.776	ng/u1	99

Data Path : Z:\svoasrv\HPCHEM1\BNA_G\Data\BG032522\
 Data File : BG052866.D
 Acq On : 26 Mar 2022 16:03
 Operator : CG/JU
 Sample : PB143594BS
 Misc :
 ALS Vial : 44 Sample Multiplier: 1

Instrument :

BNA_G

ClientSampleId :

SLCS594

Manual Integrations

APPROVED

Reviewed By :Jagrut Upadhyay 03/28/2022
 Supervised By :mohammad ahmed 03/29/2022

Quant Time: Mar 28 01:01:45 2022

Quant Method : Z:\svoasrv\HPCHEM1\BNA_G\Methods\SFAM-EPA-BG031822.M

Quant Title : SVOA CALIBRATION

QLast Update : Wed Mar 23 13:16:32 2022

Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
37) 1-Methylnaphthalene	12.819	142	205306	38.617	ng/ul	95
39) 1,2,4,5-Tetrachloroben...	12.966	216	133083	35.075	ng/ul	98
40) Hexachlorocyclopentadiene	12.949	237	77497	29.503	ng/ul	98
41) 2,4,6-Trichlorophenol	13.195	196	91513	39.471	ng/ul	97
42) 2,4,5-Trichlorophenol	13.266	196	98464	39.326	ng/ul	97
43) 1,1'-Biphenyl	13.595	154	284927	34.735	ng/ul	99
44) 2-Chloronaphthalene	13.642	162	227542	34.642	ng/ul	96
45) 2-Nitroaniline	13.836	65	88380	45.902	ng/ul	90
47) Dimethylphthalate	14.206	163	319584	37.975	ng/ul	98
48) 2,6-Dinitrotoluene	14.323	165	68051	44.638	ng/ul	96
50) Acenaphthylene	14.488	152	375769	36.308	ng/ul	99
51) 3-Nitroaniline	14.652	138	69277	45.020	ng/ul	90
52) Acenaphthene	14.828	153	251152	37.146	ng/ul	98
53) 2,4-Dinitrophenol	14.858	184	42336	44.701	ng/ul#	58
55) 4-Nitrophenol	14.952	109	66530	42.635	ng/ul	94
56) Dibenzofuran	15.157	168	371395	37.052	ng/ul	100
57) 2,4-Dinitrotoluene	15.110	165	97698	45.167	ng/ul#	93
58) 2,3,4,6-Tetrachlorophenol	15.381	232	94881	42.041	ng/ul#	97
59) Diethylphthalate	15.569	149	335326	39.342	ng/ul	99
61) Fluorene	15.809	166	307361	36.994	ng/ul	99
62) 4-Chlorophenyl-phenyle...	15.798	204	169619	38.366	ng/ul	99
63) 4-Nitroaniline	15.815	138	69599m	45.993	ng/ul	
66) 4,6-Dinitro-2-methylph...	15.874	198	60324	43.623	ng/ul	92
67) N-Nitrosodiphenylamine	16.003	169	274629	36.573	ng/ul	96
68) 4-Bromophenyl-phenylether	16.691	248	122756	44.640	ng/ul	94
69) Hexachlorobenzene	16.814	284	136607	38.457	ng/ul	95
70) Atrazine	16.949	200	106938	34.749	ng/ul	97
71) Pentachlorophenol	17.155	266	84875	39.292	ng/ul	94
72) Phenanthrene	17.548	178	546451	37.490	ng/ul	99
74) Anthracene	17.642	178	523548	35.927	ng/ul	100
75) 1,2,3,4-Tetrachloroben...	13.565	216	145206	35.105	ng/uL	97
76) Pentachlorobenzene	15.081	250	156206	39.362	ng/uL	98
77) Carbazole	17.901	167	495214	38.452	ng/ul	99
78) Di-n-butylphthalate	18.459	149	570338	37.241	ng/ul	99
80) Fluoranthene	19.551	202	703055	38.455	ng/ul	99
82) Pyrene	19.916	202	679867	37.413	ng/ul	97
83) Butylbenzylphthalate	20.797	149	254008	40.203	ng/ul	97
84) 3,3'-Dichlorobenzidine	21.684	252	240433	38.681	ng/ul	95
85) Benzo(a)anthracene	21.772	228	670987	37.487	ng/ul	98
86) Bis(2-ethylhexyl)phtha...	21.672	149	348705	39.848	ng/ul	99
87) Chrysene	21.843	228	636437	37.278	ng/ul	98
89) Di-n-octyl phthalate	22.923	149	593883	40.753	ng/ul	100
90) Benzo(b)fluoranthene	24.057	252	717902	39.738	ng/ul	99
91) Benzo(k)fluoranthene	24.128	252	645489	38.714	ng/ul#	98
93) Benzo(a)pyrene	24.962	252	657770	38.564	ng/ul#	96
94) Indeno(1,2,3-cd)pyrene	28.898	276	781476m	40.512	ng/ul	
95) Dibenzo(a,h)anthracene	28.974	278	656919	40.067	ng/ul	99
96) Benzo(g,h,i)perylene	30.090	276	654151	40.116	ng/ul	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed


Data Path : Z:\svoasrv\HPCHEM1\BNA_G\Data\BG032522\
 Data File : BG052866.D
 Acq On : 26 Mar 2022 16:03
 Operator : CG/JU
 Sample : PB143594BS
 Misc :
 ALS Vial : 44 Sample Multiplier: 1


Instrument :
 BNA_G
 ClientSampleId :
 SLCS594

Quant Time: Mar 28 01:01:45 2022
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_G\Methods\SFAM-EPA-BG031822.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Wed Mar 23 13:16:32 2022
 Response via : Initial Calibration

Manual Integrations APPROVED

Reviewed By :Jagrut Upadhyay 03/28/2022
 Supervised By :mohammad ahmed 03/29/2022









#2
1,4-Dioxane
Concen: 9.936 ng/uL m
RT: 3.585 min Scan# 33
Delta R.T. -0.009 min
Lab File: BG052866.D
Acq: 26 Mar 2022 16:03

Instrument :
BNA_G
ClientSampleId :
SLCS594

Tgt Ion: 88 Resp: 10540
Ion Ratio Lower Upper
88 100
43 36.9 30.6 46.0
58 73.3 58.6 87.8




Manual Integrations
APPROVED


Reviewed By :Jagrut Upadhyay 03/28/2022
Supervised By :mohammad ahmed 03/29/2022



#5
Pyridine
Concen: 34.116 ng/uL
RT: 3.990 min Scan# 102
Delta R.T. -0.003 min
Lab File: BG052866.D
Acq: 26 Mar 2022 16:03


Tgt Ion: 79 Resp: 79678
Ion Ratio Lower Upper
79 100
52 74.8 57.2 85.8
51 38.8 31.9 47.9





#6
Benzaldehyde
Concen: 41.056 ng/ul m
RT: 7.274 min Scan# 661
Delta R.T. 0.003 min
Lab File: BG052866.D
Acq: 26 Mar 2022 16:03




Instrument :
BNA_G
ClientSampleId :
SLCS594



Tgt Ion: 77 Resp: 7476
Ion Ratio Lower Upper
77 100
105 84.9 72.9 109.3
106 78.2 66.5 99.7




Manual Integrations
APPROVED


Reviewed By :Jagrut Upadhyay 03/28/2022
Supervised By :mohammad ahmed 03/29/2022



#8
Phenol
Concen: 42.361 ng/ul
RT: 7.315 min Scan# 668
Delta R.T. 0.003 min
Lab File: BG052866.D
Acq: 26 Mar 2022 16:03

Tgt Ion: 94 Resp: 111828
Ion Ratio Lower Upper
94 100
65 32.1 25.9 38.9
66 45.5 35.4 53.2









#10
Bis(2-Chloroethyl)ether
Concen: 38.158 ng/ul
RT: 7.550 min Scan# 708
Delta R.T. -0.003 min
Lab File: BG052866.D
Acq: 26 Mar 2022 16:03

Instrument :
BNA_G
ClientSampleId :
SLCS594

Tgt Ion: 93 Resp: 7599
Ion Ratio Lower Upper
93 100
63 81.5 68.2 102.4
95 33.9 24.7 37.1




Manual Integrations
APPROVED


Reviewed By :Jagrut Upadhyay 03/28/2022
Supervised By :mohammad ahmed 03/29/2022



#12
2-Chlorophenol
Concen: 38.803 ng/ul
RT: 7.703 min Scan# 734
Delta R.T. -0.003 min
Lab File: BG052866.D
Acq: 26 Mar 2022 16:03


Tgt Ion:128 Resp: 73736
Ion Ratio Lower Upper
128 100
64 55.0 39.8 59.6
130 31.0 25.8 38.8





#13
2-Methylphenol
Concen: 41.507 ng/ul
RT: 8.572 min Scan# 882
Delta R.T. -0.003 min
Lab File: BG052866.D
Acq: 26 Mar 2022 16:03




Instrument :
BNA_G
ClientSampleId :
SLCS594



Tgt Ion:108 Resp: 85160
Ion Ratio Lower Upper
108 100
107 85.3 68.3 102.5




Manual Integrations
APPROVED


Reviewed By :Jagrut Upadhyay 03/28/2022
Supervised By :mohammad ahmed 03/29/2022



#14
2,2'-oxybis(1-Chloropropane)
Concen: 39.308 ng/ul
RT: 8.666 min Scan# 898
Delta R.T. 0.003 min
Lab File: BG052866.D
Acq: 26 Mar 2022 16:03

Tgt Ion: 45 Resp: 127245
Ion Ratio Lower Upper
45 100
77 14.6 11.0 16.4
79 10.4 9.3 13.9









#16
Acetophenone
Concen: 43.118 ng/ul
RT: 8.960 min Scan# 948
Delta R.T. -0.003 min
Lab File: BG052866.D
Acq: 26 Mar 2022 16:03

Instrument :
BNA_G
ClientSampleId :
SLCS594

Tgt Ion:105 Resp: 13346
Ion Ratio Lower Upper
105 100
77 85.7 72.9 109.3
51 30.4 25.0 37.6




Manual Integrations
APPROVED


Reviewed By :Jagrut Upadhyay 03/28/2022
Supervised By :mohammad ahmed 03/29/2022



#17
N-Nitroso-di-n-propylamine
Concen: 44.439 ng/ul
RT: 8.948 min Scan# 946
Delta R.T. 0.003 min
Lab File: BG052866.D
Acq: 26 Mar 2022 16:03


Tgt Ion: 70 Resp: 80191
Ion Ratio Lower Upper
70 100
42 61.1 47.8 71.6
101 8.3 6.8 10.2
130 16.3 14.7 22.1





#18
4-Methylphenol
Concen: 43.999 ng/ul
RT: 8.901 min Scan# 91
Delta R.T. 0.003 min
Lab File: BG052866.D
Acq: 26 Mar 2022 16:03




Instrument :
BNA_G
ClientSampleId :
SLCS594




Tgt Ion:108 Resp: 94052
Ion Ratio Lower Upper
108 100
107 108.1 82.7 124.1

Manual Integrations
APPROVED



Reviewed By :Jagrut Upadhyay 03/28/2022
Supervised By :mohammad ahmed 03/29/2022




#19
Hexachloroethane
Concen: 37.226 ng/ul
RT: 9.230 min Scan# 994
Delta R.T. -0.009 min
Lab File: BG052866.D
Acq: 26 Mar 2022 16:03



Tgt Ion:117 Resp: 31386
Ion Ratio Lower Upper
117 100
201 107.5 83.9 125.9
199 70.8 51.4 77.2









#22
Nitrobenzene
Concen: 38.852 ng/ul
RT: 9.342 min Scan# 1103
Delta R.T. -0.003 min
Lab File: BG052866.D
Acq: 26 Mar 2022 16:03

Instrument :
BNA_G
ClientSampleId :
SLCS594

Tgt Ion: 77 Resp: 117504
Ion Ratio Lower Upper
77 100
123 38.5 25.7 38.5
65 13.4 9.7 14.5




Manual Integrations
APPROVED


Reviewed By :Jagrut Upadhyay 03/28/2022
Supervised By :mohammad ahmed 03/29/2022



#23
Isophorone
Concen: 40.259 ng/ul
RT: 9.870 min Scan# 1103
Delta R.T. 0.003 min
Lab File: BG052866.D
Acq: 26 Mar 2022 16:03


Tgt Ion: 82 Resp: 232738
Ion Ratio Lower Upper
82 100
95 7.5 6.1 9.1
138 15.2 11.0 16.4





#25
2-Nitrophenol
Concen: 41.846 ng/ul
RT: 10.058 min Scan# 1135
Delta R.T. -0.003 min
Lab File: BG052866.D
Acq: 26 Mar 2022 16:03




Instrument :
BNA_G
ClientSampleId :
SLCS594




Tgt Ion:139 Resp: 47379
Ion Ratio Lower Upper
139 100
65 71.9 48.9 73.3
109 51.3 29.0 43.6

Manual Integrations
APPROVED



Reviewed By :Jagrut Upadhyay 03/28/2022
Supervised By :mohammad ahmed 03/29/2022




#26
2,4-Dimethylphenol
Concen: 36.288 ng/ul
RT: 10.105 min Scan# 1143
Delta R.T. -0.009 min
Lab File: BG052866.D
Acq: 26 Mar 2022 16:03




Tgt Ion:107 Resp: 101798
Ion Ratio Lower Upper
107 100
121 44.5 37.4 56.0
122 75.2 63.0 94.4





#27
 Bis(2-Chloroethoxy)methane
 Concen: 37.672 ng/ul
 RT: 10.346 min Scan# 1184
 Delta R.T. -0.003 min
 Lab File: BG052866.D
 Acq: 26 Mar 2022 16:03




Instrument :
 BNA_G
 ClientSampleId :
 SLCS594




Tgt Ion: 93 Resp: 12110
 Ion Ratio Lower Upper
 93 100
 95 33.0 25.4 38.2
 123 8.7 8.2 12.2

Manual Integrations
 APPROVED



Reviewed By :Jagrut Upadhyay 03/28/2022
 Supervised By :mohammad ahmed 03/29/2022




#29
 2,4-Dichlorophenol
 Concen: 39.733 ng/ul
 RT: 10.593 min Scan# 1226
 Delta R.T. -0.003 min
 Lab File: BG052866.D
 Acq: 26 Mar 2022 16:03




Tgt Ion:162 Resp: 91753
 Ion Ratio Lower Upper
 162 100
 164 65.4 53.2 79.8
 98 37.1 29.1 43.7





#30
Naphthalene
Concen: 35.395 ng/ul
RT: 11.004 min Scan# 11
Delta R.T. -0.003 min
Lab File: BG052866.D
Acq: 26 Mar 2022 16:03




Instrument :
BNA_G
ClientSampleId :
SLCS594



Tgt Ion:128 Resp: 26338
Ion Ratio Lower Upper
128 100
129 11.1 9.0 13.6
127 14.0 9.8 14.6




Manual Integrations
APPROVED


Reviewed By :Jagrut Upadhyay 03/28/2022
Supervised By :mohammad ahmed 03/29/2022



#32
4-Chloroaniline
Concen: 35.508 ng/ul
RT: 11.104 min Scan# 1313
Delta R.T. 0.003 min
Lab File: BG052866.D
Acq: 26 Mar 2022 16:03

Tgt Ion:127 Resp: 109525
Ion Ratio Lower Upper
127 100
129 32.4 25.3 37.9









#33
Hexachlorobutadiene
Concen: 35.124 ng/ul
RT: 11.292 min Scan# 11
Delta R.T. -0.009 min
Lab File: BG052866.D
Acq: 26 Mar 2022 16:03

Instrument :
BNA_G
ClientSampleId :
SLCS594

Tgt Ion:225 Resp: 6624
Ion Ratio Lower Upper
225 100
223 62.3 51.7 77.5
227 62.3 49.6 74.4




Manual Integrations
APPROVED


Reviewed By :Jagrut Upadhyay 03/28/2022
Supervised By :mohammad ahmed 03/29/2022



#34
Caprolactam
Concen: 50.733 ng/ul m
RT: 11.856 min Scan# 1441
Delta R.T. 0.003 min
Lab File: BG052866.D
Acq: 26 Mar 2022 16:03


Tgt Ion:113 Resp: 35284
Ion Ratio Lower Upper
113 100
55 199.0 142.0 213.0
56 138.6 96.1 144.1





#35
4-Chloro-3-methylphenol
Concen: 44.713 ng/ul
RT: 12.214 min Scan# 1502
Delta R.T. 0.003 min
Lab File: BG052866.D
Acq: 26 Mar 2022 16:03




Instrument :
BNA_G
ClientSampleId :
SLCS594




Tgt Ion:107 Resp: 11200
Ion Ratio Lower Upper
107 100
144 23.4 21.8 32.6
142 72.0 64.6 97.0

Manual Integrations
APPROVED



Reviewed By :Jagrut Upadhyay 03/28/2022
Supervised By :mohammad ahmed 03/29/2022




#36
2-Methylnaphthalene
Concen: 37.776 ng/ul
RT: 12.602 min Scan# 1568
Delta R.T. -0.003 min
Lab File: BG052866.D
Acq: 26 Mar 2022 16:03




Tgt Ion:142 Resp: 198380
Ion Ratio Lower Upper
142 100
141 89.7 71.2 106.8





#37
1-Methylnaphthalene
Concen: 38.617 ng/ul
RT: 12.819 min Scan# 1605
Delta R.T. -0.003 min
Lab File: BG052866.D
Acq: 26 Mar 2022 16:03




Instrument :
BNA_G
ClientSampleId :
SLCS594




Tgt Ion:142 Resp: 205300
Ion Ratio Lower Upper
142 100
141 94.5 71.4 107.0
116 4.6 3.5 5.3

Manual Integrations
APPROVED



Reviewed By :Jagrut Upadhyay 03/28/2022
Supervised By :mohammad ahmed 03/29/2022




#39
1,2,4,5-Tetrachlorobenzene
Concen: 35.075 ng/ul
RT: 12.966 min Scan# 1630
Delta R.T. -0.003 min
Lab File: BG052866.D
Acq: 26 Mar 2022 16:03




Tgt Ion:216 Resp: 133083
Ion Ratio Lower Upper
216 100
214 79.8 62.0 93.0
179 20.0 15.7 23.5
108 15.1 12.9 19.3





#40
Hexachlorocyclopentadiene
Concen: 29.503 ng/ul
RT: 12.949 min Scan# 1627
Delta R.T. -0.003 min
Lab File: BG052866.D
Acq: 26 Mar 2022 16:03




Instrument :
BNA_G
ClientSampleId :
SLCS594




Tgt Ion:237 Resp: 7749
Ion Ratio Lower Upper
237 100
235 61.2 50.1 75.1
272 12.3 10.2 15.4

Manual Integrations
APPROVED



Reviewed By :Jagrut Upadhyay 03/28/2022
Supervised By :mohammad ahmed 03/29/2022




#41
2,4,6-Trichlorophenol
Concen: 39.471 ng/ul
RT: 13.195 min Scan# 1669
Delta R.T. -0.003 min
Lab File: BG052866.D
Acq: 26 Mar 2022 16:03




Tgt Ion:196 Resp: 91513
Ion Ratio Lower Upper
196 100
198 91.9 75.4 113.2
200 28.8 24.9 37.3





#42
2,4,5-Trichlorophenol
Concen: 39.326 ng/ul
RT: 13.266 min Scan# 1681
Delta R.T. -0.003 min
Lab File: BG052866.D
Acq: 26 Mar 2022 16:03




Instrument :
BNA_G
ClientSampleId :
SLCS594



Tgt Ion:196 Resp: 98464
Ion Ratio Lower Upper
196 100
198 100.2 77.0 115.4
200 29.9 24.2 36.4




Manual Integrations
APPROVED


Reviewed By :Jagrut Upadhyay 03/28/2022
Supervised By :mohammad ahmed 03/29/2022



#43
1,1'-Biphenyl
Concen: 34.735 ng/ul
RT: 13.595 min Scan# 1737
Delta R.T. -0.003 min
Lab File: BG052866.D
Acq: 26 Mar 2022 16:03


Tgt Ion:154 Resp: 284927
Ion Ratio Lower Upper
154 100
153 40.4 33.0 49.4
76 13.4 10.6 15.8





#44
2-Chloronaphthalene
Concen: 34.642 ng/ul
RT: 13.642 min Scan# 1745
Delta R.T. -0.003 min
Lab File: BG052866.D
Acq: 26 Mar 2022 16:03




Instrument :
BNA_G
ClientSampleId :
SLCS594



Tgt Ion:162 Resp: 22754
Ion Ratio Lower Upper
162 100
164 31.9 26.6 40.0
127 38.1 28.2 42.2




Manual Integrations
APPROVED


Reviewed By :Jagrut Upadhyay 03/28/2022
Supervised By :mohammad ahmed 03/29/2022



#45
2-Nitroaniline
Concen: 45.902 ng/ul
RT: 13.836 min Scan# 1778
Delta R.T. 0.003 min
Lab File: BG052866.D
Acq: 26 Mar 2022 16:03


Tgt Ion: 65 Resp: 88380
Ion Ratio Lower Upper
65 100
92 60.6 54.0 81.0
138 84.2 76.0 114.0





#47
Dimethylphthalate
Concen: 37.975 ng/ul
RT: 14.206 min Scan# 1841
Delta R.T. -0.003 min
Lab File: BG052866.D
Acq: 26 Mar 2022 16:03




Instrument :
BNA_G
ClientSampleId :
SLCS594




Tgt Ion:163 Resp: 319584
Ion Ratio Lower Upper
163 100
194 3.6 2.9 4.3
164 10.4 7.7 11.5

Manual Integrations
APPROVED



Reviewed By :Jagrut Upadhyay 03/28/2022
Supervised By :mohammad ahmed 03/29/2022




#48
2,6-Dinitrotoluene
Concen: 44.638 ng/ul
RT: 14.323 min Scan# 1861
Delta R.T. -0.003 min
Lab File: BG052866.D
Acq: 26 Mar 2022 16:03




Tgt Ion:165 Resp: 68051
Ion Ratio Lower Upper
165 100
89 68.0 53.4 80.2
121 21.6 21.2 31.8





#50
Acenaphthylene
Concen: 36.308 ng/ul
RT: 14.488 min Scan# 1889
Delta R.T. 0.003 min
Lab File: BG052866.D
Acq: 26 Mar 2022 16:03




Instrument :
BNA_G
ClientSampleId :
SLCS594




Tgt Ion:152 Resp: 375769
Ion Ratio Lower Upper
152 100
151 20.2 16.4 24.6
153 13.0 11.2 16.8

Manual Integrations
APPROVED



Reviewed By :Jagrut Upadhyay 03/28/2022
Supervised By :mohammad ahmed 03/29/2022




#51
3-Nitroaniline
Concen: 45.020 ng/ul
RT: 14.652 min Scan# 1917
Delta R.T. -0.003 min
Lab File: BG052866.D
Acq: 26 Mar 2022 16:03



Tgt Ion:138 Resp: 69277
Ion Ratio Lower Upper
138 100
108 9.2 7.8 11.8
92 117.0 103.5 155.3









#52
Acenaphthene
Concen: 37.146 ng/ul
RT: 14.828 min Scan# 1946
Delta R.T. 0.003 min
Lab File: BG052866.D
Acq: 26 Mar 2022 16:03

Instrument :
BNA_G
ClientSampleId :
SLCS594

Tgt Ion:153 Resp: 251152
Ion Ratio Lower Upper
153 100
152 46.6 38.0 57.0
154 85.7 70.0 105.0




Manual Integrations
APPROVED


Reviewed By :Jagrut Upadhyay 03/28/2022
Supervised By :mohammad ahmed 03/29/2022



#53
2,4-Dinitrophenol
Concen: 44.701 ng/ul
RT: 14.858 min Scan# 1952
Delta R.T. 0.003 min
Lab File: BG052866.D
Acq: 26 Mar 2022 16:03


Tgt Ion:184 Resp: 42336
Ion Ratio Lower Upper
184 100
63 69.0 49.3 73.9
154 67.4 108.6 162.8#





#55
4-Nitrophenol
Concen: 42.635 ng/ul
RT: 14.952 min Scan# 1968
Delta R.T. 0.003 min
Lab File: BG052866.D
Acq: 26 Mar 2022 16:03




Instrument :
BNA_G
ClientSampleId :
SLCS594



Tgt Ion:109 Resp: 66530
Ion Ratio Lower Upper
109 100
139 90.3 75.4 113.2
65 121.9 91.4 137.2




Manual Integrations
APPROVED


Reviewed By :Jagrut Upadhyay 03/28/2022
Supervised By :mohammad ahmed 03/29/2022



#56
Dibenzofuran
Concen: 37.052 ng/ul
RT: 15.157 min Scan# 2003
Delta R.T. -0.003 min
Lab File: BG052866.D
Acq: 26 Mar 2022 16:03


Tgt Ion:168 Resp: 371395
Ion Ratio Lower Upper
168 100
139 39.5 31.8 47.6





#57
2,4-Dinitrotoluene
Concen: 45.167 ng/ul
RT: 15.110 min Scan# 1995
Delta R.T. 0.003 min
Lab File: BG052866.D
Acq: 26 Mar 2022 16:03




Instrument :
BNA_G
ClientSampleId :
SLCS594




Tgt Ion:165 Resp: 97698
Ion Ratio Lower Upper
165 100
63 45.8 33.0 49.4
182 2.5 3.4 5.0

Manual Integrations
APPROVED



Reviewed By :Jagrut Upadhyay 03/28/2022
Supervised By :mohammad ahmed 03/29/2022




#58
2,3,4,6-Tetrachlorophenol
Concen: 42.041 ng/ul
RT: 15.381 min Scan# 2041
Delta R.T. -0.003 min
Lab File: BG052866.D
Acq: 26 Mar 2022 16:03



Tgt Ion:232 Resp: 94881
Ion Ratio Lower Upper
232 100
131 40.0 29.8 44.8
130 2.6 1.0 1.4#
166 27.0 22.0 33.0





#59
Diethylphthalate
Concen: 39.342 ng/ul
RT: 15.569 min Scan# 2073
Delta R.T. -0.003 min
Lab File: BG052866.D
Acq: 26 Mar 2022 16:03

Instrument :

BNA_G

ClientSampleId :





SLCS594

Tgt Ion:149 Resp: 335320
Ion Ratio Lower Upper
149 100
177 21.9 17.5 26.3
150 12.5 9.5 14.3

Manual Integrations




APPROVED


Reviewed By :Jagrut Upadhyay 03/28/2022
Supervised By :mohammad ahmed 03/29/2022



#61
Fluorene
Concen: 36.994 ng/ul
RT: 15.809 min Scan# 2114
Delta R.T. 0.003 min
Lab File: BG052866.D
Acq: 26 Mar 2022 16:03


Tgt Ion:166 Resp: 307361
Ion Ratio Lower Upper
166 100
165 94.1 75.7 113.5
167 13.3 10.3 15.5





#62
4-Chlorophenyl-phenylether
Concen: 38.366 ng/ul
RT: 15.798 min Scan# 2112
Delta R.T. 0.003 min
Lab File: BG052866.D
Acq: 26 Mar 2022 16:03




Instrument :
BNA_G
ClientSampleId :
SLCS594




Tgt Ion:204 Resp: 169619
Ion Ratio Lower Upper
204 100
206 33.1 27.4 41.0
141 54.6 44.2 66.2

Manual Integrations
APPROVED



Reviewed By :Jagrut Upadhyay 03/28/2022
Supervised By :mohammad ahmed 03/29/2022




#63
4-Nitroaniline
Concen: 45.993 ng/ul m
RT: 15.815 min Scan# 2115
Delta R.T. 0.003 min
Lab File: BG052866.D
Acq: 26 Mar 2022 16:03



Tgt Ion:138 Resp: 69599
Ion Ratio Lower Upper
138 100
92 57.4 43.0 64.6
108 115.2 84.6 126.8









#66
4,6-Dinitro-2-methylphenol
Concen: 43.623 ng/ul
RT: 15.874 min Scan# 2125
Delta R.T. 0.003 min
Lab File: BG052866.D
Acq: 26 Mar 2022 16:03

Instrument :
BNA_G
ClientSampleId :
SLCS594

Tgt Ion:198 Resp: 60324
Ion Ratio Lower Upper
198 100
182 3.3 3.0 4.6
77 29.1 19.7 29.5




Manual Integrations
APPROVED


Reviewed By :Jagrut Upadhyay 03/28/2022
Supervised By :mohammad ahmed 03/29/2022



#67
N-Nitrosodiphenylamine
Concen: 36.573 ng/ul
RT: 16.003 min Scan# 2147
Delta R.T. -0.003 min
Lab File: BG052866.D
Acq: 26 Mar 2022 16:03


Tgt Ion:169 Resp: 274629
Ion Ratio Lower Upper
169 100
168 70.2 57.4 86.0
167 39.7 28.1 42.1





#68
4-Bromophenyl-phenylether
Concen: 44.640 ng/ul
RT: 16.691 min Scan# 211
Delta R.T. -0.003 min
Lab File: BG052866.D
Acq: 26 Mar 2022 16:03




Instrument :
BNA_G
ClientSampleId :
SLCS594




Tgt Ion:248 Resp: 122750
Ion Ratio Lower Upper
248 100
250 101.3 76.3 114.5
141 68.3 58.7 88.1

Manual Integrations
APPROVED



Reviewed By :Jagrut Upadhyay 03/28/2022
Supervised By :mohammad ahmed 03/29/2022




#69
Hexachlorobenzene
Concen: 38.457 ng/ul
RT: 16.814 min Scan# 2285
Delta R.T. -0.003 min
Lab File: BG052866.D
Acq: 26 Mar 2022 16:03




Tgt Ion:284 Resp: 136607
Ion Ratio Lower Upper
284 100
142 29.6 22.3 33.5
249 31.2 21.9 32.9





#70
 Atrazine
 Concen: 34.749 ng/ul
 RT: 16.949 min Scan# 21
 Delta R.T. -0.003 min
 Lab File: BG052866.D
 Acq: 26 Mar 2022 16:03




Instrument :
 BNA_G
 ClientSampleId :
 SLCS594




Tgt Ion:200 Resp: 106938
 Ion Ratio Lower Upper
 200 100
 173 21.8 19.8 29.6
 215 48.6 39.5 59.3

Manual Integrations
 APPROVED



Reviewed By :Jagrut Upadhyay 03/28/2022
 Supervised By :mohammad ahmed 03/29/2022




#71
 Pentachlorophenol
 Concen: 39.292 ng/ul
 RT: 17.155 min Scan# 2343
 Delta R.T. -0.003 min
 Lab File: BG052866.D
 Acq: 26 Mar 2022 16:03




Tgt Ion:266 Resp: 84875
 Ion Ratio Lower Upper
 266 100
 264 63.1 50.3 75.5
 268 65.6 45.8 68.8





#72
Phenanthrene
Concen: 37.490 ng/u1
RT: 17.548 min Scan# 2410
Delta R.T. -0.003 min
Lab File: BG052866.D
Acq: 26 Mar 2022 16:03




Instrument :
BNA_G
ClientSampleId :
SLCS594




Tgt Ion:178 Resp: 54645
Ion Ratio Lower Upper
178 100
179 15.3 11.9 17.9
176 19.2 15.0 22.6

Manual Integrations
APPROVED



Reviewed By :Jagrut Upadhyay 03/28/2022
Supervised By :mohammad ahmed 03/29/2022




#74
Anthracene
Concen: 35.927 ng/u1
RT: 17.642 min Scan# 2426
Delta R.T. 0.003 min
Lab File: BG052866.D
Acq: 26 Mar 2022 16:03



Tgt Ion:178 Resp: 523548
Ion Ratio Lower Upper
178 100
179 15.9 12.8 19.2
176 18.4 14.7 22.1









#75
1,2,3,4-Tetrachlorobenzene
Concen: 35.105 ng/uL
RT: 13.565 min Scan# 1732
Delta R.T. -0.003 min
Lab File: BG052866.D
Acq: 26 Mar 2022 16:03

Instrument :
BNA_G
ClientSampleId :
SLCS594

Tgt Ion:216 Resp: 145200
Ion Ratio Lower Upper
216 100
214 83.0 63.8 95.8
218 48.4 38.3 57.5




Manual Integrations
APPROVED


Reviewed By :Jagrut Upadhyay 03/28/2022
Supervised By :mohammad ahmed 03/29/2022



#76
Pentachlorobenzene
Concen: 39.362 ng/uL
RT: 15.081 min Scan# 1990
Delta R.T. -0.003 min
Lab File: BG052866.D
Acq: 26 Mar 2022 16:03


Tgt Ion:250 Resp: 156206
Ion Ratio Lower Upper
250 100
248 64.0 50.4 75.6
252 63.0 49.2 73.8





#77
 Carbazole
 Concen: 38.452 ng/ul
 RT: 17.901 min Scan# 2470
 Delta R.T. -0.003 min
 Lab File: BG052866.D
 Acq: 26 Mar 2022 16:03




Instrument :
 BNA_G
 ClientSampleId :
 SLCS594




Tgt Ion:167 Resp: 495214
 Ion Ratio Lower Upper
 167 100
 166 22.3 17.3 25.9
 139 11.7 9.4 14.0

Manual Integrations
 APPROVED



Reviewed By :Jagrut Upadhyay 03/28/2022
 Supervised By :mohammad ahmed 03/29/2022




#78
 Di-n-butylphthalate
 Concen: 37.241 ng/ul
 RT: 18.459 min Scan# 2565
 Delta R.T. -0.003 min
 Lab File: BG052866.D
 Acq: 26 Mar 2022 16:03




Tgt Ion:149 Resp: 570338
 Ion Ratio Lower Upper
 149 100
 150 9.2 7.5 11.3
 104 6.2 4.6 6.8





#80
Fluoranthene
Concen: 38.455 ng/u1
RT: 19.551 min Scan# 2751
Delta R.T. -0.003 min
Lab File: BG052866.D
Acq: 26 Mar 2022 16:03




Instrument :
BNA_G
ClientSampleId :
SLCS594



Tgt Ion:202 Resp: 70305
Ion Ratio Lower Upper
202 100
101 7.5 6.1 9.1
100 6.3 5.4 8.0




Manual Integrations
APPROVED


Reviewed By :Jagrut Upadhyay 03/28/2022
Supervised By :mohammad ahmed 03/29/2022



#82
Pyrene
Concen: 37.413 ng/u1
RT: 19.916 min Scan# 2813
Delta R.T. 0.003 min
Lab File: BG052866.D
Acq: 26 Mar 2022 16:03


Tgt Ion:202 Resp: 679867
Ion Ratio Lower Upper
202 100
101 8.7 8.2 12.4
100 7.6 6.8 10.2





#83
Butylbenzylphthalate
Concen: 40.203 ng/ul
RT: 20.797 min Scan# 2962
Delta R.T. 0.003 min
Lab File: BG052866.D
Acq: 26 Mar 2022 16:03




Instrument :
BNA_G
ClientSampleId :
SLCS594




Tgt Ion:149 Resp: 254008
Ion Ratio Lower Upper
149 100
91 74.0 57.1 85.7
206 22.1 17.4 26.2

Manual Integrations
APPROVED



Reviewed By :Jagrut Upadhyay 03/28/2022
Supervised By :mohammad ahmed 03/29/2022




#84
3,3'-Dichlorobenzidine
Concen: 38.681 ng/ul
RT: 21.684 min Scan# 3114
Delta R.T. 0.003 min
Lab File: BG052866.D
Acq: 26 Mar 2022 16:03




Tgt Ion:252 Resp: 240433
Ion Ratio Lower Upper
252 100
254 64.7 48.6 73.0
126 7.5 6.2 9.2





#85
Benzo(a)anthracene
Concen: 37.487 ng/ul
RT: 21.772 min Scan# 3129
Delta R.T. -0.003 min
Lab File: BG052866.D
Acq: 26 Mar 2022 16:03




Instrument :
BNA_G
ClientSampleId :
SLCS594



Tgt Ion:228 Resp: 67098
Ion Ratio Lower Upper
228 100
229 20.5 15.4 23.2
226 27.3 21.4 32.2




Manual Integrations
APPROVED


Reviewed By :Jagrut Upadhyay 03/28/2022
Supervised By :mohammad ahmed 03/29/2022



#86
Bis(2-ethylhexyl)phthalate
Concen: 39.848 ng/ul
RT: 21.672 min Scan# 3112
Delta R.T. -0.003 min
Lab File: BG052866.D
Acq: 26 Mar 2022 16:03

Tgt Ion:149 Resp: 348705
Ion Ratio Lower Upper
149 100
167 26.2 21.4 32.2
279 4.7 3.4 5.2









#87
Chrysene
Concen: 37.278 ng/ul
RT: 21.843 min Scan# 3141
Delta R.T. 0.003 min
Lab File: BG052866.D
Acq: 26 Mar 2022 16:03

Instrument :
BNA_G
ClientSampleId :
SLCS594

Tgt Ion:228 Resp: 63643
Ion Ratio Lower Upper
228 100
226 31.4 23.9 35.9
229 20.1 16.0 24.0




Manual Integrations
APPROVED


Reviewed By :Jagrut Upadhyay 03/28/2022
Supervised By :mohammad ahmed 03/29/2022



#89
Di-n-octyl phthalate
Concen: 40.753 ng/ul
RT: 22.923 min Scan# 3325
Delta R.T. 0.003 min
Lab File: BG052866.D
Acq: 26 Mar 2022 16:03


Tgt Ion:149 Resp: 593883





#90
Benzo(b)fluoranthene
Concen: 39.738 ng/ul
RT: 24.057 min Scan# 3517
Delta R.T. 0.003 min
Lab File: BG052866.D
Acq: 26 Mar 2022 16:03




Instrument :
BNA_G
ClientSampleId :
SLCS594




Tgt Ion:252 Resp: 71790
Ion Ratio Lower Upper
252 100
253 22.6 17.8 26.6
125 5.9 5.5 8.3

Manual Integrations
APPROVED



Reviewed By :Jagrut Upadhyay 03/28/2022
Supervised By :mohammad ahmed 03/29/2022




#91
Benzo(k)fluoranthene
Concen: 38.714 ng/ul
RT: 24.128 min Scan# 3530
Delta R.T. 0.009 min
Lab File: BG052866.D
Acq: 26 Mar 2022 16:03




Tgt Ion:252 Resp: 645489
Ion Ratio Lower Upper
252 100
253 22.1 17.4 26.0
125 5.9 6.0 9.0





#93
Benzo(a)pyrene
Concen: 38.564 ng/ul
RT: 24.962 min Scan# 30
Delta R.T. 0.009 min
Lab File: BG052866.D
Acq: 26 Mar 2022 16:03




Instrument :
BNA_G
ClientSampleId :
SLCS594




Tgt Ion:252 Resp: 657770
Ion Ratio Lower Upper
252 100
253 22.1 16.5 24.7
125 6.0 6.0 9.0

Manual Integrations
APPROVED



Reviewed By :Jagrut Upadhyay 03/28/2022
Supervised By :mohammad ahmed 03/29/2022




#94
Indeno(1,2,3-cd)pyrene
Concen: 40.512 ng/ul m
RT: 28.898 min Scan# 4342
Delta R.T. 0.003 min
Lab File: BG052866.D
Acq: 26 Mar 2022 16:03




Tgt Ion:276 Resp: 781476
Ion Ratio Lower Upper
276 100
138 12.2 9.3 13.9
277 23.9 19.4 29.2





#95
Dibenzo(a,h)anthracene
Concen: 40.067 ng/ul
RT: 28.974 min Scan# 41
Delta R.T. 0.009 min
Lab File: BG052866.D
Acq: 26 Mar 2022 16:03




Instrument :
BNA_G
ClientSampleId :
SLCS594



Tgt Ion:278 Resp: 656919
Ion Ratio Lower Upper
278 100
139 9.5 8.1 12.1
279 23.3 19.0 28.4

Manual Integrations
APPROVED

Reviewed By :Jagrut Upadhyay 03/28/2022
Supervised By :mohammad ahmed 03/29/2022



#96
Benzo(g,h,i)perylene
Concen: 40.116 ng/ul
RT: 30.090 min Scan# 4545
Delta R.T. 0.009 min
Lab File: BG052866.D
Acq: 26 Mar 2022 16:03

Tgt Ion:276 Resp: 654151
Ion Ratio Lower Upper
276 100
138 13.3 10.5 15.7
277 24.4 19.0 28.6

