

Data Path : Z:\SVOASRV\HPCHEM1\BNA_G\DATA\BG032919\
 Data File : BG040248.D
 Acq On : 30 Mar 2019 5:56
 Operator : JU/SJ
 Sample : PB118419BL
 Misc :
 ALS Vial : 18 Sample Multiplier: 1

Instrument :
 BNA_G
 ClientSampleId :
 PB118419BL

Quant Time: Mar 30 07:04:05 2019
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_G\METHODS\8270-BG032719.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Thu Mar 28 05:47:00 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	8.06	152	54612	20.00	ng	-0.01
21) Naphthalene-d8	10.87	136	227108	20.00	ng	-0.01
39) Acenaphthene-d10	14.69	164	153959	20.00	ng	-0.02
64) Phenanthrene-d10	17.43	188	418896	20.00	ng	-0.02
76) Chrysene-d12	21.71	240	443381	20.00	ng	-0.02
87) Perylene-d12	25.01	264	439302	20.00	ng	-0.02
System Monitoring Compounds						
5) 2-Fluorophenol	5.61	112	397931	121.13	ng	-0.01
7) Phenol-d6	7.21	99	554025	122.03	ng	-0.02
23) Nitrobenzene-d5	9.23	82	313234	73.31	ng	-0.02
42) 2,4,6-Tribromophenol	16.18	330	216304	130.00	ng	-0.01
45) 2-Fluorobiphenyl	13.31	172	776353	74.72	ng	-0.02
79) Terphenyl-d14	20.03	244	1397491	70.91	ng	-0.02


Target Compounds	Qvalue
------------------	--------

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\SVOASRV\HPCHEM1\BNA_G\DATA\BG032919\
Data File : BG040248.D
Acq On : 30 Mar 2019 5:56
Operator : JU/SJ
Sample : PB118419BL
Misc :
ALS Vial : 18 Sample Multiplier: 1

Instrument :
BNA_G
ClientSampleId :
PB118419BL

Quant Time: Mar 30 07:04:05 2019
Quant Method : Z:\SVOASRV\HPCHEM1\BNA_G\METHODS\8270-BG032719.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Thu Mar 28 05:47:00 2019
Response via : Initial Calibration






#1



1,4-Dichlorobenzene-d4
Concen: 20.000 ng
RT: 8.06 min Scan# 846
Delta R.T. -0.01 min
Lab File: BG040248.D
Acq: 30 Mar 2019 5:56

Instrument :
BNA_G
ClientSampleId :
PB118419BL

Tgt Ion:152 Resp: 54612
Ion Ratio Lower Upper
152 100
150 154.5 123.7 185.5
115 54.3 46.9 70.3




Abundance Ion 152.00 (151.70 to 152.70): E
Ion 150.00 (149.70 to 150.70): E
Ion 115.00 (114.70 to 115.70): E





#5

2-Fluorophenol
Concen: 121.132 ng
RT: 5.61 min Scan# 429
Delta R.T. -0.01 min
Lab File: BG040248.D
Acq: 30 Mar 2019 5:56

Tgt Ion:112 Resp: 397931
Ion Ratio Lower Upper
112 100
64 65.7 54.2 81.2
63 32.3 26.4 39.6



Abundance Ion 112.00 (111.70 to 112.70): E
Ion 64.00 (63.70 to 64.70): BG
Ion 63.00 (62.70 to 63.70): BG





#7

Phenol-d6

Concen: 122.030 ng

RT: 7.21 min Scan# 701

Delta R.T. -0.02 min

Lab File: BG040248.D

Acq: 30 Mar 2019 5:56

Instrument :

BNA_G

ClientSampleId :

PB118419BL


Tgt Ion: 99 Resp: 554025

Ion Ratio Lower Upper



99 100

42 17.6 16.2 24.2

71 32.4 26.8 40.2



Abundance Ion 99.00 (98.70 to 99.70): BGC
Ion 42.00 (41.70 to 42.70): BGC
Ion 71.00 (70.70 to 71.70): BGC



#21

Naphthalene-d8

Concen: 20.000 ng

RT: 10.87 min Scan# 1325

Delta R.T. -0.01 min

Lab File: BG040248.D

Acq: 30 Mar 2019 5:56

Tgt Ion: 136 Resp: 227108


Ion Ratio Lower Upper

136 100



137 10.9 8.9 13.3


54 10.0 9.6 14.4

68 6.8 5.1 7.7



Abundance Ion 136.00 (135.70 to 136.70): BGC
Ion 137.00 (136.70 to 137.70): BGC
Ion 54.00 (53.70 to 54.70): BGC
Ion 68.00 (67.70 to 68.70): BGC









#23
Nitrobenzene-d5
Concen: 73.311 ng
RT: 9.23 min Scan# 1045
Delta R.T. -0.02 min
Lab File: BG040248.D
Acq: 30 Mar 2019 5:56

Instrument :
BNA_G
ClientSampleId :
PB118419BL

Tgt Ion: 82 Resp: 313234
Ion Ratio Lower Upper
82 100
128 43.8 35.8 53.8
54 55.1 46.2 69.2




Abundance Ion 82.00 (81.70 to 82.70): BGC
Ion 128.00 (127.70 to 128.70): BGC
Ion 54.00 (53.70 to 54.70): BGC






#39
Acenaphthene-d10
Concen: 20.000 ng
RT: 14.69 min Scan# 1974
Delta R.T. -0.02 min
Lab File: BG040248.D
Acq: 30 Mar 2019 5:56

Tgt Ion: 164 Resp: 153959
Ion Ratio Lower Upper
164 100
162 98.1 77.4 116.2
160 45.4 35.3 52.9



Abundance Ion 164.00 (163.70 to 164.70): BGC
Ion 162.00 (161.70 to 162.70): BGC
Ion 160.00 (159.70 to 160.70): BGC









#42
 2,4,6-Tribromophenol
 Concen: 129.999 ng
 RT: 16.18 min Scan# 2228
 Delta R.T. -0.01 min
 Lab File: BG040248.D
 Acq: 30 Mar 2019 5:56

Instrument :
 BNA_G
 ClientSampleId :
 PB118419BL

Tgt Ion:330 Resp: 216304
 Ion Ratio Lower Upper
 330 100
 332 97.9 76.6 115.0
 141 38.8 32.6 48.8




Abundance Ion 329.80 (329.50 to 330.50): E
 Ion 331.80 (331.50 to 332.50): E
 Ion 141.00 (140.70 to 141.70): E






#45
 2-Fluorobiphenyl
 Concen: 74.719 ng
 RT: 13.31 min Scan# 1739
 Delta R.T. -0.02 min
 Lab File: BG040248.D
 Acq: 30 Mar 2019 5:56

Tgt Ion:172 Resp: 776353
 Ion Ratio Lower Upper
 172 100
 171 37.1 30.5 45.7
 170 24.6 20.1 30.1



Abundance Ion 172.00 (171.70 to 172.70): E
 Ion 171.00 (170.70 to 171.70): E
 Ion 170.00 (169.70 to 170.70): E









#64
 Phenanthrene-d10
 Concen: 20.000 ng
 RT: 17.43 min Scan# 2441
 Delta R.T. -0.02 min
 Lab File: BG040248.D
 Acq: 30 Mar 2019 5:56

Instrument :
 BNA_G
 ClientSampleId :
 PB118419BL

Tgt Ion:188 Resp: 418896
 Ion Ratio Lower Upper
 188 100
 94 8.6 7.0 10.4
 80 10.1 8.6 13.0




Abundance Ion 188.00 (187.70 to 188.70): E
 Ion 94.00 (93.70 to 94.70): BG
 Ion 80.00 (79.70 to 80.70): BG





#76
 Chrysene-d12
 Concen: 20.000 ng
 RT: 21.71 min Scan# 3169
 Delta R.T. -0.02 min
 Lab File: BG040248.D
 Acq: 30 Mar 2019 5:56

Tgt Ion:240 Resp: 443381
 Ion Ratio Lower Upper
 240 100
 120 9.2 8.4 12.6
 236 25.7 20.8 31.2



Abundance Ion 240.00 (239.70 to 240.70): E
 Ion 120.00 (119.70 to 120.70): E
 Ion 236.00 (235.70 to 236.70): E





#79

Terphenyl-d14

Concen: 70.907 ng

RT: 20.03 min Scan# 2884

Delta R.T. -0.02 min

Lab File: BG040248.D

Acq: 30 Mar 2019 5:56

Instrument :

BNA_G

ClientSampleId :

PB118419BL


Tgt Ion:244 Resp: 1397491

Ion Ratio Lower Upper




244 100

212 8.7 7.0 10.6

122 11.7 8.5 12.7



Abundance Ion 244.00 (243.70 to 244.70): E
Ion 212.00 (211.70 to 212.70): E
Ion 122.00 (121.70 to 122.70): E



#87

Perylene-d12

Concen: 20.000 ng

RT: 25.01 min Scan# 3731

Delta R.T. -0.02 min

Lab File: BG040248.D

Acq: 30 Mar 2019 5:56


Tgt Ion:264 Resp: 439302

Ion Ratio Lower Upper

264 100

260 23.0 17.9 26.9

265 22.0 16.8 25.2



Abundance Ion 264.00 (263.70 to 264.70): E
Ion 260.00 (259.70 to 260.70): E
Ion 265.00 (264.70 to 265.70): E

