

Data Path : Z:\svoasrv\HPCHEM1\BNA_G\Data\BG040121\
 Data File : BG048536.D
 Acq On : 1 Apr 2021 9:16
 Operator : CG/JU
 Sample : SSTDCCC040
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_G
 ClientSampleId :
 SSTDCCC040

Quant Time: Apr 01 11:20:11 2021
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_G\METHODS\8270-BG033021.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Thu Apr 01 11:19:59 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	8.093	152	21343	20.00	ng	# 0.00
21) Naphthalene-d8	10.925	136	88413	20.00	ng	# 0.00
39) Acenaphthene-d10	14.750	164	63637	20.00	ng	0.00
64) Phenanthrene-d10	17.505	188	149871	20.00	ng	0.00
76) Chrysene-d12	21.800	240	142999	20.00	ng	# 0.00
86) Perylene-d12	25.143	264	143798	20.00	ng	# 0.00
System Monitoring Compounds						
5) 2-Fluorophenol	5.637	112	101229	83.74	ng	0.00
7) Phenol-d6	7.247	99	145286	82.87	ng	0.00
23) Nitrobenzene-d5	9.268	82	149434	81.83	ng	0.00
42) 2,4,6-Tribromophenol	16.242	330	68903	82.37	ng	0.00
45) 2-Fluorobiphenyl	13.369	172	343885	80.32	ng	0.00
79) Terphenyl-d14	20.114	244	621234	79.45	ng	0.00
Target Compounds						Qvalue
2) 1,4-Dioxane	3.510	88	21122	43.43	ng	95
3) Pyridine	3.915	79	56330	46.20	ng	95
4) n-Nitrosodimethylamine	3.827	42	34828	43.72	ng	93
6) Aniline	7.417	93	84744	42.09	ng	97
8) 2-Chlorophenol	7.658	128	56559	41.40	ng	94
9) Benzaldehyde	7.229	77	46756	41.72	ng	96
10) Phenol	7.276	94	69604	39.65	ng	92
11) bis(2-Chloroethyl)ether	7.511	93	51617	42.38	ng	98
12) 1,3-Dichlorobenzene	7.987	146	62877	40.86	ng	96
13) 1,4-Dichlorobenzene	8.134	146	64539	41.59	ng	97
14) 1,2-Dichlorobenzene	8.457	146	59367	39.95	ng	95
15) Benzyl Alcohol	8.334	79	59467	41.41	ng	95
16) 2,2'-oxybis(1-Chloropr...	8.621	45	47616	40.53	ng	99
17) 2-Methylphenol	8.539	107	44670	39.33	ng	91
18) Hexachloroethane	9.191	117	24007	41.64	ng	# 81
19) n-Nitroso-di-n-propyla...	8.903	70	49975	41.35	ng	96
20) 3+4-Methylphenols	8.862	107	63993	40.59	ng	93
22) Acetophenone	8.921	105	89090	39.08	ng	# 97
24) Nitrobenzene	9.309	77	74453	41.54	ng	93
25) Isophorone	9.832	82	134192	39.79	ng	# 98
26) 2-Nitrophenol	10.026	139	33680	40.90	ng	94
27) 2,4-Dimethylphenol	10.079	122	51698	39.90	ng	96
28) bis(2-Chloroethoxy)met...	10.319	93	61242	39.62	ng	95
29) 2,4-Dichlorophenol	10.566	162	60050	40.93	ng	95
30) 1,2,4-Trichlorobenzene	10.784	180	69145	41.01	ng	93
31) Naphthalene	10.977	128	185179	40.74	ng	99
32) Benzoic acid	10.214	122	41876	42.01	ng	96
33) 4-Chloroaniline	11.077	127	78063	40.69	ng	96
34) Hexachlorobutadiene	11.259	225	50635	41.13	ng	91
35) Caprolactam	11.847	113	21617	40.39	ng	# 69
36) 4-Chloro-3-methylphenol	12.200	107	65888	39.99	ng	98
37) 2-Methylnaphthalene	12.581	142	144578	39.95	ng	95
38) 1-Methylnaphthalene	12.799	142	137373	39.80	ng	99
40) 1,2,4,5-Tetrachloroben...	12.946	216	80683	40.90	ng	98
41) Hexachlorocyclopentadiene	12.922	237	49574	43.38	ng	89

Data Path : Z:\svoasrv\HPCHEM1\BNA_G\Data\BG040121\
 Data File : BG048536.D
 Acq On : 1 Apr 2021 9:16
 Operator : CG/JU
 Sample : SSTDCCC040
 Misc :
 ALS Vial : 2 Sample Multiplier: 1


Instrument :
 BNA_G
 ClientSampleId :
 SSTDCCC040


Quant Time: Apr 01 11:20:11 2021
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_G\METHODS\8270-BG033021.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Thu Apr 01 11:19:59 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
43) 2,4,6-Trichlorophenol	13.181	196	56842	41.92	ng	95
44) 2,4,5-Trichlorophenol	13.251	196	60295	41.56	ng	97
46) 1,1'-Biphenyl	13.580	154	188052	40.44	ng	97
47) 2-Chloronaphthalene	13.627	162	142090	40.11	ng	96
48) 2-Nitroaniline	13.821	65	47576	42.25	ng	# 82
49) Acenaphthylene	14.473	152	224086	40.35	ng	97
50) Dimethylphthalate	14.197	163	191244	40.49	ng	99
51) 2,6-Dinitrotoluene	14.315	165	44230	40.93	ng	97
52) Acenaphthene	14.814	154	148429	36.95	ng	97
53) 3-Nitroaniline	14.644	138	41984	39.62	ng	89
54) 2,4-Dinitrophenol	14.849	184	26246	43.19	ng	87
55) Dibenzofuran	15.149	168	230307	39.26	ng	96
56) 4-Nitrophenol	14.949	139	36415	42.57	ng	88
57) 2,4-Dinitrotoluene	15.102	165	62352	40.79	ng	96
58) Fluorene	15.801	166	192245	39.15	ng	99
59) 2,3,4,6-Tetrachlorophenol	15.372	232	57660	40.74	ng	99
60) Diethylphthalate	15.560	149	197217	40.64	ng	# 97
61) 4-Chlorophenyl-phenyle...	15.789	204	107049	40.52	ng	97
62) 4-Nitroaniline	15.813	138	47580	42.93	ng	95
63) Azobenzene	16.083	77	187837	39.34	ng	94
65) 4,6-Dinitro-2-methylph...	15.866	198	39893	44.30	ng	96
66) n-Nitrosodiphenylamine	16.001	169	173198	40.62	ng	98
67) 4-Bromophenyl-phenylether	16.688	248	69015	41.54	ng	93
68) Hexachlorobenzene	16.806	284	71309	41.11	ng	95
69) Atrazine	16.947	200	72031	41.42	ng	98
70) Pentachlorophenol	17.147	266	46007	42.67	ng	94
71) Phenanthrene	17.546	178	316401	40.67	ng	98
72) Anthracene	17.640	178	318040	41.02	ng	100
73) Carbazole	17.905	167	296381	40.31	ng	97
74) Di-n-butylphthalate	18.457	149	333875	40.65	ng	97
75) Fluoranthene	19.556	202	387214	40.15	ng	97
77) Benzidine	19.732	184	199403	41.18	ng	100
78) Pyrene	19.920	202	389425	39.61	ng	99
80) Butylbenzylphthalate	20.795	149	146497	40.12	ng	90
81) Benzo(a)anthracene	21.782	228	383029	40.09	ng	95
82) 3,3'-Dichlorobenzidine	21.688	252	127575	38.87	ng	97
83) Chrysene	21.847	228	362134	39.72	ng	95
84) Bis(2-ethylhexyl)phtha...	21.677	149	209544	41.19	ng	98
85) Di-n-octyl phthalate	22.934	149	361564	40.77	ng	99
87) Indeno(1,2,3-cd)pyrene	28.974	276	418433	41.51	ng	# 90
88) Benzo(b)fluoranthene	24.150	252	368667	39.47	ng	96
89) Benzo(k)fluoranthene	24.150	252	368667	41.06	ng	97
90) Benzo(a)pyrene	24.985	252	354610	41.20	ng	99
91) Dibenzo(a,h)anthracene	29.062	278	355115	41.27	ng	98
92) Benzo(g,h,i)perylene	30.178	276	347120	41.19	ng	92

(#) = qualifier out of range (m) = manual integration (+) = signals summed


Quant Time: Apr 01 11:20:11 2021
Quant Method : Z:\SVOASRV\HPCHEM1\BNA_G\METHODS\8270-BG033021.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Thu Apr 01 11:19:59 2021
Response via : Initial Calibration








1,4-Dichlorobenzene-d4
 Concen: 20.00 ng
 RT: 8.093 min Scan# 809
 Delta R.T. -0.000 min
 Lab File: BG048536.D
 Acq: 1 Apr 2021 9:16

Instrument :
 BNA_G
 ClientSampleId :
 SSTDCCC040







Tgt Ion:	152	Resp:	21343
Ion	Ratio	Lower	Upper
152	100		
150	156.4	101.3	151.9#
115	66.5	42.3	63.5#



1,4-Dioxane
 Concen: 43.43 ng
 RT: 3.510 min Scan# 29
 Delta R.T. 0.000 min
 Lab File: BG048536.D
 Acq: 1 Apr 2021 9:16

Tgt Ion:	88	Resp:	21122
Ion	Ratio	Lower	Upper
88	100		
58	63.5	55.0	82.6
43	31.6	26.2	39.4









Pyridine
 Concen: 46.20 ng
 RT: 3.915 min Scan# 98
 Delta R.T. 0.000 min
 Lab File: BG048536.D
 Acq: 1 Apr 2021 9:16




Instrument :
 BNA_G
 ClientSampleId :
 SSTDCCC040


Tgt Ion:	79	Resp:	56330
Ion	Ratio	Lower	Upper
79	100		
52	65.6	48.2	72.2
51	43.5	33.2	49.8



n-Nitrosodimethylamine
 Concen: 43.72 ng
 RT: 3.827 min Scan# 83
 Delta R.T. 0.000 min
 Lab File: BG048536.D
 Acq: 1 Apr 2021 9:16

Tgt Ion:	42	Resp:	34828
Ion	Ratio	Lower	Upper
42	100		
74	95.4	70.6	105.8
44	4.5	3.5	5.3









2-Fluorophenol
 Concen: 83.74 ng
 RT: 5.637 min Scan# 391
 Delta R.T. 0.000 min
 Lab File: BG048536.D
 Acq: 1 Apr 2021 9:16




Instrument :
 BNA_G
 ClientSampleId :
 SSTDCCC040

Tgt Ion:	112	Resp:	101229
Ion	Ratio	Lower	Upper
112	100		
64	54.5	50.4	75.6
63	35.0	31.4	47.2




Aniline
 Concen: 42.09 ng
 RT: 7.417 min Scan# 694
 Delta R.T. 0.000 min
 Lab File: BG048536.D
 Acq: 1 Apr 2021 9:16

Tgt Ion:	93	Resp:	84744
Ion	Ratio	Lower	Upper
93	100		
66	38.4	31.8	47.6
65	16.9	15.2	22.8





Abundance Scan 665 (7.248 min): BG048516.D\data.ms (-656) #7




Phenol-d6
Concen: 82.87 ng
RT: 7.247 min Scan# 665
Delta R.T. 0.000 min
Lab File: BG048536.D
Acq: 1 Apr 2021 9:16

Instrument :
BNA_G
ClientSampleId :
SSTDCCC040


Tgt Ion:	99	Resp:	145286
Ion	Ratio	Lower	Upper
99	100		
42	23.1	15.5	23.3
71	46.6	36.6	54.8



Abundance Scan 665 (7.247 min): BG048536.D\data.ms (-575) (-)





Abundance Scan 735 (7.659 min): BG048516.D\data.ms (-727) #8





2-Chlorophenol
Concen: 41.40 ng
RT: 7.658 min Scan# 735
Delta R.T. 0.000 min
Lab File: BG048536.D
Acq: 1 Apr 2021 9:16

Tgt Ion:	128	Resp:	56559
Ion	Ratio	Lower	Upper
128	100		
130	30.6	12.4	52.4
64	42.6	27.6	67.6



Abundance Scan 735 (7.658 min): BG048536.D\data.ms (-645) (-)









Benzaldehyde
 Concen: 41.72 ng
 RT: 7.229 min Scan# 662
 Delta R.T. 0.000 min
 Lab File: BG048536.D
 Acq: 1 Apr 2021 9:16




Instrument :
 BNA_G
ClientSampleId :
 SSTDCCC040

Tgt Ion:	77	Resp:	46756
Ion	Ratio	Lower	Upper
77	100		
105	90.0	75.2	115.2
106	84.9	66.5	106.5



Phenol
 Concen: 39.65 ng
 RT: 7.276 min Scan# 670
 Delta R.T. 0.000 min
 Lab File: BG048536.D
 Acq: 1 Apr 2021 9:16


Tgt Ion:	94	Resp:	69604
Ion	Ratio	Lower	Upper
94	100		
65	31.5	15.1	55.1
66	55.2	28.9	68.9






Abundance Scan 709 (7.507 min): BG048516.D\data.ms (-703) #11

bis(2-Chloroethyl)ether
Concen: 42.38 ng
RT: 7.511 min Scan# 710
Delta R.T. 0.000 min
Lab File: BG048536.D
Acq: 1 Apr 2021 9:16

Instrument :
BNA_G
ClientSampleId :
SSTDCCC040







Tgt Ion:	93	Resp:	51617
Ion	Ratio	Lower	Upper
93	100		
63	69.7	47.5	87.5
95	30.6	11.1	51.1




Abundance Scan 790 (7.983 min): BG048516.D\data.ms (-783) #12

1,3-Dichlorobenzene
Concen: 40.86 ng
RT: 7.987 min Scan# 791
Delta R.T. 0.000 min
Lab File: BG048536.D
Acq: 1 Apr 2021 9:16

Tgt Ion:	146	Resp:	62877
Ion	Ratio	Lower	Upper
146	100		
148	66.2	49.7	74.5
75	39.5	30.9	46.3






Abundance Scan 815 (8.130 min): BG048516.D\data.ms (-806) #13




1,4-Dichlorobenzene
Concen: 41.59 ng
RT: 8.134 min Scan# 816
Delta R.T. 0.000 min
Lab File: BG048536.D
Acq: 1 Apr 2021 9:16

Instrument :
BNA_G
ClientSampleId :
SSTDCCC040

Tgt Ion:	146	Resp:	64539
Ion	Ratio	Lower	Upper
146	100		
148	58.5	49.7	74.5
111	37.6	30.5	45.7






Abundance Scan 870 (8.453 min): BG048516.D\data.ms (-864) #14




1,2-Dichlorobenzene
Concen: 39.95 ng
RT: 8.457 min Scan# 871
Delta R.T. 0.000 min
Lab File: BG048536.D
Acq: 1 Apr 2021 9:16

Tgt Ion:	146	Resp:	59367
Ion	Ratio	Lower	Upper
146	100		
148	65.6	51.5	77.3
111	45.1	41.8	62.8




Abundance Scan 850 (8.335 min): BG048516.D\data.ms (-842) #45





Benzyl Alcohol
Concen: 41.41 ng
RT: 8.334 min Scan# 850
Delta R.T. 0.000 min
Lab File: BG048536.D
Acq: 1 Apr 2021 9:16


Instrument :
BNA_G
ClientSampleId :
SSTDCCC040



Tgt Ion:	79	Resp:	59467
Ion	Ratio	Lower	Upper
79	100		
108	66.3	54.1	81.1
77	71.7	52.6	79.0






Abundance Scan 899 (8.623 min): BG048516.D\data.ms (-891) #46



2,2'-oxybis(1-Chloropropane)
Concen: 40.53 ng
RT: 8.621 min Scan# 899
Delta R.T. 0.000 min
Lab File: BG048536.D
Acq: 1 Apr 2021 9:16


Tgt Ion:	45	Resp:	47616
Ion	Ratio	Lower	Upper
45	100		
77	19.4	0.0	38.4
79	14.4	0.0	34.5






Abundance Scan 884 (8.535 min): BG048516.D\data.ms (-876) #17

2-Methylphenol
Concen: 39.33 ng
RT: 8.539 min Scan# 885
Delta R.T. 0.000 min
Lab File: BG048536.D
Acq: 1 Apr 2021 9:16

Instrument :
BNA_G
ClientSampleId :
SSTDCCC040




Tgt Ion:	107	Resp:	44670
Ion Ratio	Lower	Upper	
107	100		
108	112.3	101.7	152.5
77	62.7	54.4	81.6
79	57.5	46.8	70.2







Abundance Scan 996 (9.193 min): BG048516.D\data.ms (-989) #18

Hexachloroethane
Concen: 41.64 ng
RT: 9.191 min Scan# 996
Delta R.T. 0.000 min
Lab File: BG048536.D
Acq: 1 Apr 2021 9:16



Tgt Ion:	117	Resp:	24007
Ion Ratio	Lower	Upper	
117	100		
119	95.9	67.0	100.6
201	123.1	79.8	119.6#









n-Nitroso-di-n-propylamine
 Concen: 41.35 ng
 RT: 8.903 min Scan# 947
 Delta R.T. 0.000 min
 Lab File: BG048536.D
 Acq: 1 Apr 2021 9:16




Instrument :
 BNA_G
 ClientSampleId :
 SSTDCCC040


Tgt Ion:	70	Resp:	49975
Ion	Ratio	Lower	Upper
70	100		
42	60.2	50.8	76.2
101	10.0	9.7	14.5
130	23.1	17.4	26.0



3+4-Methylphenols
 Concen: 40.59 ng
 RT: 8.862 min Scan# 940
 Delta R.T. 0.000 min
 Lab File: BG048536.D
 Acq: 1 Apr 2021 9:16

Tgt Ion:	107	Resp:	63993
Ion	Ratio	Lower	Upper
107	100		
108	91.3	62.3	102.3
77	39.4	21.3	61.3
79	40.3	18.1	58.1









Naphthalene-d8
 Concen: 20.00 ng
 RT: 10.925 min Scan# 1291
 Delta R.T. 0.000 min
 Lab File: BG048536.D
 Acq: 1 Apr 2021 9:16




Instrument :
 BNA_G
 ClientSampleId :
 SSTDCCC040

Tgt Ion:	136	Resp:	88413
Ion	Ratio	Lower	Upper
136	100		
137	12.9	8.6	12.8#
54	5.7	4.2	6.4
68	2.9	3.0	4.4#




Acetophenone
 Concen: 39.08 ng
 RT: 8.921 min Scan# 950
 Delta R.T. 0.000 min
 Lab File: BG048536.D
 Acq: 1 Apr 2021 9:16

Tgt Ion:	105	Resp:	89090
Ion	Ratio	Lower	Upper
105	100		
71	3.3	1.1	1.7#
51	31.4	26.1	39.1
120	19.9	17.5	26.3





Abundance Scan 1008 (9.263 min): BG048516.D\data.ms (-10023-)




Nitrobenzene-d5
Concen: 81.83 ng
RT: 9.268 min Scan# 1009
Delta R.T. 0.000 min
Lab File: BG048536.D
Acq: 1 Apr 2021 9:16

Instrument :
BNA_G
ClientSampleId :
SSTDCCC040


Tgt Ion:	82	Resp:	149434
Ion	Ratio	Lower	Upper
82	100		
128	39.0	26.2	39.4
54	47.0	38.9	58.3



Abundance Scan 1009 (9.268 min): BG048536.D\data.ms (-919) (-)





Abundance Scan 1016 (9.310 min): BG048516.D\data.ms (-10024-)




Nitrobenzene
Concen: 41.54 ng
RT: 9.309 min Scan# 1016
Delta R.T. 0.000 min
Lab File: BG048536.D
Acq: 1 Apr 2021 9:16


Tgt Ion:	77	Resp:	74453
Ion	Ratio	Lower	Upper
77	100		
123	35.9	33.4	50.2
65	17.5	13.3	19.9



Abundance Scan 1016 (9.309 min): BG048536.D\data.ms (-926) (-)





Abundance Scan 1105 (9.833 min): BG048516.D\data.ms (-1075) (-)




Isophorone
Concen: 39.79 ng
RT: 9.832 min Scan# 1105
Delta R.T. 0.000 min
Lab File: BG048536.D
Acq: 1 Apr 2021 9:16

Instrument :
BNA_G
ClientSampleId :
SSTDCCC040


Tgt Ion:	82	Resp:	134192
Ion	Ratio	Lower	Upper
82	100		
95	10.8	7.1	10.7#
138	18.1	14.8	22.2



Abundance Scan 1105 (9.832 min): BG048536.D\data.ms (-1015) (-)





Abundance Scan 1138 (10.027 min): BG048516.D\data.ms (-1126) (-)




2-Nitrophenol
Concen: 40.90 ng
RT: 10.026 min Scan# 1138
Delta R.T. 0.000 min
Lab File: BG048536.D
Acq: 1 Apr 2021 9:16

Tgt Ion:	139	Resp:	33680
Ion	Ratio	Lower	Upper
139	100		
109	42.5	33.7	50.5
65	66.0	46.8	70.2




Abundance Scan 1138 (10.026 min): BG048536.D\data.ms (-1048) (-)






Abundance Scan 1147 (10.080 min): BG048516.D\data.ms (-121) (-)

2,4-Dimethylphenol
Concen: 39.90 ng
RT: 10.079 min Scan# 1147
Delta R.T. 0.000 min
Lab File: BG048536.D
Acq: 1 Apr 2021 9:16

Instrument :
BNA_G
ClientSampleId :
SSTDCCC040




Tgt Ion:	122	Resp:	51698
Ion	Ratio	Lower	Upper
122	100		
107	126.3	96.6	145.0
121	57.6	45.3	67.9






Abundance Scan 1187 (10.315 min): BG048516.D\data.ms (-123) (-)

bis(2-Chloroethoxy)methane
Concen: 39.62 ng
RT: 10.319 min Scan# 1188
Delta R.T. 0.000 min
Lab File: BG048536.D
Acq: 1 Apr 2021 9:16




Tgt Ion:	93	Resp:	61242
Ion	Ratio	Lower	Upper
93	100		
95	29.7	27.0	40.6
123	11.4	9.0	13.6






Abundance Scan 1230 (10.568 min): BG048516.D\data.ms (-1229) (-)

2,4-Dichlorophenol
Concen: 40.93 ng
RT: 10.566 min Scan# 1230
Delta R.T. 0.000 min
Lab File: BG048536.D
Acq: 1 Apr 2021 9:16

Instrument :
BNA_G
ClientSampleId :
SSTDCCC040




Tgt Ion:	162	Resp:	60050
Ion Ratio	Lower	Upper	
162	100		
164	64.9	51.2	91.2
98	33.6	13.6	53.6







Abundance Scan 1267 (10.785 min): BG048516.D\data.ms (-1230) (-)

1,2,4-Trichlorobenzene
Concen: 41.01 ng
RT: 10.784 min Scan# 1267
Delta R.T. 0.000 min
Lab File: BG048536.D
Acq: 1 Apr 2021 9:16



Tgt Ion:	180	Resp:	69145
Ion Ratio	Lower	Upper	
180	100		
182	89.5	78.8	118.2
145	30.4	24.0	36.0









Naphthalene
Concen: 40.74 ng
RT: 10.977 min Scan# 1300
Delta R.T. 0.000 min
Lab File: BG048536.D
Acq: 1 Apr 2021 9:16




Instrument :
BNA_G
ClientSampleId :
SSTDCCC040

Tgt Ion:	128	Resp:	185179
Ion	Ratio	Lower	Upper
128	100		
129	10.2	8.2	12.2
127	13.7	10.6	16.0



Benzoic acid
Concen: 42.01 ng
RT: 10.214 min Scan# 1170
Delta R.T. 0.000 min
Lab File: BG048536.D
Acq: 1 Apr 2021 9:16


Tgt Ion:	122	Resp:	41876
Ion	Ratio	Lower	Upper
122	100		
105	125.0	97.1	137.1
77	97.4	76.8	116.8





Abundance Scan 1316 (11.073 min): BG048516.D\data.ms (-1339) (-)

4-Chloroaniline
Concen: 40.69 ng
RT: 11.077 min Scan# 1317
Delta R.T. 0.000 min
Lab File: BG048536.D
Acq: 1 Apr 2021 9:16


Instrument :
BNA_G
ClientSampleId :
SSTDCCC040



Tgt Ion:	127	Resp:	78063
Ion	Ratio	Lower	Upper
127	100		
129	33.3	24.3	36.5
65	35.9	27.0	40.4
92	19.4	16.4	24.6




Abundance Scan 1317 (11.077 min): BG048536.D\data.ms (-1227) (-)





Abundance Scan 1348 (11.261 min): BG048516.D\data.ms (-1549) (-)


Hexachlorobutadiene
Concen: 41.13 ng
RT: 11.259 min Scan# 1348
Delta R.T. 0.000 min
Lab File: BG048536.D
Acq: 1 Apr 2021 9:16




Tgt Ion:	225	Resp:	50635
Ion	Ratio	Lower	Upper
225	100		
223	60.7	44.8	67.2
227	64.9	45.4	68.0



Abundance Scan 1348 (11.259 min): BG048536.D\data.ms (-1258) (-)









Caprolactam
Concen: 40.39 ng
RT: 11.847 min Scan# 1448
Delta R.T. 0.000 min
Lab File: BG048536.D
Acq: 1 Apr 2021 9:16




Instrument :
BNA_G
ClientSampleId :
SSTDCCC040

Tgt Ion:113 Resp: 21617
Ion Ratio Lower Upper
113 100
55 134.8 76.8 116.8#
56 113.3 71.1 111.1#




4-Chloro-3-methylphenol
Concen: 39.99 ng
RT: 12.200 min Scan# 1508
Delta R.T. 0.000 min
Lab File: BG048536.D
Acq: 1 Apr 2021 9:16

Tgt Ion:107 Resp: 65888
Ion Ratio Lower Upper
107 100
144 27.0 20.6 31.0
142 85.4 67.0 100.6






Abundance Scan 1572 (12.577 min): BG048516.D\data.ms (-1537) (-)




2-Methylnaphthalene
Concen: 39.95 ng
RT: 12.581 min Scan# 1573
Delta R.T. 0.000 min
Lab File: BG048536.D
Acq: 1 Apr 2021 9:16

Instrument :
BNA_G
ClientSampleId :
SSTDCCC040

Tgt Ion:	142	Resp:	144578
Ion	Ratio	Lower	Upper
142	100		
141	89.9	68.2	102.4
115	41.3	30.8	46.2






Abundance Scan 1610 (12.801 min): BG048516.D\data.ms (-1635) (-)




1-Methylnaphthalene
Concen: 39.80 ng
RT: 12.799 min Scan# 1610
Delta R.T. 0.000 min
Lab File: BG048536.D
Acq: 1 Apr 2021 9:16

Tgt Ion:	142	Resp:	137373
Ion	Ratio	Lower	Upper
142	100		
141	89.9	71.2	106.8
116	3.9	3.4	5.0






Abundance Scan 1942 (14.751 min): BG048516.D\data.ms (-1539) (-)




Acenaphthene-d10
Concen: 20.00 ng
RT: 14.750 min Scan# 1942
Delta R.T. 0.000 min
Lab File: BG048536.D
Acq: 1 Apr 2021 9:16

Instrument :
BNA_G
ClientSampleId :
SSTDCCC040

Tgt Ion:	164	Resp:	63637
Ion	Ratio	Lower	Upper
164	100		
162	109.9	87.0	130.6
160	43.9	38.4	57.6






Abundance Scan 1634 (12.942 min): BG048516.D\data.ms (-1641) (-)



1,2,4,5-Tetrachlorobenzene
Concen: 40.90 ng
RT: 12.946 min Scan# 1635
Delta R.T. 0.000 min
Lab File: BG048536.D
Acq: 1 Apr 2021 9:16


Tgt Ion:	216	Resp:	80683
Ion	Ratio	Lower	Upper
216	100		
214	78.8	61.1	91.7
179	21.5	17.3	25.9
108	17.7	15.2	22.8






Abundance Scan 1631 (12.924 min): BG048516.D\data.ms (-1641) (-)

Hexachlorocyclopentadiene
Concen: 43.38 ng
RT: 12.922 min Scan# 1631
Delta R.T. 0.000 min
Lab File: BG048536.D
Acq: 1 Apr 2021 9:16

Instrument :
BNA_G
ClientSampleId :
SSTDCCC040




Tgt Ion:	237	Resp:	49574
Ion	Ratio	Lower	Upper
237	100		
235	68.9	39.7	79.7
272	10.8	0.0	28.8






Abundance Scan 2195 (16.238 min): BG048516.D\data.ms (-2195) (-)


2,4,6-Tribromophenol
Concen: 82.37 ng
RT: 16.242 min Scan# 2196
Delta R.T. 0.000 min
Lab File: BG048536.D
Acq: 1 Apr 2021 9:16



Tgt Ion:	330	Resp:	68903
Ion	Ratio	Lower	Upper
330	100		
332	94.3	77.6	116.4
141	41.8	34.6	52.0






Abundance Scan 1675 (13.182 min): BG048516.D\data.ms (-1642) (-)




2,4,6-Trichlorophenol
Concen: 41.92 ng
RT: 13.181 min Scan# 1675
Delta R.T. 0.000 min
Lab File: BG048536.D
Acq: 1 Apr 2021 9:16

Instrument :
BNA_G
ClientSampleId :
SSTDCCC040

Tgt Ion:	196	Resp:	56842
Ion	Ratio	Lower	Upper
196	100		
198	101.6	75.8	113.6
200	33.5	27.0	40.6






Abundance Scan 1687 (13.253 min): BG048516.D\data.ms (-1641) (-)




2,4,5-Trichlorophenol
Concen: 41.56 ng
RT: 13.251 min Scan# 1687
Delta R.T. 0.000 min
Lab File: BG048536.D
Acq: 1 Apr 2021 9:16

Tgt Ion:	196	Resp:	60295
Ion	Ratio	Lower	Upper
196	100		
198	94.4	77.1	115.7
97	50.5	37.5	56.3
132	21.6	17.3	25.9






Abundance Scan 1707 (13.370 min): BG048516.D\data.ms (-1695) (-)




2-Fluorobiphenyl
Concen: 80.32 ng
RT: 13.369 min Scan# 1707
Delta R.T. 0.000 min
Lab File: BG048536.D
Acq: 1 Apr 2021 9:16

Instrument :
BNA_G
ClientSampleId :
SSTDCCC040

Tgt Ion:	172	Resp:	343885
Ion	Ratio	Lower	Upper
172	100		
171	34.3	28.5	42.7
170	23.5	19.8	29.6






Abundance Scan 1743 (13.582 min): BG048516.D\data.ms (-1745) (-)




1,1'-Biphenyl
Concen: 40.44 ng
RT: 13.580 min Scan# 1743
Delta R.T. 0.000 min
Lab File: BG048536.D
Acq: 1 Apr 2021 9:16

Tgt Ion:	154	Resp:	188052
Ion	Ratio	Lower	Upper
154	100		
153	39.0	17.3	57.3
76	7.3	0.0	28.6





Abundance Scan 1751 (13.629 min): BG048516.D\data.ms (-1741) (-)




2-Chloronaphthalene
Concen: 40.11 ng
RT: 13.627 min Scan# 1751
Delta R.T. 0.000 min
Lab File: BG048536.D
Acq: 1 Apr 2021 9:16

Instrument :
BNA_G
ClientSampleId :
SSTDCCC040


Tgt Ion:	162	Resp:	142090
Ion	Ratio	Lower	Upper
162	100		
127	38.5	31.8	47.8
164	31.4	27.8	41.6



Abundance Scan 1751 (13.627 min): BG048536.D\data.ms (-1661) (-)





Abundance Scan 1784 (13.823 min): BG048516.D\data.ms (-1746) (-)




2-Nitroaniline
Concen: 42.25 ng
RT: 13.821 min Scan# 1784
Delta R.T. 0.000 min
Lab File: BG048536.D
Acq: 1 Apr 2021 9:16


Tgt Ion:	65	Resp:	47576
Ion	Ratio	Lower	Upper
65	100		
92	63.8	56.7	85.1
138	85.8	89.6	134.4#



Abundance Scan 1784 (13.821 min): BG048536.D\data.ms (-1694) (-)






Abundance Scan 1894 (14.469 min): BG048516.D\data.ms (-1849) (-)




Acenaphthylene
Concen: 40.35 ng
RT: 14.473 min Scan# 1895
Delta R.T. 0.000 min
Lab File: BG048536.D
Acq: 1 Apr 2021 9:16

Instrument :
BNA_G
ClientSampleId :
SSTDCCC040

Tgt Ion:	152	Resp:	224086
Ion	Ratio	Lower	Upper
152	100		
151	20.1	15.0	22.6
153	12.5	11.0	16.6





Abundance Scan 1847 (14.193 min): BG048516.D\data.ms (-1850) (-)




Dimethylphthalate
Concen: 40.49 ng
RT: 14.197 min Scan# 1848
Delta R.T. 0.000 min
Lab File: BG048536.D
Acq: 1 Apr 2021 9:16

Tgt Ion:	163	Resp:	191244
Ion	Ratio	Lower	Upper
163	100		
194	4.2	4.1	6.1
164	11.0	8.6	12.8




Abundance Scan 1848 (14.197 min): BG048536.D\data.ms (-1758) (-)





Abundance Scan 1868 (14.316 min): BG048516.D\data.ms (-1751) (-)

2,6-Dinitrotoluene
Concen: 40.93 ng
RT: 14.315 min Scan# 1868
Delta R.T. 0.000 min
Lab File: BG048536.D
Acq: 1 Apr 2021 9:16


Instrument :
BNA_G
ClientSampleId :
SSTDCCC040



Tgt Ion:	165	Resp:	44230
Ion	Ratio	Lower	Upper
165	100		
63	65.0	53.8	80.6
89	60.9	50.8	76.2




Abundance Scan 1868 (14.315 min): BG048536.D\data.ms (-1778) (-)





Abundance Scan 1953 (14.816 min): BG048516.D\data.ms (-1752) (-)

Acenaphthene
Concen: 36.95 ng
RT: 14.814 min Scan# 1953
Delta R.T. 0.000 min
Lab File: BG048536.D
Acq: 1 Apr 2021 9:16

Tgt Ion:	154	Resp:	148429
Ion	Ratio	Lower	Upper
154	100		
153	112.1	86.6	129.8
152	52.0	41.5	62.3




Abundance Scan 1953 (14.814 min): BG048536.D\data.ms (-1863) (-)






Abundance Scan 1924 (14.645 min): BG048516.D\data.ms (-1952) (-)

3-Nitroaniline
Concen: 39.62 ng
RT: 14.644 min Scan# 1924
Delta R.T. 0.000 min
Lab File: BG048536.D
Acq: 1 Apr 2021 9:16

Instrument :
BNA_G
ClientSampleId :
SSTDCCC040




Tgt Ion:	138	Resp:	41984
Ion	Ratio	Lower	Upper
138	100		
108	10.9	9.2	13.8
92	115.5	103.8	155.8






Abundance Scan 1959 (14.851 min): BG048516.D\data.ms (-1952) (-)


2,4-Dinitrophenol
Concen: 43.19 ng
RT: 14.849 min Scan# 1959
Delta R.T. 0.000 min
Lab File: BG048536.D
Acq: 1 Apr 2021 9:16



Tgt Ion:	184	Resp:	26246
Ion	Ratio	Lower	Upper
184	100		
63	80.7	56.6	84.8
154	80.3	54.9	82.3






Abundance Scan 2009 (15.145 min): BG048516.D\data.ms (-1920) (-)




Dibenzenofuran
Concen: 39.26 ng
RT: 15.149 min Scan# 2010
Delta R.T. 0.000 min
Lab File: BG048536.D
Acq: 1 Apr 2021 9:16

Instrument :
BNA_G
ClientSampleId :
SSTDCCC040

Tgt Ion:168 Resp: 230307
Ion Ratio Lower Upper
168 100
139 38.9 33.7 50.5
169 14.3 12.5 18.7







Abundance Scan 1976 (14.951 min): BG048516.D\data.ms (-1956) (-)



4-Nitrophenol
Concen: 42.57 ng
RT: 14.949 min Scan# 1976
Delta R.T. 0.000 min
Lab File: BG048536.D
Acq: 1 Apr 2021 9:16

Tgt Ion:139 Resp: 36415
Ion Ratio Lower Upper
139 100
109 117.1 113.0 153.0
65 117.8 109.4 149.4









2,4-Dinitrotoluene
Concen: 40.79 ng
RT: 15.102 min Scan# 2002
Delta R.T. 0.000 min
Lab File: BG048536.D
Acq: 1 Apr 2021 9:16




Instrument :
BNA_G
ClientSampleId :
SSTDCCC040

Tgt Ion:	165	Resp:	62352
Ion	Ratio	Lower	Upper
165	100		
63	49.0	38.3	57.5
89	75.2	56.5	84.7
182	2.5	1.8	2.6




Fluorene
Concen: 39.15 ng
RT: 15.801 min Scan# 2121
Delta R.T. 0.000 min
Lab File: BG048536.D
Acq: 1 Apr 2021 9:16

Tgt Ion:	166	Resp:	192245
Ion	Ratio	Lower	Upper
166	100		
165	95.3	76.8	115.2
167	13.5	11.1	16.7






Abundance Scan 2047 (15.368 min): BG048516.D\data.ms (-2059) (-)




2,3,4,6-Tetrachlorophenol
Concen: 40.74 ng
RT: 15.372 min Scan# 2048
Delta R.T. 0.000 min
Lab File: BG048536.D
Acq: 1 Apr 2021 9:16

Instrument :
BNA_G
ClientSampleId :
SSTDCCC040

Tgt Ion:	232	Resp:	57660
Ion Ratio	Lower	Upper	
232	100		
131	42.4	35.0	52.4
130	2.8	2.4	3.6
166	29.4	23.6	35.4






Abundance Scan 2079 (15.556 min): BG048516.D\data.ms (-2060) (-)




Diethylphthalate
Concen: 40.64 ng
RT: 15.560 min Scan# 2080
Delta R.T. 0.000 min
Lab File: BG048536.D
Acq: 1 Apr 2021 9:16

Tgt Ion:	149	Resp:	197217
Ion Ratio	Lower	Upper	
149	100		
177	22.9	18.1	27.1
150	14.4	9.1	13.7#





Abundance Scan 2118 (15.785 min): BG048516.D\data.ms (-2161) (-)




4-Chlorophenyl-phenylether
Concen: 40.52 ng
RT: 15.789 min Scan# 2119
Delta R.T. 0.000 min
Lab File: BG048536.D
Acq: 1 Apr 2021 9:16

Instrument :
BNA_G
ClientSampleId :
SSTDCCC040


Tgt Ion:	204	Resp:	107049
Ion Ratio	Lower	Upper	
204	100		
206	29.9	23.8	35.8
141	51.3	44.0	66.0



Abundance Scan 2119 (15.789 min): BG048536.D\data.ms (-2029) (-)





Abundance Scan 2122 (15.809 min): BG048516.D\data.ms (-2162) (-)




4-Nitroaniline
Concen: 42.93 ng
RT: 15.813 min Scan# 2123
Delta R.T. 0.000 min
Lab File: BG048536.D
Acq: 1 Apr 2021 9:16


Tgt Ion:	138	Resp:	47580
Ion Ratio	Lower	Upper	
138	100		
92	56.7	32.2	72.2
108	118.4	104.1	144.1



Abundance Scan 2123 (15.813 min): BG048536.D\data.ms (-2033) (-)





Abundance Scan 2168 (16.079 min): BG048516.D\data.ms (-2168) (-)




Azobenzene
Concen: 39.34 ng
RT: 16.083 min Scan# 2169
Delta R.T. 0.000 min
Lab File: BG048536.D
Acq: 1 Apr 2021 9:16

Instrument :
BNA_G
ClientSampleId :
SSTDCCC040


Tgt Ion:	77	Resp:	187837
Ion	Ratio	Lower	Upper
77	100		
182	29.2	5.3	45.3
105	20.0	0.0	37.8
51	40.4	17.0	57.0



Abundance Scan 2169 (16.083 min): BG048536.D\data.ms (-2079) (-)





Abundance Scan 2410 (17.501 min): BG048516.D\data.ms (-2410) (-)




Phenanthrene-d10
Concen: 20.00 ng
RT: 17.505 min Scan# 2411
Delta R.T. 0.000 min
Lab File: BG048536.D
Acq: 1 Apr 2021 9:16

Tgt Ion:	188	Resp:	149871
Ion	Ratio	Lower	Upper
188	100		
94	4.5	3.0	4.6
80	5.8	4.5	6.7




Abundance Scan 2411 (17.505 min): BG048536.D\data.ms (-2321) (-)





Abundance Scan 2132 (15.866 min): BG048516.D\data.ms (-2132) (-)

4,6-Dinitro-2-methylphenol
Concen: 44.30 ng
RT: 15.866 min Scan# 2132
Delta R.T. 0.000 min
Lab File: BG048536.D
Acq: 1 Apr 2021 9:16


Instrument :
BNA_G
ClientSampleId :
SSTDCCC040



Tgt Ion:	198	Resp:	39893
Ion	Ratio	Lower	Upper
198	100		
51	43.0	26.5	66.5
105	37.8	16.0	56.0




Abundance Scan 2132 (15.866 min): BG048536.D\data.ms (-2042) (-)





Abundance Scan 2155 (16.003 min): BG048516.D\data.ms (-2155) (-)


n-Nitrosodiphenylamine
Concen: 40.62 ng
RT: 16.001 min Scan# 2155
Delta R.T. 0.000 min
Lab File: BG048536.D
Acq: 1 Apr 2021 9:16



Tgt Ion:	169	Resp:	173198
Ion	Ratio	Lower	Upper
169	100		
168	67.3	52.7	79.1
167	36.6	27.8	41.8




Abundance Scan 2155 (16.001 min): BG048536.D\data.ms (-2065) (-)






Abundance Scan 2271 (16.684 min): BG048516.D\data.ms (-2265) (-)

4-Bromophenyl-phenylether
Concen: 41.54 ng
RT: 16.688 min Scan# 2272
Delta R.T. 0.000 min
Lab File: BG048536.D
Acq: 1 Apr 2021 9:16

Instrument :
BNA_G
ClientSampleId :
SSTDCCC040




Tgt Ion:	248	Resp:	69015
Ion Ratio	Lower	Upper	
248	100		
250	97.7	82.6	124.0
141	69.0	61.0	91.6







Abundance Scan 2291 (16.802 min): BG048516.D\data.ms (-2264) (-)

Hexachlorobenzene
Concen: 41.11 ng
RT: 16.806 min Scan# 2292
Delta R.T. 0.000 min
Lab File: BG048536.D
Acq: 1 Apr 2021 9:16



Tgt Ion:	284	Resp:	71309
Ion Ratio	Lower	Upper	
284	100		
142	32.5	29.1	43.7
249	33.0	27.7	41.5









Atrazine
Concen: 41.42 ng
RT: 16.947 min Scan# 2316
Delta R.T. 0.000 min
Lab File: BG048536.D
Acq: 1 Apr 2021 9:16




Instrument :
BNA_G
ClientSampleId :
SSTDCCC040


Tgt Ion:	200	Resp:	72031
Ion	Ratio	Lower	Upper
200	100		
173	23.0	2.8	42.8
215	49.8	31.7	71.7



Pentachlorophenol
Concen: 42.67 ng
RT: 17.147 min Scan# 2350
Delta R.T. 0.000 min
Lab File: BG048536.D
Acq: 1 Apr 2021 9:16

Tgt Ion:	266	Resp:	46007
Ion	Ratio	Lower	Upper
266	100		
268	65.0	48.6	73.0
264	64.7	56.4	84.6









Phenanthrene
Concen: 40.67 ng
RT: 17.546 min Scan# 2418
Delta R.T. 0.000 min
Lab File: BG048536.D
Acq: 1 Apr 2021 9:16




Instrument :
BNA_G
ClientSampleId :
SSTDCCC040

Tgt Ion:	178	Resp:	316401
Ion	Ratio	Lower	Upper
178	100		
176	19.3	15.6	23.4
179	14.6	12.9	19.3




Anthracene
Concen: 41.02 ng
RT: 17.640 min Scan# 2434
Delta R.T. 0.000 min
Lab File: BG048536.D
Acq: 1 Apr 2021 9:16

Tgt Ion:	178	Resp:	318040
Ion	Ratio	Lower	Upper
178	100		
176	17.9	14.1	21.1
179	14.8	11.9	17.9





Abundance Scan 2479 (17.906 min): BG048516.D\data.ms (-2479) (-)




Carbazole
Concen: 40.31 ng
RT: 17.905 min Scan# 2479
Delta R.T. 0.000 min
Lab File: BG048536.D
Acq: 1 Apr 2021 9:16

Instrument :
BNA_G
ClientSampleId :
SSTDCCC040


Tgt Ion:	167	Resp:	296381
Ion	Ratio	Lower	Upper
167	100		
166	21.4	19.0	28.6
139	12.6	10.0	15.0



Abundance Scan 2479 (17.905 min): BG048536.D\data.ms (-2389) (-)





Abundance Scan 2573 (18.459 min): BG048516.D\data.ms (-2573) (-)





Di-n-butylphthalate
Concen: 40.65 ng
RT: 18.457 min Scan# 2573
Delta R.T. 0.000 min
Lab File: BG048536.D
Acq: 1 Apr 2021 9:16

Tgt Ion:	149	Resp:	333875
Ion	Ratio	Lower	Upper
149	100		
150	9.5	7.0	10.4
104	6.4	5.9	8.9



Abundance Scan 2573 (18.457 min): BG048536.D\data.ms (-2483) (-)









Fluoranthene
 Concen: 40.15 ng
 RT: 19.556 min Scan# 2760
 Delta R.T. 0.000 min
 Lab File: BG048536.D
 Acq: 1 Apr 2021 9:16




Instrument :
 BNA_G
 ClientSampleId :
 SSTCCCC040

Tgt Ion:	202	Resp:	387214
Ion	Ratio	Lower	Upper
202	100		
101	5.5	0.0	25.8
203	17.9	0.0	36.4




Chrysene-d12
 Concen: 20.00 ng
 RT: 21.800 min Scan# 3142
 Delta R.T. 0.000 min
 Lab File: BG048536.D
 Acq: 1 Apr 2021 9:16

Tgt Ion:	240	Resp:	142999
Ion	Ratio	Lower	Upper
240	100		
120	4.5	2.6	3.8#
236	25.6	19.8	29.6






Abundance Scan 2790 (19.734 min): BG048516.D\data.ms (-277) (-)




Benzidine
Concen: 41.18 ng
RT: 19.732 min Scan# 2790
Delta R.T. 0.000 min
Lab File: BG048536.D
Acq: 1 Apr 2021 9:16

Instrument :
BNA_G
ClientSampleId :
SSTDCCC040

Tgt Ion:	184	Resp:	199403
Ion Ratio	Lower	Upper	
184	100		
185	13.7	11.0	16.4
183	11.4	9.2	13.8






Abundance Scan 2821 (19.916 min): BG048516.D\data.ms (-277) (-)




Pyrene
Concen: 39.61 ng
RT: 19.920 min Scan# 2822
Delta R.T. 0.000 min
Lab File: BG048536.D
Acq: 1 Apr 2021 9:16

Tgt Ion:	202	Resp:	389425
Ion Ratio	Lower	Upper	
202	100		
200	20.5	16.1	24.1
203	17.5	13.6	20.4






Abundance Scan 2854 (20.110 min): BG048516.D\data.ms (-2679) (-)




Terphenyl-d14
Concen: 79.45 ng
RT: 20.114 min Scan# 2855
Delta R.T. 0.000 min
Lab File: BG048536.D
Acq: 1 Apr 2021 9:16

Instrument :
BNA_G
ClientSampleId :
SSTDCCC040

Tgt Ion:	244	Resp:	621234
Ion	Ratio	Lower	Upper
244	100		
212	7.5	6.0	9.0
122	5.9	5.8	8.6







Abundance Scan 2971 (20.797 min): BG048516.D\data.ms (-2686) (-)



Butylbenzylphthalate
Concen: 40.12 ng
RT: 20.795 min Scan# 2971
Delta R.T. 0.000 min
Lab File: BG048536.D
Acq: 1 Apr 2021 9:16

Tgt Ion:	149	Resp:	146497
Ion	Ratio	Lower	Upper
149	100		
91	76.8	53.6	80.4
206	24.1	18.3	27.5









Scan 3138 (21.778 min): BG048516.D\data.ms (-3129) (-)
 Benzo(a)anthracene
 Concen: 40.09 ng
 RT: 21.782 min Scan# 3139
 Delta R.T. 0.000 min
 Lab File: BG048536.D
 Acq: 1 Apr 2021 9:16




Instrument :
 BNA_G
 ClientSampleId :
 SSTDCCC040

Tgt Ion:	228	Resp:	383029
Ion	Ratio	Lower	Upper
228	100		
226	28.4	20.6	30.8
229	21.5	15.8	23.8




Scan 3123 (21.690 min): BG048516.D\data.ms (-3123) (-)
 3,3'-Dichlorobenzidine
 Concen: 38.87 ng
 RT: 21.688 min Scan# 3123
 Delta R.T. 0.000 min
 Lab File: BG048536.D
 Acq: 1 Apr 2021 9:16

Tgt Ion:	252	Resp:	127575
Ion	Ratio	Lower	Upper
252	100		
254	61.8	47.4	71.2
126	8.3	6.1	9.1






Abundance Scan 3150 (21.849 min): BG048516.D\data.ms (-3184) (-)




Chrysene
Concen: 39.72 ng
RT: 21.847 min Scan# 3150
Delta R.T. 0.000 min
Lab File: BG048536.D
Acq: 1 Apr 2021 9:16

Instrument :
BNA_G
ClientSampleId :
SSTDCCC040

Tgt Ion:	228	Resp:	362134
Ion	Ratio	Lower	Upper
228	100		
226	29.5	20.6	31.0
229	19.6	16.3	24.5






Abundance Scan 3121 (21.678 min): BG048516.D\data.ms (-3184) (-)




Bis(2-ethylhexyl)phthalate
Concen: 41.19 ng
RT: 21.677 min Scan# 3121
Delta R.T. 0.000 min
Lab File: BG048536.D
Acq: 1 Apr 2021 9:16

Tgt Ion:	149	Resp:	209544
Ion	Ratio	Lower	Upper
149	100		
167	28.9	22.4	33.6
279	4.4	3.4	5.2





Abundance Scan 3335 (22.936 min): BG048516.D\data.ms (-3325) (-)




Di-n-octyl phthalate
Concen: 40.77 ng
RT: 22.934 min Scan# 3335
Delta R.T. 0.000 min
Lab File: BG048536.D
Acq: 1 Apr 2021 9:16

Instrument :
BNA_G
ClientSampleId :
SSTDCCC040


Tgt Ion:	149	Resp:	361564
Ion	Ratio	Lower	Upper
149	100		
167	1.5	1.2	1.8
43	8.9	7.4	11.0



Abundance Scan 3335 (22.934 min): BG048536.D\data.ms (-3245) (-)





Abundance Scan 3711 (25.145 min): BG048516.D\data.ms (-3695) (-)




Perylene-d12
Concen: 20.00 ng
RT: 25.143 min Scan# 3711
Delta R.T. 0.000 min
Lab File: BG048536.D
Acq: 1 Apr 2021 9:16


Tgt Ion:	264	Resp:	143798
Ion	Ratio	Lower	Upper
264	100		
260	23.3	19.1	28.7
265	24.3	16.1	24.1#



Abundance Scan 3711 (25.143 min): BG048536.D\data.ms (-3621) (-)






Abundance Scan 4363 (28.976 min): BG048516.D\data.ms (-4363) (-)




Indeno(1,2,3-cd)pyrene
Concen: 41.51 ng
RT: 28.974 min Scan# 4363
Delta R.T. 0.000 min
Lab File: BG048536.D
Acq: 1 Apr 2021 9:16

Instrument :
BNA_G
ClientSampleId :
SSTDCCC040

Tgt Ion	Ratio	Lower	Upper
276	100		
138	6.0	7.8	11.8#
277	22.2	0.0	0.0#







Abundance Scan 3529 (24.076 min): BG048516.D\data.ms (-3529) (-)



Benzo(b)fluoranthene
Concen: 39.47 ng
RT: 24.150 min Scan# 3542
Delta R.T. 0.071 min
Lab File: BG048536.D
Acq: 1 Apr 2021 9:16

Tgt Ion	Ratio	Lower	Upper
252	100		
253	21.0	18.2	27.4
125	5.3	5.3	7.9









Benzo(k)fluoranthene
Concen: 41.06 ng
RT: 24.150 min Scan# 3542
Delta R.T. 0.000 min
Lab File: BG048536.D
Acq: 1 Apr 2021 9:16




Instrument :
BNA_G
ClientSampleId :
SSTDCCC040

Tgt Ion	Ratio	Lower	Upper
252	100		
253	21.0	17.9	26.9
125	5.3	3.9	5.9



Benzo(a)pyrene
Concen: 41.20 ng
RT: 24.985 min Scan# 3684
Delta R.T. 0.000 min
Lab File: BG048536.D
Acq: 1 Apr 2021 9:16


Tgt Ion	Ratio	Lower	Upper
252	100		
253	22.1	17.1	25.7
125	5.1	4.4	6.6






Abundance Scan 4374 (29.040 min): BG048516.D\data.ms (-4374) (-)

Dibenzo(a,h)anthracene
Concen: 41.27 ng
RT: 29.062 min Scan# 4378
Delta R.T. 0.000 min
Lab File: BG048536.D
Acq: 1 Apr 2021 9:16

Instrument :
BNA_G
ClientSampleId :
SSTDCCC040




Tgt Ion:	278	Resp:	355115
Ion Ratio	Lower	Upper	
278	100		
139	5.9	5.8	8.6
279	22.4	18.6	28.0



Abundance Scan 4567 (30.174 min): BG048516.D\data.ms (-4567) (-)

Benzo(g,h,i)perylene
Concen: 41.19 ng
RT: 30.178 min Scan# 4568
Delta R.T. 0.000 min
Lab File: BG048536.D
Acq: 1 Apr 2021 9:16



Tgt Ion:	276	Resp:	347120
Ion Ratio	Lower	Upper	
276	100		
277	20.4	20.4	30.6
138	10.8	8.1	12.1

