

Data Path : Z:\SVOASRV\HPCHEM1\BNA_G\DATA\BG040219\
 Data File : BG040303.D
 Acq On : 2 Apr 2019 19:18
 Operator : JU/SJ
 Sample : K2179-12
 Misc :
 ALS Vial : 5 Sample Multiplier: 1

Instrument :
 BNA_G
 ClientSampleId :
 TP-8S

Quant Time: Apr 03 01:26:55 2019
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_G\METHODS\8270-BG032719.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Thu Mar 28 05:47:00 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	8.05	152	73322	20.00	ng	-0.01
21) Naphthalene-d8	10.87	136	312539	20.00	ng	-0.01
39) Acenaphthene-d10	14.68	164	219023	20.00	ng	-0.02
64) Phenanthrene-d10	17.43	188	540534	20.00	ng	-0.02
76) Chrysene-d12	21.71	240	532089	20.00	ng	-0.02
87) Perylene-d12	25.02	264	536177	20.00	ng	0.00
System Monitoring Compounds						
5) 2-Fluorophenol	5.61	112	468528	106.23	ng	0.00
7) Phenol-d6	7.21	99	581825	95.45	ng	-0.01
23) Nitrobenzene-d5	9.23	82	487635	82.93	ng	-0.01
42) 2,4,6-Tribromophenol	16.17	330	288784	122.00	ng	-0.01
45) 2-Fluorobiphenyl	13.31	172	1119491	75.74	ng	-0.01
79) Terphenyl-d14	20.04	244	1835245	77.59	ng	-0.01


Target Compounds	Qvalue
------------------	--------

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\SVOASRV\HPCHEM1\BNA_G\DATA\BG040219\
Data File : BG040303.D
Acq On : 2 Apr 2019 19:18
Operator : JU/SJ
Sample : K2179-12
Misc :
ALS Vial : 5 Sample Multiplier: 1

Instrument :
BNA_G
ClientSampleId :
TP-8S

Quant Time: Apr 03 01:26:55 2019
Quant Method : Z:\SVOASRV\HPCHEM1\BNA_G\METHODS\8270-BG032719.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Thu Mar 28 05:47:00 2019
Response via : Initial Calibration








#1
 1,4-Dichlorobenzene-d4
 Concen: 20.000 ng
 RT: 8.05 min Scan# 845
 Delta R.T. -0.01 min
 Lab File: BG040303.D
 Acq: 2 Apr 2019 19:18

Instrument :
 BNA_G
 ClientSampleId :
 TP-8S

Tgt Ion:152 Resp: 73322
 Ion Ratio Lower Upper
 152 100
 150 148.5 123.7 185.5
 115 60.4 46.9 70.3




Abundance Ion 152.00 (151.70 to 152.70): E
 Ion 150.00 (149.70 to 150.70): E
 Ion 115.00 (114.70 to 115.70): E






#5
 2-Fluorophenol
 Concen: 106.228 ng
 RT: 5.61 min Scan# 429
 Delta R.T. -0.01 min
 Lab File: BG040303.D
 Acq: 2 Apr 2019 19:18

Tgt Ion:112 Resp: 468528
 Ion Ratio Lower Upper
 112 100
 64 67.7 54.2 81.2
 63 32.9 26.4 39.6



Abundance Ion 112.00 (111.70 to 112.70): E
 Ion 64.00 (63.70 to 64.70): BGC
 Ion 63.00 (62.70 to 63.70): BGC





#7

Phenol-d6

Concen: 95.452 ng

RT: 7.21 min Scan# 701

Delta R.T. -0.01 min

Lab File: BG040303.D

Acq: 2 Apr 2019 19:18

Instrument :

BNA_G

ClientSampleId :

TP-8S


Tgt Ion: 99 Resp: 581825

Ion Ratio Lower Upper




99 100

42 18.4 16.2 24.2

71 33.8 26.8 40.2



Abundance Ion 99.00 (98.70 to 99.70): BGC
Ion 42.00 (41.70 to 42.70): BGC
Ion 71.00 (70.70 to 71.70): BGC



#21

Naphthalene-d8

Concen: 20.000 ng

RT: 10.87 min Scan# 1324

Delta R.T. -0.01 min

Lab File: BG040303.D

Acq: 2 Apr 2019 19:18

Tgt Ion: 136 Resp: 312539


Ion Ratio Lower Upper

136 100



137 11.1 8.9 13.3


54 10.3 9.6 14.4

68 7.2 5.1 7.7



Abundance Ion 136.00 (135.70 to 136.70): BGC
Ion 137.00 (136.70 to 137.70): BGC
Ion 54.00 (53.70 to 54.70): BGC
Ion 68.00 (67.70 to 68.70): BGC









#23
 Nitrobenzene-d5
 Concen: 82.932 ng
 RT: 9.23 min Scan# 1045
 Delta R.T. -0.01 min
 Lab File: BG040303.D
 Acq: 2 Apr 2019 19:18




Instrument :
 BNA_G
 ClientSampleId :
 TP-8S

Tgt Ion: 82 Resp: 487635
 Ion Ratio Lower Upper
 82 100
 128 45.0 35.8 53.8
 54 52.9 46.2 69.2



#39
 Acenaphthene-d10
 Concen: 20.000 ng
 RT: 14.68 min Scan# 1973
 Delta R.T. -0.02 min
 Lab File: BG040303.D
 Acq: 2 Apr 2019 19:18

Tgt Ion: 164 Resp: 219023
 Ion Ratio Lower Upper
 164 100
 162 101.4 77.4 116.2
 160 47.1 35.3 52.9








#42
 2,4,6-Tribromophenol
 Concen: 122.001 ng
 RT: 16.17 min Scan# 2227
 Delta R.T. -0.01 min
 Lab File: BG040303.D
 Acq: 2 Apr 2019 19:18

Instrument :
 BNA_G
 ClientSampleID :
 TP-8S

Tgt Ion:330 Resp: 288784
 Ion Ratio Lower Upper
 330 100
 332 97.1 76.6 115.0
 141 38.9 32.6 48.8




Abundance Ion 329.80 (329.50 to 330.50): E
 Ion 331.80 (331.50 to 332.50): E
 Ion 141.00 (140.70 to 141.70): E






#45
 2-Fluorobiphenyl
 Concen: 75.737 ng
 RT: 13.31 min Scan# 1739
 Delta R.T. -0.01 min
 Lab File: BG040303.D
 Acq: 2 Apr 2019 19:18

Tgt Ion:172 Resp: 1119491
 Ion Ratio Lower Upper
 172 100
 171 37.9 30.5 45.7
 170 25.6 20.1 30.1



Abundance Ion 172.00 (171.70 to 172.70): E
 Ion 171.00 (170.70 to 171.70): E
 Ion 170.00 (169.70 to 170.70): E





#64

Phenanthrene-d10

Concen: 20.000 ng

RT: 17.43 min Scan# 2440

Delta R.T. -0.02 min

Lab File: BG040303.D

Acq: 2 Apr 2019 19:18

Instrument :

BNA_G

ClientSampleId :

TP-8S


Tgt Ion:188 Resp: 540534

Ion Ratio Lower Upper

188 100

94 9.3 7.0 10.4




80 10.0 8.6 13.0



Abundance Ion 188.00 (187.70 to 188.70): E

Ion 94.00 (93.70 to 94.70): BGC

Ion 80.00 (79.70 to 80.70): BGC



#76

Chrysene-d12

Concen: 20.000 ng

RT: 21.71 min Scan# 3169

Delta R.T. -0.02 min

Lab File: BG040303.D

Acq: 2 Apr 2019 19:18


Tgt Ion:240 Resp: 532089

Ion Ratio Lower Upper

240 100

120 9.5 8.4 12.6



236 26.1 20.8 31.2



Abundance Ion 240.00 (239.70 to 240.70): E

Ion 120.00 (119.70 to 120.70): E

Ion 236.00 (235.70 to 236.70): E





#79

Terphenyl-d14

Concen: 77.593 ng

RT: 20.04 min Scan# 2884

Delta R.T. -0.01 min

Lab File: BG040303.D

Acq: 2 Apr 2019 19:18

Instrument :

BNA_G

ClientSampleId :

TP-8S


Tgt Ion:244 Resp: 1835245

Ion Ratio Lower Upper

244 100

212 9.1 7.0 10.6




122 11.6 8.5 12.7



Abundance Ion 244.00 (243.70 to 244.70): E

Ion 212.00 (211.70 to 212.70): E

Ion 122.00 (121.70 to 122.70): E



#87

Perylene-d12

Concen: 20.000 ng

RT: 25.02 min Scan# 3733

Delta R.T. -0.01 min

Lab File: BG040303.D

Acq: 2 Apr 2019 19:18


Tgt Ion:264 Resp: 536177

Ion Ratio Lower Upper

264 100

260 23.8 17.9 26.9

265 21.8 16.8 25.2



Abundance Ion 264.00 (263.70 to 264.70): E

Ion 260.00 (259.70 to 260.70): E

Ion 265.00 (264.70 to 265.70): E

