

Data Path : Z:\svoasrv\HPCHEM1\BNA_G\Data\BG040224\
 Data File : BG060740.D
 Acq On : 2 Apr 2024 16:37
 Operator : MA/JU
 Sample : SSTDCCC040
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_G
 LabSampleId :
 SSTDCCC040

Quant Time: Apr 02 17:15:37 2024
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_G\Methods\8270-BG032924.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Sat Mar 30 03:12:26 2024
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	20.000	20.000	0.0	129	0.00
2	1,4-Dioxane	40.000	40.053	-0.1	133	0.00
3	Pyridine	40.000	42.838	-7.1	128	0.00
4	n-Nitrosodimethylamine	40.000	44.296	-10.7	136	0.00
5 S	2-Fluorophenol	80.000	80.615	-0.8	129	0.00
6	Aniline	40.000	39.295	1.8	124	0.00
7 S	Phenol-d6	80.000	79.661	0.4	126	0.00
8	2-Chlorophenol	40.000	40.072	-0.2	127	0.00
9	Benzaldehyde	40.000	37.528	6.2	125	0.00
10 C	Phenol	40.000	39.598	1.0	123	0.00
11	bis(2-Chloroethyl)ether	40.000	38.048	4.9	122	0.00
12	1,3-Dichlorobenzene	40.000	40.632	-1.6	133	0.00
13 C	1,4-Dichlorobenzene	40.000	39.806	0.5	130	0.00
14	1,2-Dichlorobenzene	40.000	41.317	-3.3	135	0.00
15	Benzyl Alcohol	40.000	42.323	-5.8	133	0.00
16	2,2'-oxybis(1-Chloropropane	40.000	40.483	-1.2	131	0.00
17	2-Methylphenol	40.000	40.510	-1.3	129	0.00
18	Hexachloroethane	40.000	41.098	-2.7	137	0.00
19 P	n-Nitroso-di-n-propylamine	40.000	40.508	-1.3	128	0.00
20	3+4-Methylphenols	40.000	40.862	-2.2	129	0.00
21 I	Naphthalene-d8	20.000	20.000	0.0	123	0.00
22	Acetophenone	40.000	40.769	-1.9	128	0.00
23 S	Nitrobenzene-d5	80.000	94.419	-18.0	149	0.00
24	Nitrobenzene	40.000	42.950	-7.4	140	0.00
25	Isophorone	40.000	38.251	4.4	118	0.00
26 C	2-Nitrophenol	40.000	46.891	-17.2	158	0.00
27	2,4-Dimethylphenol	40.000	39.107	2.2	122	0.00
28	bis(2-Chloroethoxy)methane	40.000	38.705	3.2	121	0.00
29 C	2,4-Dichlorophenol	40.000	42.137	-5.3	126	0.00
30	1,2,4-Trichlorobenzene	40.000	39.997	0.0	125	0.00
31	Naphthalene	40.000	39.251	1.9	121	0.00
32	Benzoic acid	40.000	45.689	-14.2	151	0.01
33	4-Chloroaniline	40.000	39.306	1.7	119	0.00
34 C	Hexachlorobutadiene	40.000	39.590	1.0	125	0.00
35	Caprolactam	40.000	41.748	-4.4	125	0.00
36 C	4-Chloro-3-methylphenol	40.000	40.954	-2.4	122	0.00
37	2-Methylnaphthalene	40.000	39.550	1.1	123	0.00
38	1-Methylnaphthalene	40.000	39.922	0.2	127	0.00
39 I	Acenaphthene-d10	20.000	20.000	0.0	126	0.00
40	1,2,4,5-Tetrachlorobenzene	40.000	40.218	-0.5	131	0.00
41 P	Hexachlorocyclopentadiene	40.000	43.267	-8.2	139	0.00
42 S	2,4,6-Tribromophenol	80.000	95.719	-19.6	151	0.00
43 C	2,4,6-Trichlorophenol	40.000	43.344	-8.4	135	0.00
44	2,4,5-Trichlorophenol	40.000	43.577	-8.9	136	0.00
45 S	2-Fluorobiphenyl	80.000	76.867	3.9	125	0.00
46	1,1'-Biphenyl	40.000	38.955	2.6	127	0.00
47	2-Chloronaphthalene	40.000	39.025	2.4	126	0.00

Data Path : Z:\svoasrv\HPCHEM1\BNA_G\Data\BG040224\
 Data File : BG060740.D
 Acq On : 2 Apr 2024 16:37
 Operator : MA/JU
 Sample : SSTDCCC040
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_G
 LabSampleId :
 SSTDCCC040

Quant Time: Apr 02 17:15:37 2024
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_G\Methods\8270-BG032924.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Sat Mar 30 03:12:26 2024
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
48	2-Nitroaniline	40.000	45.974	-14.9	155	0.00
49	Acenaphthylene	40.000	39.480	1.3	127	0.00
50	Dimethylphthalate	40.000	38.848	2.9	123	0.00
51	2,6-Dinitrotoluene	40.000	44.303	-10.8	148	0.00
52 C	Acenaphthene	40.000	40.567	-1.4	130	0.00
53	3-Nitroaniline	40.000	44.558	-11.4	144	0.00
54 P	2,4-Dinitrophenol	40.000	55.447	-38.6#	178	0.00
55	Dibenzofuran	40.000	39.271	1.8	127	0.00
56 P	4-Nitrophenol	40.000	47.428	-18.6	146	0.00
57	2,4-Dinitrotoluene	40.000	46.014	-15.0	153	0.00
58	Fluorene	40.000	39.242	1.9	125	0.00
59	2,3,4,6-Tetrachlorophenol	40.000	44.435	-11.1	134	0.00
60	Diethylphthalate	40.000	38.590	3.5	124	0.00
61	4-Chlorophenyl-phenylether	40.000	39.345	1.6	127	0.00
62	4-Nitroaniline	40.000	48.348	-20.9	145	0.00
63	Azobenzene	40.000	38.516	3.7	122	0.00
64 I	Phenanthrene-d10	20.000	20.000	0.0	126	0.00
65	4,6-Dinitro-2-methylphenol	40.000	46.645	-16.6	167	0.00
66 c	n-Nitrosodiphenylamine	40.000	39.707	0.7	126	0.00
67	4-Bromophenyl-phenylether	40.000	42.394	-6.0	134	0.00
68	Hexachlorobenzene	40.000	39.984	0.0	128	0.00
69	Atrazine	40.000	39.774	0.6	125	0.00
70 C	Pentachlorophenol	40.000	46.158	-15.4	143	0.00
71	Phenanthrene	40.000	39.137	2.2	127	0.00
72	Anthracene	40.000	39.807	0.5	128	0.00
73	Carbazole	40.000	39.407	1.5	127	0.00
74	Di-n-butylphthalate	40.000	38.937	2.7	123	0.00
75 C	Fluoranthene	40.000	38.937	2.7	127	0.00
76 I	Chrysene-d12	20.000	20.000	0.0	123	0.00
77	Benzydine	40.000	39.863	0.3	108	0.00
78	Pyrene	40.000	40.508	-1.3	127	0.00
79 S	Terphenyl-d14	80.000	78.409	2.0	124	0.00
80	Butylbenzylphthalate	40.000	44.563	-11.4	134	0.00
81	Benzo(a)anthracene	40.000	40.049	-0.1	124	0.00
82	3,3'-Dichlorobenzidine	40.000	43.224	-8.1	128	0.00
83	Chrysene	40.000	39.670	0.8	123	0.00
84	Bis(2-ethylhexyl)phthalate	40.000	41.304	-3.3	123	0.00
85 c	Di-n-octyl phthalate	40.000	43.918	-9.8	129	-0.02
86 I	Perylene-d12	20.000	20.000	0.0	120	0.00
87	Indeno(1,2,3-cd)pyrene	40.000	40.237	-0.6	123	0.00
88	Benzo(b)fluoranthene	40.000	40.150	-0.4	122	0.00
89	Benzo(k)fluoranthene	40.000	39.756	0.6	119	0.00
90 C	Benzo(a)pyrene	40.000	40.795	-2.0	122	0.00
91	Dibenzo(a,h)anthracene	40.000	40.078	-0.2	122	0.00
92	Benzo(g,h,i)perylene	40.000	40.085	-0.2	122	0.00

Data Path : Z:\svoasrv\HPCHEM1\BNA_G\Data\BG040224\
Data File : BG060740.D
Acq On : 2 Apr 2024 16:37
Operator : MA/JU
Sample : SSTDCCC040
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_G
LabSampleId :
SSTDCCC040

Quant Time: Apr 02 17:15:37 2024
Quant Method : Z:\svoasrv\HPCHEM1\BNA_G\Methods\8270-BG032924.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Sat Mar 30 03:12:26 2024
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

Compound	Amount	Calc.	%Dev	Area%	Dev(min)
----------	--------	-------	------	-------	----------

(#) = Out of Range

SPCC's out = 0 CCC's out = 0