

Data Path : Z:\svoasrv\HPCHEM1\BNA\_G\Data\BG040722\  
 Data File : BG053074.D  
 Acq On : 7 Apr 2022 15:15  
 Operator : CG/JU  
 Sample : N2237-05  
 Misc :  
 ALS Vial : 6 Sample Multiplier: 2

Instrument :  
 BNA\_G  
 ClientSampleId :  
 A12040

Quant Time: Apr 08 07:22:56 2022  
 Quant Method : Z:\svoasrv\HPCHEM1\BNA\_G\Methods\8270-BG031522.M  
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION  
 QLast Update : Wed Apr 06 14:35:39 2022  
 Response via : Initial Calibration

| Compound                    | R.T.   | QIon | Response | Conc   | Units | Dev(Min) |
|-----------------------------|--------|------|----------|--------|-------|----------|
| -----                       |        |      |          |        |       |          |
| Internal Standards          |        |      |          |        |       |          |
| 1) 1,4-Dichlorobenzene-d4   | 8.131  | 152  | 75786    | 40.000 | ng    | 0.00     |
| 21) Naphthalene-d8          | 10.939 | 136  | 298537   | 40.000 | ng    | # 0.00   |
| 39) Acenaphthene-d10        | 14.751 | 164  | 210734   | 40.000 | ng    | 0.00     |
| 64) Phenanthrene-d10        | 17.495 | 188  | 452314   | 40.000 | ng    | 0.00     |
| 76) Chrysene-d12            | 21.777 | 240  | 430790   | 40.000 | ng    | 0.00     |
| 86) Perylene-d12            | 25.090 | 264  | 437452   | 40.000 | ng    | 0.00     |
| System Monitoring Compounds |        |      |          |        |       |          |
| 5) 2-Fluorophenol           | 5.693  | 112  | 207144   | 95.201 | ng    | 0.00     |
| 7) Phenol-d6                | 7.285  | 99   | 303897   | 92.642 | ng    | 0.00     |
| 23) Nitrobenzene-d5         | 9.288  | 82   | 222569   | 74.837 | ng    | 0.00     |
| 42) 2,4,6-Tribromophenol    | 16.232 | 330  | 134104   | 94.716 | ng    | 0.00     |
| 45) 2-Fluorobiphenyl        | 13.371 | 172  | 489278   | 68.485 | ng    | 0.00     |
| 79) Terphenyl-d14           | 20.097 | 244  | 739719   | 61.090 | ng    | 0.00     |

|                  |        |
|------------------|--------|
| Target Compounds | Qvalue |
| -----            |        |

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA\_G\Data\BG040722\  
Data File : BG053074.D  
Acq On : 7 Apr 2022 15:15  
Operator : CG/JU  
Sample : N2237-05  
Misc :  
ALS Vial : 6 Sample Multiplier: 2

Instrument :  
BNA\_G  
ClientSampleId :  
A12040

Quant Time: Apr 08 07:22:56 2022  
Quant Method : Z:\svoasrv\HPCHEM1\BNA\_G\Methods\8270-BG031522.M  
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION  
QLast Update : Wed Apr 06 14:35:39 2022  
Response via : Initial Calibration

