


Data Path : Z:\SVOASRV\HPCHEM1\BNA_G\DATA\BG041519\
 Data File : BG040476.D
 Acq On : 15 Apr 2019 12:17
 Operator : JU/SJ
 Sample : DFTPP
 Misc :
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 BNA_G
 ClientSampleId :
 DFTPP

Integration File: rteint.p

Method : Z:\SVOASRV\HPCHEM1\BNA_G\METHODS\8270-BG032719.M
 Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 Last Update : Sat Apr 13 00:33:11 2019



AutoFind: Scans 2506, 2507, 2508; Background Corrected with Scan 2500

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
51	198	10	80	39.7	52360	PASS
68	69	0.00	2	1.4	665	PASS
69	198	0.00	100	35.0	46178	PASS
70	69	0.00	2	0.5	209	PASS
127	198	10	80	44.6	58845	PASS
197	198	0.00	2	0.0	0	PASS
198	198	100	100	100.0	131938	PASS
199	198	5	9	7.1	9399	PASS
275	198	10	60	24.2	31880	PASS
365	198	1	100	3.2	4195	PASS
441	198	0.01	100	8.6	11349	PASS
442	198	50	100	52.2	68856	PASS
443	442	15	24	19.3	13322	PASS