


Data Path : Z:\SVOASRV\HPCHEM1\BNA_G\DATA\BG041619\
 Data File : BG040517.D
 Acq On : 17 Apr 2019 00:53
 Operator : JU/SJ
 Sample : DFTPP
 Misc :
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 BNA_G
 ClientSampleId :
 DFTPP

Integration File: rteint.p

Method : Z:\SVOASRV\HPCHEM1\BNA_G\METHODS\8270-BG032719.M
 Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 Last Update : Sat Apr 13 00:33:11 2019



AutoFind: Scans 2505, 2506, 2507; Background Corrected with Scan 2499

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
51	198	10	80	40.4	30837	PASS
68	69	0.00	2	0.7	192	PASS
69	198	0.00	100	33.7	25720	PASS
70	69	0.00	2	0.3	73	PASS
127	198	10	80	43.2	32994	PASS
197	198	0.00	2	0.3	249	PASS
198	198	100	100	100.0	76413	PASS
199	198	5	9	7.1	5432	PASS
275	198	10	60	26.1	19940	PASS
365	198	1	100	3.2	2429	PASS
441	198	0.01	100	9.8	7489	PASS
442	198	50	100	57.0	43584	PASS
443	442	15	24	18.4	8026	PASS