

Data Path : Z:\svoasrv\HPCHEM1\BNA_G\Data\BG041621\
 Data File : BG048728.D
 Acq On : 16 Apr 2021 17:24
 Operator : CG/JU
 Sample : M2026-03
 Misc :
 ALS Vial : 5 Sample Multiplier: 1

Instrument :
 BNA_G
 ClientSampleId :
 GMW-5

Quant Time: Apr 16 18:03:56 2021
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_G\METHODS\8270-BG041421.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Fri Apr 16 15:59:00 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	8.081	152	19633	20.00	ng	# 0.00
21) Naphthalene-d8	10.908	136	81018	20.00	ng	# 0.00
39) Acenaphthene-d10	14.738	164	54858	20.00	ng	0.00
64) Phenanthrene-d10	17.494	188	140408	20.00	ng	# 0.00
76) Chrysene-d12	21.795	240	138438	20.00	ng	# 0.00
86) Perylene-d12	25.138	264	133664	20.00	ng	0.00
System Monitoring Compounds						
5) 2-Fluorophenol	5.625	112	68941	58.73	ng	0.00
7) Phenol-d6	7.235	99	53270	32.08	ng	0.00
23) Nitrobenzene-d5	9.251	82	169511	93.67	ng	0.00
42) 2,4,6-Tribromophenol	16.231	330	108376	147.34	ng	0.00
45) 2-Fluorobiphenyl	13.358	172	382715	100.75	ng	0.00
79) Terphenyl-d14	20.108	244	635370	78.99	ng	0.00
Target Compounds						
50) Dimethylphthalate	14.180	163	23463	5.70	ng	Qvalue # 94

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_G\Data\BG041621\
Data File : BG048728.D
Acq On : 16 Apr 2021 17:24
Operator : CG/JU
Sample : M2026-03
Misc :
ALS Vial : 5 Sample Multiplier: 1

Instrument :
BNA_G
ClientSampleId :
GMW-5

Quant Time: Apr 16 18:03:56 2021
Quant Method : Z:\SVOASRV\HPCHEM1\BNA_G\METHODS\8270-BG041421.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Fri Apr 16 15:59:00 2021
Response via : Initial Calibration

