




Data Path : Z:\svoasrv\HPCHEM1\BNA_G\Data\BG041621\
Data File : BG048718.D
Acq On : 16 Apr 2021 9:25
Operator : CG/JU
Sample : DFTPP
Misc :
ALS Vial : 1 Sample Multiplier: 1

Quant Time: Apr 17 05:49:53 2021
Quant Method : Z:\SVOASRV\HPCHEM1\BNA_G\METHODS\8270-BG041421.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Fri Apr 16 15:59:00 2021
Response via : Initial Calibration



TIC: BG048718.D\data.ms

(70) Pentachlorophenol (C)




17.146min (-0.000) 75897.69 ng m

response 49507

Ion	Exp%	Act%
265.70	100.00	100.00
268.00	61.20	55.48
264.00	61.30	62.66
0.00	0.00	0.00

Data Path : Z:\svoasrv\HPCHEM1\BNA_G\Data\BG041621\
 Data File : BG048718.D
 Acq On : 16 Apr 2021 9:25
 Operator : CG/JU
 Sample : DFTPP
 Misc :
 ALS Vial : 1 Sample Multiplier: 1

Quant Time: Apr 17 05:49:53 2021
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_G\METHODS\8270-BG041421.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Fri Apr 16 15:59:00 2021
 Response via : Initial Calibration



TIC: BG048718.D\data.ms

(77) Benzidine

19.737min (+ 0.006) 320487.03 ng m

response 469975

Ion	Exp%	Act%
184.00	100.00	100.00
185.00	13.40	14.12
183.00	11.70	12.98
0.00	0.00	0.00