

Data Path : Z:\SVOASRV\HPCHEM1\BNA\_G\DATA\BG041719\  
 Data File : BG040538.D  
 Acq On : 17 Apr 2019 20:29  
 Operator : JU/SJ  
 Sample : K2196-06  
 Misc :  
 ALS Vial : 11 Sample Multiplier: 1

Instrument :  
 BNA\_G  
 ClientSampleId :  
 HD-01-041619

Quant Time: Apr 18 04:25:58 2019  
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA\_G\METHODS\8270-BG032719.M  
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION  
 QLast Update : Sat Apr 13 00:33:11 2019  
 Response via : Initial Calibration

| Internal Standards          | R.T.  | QIon | Response | Conc   | Units | Dev(Min) |
|-----------------------------|-------|------|----------|--------|-------|----------|
| 1) 1,4-Dichlorobenzene-d4   | 8.04  | 152  | 47639    | 20.00  | ng    | 0.00     |
| 21) Naphthalene-d8          | 10.86 | 136  | 180119   | 20.00  | ng    | 0.00     |
| 39) Acenaphthene-d10        | 14.67 | 164  | 108533   | 20.00  | ng    | 0.00     |
| 64) Phenanthrene-d10        | 17.42 | 188  | 299501   | 20.00  | ng    | 0.00     |
| 76) Chrysene-d12            | 21.69 | 240  | 309318   | 20.00  | ng    | 0.00     |
| 87) Perylene-d12            | 24.96 | 264  | 327939   | 20.00  | ng    | 0.00     |
| System Monitoring Compounds |       |      |          |        |       |          |
| 5) 2-Fluorophenol           | 5.60  | 112  | 370039   | 129.13 | ng    | 0.00     |
| 7) Phenol-d6                | 7.20  | 99   | 451437   | 113.99 | ng    | 0.00     |
| 23) Nitrobenzene-d5         | 9.22  | 82   | 348608   | 102.87 | ng    | 0.00     |
| 42) 2,4,6-Tribromophenol    | 16.17 | 330  | 192930   | 164.48 | ng    | 0.00     |
| 45) 2-Fluorobiphenyl        | 13.29 | 172  | 732993   | 100.07 | ng    | 0.00     |
| 79) Terphenyl-d14           | 20.02 | 244  | 1384821  | 100.72 | ng    | 0.00     |

| Target Compounds | Qvalue |
|------------------|--------|
|------------------|--------|

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\SVOASRV\HPCHEM1\BNA\_G\DATA\BG041719\  
Data File : BG040538.D  
Acq On : 17 Apr 2019 20:29  
Operator : JU/SJ  
Sample : K2196-06  
Misc :  
ALS Vial : 11 Sample Multiplier: 1

Instrument :  
BNA\_G  
ClientSampleId :  
HD-01-041619

Quant Time: Apr 18 04:25:58 2019  
Quant Method : Z:\SVOASRV\HPCHEM1\BNA\_G\METHODS\8270-BG032719.M  
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION  
QLast Update : Sat Apr 13 00:33:11 2019  
Response via : Initial Calibration

