

Data Path : Z:\SVOASRV\HPCHEM1\BNA_G\DATA\BG041719\
 Data File : BG040550.D
 Acq On : 18 Apr 2019 4:34
 Operator : JU/SJ
 Sample : PB118988BL
 Misc :
 ALS Vial : 21 Sample Multiplier: 1

Instrument :
 BNA_G
 ClientSampleId :
 PB118988BL

Quant Time: Apr 18 06:05:11 2019
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_G\METHODS\8270-BG032719.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Sat Apr 13 00:33:11 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	8.04	152	40735	20.00	ng	0.00
21) Naphthalene-d8	10.86	136	151812	20.00	ng	0.00
39) Acenaphthene-d10	14.67	164	90389	20.00	ng	0.00
64) Phenanthrene-d10	17.42	188	253168	20.00	ng	0.00
76) Chrysene-d12	21.69	240	311472	20.00	ng	0.00
87) Perylene-d12	24.97	264	340761	20.00	ng	0.00
System Monitoring Compounds						
5) 2-Fluorophenol	5.60	112	299567	122.25	ng	0.00
7) Phenol-d6	7.20	99	401437	118.54	ng	0.00
23) Nitrobenzene-d5	9.22	82	237032	82.99	ng	0.00
42) 2,4,6-Tribromophenol	16.16	330	120235	123.08	ng	0.00
45) 2-Fluorobiphenyl	13.29	172	518432	84.99	ng	0.00
79) Terphenyl-d14	20.02	244	1146274	82.79	ng	0.00


Target Compounds	Qvalue
------------------	--------


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\SVOASRV\HPCHEM1\BNA_G\DATA\BG041719\
Data File : BG040550.D
Acq On : 18 Apr 2019 4:34
Operator : JU/SJ
Sample : PB118988BL
Misc :
ALS Vial : 21 Sample Multiplier: 1

Instrument :
BNA_G
ClientSampleId :
PB118988BL

Quant Time: Apr 18 06:05:11 2019
Quant Method : Z:\SVOASRV\HPCHEM1\BNA_G\METHODS\8270-BG032719.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Sat Apr 13 00:33:11 2019
Response via : Initial Calibration









#1

1,4-Dichlorobenzene-d4
Concen: 20.000 ng
RT: 8.04 min Scan# 843
Delta R.T. -0.00 min
Lab File: BG040550.D
Acq: 18 Apr 2019 4:34

Instrument :
BNA_G
ClientSampleId :
PB118988BL




Tgt Ion:152 Resp: 40735
Ion Ratio Lower Upper
152 100
150 159.8 123.7 185.5
115 60.9 46.9 70.3




#5

2-Fluorophenol
Concen: 122.255 ng
RT: 5.60 min Scan# 427
Delta R.T. -0.00 min
Lab File: BG040550.D
Acq: 18 Apr 2019 4:34

Tgt Ion:112 Resp: 299567
Ion Ratio Lower Upper
112 100
64 67.4 54.2 81.2
63 34.5 26.4 39.6





#7

Phenol-d6

Concen: 118.543 ng

RT: 7.20 min Scan# 699

Delta R.T. -0.00 min

Lab File: BG040550.D

Acq: 18 Apr 2019 4:34

Instrument :

BNA_G

ClientSampleId :

PB118988BL


Tgt Ion: 99 Resp: 401437

Ion Ratio Lower Upper




99 100

42 19.7 16.2 24.2

71 33.6 26.8 40.2



Abundance Ion 99.00 (98.70 to 99.70): BGC
Ion 42.00 (41.70 to 42.70): BGC
Ion 71.00 (70.70 to 71.70): BGC



#21

Naphthalene-d8

Concen: 20.000 ng

RT: 10.86 min Scan# 1322

Delta R.T. -0.00 min

Lab File: BG040550.D

Acq: 18 Apr 2019 4:34

Tgt Ion: 136 Resp: 151812


Ion Ratio Lower Upper

136 100



137 11.8 8.9 13.3


54 10.2 9.6 14.4

68 6.9 5.1 7.7



Abundance Ion 136.00 (135.70 to 136.70): BGC
Ion 137.00 (136.70 to 137.70): BGC
Ion 54.00 (53.70 to 54.70): BGC
Ion 68.00 (67.70 to 68.70): BGC









#23
 Nitrobenzene-d5
 Concen: 82.991 ng
 RT: 9.22 min Scan# 1043
 Delta R.T. -0.00 min
 Lab File: BG040550.D
 Acq: 18 Apr 2019 4:34

Instrument :
 BNA_G
 ClientSampleId :
 PB118988BL

Tgt Ion: 82 Resp: 237032
 Ion Ratio Lower Upper
 82 100
 128 41.2 35.8 53.8
 54 56.7 46.2 69.2




Abundance Ion 82.00 (81.70 to 82.70): BGC
 Ion 128.00 (127.70 to 128.70): BGC
 Ion 54.00 (53.70 to 54.70): BGC






#39
 Acenaphthene-d10
 Concen: 20.000 ng
 RT: 14.67 min Scan# 1972
 Delta R.T. -0.00 min
 Lab File: BG040550.D
 Acq: 18 Apr 2019 4:34

Tgt Ion: 164 Resp: 90389
 Ion Ratio Lower Upper
 164 100
 162 102.3 77.4 116.2
 160 45.2 35.3 52.9



Abundance Ion 164.00 (163.70 to 164.70): BGC
 Ion 162.00 (161.70 to 162.70): BGC
 Ion 160.00 (159.70 to 160.70): BGC





#42

2,4,6-Tribromophenol

Concen: 123.082 ng

RT: 16.16 min Scan# 2225

Delta R.T. -0.00 min

Lab File: BG040550.D

Acq: 18 Apr 2019 4:34

Instrument :

BNA_G

ClientSampleId :

PB118988BL


Tgt Ion:330 Resp: 120235

Ion Ratio Lower Upper

330 100

332 93.7 76.6 115.0




141 38.8 32.6 48.8



Abundance Ion 329.80 (329.50 to 330.50): E

Ion 331.80 (331.50 to 332.50): E

Ion 141.00 (140.70 to 141.70): E



#45

2-Fluorobiphenyl

Concen: 84.987 ng

RT: 13.29 min Scan# 1737

Delta R.T. -0.00 min

Lab File: BG040550.D

Acq: 18 Apr 2019 4:34


Tgt Ion:172 Resp: 518432

Ion Ratio Lower Upper

172 100

171 34.7 30.5 45.7



170 24.3 20.1 30.1




Abundance Ion 172.00 (171.70 to 172.70): E

Ion 171.00 (170.70 to 171.70): E

Ion 170.00 (169.70 to 170.70): E









#64
Phenanthrene-d10
Concen: 20.000 ng
RT: 17.42 min Scan# 2439
Delta R.T. -0.00 min
Lab File: BG040550.D
Acq: 18 Apr 2019 4:34

Instrument :
BNA_G
ClientSampleId :
PB118988BL

Tgt Ion:188 Resp: 253168
Ion Ratio Lower Upper
188 100
94 9.5 7.0 10.4
80 10.5 8.6 13.0




Abundance Ion 188.00 (187.70 to 188.70): E
Ion 94.00 (93.70 to 94.70): BGC
Ion 80.00 (79.70 to 80.70): BGC






#76
Chrysene-d12
Concen: 20.000 ng
RT: 21.69 min Scan# 3166
Delta R.T. -0.01 min
Lab File: BG040550.D
Acq: 18 Apr 2019 4:34

Tgt Ion:240 Resp: 311472
Ion Ratio Lower Upper
240 100
120 10.7 8.4 12.6
236 25.7 20.8 31.2



Abundance Ion 240.00 (239.70 to 240.70): E
Ion 120.00 (119.70 to 120.70): E
Ion 236.00 (235.70 to 236.70): E





#79

Terphenyl-d14

Concen: 82.791 ng

RT: 20.02 min Scan# 2881

Delta R.T. -0.01 min

Lab File: BG040550.D

Acq: 18 Apr 2019 4:34

Instrument :

BNA_G

ClientSampleId :

PB118988BL


Tgt Ion:244 Resp: 1146274

Ion Ratio Lower Upper

244 100

212 8.3 7.0 10.6




122 11.9 8.5 12.7



Abundance Ion 244.00 (243.70 to 244.70): E

Ion 212.00 (211.70 to 212.70): E

Ion 122.00 (121.70 to 122.70): E



#87

Perylene-d12

Concen: 20.000 ng

RT: 24.97 min Scan# 3724

Delta R.T. -0.00 min

Lab File: BG040550.D

Acq: 18 Apr 2019 4:34


Tgt Ion:264 Resp: 340761

Ion Ratio Lower Upper

264 100

260 23.8 17.9 26.9

265 20.0 16.8 25.2



Abundance Ion 264.00 (263.70 to 264.70): E

Ion 260.00 (259.70 to 260.70): E

Ion 265.00 (264.70 to 265.70): E

