




Data Path : Z:\svoasrv\HPCHEM1\BNA_G\Data\BG041922\
Data File : BG053192.D
Acq On : 19 Apr 2022 11:24
Operator : CG/JU
Sample : DFTPP
Misc :
ALS Vial : 1 Sample Multiplier: 1

Quant Time: Apr 20 06:41:39 2022
Quant Method : Z:\svoasrv\HPCHEM1\BNA_G\Methods\8270-BG040822.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Fri Apr 08 15:31:24 2022
Response via : Initial Calibration



TIC: BG053192.D\data.ms

(70) Pentachlorophenol (C)




17.136min (-0.009) 17375.65 ng

response 89738

Ion	Exp%	Act%
265.70	100.00	100.00
268.00	57.40	58.76
264.00	56.60	62.15
0.00	0.00	0.00

Data Path : Z:\svoasrv\HPCHEM1\BNA_G\Data\BG041922\
Data File : BG053192.D
Acq On : 19 Apr 2022 11:24
Operator : CG/JU
Sample : DFTPP
Misc :
ALS Vial : 1 Sample Multiplier: 1

Quant Time: Apr 20 06:41:39 2022
Quant Method : Z:\svoasrv\HPCHEM1\BNA_G\Methods\8270-BG040822.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Fri Apr 08 15:31:24 2022
Response via : Initial Calibration



TIC: BG053192.D\data.ms

(77) Benzidine**19.703min (-0.009) 385893.91 ng****response 649600**

Ion	Exp%	Act%
184.00	100.00	100.00
185.00	14.40	14.95
183.00	11.80	12.11
0.00	0.00	0.00