

Data Path : Z:\svoasrv\HPCHEM1\BNA_G\Data\BG041924\
 Data File : BG060955.D
 Acq On : 19 Apr 2024 11:24
 Operator : MA/JU
 Sample : SSTDCCC040
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_G
 LabSampleId :
 SSTDCCC040

Quant Time: Apr 19 12:07:51 2024
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_G\Methods\8270-BG041724.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Thu Apr 18 02:02:19 2024
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	20.000	20.000	0.0	55	0.00
2	1,4-Dioxane	40.000	37.767	5.6	58	0.00
3	Pyridine	40.000	30.350	24.1	45	0.00
4	n-Nitrosodimethylamine	40.000	40.014	-0.0	60	0.00
5 S	2-Fluorophenol	80.000	73.256	8.4	55	0.00
6	Aniline	40.000	33.890	15.3	50	0.00
7 S	Phenol-d6	80.000	70.207	12.2	50	0.00
8	2-Chlorophenol	40.000	36.888	7.8	54	0.00
9	Benzaldehyde	40.000	34.710	13.2	52	0.00
10 C	Phenol	40.000	35.248	11.9	50	0.00
11	bis(2-Chloroethyl)ether	40.000	32.115	19.7	46	0.00
12	1,3-Dichlorobenzene	40.000	38.592	3.5	58	0.00
13 C	1,4-Dichlorobenzene	40.000	38.375	4.1	57	0.00
14	1,2-Dichlorobenzene	40.000	37.835	5.4	56	0.00
15	Benzyl Alcohol	40.000	37.236	6.9	53	0.00
16	2,2'-oxybis(1-Chloropropane	40.000	33.184	17.0	49	0.00
17	2-Methylphenol	40.000	35.690	10.8	52	0.00
18	Hexachloroethane	40.000	38.597	3.5	57	0.00
19 P	n-Nitroso-di-n-propylamine	40.000	34.331	14.2	50	0.00
20	3+4-Methylphenols	40.000	35.074	12.3	50	0.00
21 I	Naphthalene-d8	20.000	20.000	0.0	53	0.00
22	Acetophenone	40.000	37.155	7.1	52	0.00
23 S	Nitrobenzene-d5	80.000	79.663	0.4	55	0.00
24	Nitrobenzene	40.000	38.411	4.0	53	0.00
25	Isophorone	40.000	35.672	10.8	48	0.00
26 C	2-Nitrophenol	40.000	42.413	-6.0	57	0.00
27	2,4-Dimethylphenol	40.000	38.683	3.3	53	0.00
28	bis(2-Chloroethoxy)methane	40.000	35.672	10.8	50	0.00
29 C	2,4-Dichlorophenol	40.000	40.132	-0.3	56	0.00
30	1,2,4-Trichlorobenzene	40.000	40.278	-0.7	56	0.00
31	Naphthalene	40.000	38.465	3.8	53	0.00
32	Benzoic acid	40.000	40.702	-1.8	56	0.00
33	4-Chloroaniline	40.000	36.667	8.3	49	0.00
34 C	Hexachlorobutadiene	40.000	47.590	-19.0	67	0.00
35	Caprolactam	40.000	35.338	11.7	47	0.00
36 C	4-Chloro-3-methylphenol	40.000	38.368	4.1	52	0.00
37	2-Methylnaphthalene	40.000	39.739	0.7	54	0.00
38	1-Methylnaphthalene	40.000	39.109	2.2	53	0.00
39 I	Acenaphthene-d10	20.000	20.000	0.0	53	0.00
40	1,2,4,5-Tetrachlorobenzene	40.000	43.552	-8.9	61	0.00
41 P	Hexachlorocyclopentadiene	40.000	49.958	-24.9	67	0.00
42 S	2,4,6-Tribromophenol	80.000	92.208	-15.3	63	0.00
43 C	2,4,6-Trichlorophenol	40.000	40.610	-1.5	56	0.00
44	2,4,5-Trichlorophenol	40.000	40.924	-2.3	56	0.00
45 S	2-Fluorobiphenyl	80.000	81.774	-2.2	57	0.00
46	1,1'-Biphenyl	40.000	38.885	2.8	54	0.00
47	2-Chloronaphthalene	40.000	38.722	3.2	53	0.00

Data Path : Z:\svoasrv\HPCHEM1\BNA_G\Data\BG041924\
 Data File : BG060955.D
 Acq On : 19 Apr 2024 11:24
 Operator : MA/JU
 Sample : SSTDCCC040
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_G
 LabSampleId :
 SSTDCCC040

Quant Time: Apr 19 12:07:51 2024
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_G\Methods\8270-BG041724.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Thu Apr 18 02:02:19 2024
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
48	2-Nitroaniline	40.000	38.153	4.6	52	0.00
49	Acenaphthylene	40.000	38.813	3.0	54	0.00
50	Dimethylphthalate	40.000	38.632	3.4	54	0.00
51	2,6-Dinitrotoluene	40.000	40.864	-2.2	55	0.00
52 C	Acenaphthene	40.000	38.696	3.3	54	0.00
53	3-Nitroaniline	40.000	36.295	9.3	49	0.00
54 P	2,4-Dinitrophenol	40.000	41.560	-3.9	60	0.00
55	Dibenzofuran	40.000	38.509	3.7	53	0.00
56 P	4-Nitrophenol	40.000	37.826	5.4	50	0.00
57	2,4-Dinitrotoluene	40.000	40.485	-1.2	55	0.00
58	Fluorene	40.000	38.261	4.3	53	0.00
59	2,3,4,6-Tetrachlorophenol	40.000	42.060	-5.2	57	0.00
60	Diethylphthalate	40.000	38.026	4.9	52	0.00
61	4-Chlorophenyl-phenylether	40.000	40.715	-1.8	56	0.00
62	4-Nitroaniline	40.000	37.018	7.5	52	0.00
63	Azobenzene	40.000	34.992	12.5	48	0.00
64 I	Phenanthrene-d10	20.000	20.000	0.0	51	0.00
65	4,6-Dinitro-2-methylphenol	40.000	44.960	-12.4	59	0.00
66 c	n-Nitrosodiphenylamine	40.000	39.109	2.2	53	0.00
67	4-Bromophenyl-phenylether	40.000	43.218	-8.0	58	0.00
68	Hexachlorobenzene	40.000	42.878	-7.2	57	0.00
69	Atrazine	40.000	41.313	-3.3	54	0.00
70 C	Pentachlorophenol	40.000	44.483	-11.2	56	0.00
71	Phenanthrene	40.000	38.177	4.6	52	0.00
72	Anthracene	40.000	38.603	3.5	52	0.00
73	Carbazole	40.000	37.204	7.0	50	0.00
74	Di-n-butylphthalate	40.000	37.362	6.6	50	0.00
75 C	Fluoranthene	40.000	39.151	2.1	54	0.00
76 I	Chrysene-d12	20.000	20.000	0.0	53	0.00
77	Benzydine	40.000	39.923	0.2	50	0.00
78	Pyrene	40.000	38.687	3.3	52	0.00
79 S	Terphenyl-d14	80.000	85.398	-6.7	58	0.00
80	Butylbenzylphthalate	40.000	37.335	6.7	51	0.00
81	Benzo(a)anthracene	40.000	38.925	2.7	53	0.00
82	3,3'-Dichlorobenzidine	40.000	38.918	2.7	53	0.00
83	Chrysene	40.000	38.761	3.1	53	0.00
84	Bis(2-ethylhexyl)phthalate	40.000	35.496	11.3	49	0.00
85 c	Di-n-octyl phthalate	40.000	35.145	12.1	48	0.00
86 I	Perylene-d12	20.000	20.000	0.0	53	0.00
87	Indeno(1,2,3-cd)pyrene	40.000	38.138	4.7	53	0.00
88	Benzo(b)fluoranthene	40.000	38.689	3.3	53	0.00
89	Benzo(k)fluoranthene	40.000	37.779	5.6	53	0.00
90 C	Benzo(a)pyrene	40.000	37.648	5.9	53	0.00
91	Dibenzo(a,h)anthracene	40.000	38.309	4.2	52	0.00
92	Benzo(g,h,i)perylene	40.000	38.559	3.6	53	0.00

Data Path : Z:\svoasrv\HPCHEM1\BNA_G\Data\BG041924\
Data File : BG060955.D
Acq On : 19 Apr 2024 11:24
Operator : MA/JU
Sample : SSTDCCC040
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_G
LabSampleId :
SSTDCCC040

Quant Time: Apr 19 12:07:51 2024
Quant Method : Z:\svoasrv\HPCHEM1\BNA_G\Methods\8270-BG041724.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Thu Apr 18 02:02:19 2024
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

Compound	Amount	Calc.	%Dev	Area%	Dev(min)
----------	--------	-------	------	-------	----------

(#) = Out of Range SPCC's out = 0 CCC's out = 0