



Data Path : Z:\svoasrv\HPCHEM1\BNA_G\Data\BG042022\
 Data File : BG053209.D
 Acq On : 20 Apr 2022 9:58
 Operator : CG/JU
 Sample : DFTPP
 Misc :
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 BNA_G
 ClientSampleId :
 DFTPP

Integration File: rteint.p

Method : Z:\svoasrv\HPCHEM1\BNA_G\Methods\8270-BG040822.M
 Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 Last Update : Fri Apr 08 15:31:24 2022






AutoFind: Scans 2471, 2472, 2473; Background Corrected with Scan 2464

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
51	198	10	80	36.5	33352	PASS
68	69	0.00	2	1.1	388	PASS
69	198	0.00	100	38.2	34888	PASS
70	69	0.00	2	0.3	114	PASS
127	198	10	80	43.3	39573	PASS
197	198	0.00	2	0.0	0	PASS
198	198	100	100	100.0	91352	PASS
199	198	5	9	6.9	6278	PASS
275	198	10	60	32.6	29752	PASS
365	198	1	100	5.3	4863	PASS
441	198	0.01	100	15.9	14524	PASS
442	442	50	100	100.0	89523	PASS
443	442	15	24	19.8	17689	PASS

Data Path : Z:\svoasrv\HPCHEM1\BNA_G\Data\BG042022\
Data File : BG053209.D
Acq On : 20 Apr 2022 9:58
Operator : CG/JU
Sample : DFTPP
Misc :
ALS Vial : 1 Sample Multiplier: 1

Instrument :
BNA_G
ClientSampleId :
DFTPP

Quant Time: Apr 20 18:40:55 2022
Quant Method : Z:\svoasrv\HPCHEM1\BNA_G\Methods\8270-BG040822.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Fri Apr 08 15:31:24 2022
Response via : Initial Calibration



TIC: BG053209.D\data.ms

(70) Pentachlorophenol (C)

17.142min (-0.003) 23091.91 ng




response 103674

Ion	Exp%	Act%
265.70	100.00	100.00
268.00	57.40	65.75
264.00	56.60	60.82
0.00	0.00	0.00

Data Path : Z:\svoasrv\HPCHEM1\BNA_G\Data\BG042022\
Data File : BG053209.D
Acq On : 20 Apr 2022 9:58
Operator : CG/JU
Sample : DFTPP
Misc :
ALS Vial : 1 Sample Multiplier: 1

Instrument :
BNA_G
ClientSampleId :
DFTPP

Quant Time: Apr 20 18:40:55 2022
Quant Method : Z:\svoasrv\HPCHEM1\BNA_G\Methods\8270-BG040822.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Fri Apr 08 15:31:24 2022
Response via : Initial Calibration



TIC: BG053209.D\data.ms

(77) Benzidine

19.709min (-0.003) 444892.11 ng

response 710835

Ion	Exp%	Act%
184.00	100.00	100.00
185.00	14.40	15.08
183.00	11.80	12.74
0.00	0.00	0.00

P

DDT Breakdown

Date	Instrument Name	DFTPP Data File
4/20/2022	BNA_G	BG053209.D
Compound Name	Response	Retention Time
DDT	335013	20.94
DDD	6593	20.5
DDE	1197	19.99
SUM(DDD+DDE)	SUM(DDT+DDD+DDE)	% Breakdown Of DDT
7790	342803	2.27

Instrument :
BNA_G
ClientSampleId :
DFTPP