

Data Path : Z:\svoasrv\HPCHEM1\BNA_G\Data\BG042221\
 Data File : BG048849.D
 Acq On : 22 Apr 2021 16:23
 Operator : CG/JU
 Sample : M2099-03
 Misc :
 ALS Vial : 12 Sample Multiplier: 1

Instrument :
 BNA_G
 ClientSampleId :
 TP-2

Quant Time: Apr 22 17:27:42 2021
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_G\METHODS\8270-BG041421.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Tue Apr 20 17:33:30 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)


Internal Standards						
1) 1,4-Dichlorobenzene-d4	8.062	152	22488	20.00	ng	0.00
21) Naphthalene-d8	10.888	136	97802	20.00	ng	0.00
39) Acenaphthene-d10	14.725	164	60004	20.00	ng	0.00
64) Phenanthrene-d10	17.480	188	131808	20.00	ng	# 0.00
76) Chrysene-d12	21.781	240	120424	20.00	ng	0.00
86) Perylene-d12	25.112	264	122934	20.00	ng	0.00
System Monitoring Compounds						
5) 2-Fluorophenol	5.612	112	160583	119.43	ng	0.00
7) Phenol-d6	7.228	99	198567	104.41	ng	0.00
23) Nitrobenzene-d5	9.237	82	157203	71.96	ng	0.00
42) 2,4,6-Tribromophenol	16.217	330	77512	96.34	ng	0.00
45) 2-Fluorobiphenyl	13.338	172	290237	69.85	ng	0.00
79) Terphenyl-d14	20.095	244	404388	57.80	ng	0.00
Target Compounds						
						Qvalue
10) Phenol	7.251	94	3752	2.01	ng	# 73
19) n-Nitroso-di-n-propyla...	9.231	70	24963	19.45	ng	# 77
31) Naphthalene	10.941	128	13533	2.70	ng	# 94
37) 2-Methylnaphthalene	12.774	142	8900	2.22	ng	89
38) 1-Methylnaphthalene	12.774	142	8900	2.39	ng	# 88
41) Hexachlorocyclopentadiene	13.179	237	119	5.96	ng	# 31
50) Dimethylphthalate	14.161	163	9131	2.03	ng	# 94
52) Acenaphthene	14.783	154	8889	2.65	ng	93
54) 2,4-Dinitrophenol	15.089	184	53	5.36	ng	# 37
55) Dibenzofuran	15.124	168	19452	3.50	ng	# 85
67) 4-Bromophenyl-phenylether	16.223	248	5980	3.94	ng	# 1
71) Phenanthrene	17.521	178	57095	8.35	ng	96
72) Anthracene	17.615	178	17020	2.51	ng	93
75) Fluoranthene	19.537	202	84812	9.80	ng	98
77) Benzidine	20.095	184	8777	2.68	ng	# 84
78) Pyrene	19.901	202	73143	8.78	ng	97
81) Benzo(a)anthracene	21.763	228	35863	4.36	ng	98
83) Chrysene	21.763	228	35863	4.58	ng	98
87) Indeno(1,2,3-cd)pyrene	28.926	276	25566	2.83	ng	# 60
88) Benzo(b)fluoranthene	24.043	252	63009	7.56	ng	# 94
89) Benzo(k)fluoranthene	24.114	252	15724	2.00	ng	# 92
90) Benzo(a)pyrene	24.948	252	28588	3.72	ng	# 93
92) Benzo(g,h,i)perylene	30.130	276	23848	3.12	ng	# 87

(#) = qualifier out of range (m) = manual integration (+) = signals summed


Data Path : Z:\svoasrv\HPCHEM1\BNA_G\Data\BG042221\
Data File : BG048849.D
Acq On : 22 Apr 2021 16:23
Operator : CG/JU
Sample : M2099-03
Misc :
ALS Vial : 12 Sample Multiplier: 1

Instrument :
BNA_G
ClientSampleId :
TP-2

Quant Time: Apr 22 17:27:42 2021
Quant Method : Z:\SVOASRV\HPCHEM1\BNA_G\METHODS\8270-BG041421.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Tue Apr 20 17:33:30 2021
Response via : Initial Calibration






Abundance Scan 807 (8.079 min): BG048693.D\data.ms (-800)#1




1,4-Dichlorobenzene-d4
Concen: 20.00 ng
RT: 8.062 min Scan# 804
Delta R.T. 0.001 min
Lab File: BG048849.D
Acq: 22 Apr 2021 16:23

Instrument :
BNA_G
ClientSampled :
TP-2

Tgt Ion	Ratio	Lower	Upper
152	100		
150	155.3	119.4	179.0
115	68.2	60.2	90.4







Abundance Scan 390 (5.629 min): BG048693.D\data.ms (-383)#5



2-Fluorophenol
Concen: 119.43 ng
RT: 5.612 min Scan# 387
Delta R.T. 0.006 min
Lab File: BG048849.D
Acq: 22 Apr 2021 16:23

Tgt Ion	Ratio	Lower	Upper
112	100		
64	57.7	46.7	70.1
63	40.4	35.3	52.9









Phenol-d6
 Concen: 104.41 ng
 RT: 7.228 min Scan# 662
 Delta R.T. 0.006 min
 Lab File: BG048849.D
 Acq: 22 Apr 2021 16:23




Instrument :
 BNA_G
 ClientSampled :
 TP-2


Tgt Ion:	99	Resp:	198567
Ion	Ratio	Lower	Upper
99	100		
42	25.7	20.5	30.7
71	56.4	41.5	62.3



Phenol
 Concen: 2.01 ng
 RT: 7.251 min Scan# 666
 Delta R.T. 0.001 min
 Lab File: BG048849.D
 Acq: 22 Apr 2021 16:23

Tgt Ion:	94	Resp:	3752
Ion	Ratio	Lower	Upper
94	100		
65	31.7	12.9	52.9
66	90.6	39.4	79.4#









n-Nitroso-di-n-propylamine
 Concen: 19.45 ng
 RT: 9.231 min Scan# 1003
 Delta R.T. 0.359 min
 Lab File: BG048849.D
 Acq: 22 Apr 2021 16:23




Instrument :
 BNA_G
 ClientSampled :
 TP-2


Tgt Ion	Ratio	Lower	Upper
70	100		
42	56.5	54.5	81.7
101	0.0	9.0	13.6#
130	0.0	17.0	25.6#



Naphthalene-d8
 Concen: 20.00 ng
 RT: 10.888 min Scan# 1285
 Delta R.T. -0.005 min
 Lab File: BG048849.D
 Acq: 22 Apr 2021 16:23

Tgt Ion	Ratio	Lower	Upper
136	100		
137	9.3	8.2	12.2
54	7.6	5.5	8.3
68	2.7	2.3	3.5









Nitrobenzene-d5
 Concen: 71.96 ng
 RT: 9.237 min Scan# 1004
 Delta R.T. 0.001 min
 Lab File: BG048849.D
 Acq: 22 Apr 2021 16:23




Instrument :
 BNA_G
 ClientSampled :
 TP-2


Tgt Ion:	82	Resp:	157203
Ion	Ratio	Lower	Upper
82	100		
128	37.0	29.8	44.6
54	53.5	45.8	68.8



Naphthalene
 Concen: 2.70 ng
 RT: 10.941 min Scan# 1294
 Delta R.T. 0.001 min
 Lab File: BG048849.D
 Acq: 22 Apr 2021 16:23

Tgt Ion:	128	Resp:	13533
Ion	Ratio	Lower	Upper
128	100		
129	14.4	8.5	12.7#
127	11.5	10.2	15.4









2-Methylnaphthalene
Concen: 2.22 ng
RT: 12.774 min Scan# 1606
Delta R.T. 0.224 min
Lab File: BG048849.D
Acq: 22 Apr 2021 16:23




Instrument :
BNA_G
ClientSampled :
TP-2

Tgt Ion:	142	Resp:	8900
Ion Ratio	Lower	Upper	
142	100		
141	77.1	68.8	103.2
115	36.8	36.4	54.6




1-Methylnaphthalene
Concen: 2.39 ng
RT: 12.774 min Scan# 1606
Delta R.T. 0.006 min
Lab File: BG048849.D
Acq: 22 Apr 2021 16:23

Tgt Ion:	142	Resp:	8900
Ion Ratio	Lower	Upper	
142	100		
141	77.1	70.6	106.0
116	6.9	4.0	6.0#





Abundance Scan 1940 (14.736 min): BG048693.D\data.ms (-1939) (-)




Acenaphthene-d10
Concen: 20.00 ng
RT: 14.725 min Scan# 1938
Delta R.T. 0.001 min
Lab File: BG048849.D
Acq: 22 Apr 2021 16:23

Instrument :
BNA_G
ClientSampled :
TP-2


Tgt Ion:	164	Resp:	60004
Ion	Ratio	Lower	Upper
164	100		
162	100.6	81.4	122.0
160	42.5	34.9	52.3



Abundance Scan 1938 (14.725 min): BG048849.D\data.ms (-1848) (-)





Abundance Scan 1629 (12.908 min): BG048693.D\data.ms (-1621) (-)




Hexachlorocyclopentadiene
Concen: 5.96 ng
RT: 13.179 min Scan# 1675
Delta R.T. 0.288 min
Lab File: BG048849.D
Acq: 22 Apr 2021 16:23

Tgt Ion:	237	Resp:	119
Ion	Ratio	Lower	Upper
237	100		
235	0.0	36.6	76.6#
272	0.0	0.0	32.5




Abundance Scan 1675 (13.179 min): BG048849.D\data.ms (-1536) (-)






Abundance Scan 2194 (16.228 min): BG048693.D\data.ms (-1642) (-)

2,4,6-Tribromophenol
Concen: 96.34 ng
RT: 16.217 min Scan# 2192
Delta R.T. 0.001 min
Lab File: BG048849.D
Acq: 22 Apr 2021 16:23

Instrument :
BNA_G
ClientSampled :
TP-2




Tgt Ion:	330	Resp:	77512
Ion	Ratio	Lower	Upper
330	100		
332	96.4	77.3	115.9
141	44.4	35.6	53.4






Abundance Scan 1705 (13.355 min): BG048693.D\data.ms (-1645) (-)

2-Fluorobiphenyl
Concen: 69.85 ng
RT: 13.338 min Scan# 1702
Delta R.T. -0.005 min
Lab File: BG048849.D
Acq: 22 Apr 2021 16:23




Tgt Ion:	172	Resp:	290237
Ion	Ratio	Lower	Upper
172	100		
171	34.7	28.6	43.0
170	24.0	18.6	27.8






Abundance Scan 1846 (14.183 min): BG048693.D\data.ms (-1750) (-)

Dimethylphthalate
Concen: 2.03 ng
RT: 14.161 min Scan# 1842
Delta R.T. -0.011 min
Lab File: BG048849.D
Acq: 22 Apr 2021 16:23

Instrument :
BNA_G
ClientSampled :
TP-2




Tgt Ion	Ratio	Lower	Upper
163	100		
194	6.8	3.0	4.6#
164	9.0	8.6	12.8







Abundance Scan 1952 (14.806 min): BG048693.D\data.ms (-1752) (-)

Acenaphthene
Concen: 2.65 ng
RT: 14.783 min Scan# 1948
Delta R.T. -0.005 min
Lab File: BG048849.D
Acq: 22 Apr 2021 16:23



Tgt Ion	Ratio	Lower	Upper
154	100		
153	101.7	87.0	130.4
152	44.8	40.1	60.1









2,4-Dinitrophenol
 Concen: 5.36 ng
 RT: 15.089 min Scan# 2000
 Delta R.T. 0.253 min
 Lab File: BG048849.D
 Acq: 22 Apr 2021 16:23




Instrument :
 BNA_G
 ClientSampled :
 TP-2

Tgt Ion:	184	Resp:	53
Ion	Ratio	Lower	Upper
184	100		
63	121.2	68.3	102.5#
154	0.0	62.6	93.8#



Dibenzofuran
 Concen: 3.50 ng
 RT: 15.124 min Scan# 2006
 Delta R.T. 0.001 min
 Lab File: BG048849.D
 Acq: 22 Apr 2021 16:23


Tgt Ion:	168	Resp:	19452
Ion	Ratio	Lower	Upper
168	100		
139	29.2	33.4	50.2#
169	14.7	12.3	18.5






Abundance Scan 2410 (17.497 min): BG048693.D\data.ms (-2402) (-)

Phenanthrene-d10
Concen: 20.00 ng
RT: 17.480 min Scan# 2407
Delta R.T. 0.001 min
Lab File: BG048849.D
Acq: 22 Apr 2021 16:23

Instrument :
BNA_G
ClientSampled :
TP-2




Tgt Ion:	188	Resp:	131808
Ion Ratio	Lower	Upper	
188	100		
94	4.3	5.4	8.0#
80	6.3	3.7	5.5#






Abundance Scan 2270 (16.674 min): BG048693.D\data.ms (-2267) (-)


4-Bromophenyl-phenylether
Concen: 3.94 ng
RT: 16.223 min Scan# 2193
Delta R.T. -0.440 min
Lab File: BG048849.D
Acq: 22 Apr 2021 16:23



Tgt Ion:	248	Resp:	5980
Ion Ratio	Lower	Upper	
248	100		
250	188.8	76.1	114.1#
141	514.8	65.4	98.2#





Abundance Scan 2417 (17.538 min): BG048693.D\data.ms (-2479) (-)




Phenanthrene
Concen: 8.35 ng
RT: 17.521 min Scan# 2414
Delta R.T. 0.001 min
Lab File: BG048849.D
Acq: 22 Apr 2021 16:23

Instrument :
BNA_G
ClientSampled :
TP-2


Tgt Ion:	178	Resp:	57095
Ion	Ratio	Lower	Upper
178	100		
176	19.1	16.0	24.0
179	15.2	14.2	21.4



Abundance Scan 2414 (17.521 min): BG048849.D\data.ms (-2324) (-)





Abundance Scan 2432 (17.626 min): BG048693.D\data.ms (-2479) (-)





Anthracene
Concen: 2.51 ng
RT: 17.615 min Scan# 2430
Delta R.T. 0.001 min
Lab File: BG048849.D
Acq: 22 Apr 2021 16:23

Tgt Ion:	178	Resp:	17020
Ion	Ratio	Lower	Upper
178	100		
176	21.5	14.6	21.8
179	14.3	13.4	20.0



Abundance Scan 2430 (17.615 min): BG048849.D\data.ms (-2340) (-)









Fluoranthene
Concen: 9.80 ng
RT: 19.537 min Scan# 2757
Delta R.T. 0.001 min
Lab File: BG048849.D
Acq: 22 Apr 2021 16:23




Instrument :
BNA_G
ClientSampled :
TP-2

Tgt Ion:	202	Resp:	84812
Ion Ratio	Lower	Upper	
202	100		
101	4.7	0.0	24.7
203	19.7	0.0	38.5




Chrysene-d12
Concen: 20.00 ng
RT: 21.781 min Scan# 3139
Delta R.T. 0.001 min
Lab File: BG048849.D
Acq: 22 Apr 2021 16:23

Tgt Ion:	240	Resp:	120424
Ion Ratio	Lower	Upper	
240	100		
120	5.0	4.2	6.2
236	25.3	21.8	32.6






Abundance Scan 2789 (19.724 min): BG048693.D\data.ms (-2789) (-)




Benzidine
Concen: 2.68 ng
RT: 20.095 min Scan# 2852
Delta R.T. 0.383 min
Lab File: BG048849.D
Acq: 22 Apr 2021 16:23

Instrument :
BNA_G
ClientSampled :
TP-2

Tgt Ion:184 Resp: 8777
Ion Ratio Lower Upper
184 100
185 20.7 10.7 16.1#
183 6.7 9.4 14.0#






Abundance Scan 2821 (19.912 min): BG048693.D\data.ms (-2821) (-)




Pyrene
Concen: 8.78 ng
RT: 19.901 min Scan# 2819
Delta R.T. 0.001 min
Lab File: BG048849.D
Acq: 22 Apr 2021 16:23

Tgt Ion:202 Resp: 73143
Ion Ratio Lower Upper
202 100
200 18.1 15.8 23.8
203 19.3 14.6 21.8






Abundance Scan 2854 (20.106 min): BG048693.D\data.ms (-2679) (-)




Terphenyl-d14
Concen: 57.80 ng
RT: 20.095 min Scan# 2852
Delta R.T. 0.001 min
Lab File: BG048849.D
Acq: 22 Apr 2021 16:23

Instrument :
BNA_G
ClientSampled :
TP-2

Tgt Ion:	244	Resp:	404388
Ion Ratio	Lower	Upper	
244	100		
212	7.9	6.9	10.3
122	6.4	5.0	7.6






Abundance Scan 3138 (21.774 min): BG048693.D\data.ms (-3129) (-)



Benzo(a)anthracene
Concen: 4.36 ng
RT: 21.763 min Scan# 3136
Delta R.T. 0.006 min
Lab File: BG048849.D
Acq: 22 Apr 2021 16:23


Tgt Ion:	228	Resp:	35863
Ion Ratio	Lower	Upper	
228	100		
226	30.3	22.7	34.1
229	19.9	16.0	24.0






Abundance Scan 3150 (21.845 min): BG048693.D\data.ms (-3150) (-)

Chrysene
Concen: 4.58 ng
RT: 21.763 min Scan# 3136
Delta R.T. -0.064 min
Lab File: BG048849.D
Acq: 22 Apr 2021 16:23

Instrument :
BNA_G
ClientSampled :
TP-2




Tgt Ion	Ratio	Lower	Upper
228	100		
226	30.3	23.1	34.7
229	19.9	15.3	22.9






Abundance Scan 3710 (25.135 min): BG048693.D\data.ms (-3696) (-)


Perylene-d12
Concen: 20.00 ng
RT: 25.112 min Scan# 3706
Delta R.T. 0.001 min
Lab File: BG048849.D
Acq: 22 Apr 2021 16:23



Tgt Ion	Ratio	Lower	Upper
264	100		
260	24.1	19.4	29.2
265	22.7	16.9	25.3






Abundance Scan 4364 (28.978 min): BG048693.D\data.ms (-4364) (-)




Indeno(1,2,3-cd)pyrene
Concen: 2.83 ng
RT: 28.926 min Scan# 4355
Delta R.T. -0.011 min
Lab File: BG048849.D
Acq: 22 Apr 2021 16:23

Instrument :
BNA_G
ClientSampled :
TP-2

Tgt Ion	Ratio	Lower	Upper
276	100		
138	7.5	5.1	7.7
277	0.0	19.8	29.6#







Abundance Scan 3530 (24.078 min): BG048693.D\data.ms (-3530) (-)



Benzo(b)fluoranthene
Concen: 7.56 ng
RT: 24.043 min Scan# 3524
Delta R.T. -0.005 min
Lab File: BG048849.D
Acq: 22 Apr 2021 16:23

Tgt Ion	Ratio	Lower	Upper
252	100		
253	22.9	16.2	24.4
125	7.2	3.0	4.4#









Benzo(k)fluoranthene
 Concen: 2.00 ng
 RT: 24.114 min Scan# 3536
 Delta R.T. -0.005 min
 Lab File: BG048849.D
 Acq: 22 Apr 2021 16:23




Instrument :
 BNA_G
 ClientSampled :
 TP-2

Tgt Ion:	252	Resp:	15724
Ion	Ratio	Lower	Upper
252	100		
253	18.9	18.0	27.0
125	7.9	3.1	4.7#




Benzo(a)pyrene
 Concen: 3.72 ng
 RT: 24.948 min Scan# 3678
 Delta R.T. -0.005 min
 Lab File: BG048849.D
 Acq: 22 Apr 2021 16:23

Tgt Ion:	252	Resp:	28588
Ion	Ratio	Lower	Upper
252	100		
253	23.0	16.6	25.0
125	9.3	3.5	5.3#




Abundance Scan 4569 (30.182 min): BG048693.D\data.ms (-4592) (-)



Benzo(g,h,i)perylene
Concen: 3.12 ng
RT: 30.130 min Scan# 4560
Delta R.T. 0.001 min
Lab File: BG048849.D
Acq: 22 Apr 2021 16:23

Instrument :
BNA_G
ClientSampleId :
TP-2

Abundance Scan 4560 (30.130 min): BG048849.D\data.ms



Tgt Ion:	276	Resp:	23848
Ion	Ratio	Lower	Upper
276	100		
277	14.2	18.4	27.6#
138	9.2	7.0	10.6

Abundance Scan 4560 (30.130 min): BG048849.D\data.ms (-4470) (-)

