

Data Path : Z:\svoasrv\HPCHEM1\BNA_G\Data\BG042221\
 Data File : BG048846.D
 Acq On : 22 Apr 2021 14:21
 Operator : CG/JU
 Sample : M2047-05
 Misc :
 ALS Vial : 9 Sample Multiplier: 1

Instrument :
 BNA_G
 ClientSampleId :
 130-SB3-(1.5-2)

Quant Time: Apr 22 14:56:50 2021
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_G\METHODS\8270-BG041421.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Tue Apr 20 17:33:30 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	8.065	152	21719	20.00	ng	# 0.00
21) Naphthalene-d8	10.891	136	86981	20.00	ng	0.00
39) Acenaphthene-d10	14.722	164	59185	20.00	ng	0.00
64) Phenanthrene-d10	17.477	188	143477	20.00	ng	# 0.00
76) Chrysene-d12	21.778	240	128328	20.00	ng	0.00
86) Perylene-d12	25.109	264	131285	20.00	ng	0.00
System Monitoring Compounds						
5) 2-Fluorophenol	5.609	112	120336	92.67	ng	0.00
7) Phenol-d6	7.224	99	161720	88.04	ng	0.00
23) Nitrobenzene-d5	9.234	82	124758	64.21	ng	0.00
42) 2,4,6-Tribromophenol	16.220	330	71967	90.69	ng	0.00
45) 2-Fluorobiphenyl	13.341	172	253406	61.83	ng	0.00
79) Terphenyl-d14	20.092	244	417283	55.97	ng	0.00

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_G\Data\BG042221\
Data File : BG048846.D
Acq On : 22 Apr 2021 14:21
Operator : CG/JU
Sample : M2047-05
Misc :
ALS Vial : 9 Sample Multiplier: 1

Instrument :
BNA_G
ClientSampleId :
130-SB3-(1.5-2)

Quant Time: Apr 22 14:56:50 2021
Quant Method : Z:\SVOASRV\HPCHEM1\BNA_G\METHODS\8270-BG041421.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Tue Apr 20 17:33:30 2021
Response via : Initial Calibration

