

Data Path : Z:\SVOASRV\HPCHEM1\BNA G\DATA\BG042320\
 Data File : BG045144.D
 Acq On : 23 Apr 2020 16:31
 Operator : CG/JU
 Sample : L2351-06
 Misc :
 ALS Vial : 7 Sample Multiplier: 1

Instrument :
 BNA_G
 ClientSampleId :
 TP-20A

Quant Time: Apr 23 17:08:05 2020
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA G\METHODS\8270-BG041520.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Thu Apr 23 15:21:03 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	8.01	152	92262	20.00	ng	0.00
21) Naphthalene-d8	10.81	136	391233	20.00	ng	0.00
39) Acenaphthene-d10	14.64	164	280111	20.00	ng	0.00
64) Phenanthrene-d10	17.39	188	602525	20.00	ng	0.00
76) Chrysene-d12	21.65	240	507667	20.00	ng	0.00
87) Perylene-d12	24.83	264	541459	20.00	ng	0.00
System Monitoring Compounds						
5) 2-Fluorophenol	5.58	112	368061	65.86	ng	0.00
7) Phenol-d6	7.17	99	669662	80.65	ng	0.00
23) Nitrobenzene-d5	9.16	82	476179	55.54	ng	0.00
42) 2,4,6-Tribromophenol	16.13	330	317034	73.26	ng	0.00
45) 2-Fluorobiphenyl	13.27	172	1176627	61.59	ng	0.00
79) Terphenyl-d14	20.00	244	1644380	70.60	ng	0.00
Target Compounds						
50) Dimethylphthalate	14.09	163	189669	8.316	ng	Qvalue 98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\SVOASRV\HPCHEM1\BNA G\DATA\BG042320\
Data File : BG045144.D
Acq On : 23 Apr 2020 16:31
Operator : CG/JU
Sample : L2351-06
Misc :
ALS Vial : 7 Sample Multiplier: 1

Instrument :
BNA_G
ClientSampleId :
TP-20A

Quant Time: Apr 23 17:08:05 2020
Quant Method : Z:\SVOASRV\HPCHEM1\BNA G\METHODS\8270-BG041520.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Thu Apr 23 15:21:03 2020
Response via : Initial Calibration

