


Data Path : Z:\svoasrv\HPCHEM1\BNA G\Data\BG042320\
 Data File : BG045138.D
 Acq On : 23 Apr 2020 12:34
 Operator : CG/JU
 Sample : DFTPP
 Misc :
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 BNA_G
 ClientSampleId :
 DFTPP

Integration File: rteint.p

Method : Z:\SVOASRV\HPCHEM1\BNA_G\METHODS\8270-BG041520.M
 Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 Last Update : Thu Apr 23 13:58:55 2020



AutoFind: Scans 2509, 2510, 2511; Background Corrected with Scan 2502

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
51	198	10	80	29.0	18490	PASS
68	69	0.00	2	0.9	203	PASS
69	198	0.00	100	36.4	23208	PASS
70	69	0.00	2	0.3	77	PASS
127	198	10	80	46.0	29344	PASS
197	198	0.00	2	0.0	0	PASS
198	198	100	100	100.0	63760	PASS
199	198	5	9	7.3	4659	PASS
275	198	10	60	31.3	19932	PASS
365	198	1	100	4.7	2999	PASS
441	198	0.01	100	12.8	8156	PASS
442	198	50	100	82.2	52394	PASS
443	442	15	24	19.4	10173	PASS