

Data Path : Z:\SVOASRV\HPCHEM1\BNA G\DATA\BG042420\
 Data File : BG045157.D
 Acq On : 24 Apr 2020 12:22
 Operator : CG/JU
 Sample : SSTDCCC040
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_G
 LabSampleId :
 SSTDCCC040

Quant Time: Apr 24 18:00:00 2020
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA G\METHODS\8270-BG041520.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Fri Apr 24 13:00:16 2020
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvqRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	92	0.00
2	1,4-Dioxane	0.538	0.536	0.4	92	0.00
3	Pyridine	1.658	1.568	5.4	91	0.00
4	n-Nitrosodimethylamine	0.508	0.499	1.8	95	0.00
5 S	2-Fluorophenol	1.211	1.210	0.1	95	0.00
6	Aniline	2.340	2.175	7.1	86	0.00
7 S	Phenol-d6	1.800	1.736	3.6	90	0.00
8	2-Chlorophenol	1.302	1.263	3.0	91	-0.01
9	Benzaldehyde	1.067	0.953	10.7	87	0.00
10 C	Phenol	1.801	1.709	5.1	88	0.00
11	bis(2-Chloroethyl)ether	1.434	1.329	7.3	87	0.00
12	1,3-Dichlorobenzene	1.534	1.500	2.2	93	0.00
13 C	1,4-Dichlorobenzene	1.529	1.522	0.5	93	0.00
14	1,2-Dichlorobenzene	1.463	1.439	1.6	91	0.00
15	Benzyl Alcohol	1.509	1.382	8.4	85	0.00
16	2,2'-oxybis(1-Chloropropane	1.408	1.240	11.9	82	0.00
17	2-Methylphenol	1.390	1.278	8.1	87	0.00
18	Hexachloroethane	0.575	0.553	3.8	89	-0.01
19 P	n-Nitroso-di-n-propylamine	1.273	1.155	9.3	84	0.00
20	3+4-Methylphenols	1.945	1.789	8.0	85	0.00
21 I	Naphthalene-d8	1.000	1.000	0.0	82	0.00
22	Acetophenone	0.541	0.524	3.1	83	0.00
23 S	Nitrobenzene-d5	0.438	0.443	-1.1	87	-0.01
24	Nitrobenzene	0.449	0.451	-0.4	85	-0.01
25	Isophorone	0.898	0.853	5.0	79	0.00
26 C	2-Nitrophenol	0.194	0.204	-5.2	89	0.00
27	2,4-Dimethylphenol	0.307	0.301	2.0	83	0.00
28	bis(2-Chloroethoxy)methane	0.472	0.445	5.7	79	0.00
29 C	2,4-Dichlorophenol	0.355	0.348	2.0	82	0.00
30	1,2,4-Trichlorobenzene	0.401	0.408	-1.7	86	0.00
31	Naphthalene	1.094	1.095	-0.1	83	0.00
32	Benzoic acid	0.230	0.163	29.1#	60	-0.01
33	4-Chloroaniline	0.510	0.478	6.3	79	0.00
34 C	Hexachlorobutadiene	0.275	0.284	-3.3	88	0.00
35	Caprolactam	0.141	0.127	9.9	75	0.00
36 C	4-Chloro-3-methylphenol	0.417	0.399	4.3	82	0.00
37	2-Methylnaphthalene	0.849	0.823	3.1	81	0.00
38	1-Methylnaphthalene	0.802	0.766	4.5	80	0.00
39 I	Acenaphthene-d10	1.000	1.000	0.0	78	0.00
40	1,2,4,5-Tetrachlorobenzene	0.644	0.649	-0.8	80	0.00
41 P	Hexachlorocyclopentadiene	0.402	0.583	-45.0#	113	0.00
42 S	2,4,6-Tribromophenol	0.309	0.307	0.6	79	0.00
43 C	2,4,6-Trichlorophenol	0.454	0.465	-2.4	81	0.00
44	2,4,5-Trichlorophenol	0.517	0.515	0.4	79	0.00

Data Path : Z:\SVOASRV\HPCHEM1\BNA G\DATA\BG042420\
 Data File : BG045157.D
 Acq On : 24 Apr 2020 12:22
 Operator : CG/JU
 Sample : SSTDCCC040
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_G
 LabSampleId :
 SSTDCCC040

Quant Time: Apr 24 18:00:00 2020
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA G\METHODS\8270-BG041520.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Fri Apr 24 13:00:16 2020
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvqRF	CCRF	%Dev	Area%	Dev(min)
45 S	2-Fluorobiphenyl	1.364	1.409	-3.3	81	0.00
46	1,1'-Biphenyl	1.520	1.531	-0.7	80	0.00
47	2-Chloronaphthalene	1.162	1.168	-0.5	81	0.00
48	2-Nitroaniline	0.382	0.389	-1.8	83	0.00
49	Acenaphthylene	1.892	1.885	0.4	79	0.00
50	Dimethylphthalate	1.628	1.605	1.4	78	0.00
51	2,6-Dinitrotoluene	0.364	0.369	-1.4	81	0.00
52 C	Acenaphthene	1.280	1.321	-3.2	82	0.00
53	3-Nitroaniline	0.399	0.386	3.3	78	0.00
54 P	2,4-Dinitrophenol	0.217	0.288	-32.7#	109	0.00
55	Dibenzofuran	1.866	1.869	-0.2	80	0.00
56 P	4-Nitrophenol	0.310	0.304	1.9	79	0.00
57	2,4-Dinitrotoluene	0.532	0.536	-0.8	82	0.00
58	Fluorene	1.524	1.533	-0.6	80	0.00
59	2,3,4,6-Tetrachlorophenol	0.460	0.461	-0.2	81	0.00
60	Diethylphthalate	1.698	1.647	3.0	78	-0.01
61	4-Chlorophenyl-phenylether	0.877	0.882	-0.6	81	0.00
62	4-Nitroaniline	0.414	0.400	3.4	78	0.00
63	Azobenzene	1.565	1.544	1.3	80	0.00
64 I	Phenanthrene-d10	1.000	1.000	0.0	78	0.00
65	4,6-Dinitro-2-methylphenol	0.131	0.143	-9.2	89	0.00
66 c	n-Nitrosodiphenylamine	0.575	0.572	0.5	79	0.00
67	4-Bromophenyl-phenylether	0.258	0.254	1.6	78	0.00
68	Hexachlorobenzene	0.268	0.268	0.0	81	0.00
69	Atrazine	0.230	0.212	7.8	78	0.00
70 C	Pentachlorophenol	0.164	0.182	-11.0	86	0.00
71	Phenanthrene	1.028	1.030	-0.2	80	0.00
72	Anthracene	1.031	1.058	-2.6	82	0.00
73	Carbazole	1.046	1.046	0.0	82	0.00
74	Di-n-butylphthalate	1.227	1.211	1.3	80	0.00
75 C	Fluoranthene	1.305	1.328	-1.8	84	0.00
76 I	Chrysene-d12	1.000	1.000	0.0	82	0.00
77	Benzidine	0.644	0.626	2.8	83	0.00
78	Pyrene	1.379	1.360	1.4	84	0.00
79 S	Terphenyl-d14	0.918	1.002	-9.2	87	0.00
80	Butylbenzylphthalate	0.572	0.585	-2.3	85	0.00
81	Benzo(a)anthracene	1.296	1.333	-2.9	86	0.00
82	3,3'-Dichlorobenzidine	0.516	0.529	-2.5	85	0.00
83	Chrysene	1.245	1.265	-1.6	86	0.00
84	Bis(2-ethylhexyl)phthalate	0.786	0.798	-1.5	85	0.00
85 c	Di-n-octyl phthalate	1.359	1.378	-1.4	85	0.00
86	Indeno(1,2,3-cd)pyrene	1.654	1.611	2.6	83	0.00
87 I	Perylene-d12	1.000	1.000	0.0	82	0.00

Data Path : Z:\SVOASRV\HPCHEM1\BNA G\DATA\BG042420\
Data File : BG045157.D
Acq On : 24 Apr 2020 12:22
Operator : CG/JU
Sample : SSTDCCC040
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_G
LabSampleId :
SSTDCCC040

Quant Time: Apr 24 18:00:00 2020
Quant Method : Z:\SVOASRV\HPCHEM1\BNA G\METHODS\8270-BG041520.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Fri Apr 24 13:00:16 2020
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
88	Benzo(b)fluoranthene	1.253	1.239	1.1	84	0.00
89	Benzo(k)fluoranthene	1.191	1.212	-1.8	84	0.00
90 C	Benzo(a)pyrene	1.183	1.188	-0.4	84	0.00
91	Dibenzo(a,h)anthracene	1.192	1.194	-0.2	85	0.00
92	Benzo(q,h,i)perylene	1.195	1.168	2.3	83	0.01

(#) = Out of Range

SPCC's out = 0 CCC's out = 0