

Data Path : Z:\svoasrv\HPCHEM1\BNA_G\Data\BG042423\
 Data File : BG057161.D
 Acq On : 24 Apr 2023 21:32
 Operator : CG/JU
 Sample : PB152179BL
 Misc :
 ALS Vial : 11 Sample Multiplier: 1

Instrument :
 BNA_G
 ClientSampleId :
 PB152179BL

Quant Time: Apr 25 05:19:17 2023
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_G\Methods\8270-BG042423.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Tue Apr 25 04:59:02 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	8.386	152	21896	20.000	ng	0.00
21) Naphthalene-d8	11.223	136	91571	20.000	ng	0.00
39) Acenaphthene-d10	15.000	164	67711	20.000	ng	0.00
64) Phenanthrene-d10	17.738	188	177763	20.000	ng	0.00
76) Chrysene-d12	22.056	240	165163	20.000	ng	0.00
86) Perylene-d12	25.574	264	169634	20.000	ng	0.00
System Monitoring Compounds						
5) 2-Fluorophenol	5.901	112	201012	166.363	ng	0.00
7) Phenol-d6	7.522	99	280318	160.346	ng	0.00
23) Nitrobenzene-d5	9.567	82	179382	96.843	ng	0.00
42) 2,4,6-Tribromophenol	16.481	330	112198	125.587	ng	0.00
45) 2-Fluorobiphenyl	13.626	172	434201	86.179	ng	0.00
79) Terphenyl-d14	20.311	244	885664	93.794	ng	0.00

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_G\Data\BG042423\
Data File : BG057161.D
Acq On : 24 Apr 2023 21:32
Operator : CG/JU
Sample : PB152179BL
Misc :
ALS Vial : 11 Sample Multiplier: 1

Instrument :
BNA_G
ClientSampleId :
PB152179BL

Quant Time: Apr 25 05:19:17 2023
Quant Method : Z:\svoasrv\HPCHEM1\BNA_G\Methods\8270-BG042423.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Tue Apr 25 04:59:02 2023
Response via : Initial Calibration

