

Data Path : Z:\svoasrv\HPCHEM1\BNA_G\Data\BG042423\
 Data File : BG057164.D
 Acq On : 24 Apr 2023 23:32
 Operator : CG/JU
 Sample : PB152303TB
 Misc :
 ALS Vial : 14 Sample Multiplier: 1

Instrument :
 BNA_G
 ClientSampleId :
 PB152303TB

Quant Time: Apr 25 05:20:24 2023
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_G\Methods\8270-BG042423.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Tue Apr 25 04:59:02 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	8.384	152	19869	20.000	ng	0.00
21) Naphthalene-d8	11.227	136	80763	20.000	ng	0.00
39) Acenaphthene-d10	14.998	164	59547	20.000	ng	0.00
64) Phenanthrene-d10	17.736	188	156025	20.000	ng	0.00
76) Chrysene-d12	22.053	240	148613	20.000	ng	0.00
86) Perylene-d12	25.572	264	151979	20.000	ng	0.00
System Monitoring Compounds						
5) 2-Fluorophenol	5.899	112	171601	156.511	ng	0.00
7) Phenol-d6	7.520	99	243503	153.497	ng	0.00
23) Nitrobenzene-d5	9.564	82	154998	94.877	ng	0.00
42) 2,4,6-Tribromophenol	16.485	330	99950	127.216	ng	0.00
45) 2-Fluorobiphenyl	13.624	172	385088	86.911	ng	0.00
79) Terphenyl-d14	20.309	244	790104	92.992	ng	0.00

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_G\Data\BG042423\
Data File : BG057164.D
Acq On : 24 Apr 2023 23:32
Operator : CG/JU
Sample : PB152303TB
Misc :
ALS Vial : 14 Sample Multiplier: 1

Instrument :
BNA_G
ClientSampleId :
PB152303TB

Quant Time: Apr 25 05:20:24 2023
Quant Method : Z:\svoasrv\HPCHEM1\BNA_G\Methods\8270-BG042423.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Tue Apr 25 04:59:02 2023
Response via : Initial Calibration

