

Data Path : Z:\svoasrv\HPCHEM1\BNA_G\Data\BG042423\
 Data File : BG057156.D
 Acq On : 24 Apr 2023 18:11
 Operator : CG/JU
 Sample : SSTDICCC040
 Misc :
 ALS Vial : 6 Sample Multiplier: 1

Instrument :
 BNA_G
 LabSampleId :
 SSTDICCC040

Quant Time: Apr 25 04:09:28 2023
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_G\Methods\8270-BG042423.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Tue Apr 25 04:01:27 2023
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	100	0.00
2	1,4-Dioxane	0.495	0.503	-1.6	100	0.00
3	Pyridine	1.501	1.619	-7.9	100	0.00
4	n-Nitrosodimethylamine	0.541	0.597	-10.4	100	0.00
5 S	2-Fluorophenol	1.104	1.149	-4.1	100	0.00
6	Aniline	2.085	2.178	-4.5	100	0.00
7 S	Phenol-d6	1.597	1.679	-5.1	100	0.00
8	2-Chlorophenol	1.156	1.188	-2.8	100	0.00
9	Benzaldehyde	0.956	0.948	0.8	100	0.00
10 C	Phenol	1.586	1.679	-5.9	100	0.00
11	bis(2-Chloroethyl)ether	1.244	1.299	-4.4	100	0.00
12	1,3-Dichlorobenzene	1.370	1.369	0.1	100	0.00
13 C	1,4-Dichlorobenzene	1.380	1.393	-0.9	100	0.00
14	1,2-Dichlorobenzene	1.333	1.353	-1.5	100	0.00
15	Benzyl Alcohol	1.177	1.231	-4.6	100	0.00
16	2,2'-oxybis(1-Chloropropane	1.344	1.465	-9.0	100	0.00
17	2-Methylphenol	1.160	1.227	-5.8	100	0.00
18	Hexachloroethane	0.468	0.483	-3.2	100	0.00
19 P	n-Nitroso-di-n-propylamine	0.993	1.157	-16.5	100	0.00
20	3+4-Methylphenols	1.588	1.672	-5.3	100	0.00
21 I	Naphthalene-d8	1.000	1.000	0.0	100	0.00
22	Acetophenone	0.543	0.575	-5.9	100	0.00
23 S	Nitrobenzene-d5	0.405	0.431	-6.4	100	0.00
24	Nitrobenzene	0.406	0.452	-11.3	100	0.00
25	Isophorone	0.776	0.838	-8.0	100	0.00
26 C	2-Nitrophenol	0.152	0.160	-5.3	100	0.00
27	2,4-Dimethylphenol	0.292	0.311	-6.5	100	0.00
28	bis(2-Chloroethoxy)methane	0.407	0.448	-10.1	100	0.00
29 C	2,4-Dichlorophenol	0.325	0.348	-7.1	100	0.00
30	1,2,4-Trichlorobenzene	0.397	0.408	-2.8	100	0.00
31	Naphthalene	1.062	1.095	-3.1	100	0.00
32	Benzoic acid	0.122	0.134	-9.8	100	0.00
33	4-Chloroaniline	0.467	0.489	-4.7	100	0.00
34 C	Hexachlorobutadiene	0.293	0.298	-1.7	100	0.00
35	Caprolactam	0.090	0.099	-10.0	100	0.00
36 C	4-Chloro-3-methylphenol	0.350	0.371	-6.0	100	0.00
37	2-Methylnaphthalene	0.836	0.867	-3.7	100	0.00
38	1-Methylnaphthalene	0.772	0.803	-4.0	100	0.00
39 I	Acenaphthene-d10	1.000	1.000	0.0	100	0.00
40	1,2,4,5-Tetrachlorobenzene	0.732	0.735	-0.4	100	0.00
41 P	Hexachlorocyclopentadiene	0.278	0.301	-8.3	100	0.00
42 S	2,4,6-Tribromophenol	0.264	0.257	2.7	100	0.00
43 C	2,4,6-Trichlorophenol	0.396	0.420	-6.1	100	0.00
44	2,4,5-Trichlorophenol	0.478	0.503	-5.2	100	0.00
45 S	2-Fluorobiphenyl	1.488	1.495	-0.5	100	0.00
46	1,1'-Biphenyl	1.491	1.524	-2.2	100	0.00
47	2-Chloronaphthalene	1.125	1.143	-1.6	100	0.00

Data Path : Z:\svoasrv\HPCHEM1\BNA_G\Data\BG042423\
 Data File : BG057156.D
 Acq On : 24 Apr 2023 18:11
 Operator : CG/JU
 Sample : SSTDICCC040
 Misc :
 ALS Vial : 6 Sample Multiplier: 1

Instrument :
 BNA_G
 LabSampleId :
 SSTDICCC040

Quant Time: Apr 25 04:09:28 2023
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_G\Methods\8270-BG042423.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Tue Apr 25 04:01:27 2023
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
48	2-Nitroaniline	0.213	0.273	-28.2#	100	0.00
49	Acenaphthylene	1.786	1.830	-2.5	100	0.00
50	Dimethylphthalate	1.404	1.458	-3.8	100	0.00
51	2,6-Dinitrotoluene	0.293	0.322	-9.9	100	0.00
52 C	Acenaphthene	1.167	1.214	-4.0	100	0.00
53	3-Nitroaniline	0.291	0.312	-7.2	100	0.00
54 P	2,4-Dinitrophenol	0.139	0.150	-7.9	100	0.00
55	Dibenzofuran	1.891	1.912	-1.1	100	0.00
56 P	4-Nitrophenol	0.205	0.218	-6.3	100	0.00
57	2,4-Dinitrotoluene	0.399	0.428	-7.3	100	0.00
58	Fluorene	1.522	1.537	-1.0	100	0.00
59	2,3,4,6-Tetrachlorophenol	0.410	0.410	0.0	100	0.00
60	Diethylphthalate	1.353	1.391	-2.8	100	0.00
61	4-Chlorophenyl-phenylether	0.914	0.914	0.0	100	0.00
62	4-Nitroaniline	0.293	0.310	-5.8	100	0.00
63	Azobenzene	1.439	1.556	-8.1	100	0.00
64 I	Phenanthrene-d10	1.000	1.000	0.0	100	0.00
65	4,6-Dinitro-2-methylphenol	0.095	0.102	-7.4	100	0.00
66 c	n-Nitrosodiphenylamine	0.522	0.544	-4.2	100	0.00
67	4-Bromophenyl-phenylether	0.241	0.246	-2.1	100	0.00
68	Hexachlorobenzene	0.247	0.243	1.6	100	0.00
69	Atrazine	0.167	0.178	-6.6	100	0.00
70 C	Pentachlorophenol	0.116	0.125	-7.8	100	0.00
71	Phenanthrene	1.021	1.049	-2.7	100	0.00
72	Anthracene	1.032	1.066	-3.3	100	0.00
73	Carbazole	0.871	0.912	-4.7	100	0.00
74	Di-n-butylphthalate	0.811	0.867	-6.9	100	0.00
75 C	Fluoranthene	1.268	1.289	-1.7	100	0.00
76 I	Chrysene-d12	1.000	1.000	0.0	100	0.00
77	Benzydine	0.301	0.328	-9.0	100	0.00
78	Pyrene	1.358	1.429	-5.2	100	0.00
79 S	Terphenyl-d14	1.143	1.171	-2.4	100	0.00
80	Butylbenzylphthalate	0.294	0.352	-19.7	100	0.00
81	Benzo(a)anthracene	1.375	1.435	-4.4	100	0.00
82	3,3'-Dichlorobenzidine	0.360	0.383	-6.4	100	0.00
83	Chrysene	1.326	1.367	-3.1	100	0.00
84	Bis(2-ethylhexyl)phthalate	0.494	0.549	-11.1	100	0.00
85 c	Di-n-octyl phthalate	0.897	0.953	-6.2	100	0.00
86 I	Perylene-d12	1.000	1.000	0.0	100	0.00
87	Indeno(1,2,3-cd)pyrene	1.433	1.464	-2.2	100	0.00
88	Benzo(b)fluoranthene	1.254	1.300	-3.7	100	0.00
89	Benzo(k)fluoranthene	1.297	1.333	-2.8	100	0.00
90 C	Benzo(a)pyrene	1.235	1.277	-3.4	100	0.00
91	Dibenzo(a,h)anthracene	1.190	1.221	-2.6	100	0.00
92	Benzo(g,h,i)perylene	1.177	1.203	-2.2	100	0.00

Data Path : Z:\svoasrv\HPCHEM1\BNA_G\Data\BG042423\
Data File : BG057156.D
Acq On : 24 Apr 2023 18:11
Operator : CG/JU
Sample : SSTDICCC040
Misc :
ALS Vial : 6 Sample Multiplier: 1

Instrument :
BNA_G
LabSampleId :
SSTDICCC040

Quant Time: Apr 25 04:09:28 2023
Quant Method : Z:\svoasrv\HPCHEM1\BNA_G\Methods\8270-BG042423.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Tue Apr 25 04:01:27 2023
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

Compound	AvgRF	CCRF	%Dev Area	% Dev(min)
----------	-------	------	-----------	------------

(#) = Out of Range

SPCC's out = 0 CCC's out = 0