

Data Path : Z:\svoasrv\HPCHEM1\BNA_G\Data\BG042423\
 Data File : BG057156.D
 Acq On : 24 Apr 2023 18:11
 Operator : CG/JU
 Sample : SSTDICCC040
 Misc :
 ALS Vial : 6 Sample Multiplier: 1

Instrument :
 BNA_G
 LabSampleId :
 SSTDICCC040

Quant Time: Apr 25 04:09:28 2023
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_G\Methods\8270-BG042423.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Tue Apr 25 04:01:27 2023
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	20.000	20.000	0.0	100	0.00
2	1,4-Dioxane	40.000	40.687	-1.7	100	0.00
3	Pyridine	40.000	43.168	-7.9	100	0.00
4	n-Nitrosodimethylamine	40.000	44.155	-10.4	100	0.00
5 S	2-Fluorophenol	80.000	83.301	-4.1	100	0.00
6	Aniline	40.000	41.776	-4.4	100	0.00
7 S	Phenol-d6	80.000	84.120	-5.2	100	0.00
8	2-Chlorophenol	40.000	41.103	-2.8	100	0.00
9	Benzaldehyde	40.000	39.651	0.9	100	0.00
10 C	Phenol	40.000	42.360	-5.9	100	0.00
11	bis(2-Chloroethyl)ether	40.000	41.764	-4.4	100	0.00
12	1,3-Dichlorobenzene	40.000	39.988	0.0	100	0.00
13 C	1,4-Dichlorobenzene	40.000	40.370	-0.9	100	0.00
14	1,2-Dichlorobenzene	40.000	40.596	-1.5	100	0.00
15	Benzyl Alcohol	40.000	41.860	-4.6	100	0.00
16	2,2'-oxybis(1-Chloropropane	40.000	37.180	7.1	100	0.00
17	2-Methylphenol	40.000	42.316	-5.8	100	0.00
18	Hexachloroethane	40.000	41.273	-3.2	100	0.00
19 P	n-Nitroso-di-n-propylamine	40.000	46.605	-16.5	100	0.00
20	3+4-Methylphenols	40.000	42.111	-5.3	100	0.00
21 I	Naphthalene-d8	20.000	20.000	0.0	100	0.00
22	Acetophenone	40.000	42.372	-5.9	100	0.00
23 S	Nitrobenzene-d5	80.000	85.187	-6.5	100	0.00
24	Nitrobenzene	40.000	44.479	-11.2	100	0.00
25	Isophorone	40.000	43.184	-8.0	100	0.00
26 C	2-Nitrophenol	40.000	42.227	-5.6	100	0.00
27	2,4-Dimethylphenol	40.000	42.658	-6.6	100	0.00
28	bis(2-Chloroethoxy)methane	40.000	44.068	-10.2	100	0.00
29 C	2,4-Dichlorophenol	40.000	42.800	-7.0	100	0.00
30	1,2,4-Trichlorobenzene	40.000	41.089	-2.7	100	0.00
31	Naphthalene	40.000	41.230	-3.1	100	0.00
32	Benzoic acid	40.000	40.328	-0.8	100	0.00
33	4-Chloroaniline	40.000	41.892	-4.7	100	0.00
34 C	Hexachlorobutadiene	40.000	40.681	-1.7	100	0.00
35	Caprolactam	40.000	43.995	-10.0	100	0.00
36 C	4-Chloro-3-methylphenol	40.000	42.513	-6.3	100	0.00
37	2-Methylnaphthalene	40.000	41.478	-3.7	100	0.00
38	1-Methylnaphthalene	40.000	41.631	-4.1	100	0.00
39 I	Acenaphthene-d10	20.000	20.000	0.0	100	0.00
40	1,2,4,5-Tetrachlorobenzene	40.000	40.189	-0.5	100	0.00
41 P	Hexachlorocyclopentadiene	40.000	40.907	-2.3	100	0.00
42 S	2,4,6-Tribromophenol	80.000	77.957	2.6	100	0.00
43 C	2,4,6-Trichlorophenol	40.000	42.370	-5.9	100	0.00
44	2,4,5-Trichlorophenol	40.000	42.031	-5.1	100	0.00
45 S	2-Fluorobiphenyl	80.000	80.371	-0.5	100	0.00
46	1,1'-Biphenyl	40.000	40.891	-2.2	100	0.00
47	2-Chloronaphthalene	40.000	40.621	-1.6	100	0.00

Data Path : Z:\svoasrv\HPCHEM1\BNA_G\Data\BG042423\
 Data File : BG057156.D
 Acq On : 24 Apr 2023 18:11
 Operator : CG/JU
 Sample : SSTDICCC040
 Misc :
 ALS Vial : 6 Sample Multiplier: 1

Instrument :
 BNA_G
 LabSampleId :
 SSTDICCC040

Quant Time: Apr 25 04:09:28 2023
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_G\Methods\8270-BG042423.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Tue Apr 25 04:01:27 2023
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
48	2-Nitroaniline	40.000	39.741	0.6	100	0.00
49	Acenaphthylene	40.000	40.986	-2.5	100	0.00
50	Dimethylphthalate	40.000	41.552	-3.9	100	0.00
51	2,6-Dinitrotoluene	40.000	43.985	-10.0	100	0.00
52 C	Acenaphthene	40.000	41.630	-4.1	100	0.00
53	3-Nitroaniline	40.000	42.917	-7.3	100	0.00
54 P	2,4-Dinitrophenol	40.000	39.752	0.6	100	0.00
55	Dibenzofuran	40.000	40.440	-1.1	100	0.00
56 P	4-Nitrophenol	40.000	42.633	-6.6	100	0.00
57	2,4-Dinitrotoluene	40.000	42.934	-7.3	100	0.00
58	Fluorene	40.000	40.392	-1.0	100	0.00
59	2,3,4,6-Tetrachlorophenol	40.000	40.023	-0.1	100	0.00
60	Diethylphthalate	40.000	41.117	-2.8	100	0.00
61	4-Chlorophenyl-phenylether	40.000	40.008	-0.0	100	0.00
62	4-Nitroaniline	40.000	42.378	-5.9	100	0.00
63	Azobenzene	40.000	43.239	-8.1	100	0.00
64 I	Phenanthrene-d10	20.000	20.000	0.0	100	0.00
65	4,6-Dinitro-2-methylphenol	40.000	40.280	-0.7	100	0.00
66 c	n-Nitrosodiphenylamine	40.000	41.686	-4.2	100	0.00
67	4-Bromophenyl-phenylether	40.000	40.925	-2.3	100	0.00
68	Hexachlorobenzene	40.000	39.460	1.3	100	0.00
69	Atrazine	40.000	42.835	-7.1	100	0.00
70 C	Pentachlorophenol	40.000	43.052	-7.6	100	0.00
71	Phenanthrene	40.000	41.086	-2.7	100	0.00
72	Anthracene	40.000	41.288	-3.2	100	0.00
73	Carbazole	40.000	41.841	-4.6	100	0.00
74	Di-n-butylphthalate	40.000	42.762	-6.9	100	0.00
75 C	Fluoranthene	40.000	40.661	-1.7	100	0.00
76 I	Chrysene-d12	20.000	20.000	0.0	100	0.00
77	Benzydine	40.000	43.498	-8.7	100	0.00
78	Pyrene	40.000	42.105	-5.3	100	0.00
79 S	Terphenyl-d14	80.000	81.961	-2.5	100	0.00
80	Butylbenzylphthalate	40.000	40.371	-0.9	100	0.00
81	Benzo(a)anthracene	40.000	41.736	-4.3	100	0.00
82	3,3'-Dichlorobenzidine	40.000	42.593	-6.5	100	0.00
83	Chrysene	40.000	41.222	-3.1	100	0.00
84	Bis(2-ethylhexyl)phthalate	40.000	39.869	0.3	100	0.00
85 c	Di-n-octyl phthalate	40.000	42.511	-6.3	100	0.00
86 I	Perylene-d12	20.000	20.000	0.0	100	0.00
87	Indeno(1,2,3-cd)pyrene	40.000	40.883	-2.2	100	0.00
88	Benzo(b)fluoranthene	40.000	41.477	-3.7	100	0.00
89	Benzo(k)fluoranthene	40.000	41.107	-2.8	100	0.00
90 C	Benzo(a)pyrene	40.000	41.353	-3.4	100	0.00
91	Dibenzo(a,h)anthracene	40.000	41.072	-2.7	100	0.00
92	Benzo(g,h,i)perylene	40.000	40.871	-2.2	100	0.00

Data Path : Z:\svoasrv\HPCHEM1\BNA_G\Data\BG042423\
Data File : BG057156.D
Acq On : 24 Apr 2023 18:11
Operator : CG/JU
Sample : SSTDICCC040
Misc :
ALS Vial : 6 Sample Multiplier: 1

Instrument :
BNA_G
LabSampleId :
SSTDICCC040

Quant Time: Apr 25 04:09:28 2023
Quant Method : Z:\svoasrv\HPCHEM1\BNA_G\Methods\8270-BG042423.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Tue Apr 25 04:01:27 2023
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

Compound	Amount	Calc.	%Dev	Area%	Dev(min)
----------	--------	-------	------	-------	----------

(#) = Out of Range

SPCC's out = 0 CCC's out = 0