

Data Path : Z:\svoasrv\HPCHEM1\BNA_G\Data\BG042423\
 Data File : BG057160.D
 Acq On : 24 Apr 2023 20:52
 Operator : CG/JU
 Sample : SSTDICV040
 Misc :
 ALS Vial : 10 Sample Multiplier: 1

Instrument :
 BNA_G
 ClientSampleId :
 ICVBG042423

Quant Time: Apr 25 05:00:07 2023
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_G\Methods\8270-BG042423.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Tue Apr 25 04:59:02 2023
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	20.000	20.000	0.0	89	0.00
2	1,4-Dioxane	40.000	46.671	-16.7	102	0.00
3	Pyridine	40.000	48.472	-21.2	100	0.00
4	n-Nitrosodimethylamine	40.000	55.755	-39.4#	112	0.00
5 S	2-Fluorophenol	80.000	89.831	-12.3	96	0.00
6	Aniline	40.000	46.268	-15.7	99	0.00
7 S	Phenol-d6	80.000	93.290	-16.6	99	0.00
8	2-Chlorophenol	40.000	44.391	-11.0	96	0.00
9	Benzaldehyde	40.000	45.126	-12.8	101	0.00
10 C	Phenol	40.000	46.789	-17.0	98	0.00
11	bis(2-Chloroethyl)ether	40.000	46.216	-15.5	99	0.00
12	1,3-Dichlorobenzene	40.000	41.310	-3.3	92	0.00
13 C	1,4-Dichlorobenzene	40.000	41.698	-4.2	92	0.00
14	1,2-Dichlorobenzene	40.000	41.811	-4.5	92	0.00
15	Benzyl Alcohol	40.000	50.250	-25.6#	107	0.00
16	2,2'-oxybis(1-Chloropropane	40.000	46.625	-16.6	116	0.00
17	2-Methylphenol	40.000	45.769	-14.4	96	0.00
18	Hexachloroethane	40.000	43.447	-8.6	94	0.00
19 P	n-Nitroso-di-n-propylamine	40.000	54.903	-37.3#	105	0.00
20	3+4-Methylphenols	40.000	46.013	-15.0	97	0.00
21 I	Naphthalene-d8	20.000	20.000	0.0	93	0.00
22	Acetophenone	40.000	43.795	-9.5	96	0.00
23 S	Nitrobenzene-d5	80.000	93.542	-16.9	102	0.00
24	Nitrobenzene	40.000	47.702	-19.3	100	0.00
25	Isophorone	40.000	47.446	-18.6	102	0.00
26 C	2-Nitrophenol	40.000	45.680	-14.2	100	0.00
27	2,4-Dimethylphenol	40.000	44.024	-10.1	96	0.00
28	bis(2-Chloroethoxy)methane	40.000	46.349	-15.9	98	0.00
29 C	2,4-Dichlorophenol	40.000	42.003	-5.0	91	0.00
30	1,2,4-Trichlorobenzene	40.000	39.014	2.5	88	0.00
31	Naphthalene	40.000	41.730	-4.3	94	0.00
32	Benzoic acid	40.000	44.622	-11.6	105	0.01
33	4-Chloroaniline	40.000	42.291	-5.7	94	0.00
34 C	Hexachlorobutadiene	40.000	36.885	7.8	84	0.00
35	Caprolactam	40.000	47.588	-19.0	100	0.01
36 C	4-Chloro-3-methylphenol	40.000	44.980	-12.4	98	0.00
37	2-Methylnaphthalene	40.000	40.724	-1.8	91	0.00
38	1-Methylnaphthalene	40.000	41.253	-3.1	92	0.00
39 I	Acenaphthene-d10	20.000	20.000	0.0	90	0.01
40	1,2,4,5-Tetrachlorobenzene	40.000	39.616	1.0	89	0.00
41 P	Hexachlorocyclopentadiene	40.000	40.211	-0.5	88	0.00
42 S	2,4,6-Tribromophenol	80.000	78.426	2.0	91	0.00
43 C	2,4,6-Trichlorophenol	40.000	43.405	-8.5	92	0.00
44	2,4,5-Trichlorophenol	40.000	41.255	-3.1	89	0.00
45 S	2-Fluorobiphenyl	80.000	80.464	-0.6	90	0.01
46	1,1'-Biphenyl	40.000	42.097	-5.2	93	0.00
47	2-Chloronaphthalene	40.000	40.780	-2.0	91	0.00

Data Path : Z:\svoasrv\HPCHEM1\BNA_G\Data\BG042423\
 Data File : BG057160.D
 Acq On : 24 Apr 2023 20:52
 Operator : CG/JU
 Sample : SSTDICV040
 Misc :
 ALS Vial : 10 Sample Multiplier: 1

Instrument :
 BNA_G
 ClientSampleId :
 ICVBG042423

Quant Time: Apr 25 05:00:07 2023
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_G\Methods\8270-BG042423.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Tue Apr 25 04:59:02 2023
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
48	2-Nitroaniline	40.000	47.345	-18.4	110	0.00
49	Acenaphthylene	40.000	41.508	-3.8	91	0.00
50	Dimethylphthalate	40.000	43.565	-8.9	95	0.01
51	2,6-Dinitrotoluene	40.000	45.919	-14.8	94	0.00
52 C	Acenaphthene	40.000	42.590	-6.5	92	0.01
53	3-Nitroaniline	40.000	45.286	-13.2	95	0.00
54 P	2,4-Dinitrophenol	40.000	41.860	-4.6	96	0.00
55	Dibenzofuran	40.000	40.329	-0.8	90	0.00
56 P	4-Nitrophenol	40.000	45.017	-12.5	95	0.00
57	2,4-Dinitrotoluene	40.000	45.243	-13.1	95	0.00
58	Fluorene	40.000	40.389	-1.0	90	0.00
59	2,3,4,6-Tetrachlorophenol	40.000	42.185	-5.5	95	0.00
60	Diethylphthalate	40.000	44.559	-11.4	98	0.01
61	4-Chlorophenyl-phenylether	40.000	39.462	1.3	89	0.00
62	4-Nitroaniline	40.000	46.408	-16.0	99	0.00
63	Azobenzene	40.000	46.898	-17.2	98	0.00
64 I	Phenanthrene-d10	20.000	20.000	0.0	90	0.00
65	4,6-Dinitro-2-methylphenol	40.000	42.691	-6.7	96	0.00
66 c	n-Nitrosodiphenylamine	40.000	42.622	-6.6	92	0.00
67	4-Bromophenyl-phenylether	40.000	39.061	2.3	86	0.00
68	Hexachlorobenzene	40.000	37.728	5.7	86	0.00
69	Atrazine	40.000	45.689	-14.2	96	0.00
70 C	Pentachlorophenol	40.000	43.704	-9.3	91	0.00
71	Phenanthrene	40.000	41.051	-2.6	90	0.00
72	Anthracene	40.000	41.684	-4.2	91	0.01
73	Carbazole	40.000	42.037	-5.1	90	0.00
74	Di-n-butylphthalate	40.000	45.773	-14.4	96	0.00
75 C	Fluoranthene	40.000	40.594	-1.5	90	0.00
76 I	Chrysene-d12	20.000	20.000	0.0	85	0.01
77	Benzydine	40.000	45.609	-14.0	89	0.01
78	Pyrene	40.000	44.239	-10.6	90	0.00
79 S	Terphenyl-d14	80.000	84.721	-5.9	88	0.00
80	Butylbenzylphthalate	40.000	45.692	-14.2	97	0.01
81	Benzo(a)anthracene	40.000	42.017	-5.0	86	0.01
82	3,3'-Dichlorobenzidine	40.000	45.502	-13.8	91	0.01
83	Chrysene	40.000	41.729	-4.3	86	0.00
84	Bis(2-ethylhexyl)phthalate	40.000	43.631	-9.1	94	0.01
85 c	Di-n-octyl phthalate	40.000	46.191	-15.5	93	0.02
86 I	Perylene-d12	20.000	20.000	0.0	86	0.02
87	Indeno(1,2,3-cd)pyrene	40.000	41.159	-2.9	86	0.02
88	Benzo(b)fluoranthene	40.000	41.516	-3.8	86	0.01
89	Benzo(k)fluoranthene	40.000	41.938	-4.8	87	0.02
90 C	Benzo(a)pyrene	40.000	41.507	-3.8	86	0.02
91	Dibenzo(a,h)anthracene	40.000	41.213	-3.0	86	0.02
92	Benzo(g,h,i)perylene	40.000	40.676	-1.7	85	0.03

Data Path : Z:\svoasrv\HPCHEM1\BNA_G\Data\BG042423\
Data File : BG057160.D
Acq On : 24 Apr 2023 20:52
Operator : CG/JU
Sample : SSTDICV040
Misc :
ALS Vial : 10 Sample Multiplier: 1

Instrument :
BNA_G
ClientSampleId :
ICVBG042423

Quant Time: Apr 25 05:00:07 2023
Quant Method : Z:\svoasrv\HPCHEM1\BNA_G\Methods\8270-BG042423.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Tue Apr 25 04:59:02 2023
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

Compound	Amount	Calc.	%Dev	Area%	Dev(min)
----------	--------	-------	------	-------	----------

(#) = Out of Range

SPCC's out = 0 CCC's out = 0