

Data Path : Z:\svoasrv\HPCHEM1\BNA_G\Data\BG042723\
 Data File : BG057206.D
 Acq On : 27 Apr 2023 22:55
 Operator : CG/JU
 Sample : SSTDICV040
 Misc :
 ALS Vial : 10 Sample Multiplier: 1

Instrument :
 BNA_G
 ClientSampleId :
 ICVBG042723

Quant Time: Apr 28 04:41:36 2023
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_G\Methods\8270-BG042723.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Fri Apr 28 04:39:54 2023
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	20.000	20.000	0.0	109	0.00
2	1,4-Dioxane	40.000	38.992	2.5	106	0.00
3	Pyridine	40.000	40.239	-0.6	105	0.00
4	n-Nitrosodimethylamine	40.000	38.938	2.7	101	0.00
5 S	2-Fluorophenol	80.000	79.747	0.3	106	0.00
6	Aniline	40.000	40.436	-1.1	107	0.00
7 S	Phenol-d6	80.000	81.368	-1.7	108	0.00
8	2-Chlorophenol	40.000	39.637	0.9	106	0.00
9	Benzaldehyde	40.000	40.460	-1.2	110	0.00
10 C	Phenol	40.000	39.877	0.3	104	0.00
11	bis(2-Chloroethyl)ether	40.000	39.256	1.9	105	0.00
12	1,3-Dichlorobenzene	40.000	39.396	1.5	106	0.00
13 C	1,4-Dichlorobenzene	40.000	39.684	0.8	105	0.00
14	1,2-Dichlorobenzene	40.000	39.622	0.9	108	0.00
15	Benzyl Alcohol	40.000	41.658	-4.1	106	0.00
16	2,2'-oxybis(1-Chloropropane	40.000	39.286	1.8	105	0.00
17	2-Methylphenol	40.000	40.329	-0.8	106	0.00
18	Hexachloroethane	40.000	39.780	0.5	108	0.00
19 P	n-Nitroso-di-n-propylamine	40.000	39.707	0.7	106	0.00
20	3+4-Methylphenols	40.000	40.502	-1.3	108	0.00
21 I	Naphthalene-d8	20.000	20.000	0.0	104	0.00
22	Acetophenone	40.000	40.342	-0.9	104	0.00
23 S	Nitrobenzene-d5	80.000	82.930	-3.7	105	0.00
24	Nitrobenzene	40.000	40.907	-2.3	104	0.00
25	Isophorone	40.000	41.128	-2.8	106	0.00
26 C	2-Nitrophenol	40.000	43.401	-8.5	110	0.00
27	2,4-Dimethylphenol	40.000	41.611	-4.0	104	0.00
28	bis(2-Chloroethoxy)methane	40.000	40.735	-1.8	105	0.00
29 C	2,4-Dichlorophenol	40.000	42.675	-6.7	109	0.00
30	1,2,4-Trichlorobenzene	40.000	40.672	-1.7	106	0.00
31	Naphthalene	40.000	41.327	-3.3	106	0.00
32	Benzoic acid	40.000	42.022	-5.1	109	0.00
33	4-Chloroaniline	40.000	41.608	-4.0	104	0.00
34 C	Hexachlorobutadiene	40.000	41.153	-2.9	108	0.00
35	Caprolactam	40.000	42.555	-6.4	106	0.00
36 C	4-Chloro-3-methylphenol	40.000	42.136	-5.3	108	0.00
37	2-Methylnaphthalene	40.000	40.866	-2.2	105	0.00
38	1-Methylnaphthalene	40.000	41.253	-3.1	107	0.00
39 I	Acenaphthene-d10	20.000	20.000	0.0	107	0.00
40	1,2,4,5-Tetrachlorobenzene	40.000	40.499	-1.2	105	0.00
41 P	Hexachlorocyclopentadiene	40.000	40.184	-0.5	107	0.00
42 S	2,4,6-Tribromophenol	80.000	82.793	-3.5	109	0.00
43 C	2,4,6-Trichlorophenol	40.000	41.373	-3.4	105	0.00
44	2,4,5-Trichlorophenol	40.000	41.819	-4.5	106	0.00
45 S	2-Fluorobiphenyl	80.000	80.090	-0.1	105	0.00
46	1,1'-Biphenyl	40.000	40.416	-1.0	105	0.00
47	2-Chloronaphthalene	40.000	40.545	-1.4	106	0.00

Data Path : Z:\svoasrv\HPCHEM1\BNA_G\Data\BG042723\
 Data File : BG057206.D
 Acq On : 27 Apr 2023 22:55
 Operator : CG/JU
 Sample : SSTDICV040
 Misc :
 ALS Vial : 10 Sample Multiplier: 1

Instrument :
 BNA_G
 ClientSampleId :
 ICVBG042723

Quant Time: Apr 28 04:41:36 2023
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_G\Methods\8270-BG042723.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Fri Apr 28 04:39:54 2023
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
48	2-Nitroaniline	40.000	42.702	-6.8	109	0.00
49	Acenaphthylene	40.000	40.525	-1.3	106	0.00
50	Dimethylphthalate	40.000	40.079	-0.2	108	0.00
51	2,6-Dinitrotoluene	40.000	42.722	-6.8	111	0.00
52 C	Acenaphthene	40.000	40.477	-1.2	106	0.00
53	3-Nitroaniline	40.000	40.468	-1.2	106	0.00
54 P	2,4-Dinitrophenol	40.000	41.159	-2.9	109	0.00
55	Dibenzofuran	40.000	40.142	-0.4	106	0.00
56 P	4-Nitrophenol	40.000	44.030	-10.1	107	0.00
57	2,4-Dinitrotoluene	40.000	41.528	-3.8	109	0.00
58	Fluorene	40.000	40.269	-0.7	106	0.00
59	2,3,4,6-Tetrachlorophenol	40.000	41.494	-3.7	106	0.00
60	Diethylphthalate	40.000	40.079	-0.2	106	0.00
61	4-Chlorophenyl-phenylether	40.000	40.515	-1.3	106	0.00
62	4-Nitroaniline	40.000	41.306	-3.3	106	0.00
63	Azobenzene	40.000	40.554	-1.4	107	0.00
64 I	Phenanthrene-d10	20.000	20.000	0.0	109	0.00
65	4,6-Dinitro-2-methylphenol	40.000	43.290	-8.2	107	0.00
66 c	n-Nitrosodiphenylamine	40.000	40.385	-1.0	106	0.00
67	4-Bromophenyl-phenylether	40.000	40.020	-0.1	106	0.00
68	Hexachlorobenzene	40.000	41.783	-4.5	111	0.00
69	Atrazine	40.000	41.087	-2.7	109	0.00
70 C	Pentachlorophenol	40.000	42.808	-7.0	107	0.00
71	Phenanthrene	40.000	40.316	-0.8	107	0.00
72	Anthracene	40.000	40.532	-1.3	107	0.00
73	Carbazole	40.000	40.440	-1.1	108	0.00
74	Di-n-butylphthalate	40.000	41.426	-3.6	108	0.00
75 C	Fluoranthene	40.000	40.217	-0.5	107	0.00
76 I	Chrysene-d12	20.000	20.000	0.0	105	0.00
77	Benzydine	40.000	35.257	11.9	105	0.00
78	Pyrene	40.000	41.253	-3.1	108	0.00
79 S	Terphenyl-d14	80.000	81.939	-2.4	108	0.00
80	Butylbenzylphthalate	40.000	42.016	-5.0	106	0.00
81	Benzo(a)anthracene	40.000	40.412	-1.0	103	0.00
82	3,3'-Dichlorobenzidine	40.000	40.919	-2.3	103	0.00
83	Chrysene	40.000	40.737	-1.8	105	0.00
84	Bis(2-ethylhexyl)phthalate	40.000	41.423	-3.6	104	0.00
85 c	Di-n-octyl phthalate	40.000	41.911	-4.8	106	0.00
86 I	Perylene-d12	20.000	20.000	0.0	106	0.00
87	Indeno(1,2,3-cd)pyrene	40.000	40.991	-2.5	106	-0.01
88	Benzo(b)fluoranthene	40.000	40.438	-1.1	104	0.00
89	Benzo(k)fluoranthene	40.000	40.492	-1.2	105	0.00
90 C	Benzo(a)pyrene	40.000	40.505	-1.3	105	0.00
91	Dibenzo(a,h)anthracene	40.000	40.390	-1.0	104	0.00
92	Benzo(g,h,i)perylene	40.000	40.710	-1.8	106	0.00

Data Path : Z:\svoasrv\HPCHEM1\BNA_G\Data\BG042723\
Data File : BG057206.D
Acq On : 27 Apr 2023 22:55
Operator : CG/JU
Sample : SSTDICV040
Misc :
ALS Vial : 10 Sample Multiplier: 1

Instrument :
BNA_G
ClientSampleId :
ICVBG042723

Quant Time: Apr 28 04:41:36 2023
Quant Method : Z:\svoasrv\HPCHEM1\BNA_G\Methods\8270-BG042723.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Fri Apr 28 04:39:54 2023
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

Compound	Amount	Calc.	%Dev	Area%	Dev(min)
----------	--------	-------	------	-------	----------

(#) = Out of Range

SPCC's out = 0 CCC's out = 0