


Data Path : Z:\svoasrv\HPCHEM1\BNA G\Data\BG042820\
 Data File : BG045183.D
 Acq On : 28 Apr 2020 13:17
 Operator : CG/JU
 Sample : DFTPP
 Misc :
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 BNA_G
 ClientSampleId :
 DFTPP

Integration File: rteint.p

Method : Z:\SVOASRV\HPCHEM1\BNA_G\METHODS\8270-BG041520.M
 Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 Last Update : Tue Apr 28 14:34:23 2020



AutoFind: Scans 2509, 2510, 2511; Background Corrected with Scan 2502

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
51	198	10	80	26.0	15532	PASS
68	69	0.00	2	0.0	0	PASS
69	198	0.00	100	34.8	20800	PASS
70	69	0.00	2	0.3	61	PASS
127	198	10	80	45.2	27013	PASS
197	198	0.00	2	0.7	425	PASS
198	198	100	100	100.0	59802	PASS
199	198	5	9	6.7	4004	PASS
275	198	10	60	30.9	18491	PASS
365	198	1	100	4.7	2837	PASS
441	198	0.01	100	13.5	8062	PASS
442	198	50	100	91.2	54560	PASS
443	442	15	24	20.6	11245	PASS