

Data Path : Z:\SVOASRV\HPCHEM1\BNA G\DATA\BG042920\
 Data File : TEST7.D
 Acq On : 30 Apr 2020 1:38
 Operator : CG/JU
 Sample : BS
 Misc :
 ALS Vial : 16 Sample Multiplier: 1

Instrument :
 BNA_G
 ClientSampleId :

Quant Time: Apr 30 02:38:22 2020
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA G\METHODS\8270-BG041520.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed Apr 29 12:24:27 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	7.99	152	64605	20.00	ng	0.00
21) Naphthalene-d8	10.80	136	251957	20.00	ng	0.00
39) Acenaphthene-d10	14.63	164	188895	20.00	ng	0.00
64) Phenanthrene-d10	17.38	188	476841	20.00	ng	0.00
76) Chrysene-d12	21.65	240	457553	20.00	ng	0.00
87) Perylene-d12	24.82	264	510556	20.00	ng	0.00

System Monitoring Compounds						
5) 2-Fluorophenol	5.56	112	552694	141.24	ng	0.00
7) Phenol-d6	7.16	99	748887	128.81	ng	0.00
23) Nitrobenzene-d5	9.15	82	479504	86.84	ng	0.00
42) 2,4,6-Tribromophenol	16.12	330	413607	141.73	ng	0.00
45) 2-Fluorobiphenyl	13.25	172	1153039	89.51	ng	0.00
79) Terphenyl-d14	19.99	244	2107341	100.38	ng	0.00

Target Compounds				Qvalue		
2) 1,4-Dioxane	3.44	88	76633	44.065	ng	97
3) Pyridine	3.84	79	172190	32.157	ng	97
4) n-Nitrosodimethylamine	3.75	42	75474	46.012	ng	98
6) Aniline	7.31	93	251548	33.276	ng	99
8) 2-Chlorophenol	7.55	128	206874	49.190	ng	99
9) Benzaldehyde	7.12	77	170084	58.711	ng	97
10) Phenol	7.19	94	268418	46.136	ng	98
11) bis(2-Chloroethyl)ether	7.41	93	188988	40.799	ng	97
12) 1,3-Dichlorobenzene	7.88	146	224509	45.302	ng	98
13) 1,4-Dichlorobenzene	8.02	146	220434	44.639	ng	97
14) 1,2-Dichlorobenzene	8.35	146	209279	44.299	ng	95
15) Benzyl Alcohol	8.23	79	206202	42.301	ng	92
16) 2,2'-oxybis(1-Chloropropan	8.52	45	173815	38.226	ng	94
17) 2-Methylphenol	8.44	107	188407	41.972	ng	97
18) Hexachloroethane	9.08	117	85105	45.844	ng	95
19) n-Nitroso-di-n-propylamine	8.80	70	164950	40.117	ng	95
20) 3+4-Methylphenols	8.76	107	254146	40.449	ng	98
22) Acetophenone	8.81	105	310364	45.551	ng	100
24) Nitrobenzene	9.19	77	256458	45.309	ng	96
25) Isophorone	9.71	82	477566	42.232	ng	98
26) 2-Nitrophenol	9.90	139	121792	49.954	ng	97
27) 2,4-Dimethylphenol	9.97	122	195464	50.526	ng	97
28) bis(2-Chloroethoxy)methane	10.20	93	259714	43.713	ng	99
29) 2,4-Dichlorophenol	10.45	162	219910	49.230	ng	98
30) 1,2,4-Trichlorobenzene	10.66	180	234612	46.389	ng	97
31) Naphthalene	10.85	128	638019	46.289	ng	99
32) Benzoic acid	10.11	122	131791	42.945	ng	99
33) 4-Chloroaniline	10.95	127	134865	20.981	ng	97
34) Hexachlorobutadiene	11.14	225	161817	46.666	ng	96
35) Caprolactam	11.72	113	46950	26.409	ng	97
36) 4-Chloro-3-methylphenol	12.08	107	253320	48.274	ng	99
37) 2-Methylnaphthalene	12.46	142	486606	45.484	ng	95
38) 1-Methylnaphthalene	12.68	142	462077	45.752	ng	97
40) 1,2,4,5-Tetrachlorobenzene	12.83	216	275878	45.360	ng	98


Data Path : Z:\SVOASRV\HPCHEM1\BNA G\DATA\BG042920\
 Data File : TEST7.D
 Acq On : 30 Apr 2020 1:38
 Operator : CG/JU
 Sample : BS
 Misc :
 ALS Vial : 16 Sample Multiplier: 1

Instrument :
 BNA_G
 ClientSampleId :

Quant Time: Apr 30 02:38:22 2020
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA G\METHODS\8270-BG041520.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed Apr 29 12:24:27 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
41) Hexachlorocyclopentadiene	12.81	237	380894	100.201	ng	96
43) 2,4,6-Trichlorophenol	13.06	196	204180	47.568	ng	96
44) 2,4,5-Trichlorophenol	13.14	196	222588	45.557	ng	98
46) 1,1'-Biphenyl	13.46	154	636899	44.361	ng	100
47) 2-Chloronaphthalene	13.51	162	488405	44.520	ng	99
48) 2-Nitroaniline	13.70	65	166466	46.119	ng	99
49) Acenaphthylene	14.36	152	826872	46.273	ng	99
50) Dimethylphthalate	14.09	163	694961	45.184	ng	98
51) 2,6-Dinitrotoluene	14.20	165	161437	46.926	ng	97
52) Acenaphthene	14.70	154	493274	40.799	ng	99
53) 3-Nitroaniline	14.53	138	100810	26.718	ng	95
54) 2,4-Dinitrophenol	14.74	184	209085	102.209	ng	94
55) Dibenzofuran	15.03	168	815507	46.265	ng	97
56) 4-Nitrophenol	14.85	139	278694	95.263	ng	96
57) 2,4-Dinitrotoluene	14.99	165	230259	45.800	ng	# 98
58) Fluorene	15.68	166	678960	47.157	ng	97
59) 2,3,4,6-Tetrachlorophenol	15.26	232	224713	51.689	ng	97
60) Diethylphthalate	15.45	149	737574	45.995	ng	99
61) 4-Chlorophenyl-phenylether	15.67	204	376099	45.383	ng	97
62) 4-Nitroaniline	15.69	138	168670	43.100	ng	94
63) Azobenzene	15.96	77	687955	46.532	ng	98
65) 4,6-Dinitro-2-methylphenol	15.76	198	153688	49.034	ng	97
66) n-Nitrosodiphenylamine	15.89	169	622468	45.434	ng	99
67) 4-Bromophenyl-phenylether	16.57	248	265511	43.161	ng	95
68) Hexachlorobenzene	16.69	284	295034	46.244	ng	99
69) Atrazine	16.83	200	248894	50.379	ng	96
70) Pentachlorophenol	17.03	266	376385	96.167	ng	99
71) Phenanthrene	17.42	178	1135226	46.329	ng	99
72) Anthracene	17.52	178	1174580	47.787	ng	99
73) Carbazole	17.78	167	1084896	43.494	ng	99
74) Di-n-butylphthalate	18.34	149	1305396	48.557	ng	99
75) Fluoranthene	19.43	202	1439735	50.782	ng	99
77) Benzidine	19.61	184	869611	68.763	ng	98
78) Pyrene	19.80	202	1411196	44.738	ng	99
80) Butylbenzylphthalate	20.68	149	587435	44.927	ng	98
81) Benzo(a)anthracene	21.63	228	1375297	46.375	ng	99
82) 3,3'-Dichlorobenzidine	21.54	252	372347	31.545	ng	98
83) Chrysene	21.69	228	1329182	46.648	ng	100
84) Bis(2-ethylhexyl)phthalate	21.54	149	824728	45.861	ng	100
85) Di-n-octyl phthalate	22.74	149	1407609	45.265	ng	100
86) Indeno(1,2,3-cd)pyrene	28.42	276	1772329	46.849	ng	99
88) Benzo(b)fluoranthene	23.82	252	1482329	46.350	ng	98
89) Benzo(k)fluoranthene	23.88	252	1454968	47.839	ng	98
90) Benzo(a)pyrene	24.67	252	1370544	45.390	ng	98
91) Dibenzo(a,h)anthracene	28.49	278	1432670	47.087	ng	97
92) Benzo(g,h,i)perylene	29.55	276	1446254	47.412	ng	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed







#1

1,4-Dichlorobenzene-d4
Concen: 20.000 ng
RT: 7.99 min Scan# 834
Delta R.T. -0.00 min
Lab File: TEST7.D
Acq: 30 Apr 2020 1:38

Instrument :
BNA_G
ClientSampleId :




Tot Ion	Ratio	Lower	Upper
152	100		
150	147.8	123.6	185.4
115	61.4	49.1	73.7




#2

1,4-Dioxane
Concen: 44.065 ng
RT: 3.44 min Scan# 60
Delta R.T. -0.01 min
Lab File: TEST7.D
Acq: 30 Apr 2020 1:38

Tgt Ion	Ratio	Lower	Upper
88	100		
58	60.4	46.6	69.8
43	22.8	17.5	26.3





#3

Pvridine

Concen: 32.157 ng

RT: 3.84 min Scan# 128

Delta R.T. -0.01 min

Lab File: TEST7.D

Acq: 30 Apr 2020 1:38

Instrument :

BNA_G

ClientSampleId :


Tot Ion: 79 Resp: 172190

Ion Ratio Lower Upper

79 100

52 47.3 39.8 59.6




51 26.5 21.4 32.2



Abundance Ion 79.00 (78.70 to 79.70): TES

Ion 52.00 (51.70 to 52.70): TES

Ion 51.00 (50.70 to 51.70): TES



#4

n-Nitrosodimethylamine

Concen: 46.012 ng

RT: 3.75 min Scan# 113

Delta R.T. -0.01 min

Lab File: TEST7.D

Acq: 30 Apr 2020 1:38


Tgt Ion: 42 Resp: 75474

Ion Ratio Lower Upper

42 100

74 181.8 143.0 214.6



44 5.4 4.6 6.8




Abundance Ion 42.00 (41.70 to 42.70): TES

Ion 74.00 (73.70 to 74.70): TES

Ion 44.00 (43.70 to 44.70): TES





#5

2-Fluorophenol

Concen: 141.238 ng

RT: 5.56 min Scan# 421

Delta R.T. 0.00 min

Lab File: TEST7.D


Acq: 30 Apr 2020 1:38

Instrument :


BNA_G

ClientSampled :



Tot Ion:	112	Resp:	552694
Ion	Ratio	Lower	Upper
112	100		
64	58.5	42.4	63.6
63	32.2	23.1	34.7



Abundance Ion 112.00 (111.70 to 112.70): T



Time-->



#6

Aniline

Concen: 33.276 ng


RT: 7.31 min Scan# 719

Delta R.T. -0.00 min


Lab File: TEST7.D

Acq: 30 Apr 2020 1:38


Tgt Ion:	93	Resp:	251548
Ion	Ratio	Lower	Upper
93	100		
66	37.7	29.8	44.8
65	18.6	14.6	22.0




Abundance Ion 93.00 (92.70 to 93.70): TES



Time-->





#7

Phenol-d6

Concen: 128.806 ng

RT: 7.16 min Scan# 693

Delta R.T. 0.00 min

Lab File: TEST7.D

Acq: 30 Apr 2020 1:38


Instrument :

BNA_G

ClientSampled :

Tot Ion: 99 Resp: 748887



Ion	Ratio	Lower	Upper
99	100		
42	14.2	11.3	16.9
71	38.0	28.3	42.5



Abundance Ion 99.00 (98.70 to 99.70): TES

Ion 42.00 (41.70 to 42.70): TES

Ion 71.00 (70.70 to 71.70): TES



#8

2-Chlorophenol

Concen: 49.190 ng

RT: 7.55 min Scan# 760


Delta R.T. -0.01 min

Lab File: TEST7.D

Acq: 30 Apr 2020 1:38

Tgt Ion:128 Resp: 206874


Ion	Ratio	Lower	Upper
128	100		
130	33.3	12.3	52.3
64	45.0	24.4	64.4




Abundance Ion 128.00 (127.70 to 128.70): T

Ion 130.00 (129.70 to 130.70): T

Ion 64.00 (63.70 to 64.70): TES





#9

Benzaldehyde

Concen: 58.711 ng

RT: 7.12 min Scan# 686

Delta R.T. -0.00 min

Lab File: TEST7.D

Acq: 30 Apr 2020 1:38

Instrument :

BNA_G

ClientSampleId :


Tot Ion: 77 Resp: 170084

Ion Ratio Lower Upper

77 100

105 93.7 72.9 112.9




106 93.1 67.6 107.6



Abundance Ion 77.00 (76.70 to 77.70): TES

Ion 105.00 (104.70 to 105.70): TES

Ion 106.00 (105.70 to 106.70): TES



#10

Phenol

Concen: 46.136 ng

RT: 7.19 min Scan# 698

Delta R.T. 0.00 min

Lab File: TEST7.D

Acq: 30 Apr 2020 1:38


Tgt Ion: 94 Resp: 268418

Ion Ratio Lower Upper

94 100

65 26.9 8.2 48.2



66 39.6 20.4 60.4




Abundance Ion 94.00 (93.70 to 94.70): TES

Ion 65.00 (64.70 to 65.70): TES

Ion 66.00 (65.70 to 66.70): TES





#11

bis(2-Chloroethyl)ether

Concen: 40.799 ng

RT: 7.41 min Scan# 735

Delta R.T. -0.00 min

Lab File: TEST7.D

Acq: 30 Apr 2020 1:38

Instrument :

BNA_G

ClientSampleId :


Tot Ion: 93 Resp: 188988

Ion Ratio Lower Upper

93 100

63 69.2 51.8 91.8




95 34.8 13.8 53.8



Abundance Ion 93.00 (92.70 to 93.70): TES

Ion 63.00 (62.70 to 63.70): TES

Ion 95.00 (94.70 to 95.70): TES



#12

1,3-Dichlorobenzene

Concen: 45.302 ng

RT: 7.88 min Scan# 816

Delta R.T. 0.00 min

Lab File: TEST7.D

Acq: 30 Apr 2020 1:38


Tgt Ion: 146 Resp: 224509

Ion Ratio Lower Upper

146 100

148 66.4 54.6 81.8



75 34.2 26.0 39.0




Abundance Ion 146.00 (145.70 to 146.70): T

Ion 148.00 (147.70 to 148.70): T

Ion 75.00 (74.70 to 75.70): TES





#13

1,4-Dichlorobenzene

Concen: 44.639 ng

RT: 8.02 min Scan# 840

Delta R.T. -0.00 min

Lab File: TEST7.D

Acq: 30 Apr 2020 1:38

Instrument :

BNA_G

ClientSampleId :


Tot Ion:146 Resp: 220434

Ion Ratio Lower Upper




146 100

148 69.4 53.3 79.9

111 43.0 33.8 50.8



Abundance Ion 146.00 (145.70 to 146.70): T
Ion 148.00 (147.70 to 148.70): T
Ion 111.00 (110.70 to 111.70): T



#14

1,2-Dichlorobenzene

Concen: 44.299 ng

RT: 8.35 min Scan# 895

Delta R.T. -0.00 min

Lab File: TEST7.D

Acq: 30 Apr 2020 1:38


Tgt Ion:146 Resp: 209279

Ion Ratio Lower Upper



146 100


148 67.4 51.9 77.9

111 44.7 31.3 46.9



Abundance Ion 146.00 (145.70 to 146.70): T
Ion 148.00 (147.70 to 148.70): T
Ion 111.00 (110.70 to 111.70): T





#15

Benzyl Alcohol

Concen: 42.301 ng

RT: 8.23 min Scan# 875

Delta R.T. 0.00 min

Lab File: TEST7.D

Acq: 30 Apr 2020 1:38

Instrument :

BNA_G

ClientSampled :





Tot Ion: 79 Resp: 206202

Ion Ratio Lower Upper

79 100

108 74.5 52.7 79.1

77 67.4 51.3 76.9



#16

2,2'-oxybis(1-chloropropane)

Concen: 38.226 ng

RT: 8.52 min Scan# 924

Delta R.T. -0.00 min

Lab File: TEST7.D

Acq: 30 Apr 2020 1:38




Tgt Ion: 45 Resp: 173815


Ion Ratio Lower Upper

45 100

77 24.7 1.9 41.9

79 18.2 0.0 35.5





#17

2-Methylphenol

Concen: 41.972 ng

RT: 8.44 min Scan# 911

Delta R.T. 0.00 min

Lab File: TEST7.D

Acq: 30 Apr 2020 1:38

Instrument :

BNA_G

ClientSampleId :

Tot Ion:107 Resp: 188407


Ion Ratio Lower Upper

107 100

108 111.4 93.8 140.6

77 49.7 38.5 57.7

79 50.2 40.1 60.1






Abundance Ion 107.00 (106.70 to 107.70): T

Ion 108.00 (107.70 to 108.70): T

Ion 77.00 (76.70 to 77.70): TES

Ion 79.00 (78.70 to 79.70): TES



#18

Hexachloroethane

Concen: 45.844 ng

RT: 9.08 min Scan# 1020

Delta R.T. -0.00 min

Lab File: TEST7.D

Acq: 30 Apr 2020 1:38


Tgt Ion:117 Resp: 85105

Ion Ratio Lower Upper

117 100

119 102.5 78.3 117.5



201 122.2 92.9 139.3






Abundance Ion 117.00 (116.70 to 117.70): T

Ion 119.00 (118.70 to 119.70): T

Ion 201.00 (200.70 to 201.70): T









#19
n-Nitroso-di-n-propylamine
Concen: 40.117 ng
RT: 8.80 min Scan# 972
Delta R.T. -0.00 min
Lab File: TEST7.D
Acq: 30 Apr 2020 1:38


Instrument :
BNA_G
ClientSampleId :


Tot Ion:	70	Resp:	164950
Ion	Ratio	Lower	Upper
70	100		
42	45.8	33.3	49.9
101	11.6	7.8	11.8
130	19.3	15.8	23.6



#20
3+4-Methylphenols
Concen: 40.449 ng
RT: 8.76 min Scan# 966
Delta R.T. -0.00 min
Lab File: TEST7.D
Acq: 30 Apr 2020 1:38

Tgt Ion:	107	Resp:	254146
Ion	Ratio	Lower	Upper
107	100		
108	89.9	69.9	109.9
77	38.4	15.2	55.2
79	32.4	11.6	51.6









#21
Naphthalene-d8
Concen: 20.000 ng
RT: 10.80 min Scan# 1312
Delta R.T. -0.00 min
Lab File: TEST7.D
Acq: 30 Apr 2020 1:38

Instrument :
BNA_G
ClientSampleId :

Tot Ion:136	Resp:	251957
Ion Ratio	Lower	Upper
136 100		
137 10.7	8.6	13.0
54 6.0	4.8	7.2
68 5.3	5.2	7.8




Abundance Ion 136.00 (135.70 to 136.70): T
Ion 137.00 (136.70 to 137.70): T
Ion 54.00 (53.70 to 54.70): TES
Ion 68.00 (67.70 to 68.70): TES






#22
Acetophenone
Concen: 45.551 ng
RT: 8.81 min Scan# 974
Delta R.T. -0.00 min
Lab File: TEST7.D
Acq: 30 Apr 2020 1:38

Tgt	Ion:105	Resp:	310364
Ion	Ratio	Lower	Upper
105	100		
71	4.5	3.3	4.9
51	21.7	17.3	25.9
120	23.3	18.4	27.6



Abundance Ion 105.00 (104.70 to 105.70): T
Ion 71.00 (70.70 to 71.70): TES
Ion 51.00 (50.70 to 51.70): TES
Ion 120.00 (119.70 to 120.70): TES









#23
 Nitrobenzene-d5
 Concen: 86.837 ng
 RT: 9.15 min Scan# 1032
 Delta R.T. -0.00 min
 Lab File: TEST7.D
 Acq: 30 Apr 2020 1:38

Instrument :
 BNA_G
 ClientSampleId :

Tot Ion	82	Resp	479504
Ion Ratio	Lower	Upper	
82	100		
128	40.2	33.0	49.4
54	39.2	30.4	45.6




Abundance Ion 82.00 (81.70 to 82.70): TES
 Ion 128.00 (127.70 to 128.70): TES
 Ion 54.00 (53.70 to 54.70): TES






#24
 Nitrobenzene
 Concen: 45.309 ng
 RT: 9.19 min Scan# 1039
 Delta R.T. -0.00 min
 Lab File: TEST7.D
 Acq: 30 Apr 2020 1:38

Tgt Ion	77	Resp	256458
Ion Ratio	Lower	Upper	
77	100		
123	40.6	34.3	51.5
65	13.3	8.9	13.3



Abundance Ion 77.00 (76.70 to 77.70): TES
 Ion 123.00 (122.70 to 123.70): TES
 Ion 65.00 (64.70 to 65.70): TES





#25

Isophorone

Concen: 42.232 ng

RT: 9.71 min Scan# 1128

Delta R.T. -0.00 min

Lab File: TEST7.D

Acq: 30 Apr 2020 1:38

Instrument :

BNA_G

ClientSampleId :


Tot Ion: 82 Resp: 477566

Ion Ratio Lower Upper

82 100

95 7.5 6.0 9.0




138 17.3 12.7 19.1



Abundance Ion 82.00 (81.70 to 82.70): TES

Ion 95.00 (94.70 to 95.70): TES

Ion 138.00 (137.70 to 138.70): TES



#26

2-Nitrophenol

Concen: 49.954 ng

RT: 9.90 min Scan# 1160

Delta R.T. -0.00 min

Lab File: TEST7.D

Acq: 30 Apr 2020 1:38


Tgt Ion: 139 Resp: 121792

Ion Ratio Lower Upper

139 100

109 37.2 30.6 45.8



65 55.0 41.8 62.6




Abundance Ion 139.00 (138.70 to 139.70): TES

Ion 109.00 (108.70 to 109.70): TES

Ion 65.00 (64.70 to 65.70): TES









#27
2,4-Dimethylphenol
Concen: 50.526 ng
RT: 9.97 min Scan# 1172
Delta R.T. -0.00 min
Lab File: TEST7.D
Acq: 30 Apr 2020 1:38




Instrument :
BNA_G
ClientSampleId :


Tot Ion	Ratio	Lower	Upper
122	100		
107	125.4	96.8	145.2
121	58.3	45.6	68.4



#28
bis(2-Chloroethoxy)methane
Concen: 43.713 ng
RT: 10.20 min Scan# 1211
Delta R.T. -0.00 min
Lab File: TEST7.D
Acq: 30 Apr 2020 1:38

Tgt Ion	Ratio	Lower	Upper
93	100		
95	33.9	27.0	40.4
123	11.9	8.7	13.1









#29
 2,4-Dichlorophenol
 Concen: 49.230 ng
 RT: 10.45 min Scan# 1253
 Delta R.T. 0.00 min
 Lab File: TEST7.D
 Acq: 30 Apr 2020 1:38




Instrument :
 BNA_G
 ClientSampleId :


Tot Ion	Ratio	Lower	Upper
162	100		
164	67.8	45.0	85.0
98	35.1	15.2	55.2



#30
 1,2,4-Trichlorobenzene
 Concen: 46.389 ng
 RT: 10.66 min Scan# 1289
 Delta R.T. -0.00 min
 Lab File: TEST7.D
 Acq: 30 Apr 2020 1:38

Tgt Ion	Ratio	Lower	Upper
180	100		
182	95.9	73.3	109.9
145	28.1	22.6	33.8









#31
Naphthalene
Concen: 46.289 ng
RT: 10.85 min Scan# 1321
Delta R.T. -0.00 min
Lab File: TEST7.D
Acq: 30 Apr 2020 1:38




Instrument :
BNA_G
ClientSampleId :


Tot Ion	Ratio	Lower	Upper
128	100		
129	11.1	9.4	14.2
127	13.4	10.8	16.2



#32
Benzoic acid
Concen: 42.945 ng
RT: 10.11 min Scan# 1196
Delta R.T. -0.02 min
Lab File: TEST7.D
Acq: 30 Apr 2020 1:38

Tgt Ion	Ratio	Lower	Upper
122	100		
105	127.0	108.4	148.4
77	98.3	78.8	118.8









#33
4-Chloroaniline
Concen: 20.981 ng
RT: 10.95 min Scan# 1339
Delta R.T. 0.00 min
Lab File: TEST7.D
Acq: 30 Apr 2020 1:38

Instrument :
BNA_G
ClientSampleId :

Tot Ion:127	Resp:	134865	
Ion Ratio	Lower	Upper	
127 100			
129 33.0	25.7	38.5	
65 29.0	25.4	38.2	
92 19.7	16.1	24.1	




Abundance Ion 127.00 (126.70 to 127.70): T
Ion 129.00 (128.70 to 129.70): T
Ion 65.00 (64.70 to 65.70): TES
Ion 92.00 (91.70 to 92.70): TES






#34
Hexachlorobutadiene
Concen: 46.666 ng
RT: 11.14 min Scan# 1371
Delta R.T. -0.00 min
Lab File: TEST7.D
Acq: 30 Apr 2020 1:38

Tgt	Ion:225	Resp:	161817
Ion	Ratio	Lower	Upper
225	100		
223	66.5	50.7	76.1
227	66.2	50.2	75.2



Abundance Ion 225.00 (224.70 to 225.70): T
Ion 223.00 (222.70 to 223.70): T
Ion 227.00 (226.70 to 227.70): T









#35
 Caprolactam
 Concen: 26.409 ng
 RT: 11.72 min Scan# 1470
 Delta R.T. -0.00 min
 Lab File: TEST7.D
 Acq: 30 Apr 2020 1:38




Instrument :
 BNA_G
 ClientSampleId :


Tot Ion:113 Resp: 46950
 Ion Ratio Lower Upper
 113 100
 55 133.2 118.8 158.8
 56 102.5 82.6 122.6



#36
 4-Chloro-3-methylphenol
 Concen: 48.274 ng
 RT: 12.08 min Scan# 1531
 Delta R.T. -0.00 min
 Lab File: TEST7.D
 Acq: 30 Apr 2020 1:38

Tgt Ion:107 Resp: 253320
 Ion Ratio Lower Upper
 107 100
 144 25.8 19.8 29.6
 142 75.4 59.7 89.5





#37

2-Methylnaphthalene

Concen: 45.484 ng

RT: 12.46 min Scan# 1595

Delta R.T. -0.00 min

Lab File: TEST7.D

Acq: 30 Apr 2020 1:38

Instrument :

BNA_G

ClientSampleId :


Tot Ion:142 Resp: 486606

Ion Ratio Lower Upper




142 100

141 91.9 70.3 105.5

115 41.2 29.8 44.6



Abundance Ion 142.00 (141.70 to 142.70): T
Ion 141.00 (140.70 to 141.70): T
Ion 115.00 (114.70 to 115.70): T



#38

1-Methylnaphthalene

Concen: 45.752 ng

RT: 12.68 min Scan# 1632

Delta R.T. -0.00 min

Lab File: TEST7.D

Acq: 30 Apr 2020 1:38


Tgt Ion:142 Resp: 462077

Ion Ratio Lower Upper



142 100


141 94.0 72.7 109.1

116 4.7 3.1 4.7



Abundance Ion 142.00 (141.70 to 142.70): T
Ion 141.00 (140.70 to 141.70): T
Ion 116.00 (115.70 to 116.70): T





#39

Acenaphthene-d10

Concen: 20.000 ng

RT: 14.63 min Scan# 1965

Delta R.T. -0.00 min

Lab File: TEST7.D

Acq: 30 Apr 2020 1:38

Instrument :

BNA_G

ClientSampleId :


Tot Ion:164 Resp: 188895

Ion Ratio Lower Upper

164 100

162 99.5 78.7 118.1

160 42.9 32.8 49.2






Abundance

Ion 164.00 (163.70 to 164.70): T

Ion 162.00 (161.70 to 162.70): T

Ion 160.00 (159.70 to 160.70): T



#40

1,2,4,5-Tetrachlorobenzene

Concen: 45.360 ng

RT: 12.83 min Scan# 1658

Delta R.T. -0.00 min

Lab File: TEST7.D

Acq: 30 Apr 2020 1:38

Tgt Ion:216 Resp: 275878


Ion Ratio Lower Upper

216 100

214 80.6 63.0 94.6

179 20.2 15.8 23.8

108 19.1 14.2 21.2





Abundance


Ion 216.00 (215.70 to 216.70): T

Ion 214.00 (213.70 to 214.70): T

Ion 179.00 (178.70 to 179.70): T

Ion 108.00 (107.70 to 108.70): T





#41

Hexachlorocyclopentadiene

Concen: 100.201 ng

RT: 12.81 min Scan# 1655

Delta R.T. -0.00 min

Lab File: TEST7.D

Acq: 30 Apr 2020 1:38

Instrument :

BNA_G

ClientSampleId :


Tot Ion:237 Resp: 380894

Ion Ratio Lower Upper




237 100

235 64.7 41.6 81.6

272 13.1 0.0 32.1



Abundance Ion 236.80 (236.50 to 237.50): T
Ion 234.90 (234.60 to 235.60): T
Ion 271.75 (271.45 to 272.45): T



#42

2,4,6-Tribromophenol

Concen: 141.735 ng

RT: 16.12 min Scan# 2219

Delta R.T. -0.00 min

Lab File: TEST7.D

Acq: 30 Apr 2020 1:38


Tgt Ion:330 Resp: 413607

Ion Ratio Lower Upper



330 100


332 98.0 79.1 118.7

141 34.6 27.1 40.7



Abundance Ion 329.80 (329.50 to 330.50): T
Ion 331.80 (331.50 to 332.50): T
Ion 141.00 (140.70 to 141.70): T









#43
 2,4,6-Trichlorophenol
 Concen: 47.568 ng
 RT: 13.06 min Scan# 1698
 Delta R.T. -0.00 min
 Lab File: TEST7.D
 Acq: 30 Apr 2020 1:38




Instrument :
 BNA_G
 ClientSampleId :


Tot Ion	Ratio	Lower	Upper
196	100		
198	94.7	72.5	108.7
200	29.2	24.5	36.7



#44
 2,4,5-Trichlorophenol
 Concen: 45.557 ng
 RT: 13.14 min Scan# 1711
 Delta R.T. -0.00 min
 Lab File: TEST7.D
 Acq: 30 Apr 2020 1:38

Tgt Ion	Ratio	Lower	Upper
196	100		
198	96.7	75.7	113.5
97	48.0	39.6	59.4
132	24.0	19.6	29.4









#45
 2-Fluorobiphenyl
 Concen: 89.506 ng
 RT: 13.25 min Scan# 1730
 Delta R.T. -0.00 min
 Lab File: TEST7.D
 Acq: 30 Apr 2020 1:38

Instrument :
 BNA_G
 ClientSampleId :

Tot Ion:172 Resp: 1153039




Ion	Ratio	Lower	Upper
172	100		
171	37.7	30.6	46.0
170	23.7	20.5	30.7




#46
 1,1'-Biphenyl
 Concen: 44.361 ng
 RT: 13.46 min Scan# 1766
 Delta R.T. -0.00 min
 Lab File: TEST7.D
 Acq: 30 Apr 2020 1:38

Tgt Ion:154 Resp: 636899

Ion	Ratio	Lower	Upper
154	100		
153	41.8	21.9	61.9
76	12.4	0.0	33.1









#47
 2-Chloronaphthalene
 Concen: 44.520 ng
 RT: 13.51 min Scan# 1774
 Delta R.T. 0.00 min
 Lab File: TEST7.D
 Acq: 30 Apr 2020 1:38




Instrument :
 BNA_G
 ClientSampleId :


Tot Ion	Ratio	Lower	Upper
162	100		
127	39.6	31.3	46.9
164	35.1	27.8	41.8



#48
 2-Nitroaniline
 Concen: 46.119 ng
 RT: 13.70 min Scan# 1807
 Delta R.T. -0.00 min
 Lab File: TEST7.D
 Acq: 30 Apr 2020 1:38

Tgt Ion	Ratio	Lower	Upper
65	100		
92	71.5	58.3	87.5
138	105.8	85.6	128.4





#49

Acenaphthylene

Concen: 46.273 ng

RT: 14.36 min Scan# 1918

Delta R.T. -0.00 min

Lab File: TEST7.D

Acq: 30 Apr 2020 1:38

Instrument :

BNA_G

ClientSampleId :


Tot Ion:152 Resp: 826872

Ion Ratio Lower Upper




152 100

151 21.2 17.2 25.8

153 14.2 11.0 16.4



Abundance Ion 152.00 (151.70 to 152.70): T
Ion 151.00 (150.70 to 151.70): T
Ion 153.00 (152.70 to 153.70): T



#50

Dimethylphthalate

Concen: 45.184 ng

RT: 14.09 min Scan# 1872

Delta R.T. -0.00 min

Lab File: TEST7.D

Acq: 30 Apr 2020 1:38


Tgt Ion:163 Resp: 694961

Ion Ratio Lower Upper



163 100


194 4.6 3.5 5.3

164 11.1 8.2 12.4



Abundance Ion 163.00 (162.70 to 163.70): T
Ion 194.00 (193.70 to 194.70): T
Ion 164.00 (163.70 to 164.70): T









#51
 2,6-Dinitrotoluene
 Concen: 46.926 ng
 RT: 14.20 min Scan# 1892
 Delta R.T. 0.00 min
 Lab File: TEST7.D
 Acq: 30 Apr 2020 1:38




Instrument :
 BNA_G
 ClientSampleId :


Tot Ion	Ratio	Lower	Upper
165	100		
63	49.1	40.4	60.6
89	58.7	49.1	73.7



#52
 Acenaphthene
 Concen: 40.799 ng
 RT: 14.70 min Scan# 1976
 Delta R.T. -0.00 min
 Lab File: TEST7.D
 Acq: 30 Apr 2020 1:38

Tgt Ion	Ratio	Lower	Upper
154	100		
153	111.4	89.7	134.5
152	53.8	43.1	64.7









#53
 3-Nitroaniline
 Concen: 26.718 ng
 RT: 14.53 min Scan# 1947
 Delta R.T. -0.01 min
 Lab File: TEST7.D
 Acq: 30 Apr 2020 1:38

Instrument :
 BNA_G
 ClientSampleId :

Tot Ion:138 Resp: 100810
 Ion Ratio Lower Upper
 138 100
 108 10.9 7.8 11.8
 92 123.3 94.5 141.7




Abundance Ion 138.00 (137.70 to 138.70): T
 Ion 108.00 (107.70 to 108.70): T
 Ion 92.00 (91.70 to 92.70): TES






#54
 2,4-Dinitrophenol
 Concen: 102.209 ng
 RT: 14.74 min Scan# 1983
 Delta R.T. -0.00 min
 Lab File: TEST7.D
 Acq: 30 Apr 2020 1:38

Tgt Ion:184 Resp: 209085
 Ion Ratio Lower Upper
 184 100
 63 58.1 41.5 62.3
 154 68.4 52.3 78.5



Abundance Ion 184.00 (183.70 to 184.70): T
 Ion 63.00 (62.70 to 63.70): TES
 Ion 154.00 (153.70 to 154.70): T





#55

Dibenzenofuran

Concen: 46.265 ng

RT: 15.03 min Scan# 2033

Delta R.T. -0.00 min

Lab File: TEST7.D

Acq: 30 Apr 2020 1:38

Instrument :

BNA_G

ClientSampleId :





Tot Ion:168 Resp: 815507

Ion	Ratio	Lower	Upper
168	100		
139	40.1	30.4	45.6
169	13.8	11.0	16.4

168 100

139 40.1 30.4 45.6

169 13.8 11.0 16.4



#56

4-Nitrophenol

Concen: 95.263 ng

RT: 14.85 min Scan# 2002

Delta R.T. 0.00 min

Lab File: TEST7.D

Acq: 30 Apr 2020 1:38




Tgt Ion:139 Resp: 278694


Ion	Ratio	Lower	Upper
139	100		
109	99.5	73.6	113.6
65	102.5	81.2	121.2

139 100

109 99.5 73.6 113.6

65 102.5 81.2 121.2






#57
 2,4-Dinitrotoluene
 Concen: 45.800 ng
 RT: 14.99 min Scan# 2026
 Delta R.T. -0.00 min
 Lab File: TEST7.D
 Acq: 30 Apr 2020 1:38

Instrument :
 BNA_G
 ClientSampled :

Tot Ion	Ratio	Lower	Upper
165	100		
63	42.4	32.2	48.2
89	76.0	61.7	92.5
182	3.1	2.0	3.0






Abundance Ion 165.00 (164.70 to 165.70): T

Ion 63.00 (62.70 to 63.70): TES


Ion 89.00 (88.70 to 89.70): TES

Ion 182.00 (181.70 to 182.70): T



#58
 Fluorene
 Concen: 47.157 ng
 RT: 15.68 min Scan# 2144
 Delta R.T. 0.00 min
 Lab File: TEST7.D
 Acq: 30 Apr 2020 1:38



Tgt Ion	Ratio	Lower	Upper
166	100		
165	102.7	79.9	119.9
167	14.0	11.4	17.0




Abundance Ion 166.00 (165.70 to 166.70): T

Ion 165.00 (164.70 to 165.70): T

Ion 167.00 (166.70 to 167.70): T





#59

2,3,4,6-Tetrachlorophenol

Concen: 51.689 ng

RT: 15.26 min Scan# 2072

Delta R.T. -0.00 min

Lab File: TEST7.D

Acq: 30 Apr 2020 1:38

Instrument :

BNA_G

ClientSampleId :

Tot Ion:232 Resp: 224713


Ion Ratio Lower Upper

232 100

131 38.9 32.2 48.2

130 2.5 2.2 3.2

166 27.6 24.2 36.4






Abundance Ion 232.00 (231.70 to 232.70): T

Ion 131.00 (130.70 to 131.70): T

Ion 130.00 (129.70 to 130.70): T

Ion 166.00 (165.70 to 166.70): T



#60

Diethylphthalate

Concen: 45.995 ng

RT: 15.45 min Scan# 2105

Delta R.T. 0.00 min

Lab File: TEST7.D

Acq: 30 Apr 2020 1:38


Tgt Ion:149 Resp: 737574

Ion Ratio Lower Upper

149 100

177 23.1 18.9 28.3



150 13.0 10.6 15.8




Abundance Ion 149.00 (148.70 to 149.70): T

Ion 177.00 (176.70 to 177.70): T

Ion 150.00 (149.70 to 150.70): T









#61
 4-Chlorophenyl-phenylether
 Concen: 45.383 ng
 RT: 15.67 min Scan# 2142
 Delta R.T. -0.00 min
 Lab File: TEST7.D
 Acq: 30 Apr 2020 1:38




Instrument :
 BNA_G
 ClientSampleId :


Tot Ion	Ratio	Lower	Upper
204	100		
206	34.6	26.9	40.3
141	60.5	46.3	69.5



#62
 4-Nitroaniline
 Concen: 43.100 ng
 RT: 15.69 min Scan# 2146
 Delta R.T. -0.00 min
 Lab File: TEST7.D
 Acq: 30 Apr 2020 1:38

Tgt Ion	Ratio	Lower	Upper
138	100		
92	55.7	33.6	73.6
108	106.3	78.2	118.2





#63

Azobenzene

Concen: 46.532 ng

RT: 15.96 min Scan# 2192

Delta R.T. -0.00 min

Lab File: TEST7.D

Acq: 30 Apr 2020 1:38

Instrument :

BNA_G

ClientSampleId :

Tot Ion: 77 Resp: 687955


Ion Ratio Lower Upper

77 100

182 27.1 8.4 48.4

105 19.7 0.5 40.5

51 23.8 4.4 44.4






Abundance Ion 77.00 (76.70 to 77.70): TES

Ion 182.00 (181.70 to 182.70): TES

Ion 105.00 (104.70 to 105.70): TES

Ion 51.00 (50.70 to 51.70): TES



#64

Phenanthrene-d10

Concen: 20.000 ng

RT: 17.38 min Scan# 2433

Delta R.T. -0.00 min

Lab File: TEST7.D

Acq: 30 Apr 2020 1:38


Tgt Ion: 188 Resp: 476841

Ion Ratio Lower Upper

188 100

94 6.4 6.2 9.4



80 7.9 7.0 10.6




Abundance Ion 188.00 (187.70 to 188.70): TES

Ion 94.00 (93.70 to 94.70): TES

Ion 80.00 (79.70 to 80.70): TES









#65
 4,6-Dinitro-2-methylphenol
 Concen: 49.034 ng
 RT: 15.76 min Scan# 2157
 Delta R.T. 0.00 min
 Lab File: TEST7.D
 Acq: 30 Apr 2020 1:38




Instrument :
 BNA_G
 ClientSampleId :


Tot Ion	Ratio	Lower	Upper
198	100		
51	28.9	10.0	50.0
105	36.9	18.9	58.9



#66
 n-Nitrosodiphenylamine
 Concen: 45.434 ng
 RT: 15.89 min Scan# 2179
 Delta R.T. -0.00 min
 Lab File: TEST7.D
 Acq: 30 Apr 2020 1:38

Tgt Ion	Ratio	Lower	Upper
169	100		
168	69.5	55.8	83.8
167	37.6	30.8	46.2









#67
4-Bromophenyl-phenylether
Concen: 43.161 ng
RT: 16.57 min Scan# 2295
Delta R.T. 0.00 min
Lab File: TEST7.D
Acq: 30 Apr 2020 1:38




Instrument :
BNA_G
ClientSampled :


Tot Ion	Ratio	Lower	Upper
248	100		
250	101.9	75.4	113.2
141	69.6	54.4	81.6



#68
Hexachlorobenzene
Concen: 46.244 ng
RT: 16.69 min Scan# 2316
Delta R.T. -0.00 min
Lab File: TEST7.D
Acq: 30 Apr 2020 1:38

Tgt Ion	Ratio	Lower	Upper
284	100		
142	31.7	25.8	38.6
249	33.3	26.6	39.8





#69

Atrazine

Concen: 50.379 ng

RT: 16.83 min Scan# 2340

Delta R.T. -0.00 min

Lab File: TEST7.D

Acq: 30 Apr 2020 1:38

Instrument :

BNA_G

ClientSampled :


Tot Ion:200 Resp: 248894

Ion Ratio Lower Upper




200 100

173 23.7 4.5 44.5

215 46.0 30.0 70.0



Abundance Ion 200.00 (199.70 to 200.70): T
Ion 173.00 (172.70 to 173.70): T
Ion 215.00 (214.70 to 215.70): T



#70

Pentachlorophenol

Concen: 96.167 ng

RT: 17.03 min Scan# 2374

Delta R.T. -0.00 min

Lab File: TEST7.D

Acq: 30 Apr 2020 1:38


Tgt Ion:266 Resp: 376385

Ion Ratio Lower Upper



266 100


268 64.5 51.5 77.3

264 63.3 52.2 78.4



Abundance Ion 265.70 (265.40 to 266.40): T
Ion 268.00 (267.70 to 268.70): T
Ion 264.00 (263.70 to 264.70): T





#71

Phenanthrene

Concen: 46.329 ng

RT: 17.42 min Scan# 2440

Delta R.T. -0.00 min

Lab File: TEST7.D

Acq: 30 Apr 2020 1:38

Instrument :

BNA_G

ClientSampleId :


Tot Ion:178 Resp: 1135226

Ion Ratio Lower Upper

178 100

176 21.3 16.8 25.2




179 16.8 13.2 19.8



Abundance Ion 178.00 (177.70 to 178.70): T

Ion 176.00 (175.70 to 176.70): T

Ion 179.00 (178.70 to 179.70): T



#72

Anthracene

Concen: 47.787 ng

RT: 17.52 min Scan# 2456

Delta R.T. 0.00 min

Lab File: TEST7.D

Acq: 30 Apr 2020 1:38


Tgt Ion:178 Resp: 1174580

Ion Ratio Lower Upper

178 100

176 20.7 16.0 24.0



179 16.8 13.6 20.4




Abundance Ion 178.00 (177.70 to 178.70): T

Ion 176.00 (175.70 to 176.70): T

Ion 179.00 (178.70 to 179.70): T









#73
 Carbazole
 Concen: 43.494 ng
 RT: 17.78 min Scan# 2501
 Delta R.T. -0.00 min
 Lab File: TEST7.D
 Acq: 30 Apr 2020 1:38

Instrument :
 BNA_G
 ClientSampled :

Tot Ion:167 Resp: 1084896




Ion	Ratio	Lower	Upper
167	100		
166	23.8	19.8	29.6
139	13.3	10.9	16.3




#74
 Di-n-butylphthalate
 Concen: 48.557 ng
 RT: 18.34 min Scan# 2597
 Delta R.T. -0.00 min
 Lab File: TEST7.D
 Acq: 30 Apr 2020 1:38

Tgt Ion:149 Resp: 1305396

Ion	Ratio	Lower	Upper
149	100		
150	10.4	8.2	12.2
104	6.3	4.8	7.2





#75

Fluoranthene

Concen: 50.782 ng

RT: 19.43 min Scan# 2782

Delta R.T. -0.00 min

Lab File: TEST7.D

Acq: 30 Apr 2020 1:38

Instrument :

BNA_G

ClientSampleId :


Tot Ion:202 Resp: 1439735

Ion	Ratio	Lower	Upper
202	100		
101	8.9	0.0	29.6
203	19.9	0.0	39.7

202 100

101 8.9 0.0 29.6




203 19.9 0.0 39.7



Abundance Ion 202.00 (201.70 to 202.70): T

Ion 101.00 (100.70 to 101.70): T

Ion 203.00 (202.70 to 203.70): T



#76

Chrysene-d12

Concen: 20.000 ng

RT: 21.65 min Scan# 3159

Delta R.T. -0.00 min

Lab File: TEST7.D

Acq: 30 Apr 2020 1:38


Tgt Ion:240 Resp: 457553

Ion	Ratio	Lower	Upper
240	100		
120	7.6	6.6	10.0
236	26.1	22.2	33.2

240 100

120 7.6 6.6 10.0



236 26.1 22.2 33.2




Abundance Ion 240.00 (239.70 to 240.70): T

Ion 120.00 (119.70 to 120.70): T

Ion 236.00 (235.70 to 236.70): T





#77

Benzidine

Concen: 68.763 ng

RT: 19.61 min Scan# 2812

Delta R.T. -0.00 min

Lab File: TEST7.D

Acq: 30 Apr 2020 1:38


Instrument :

BNA_G

ClientSampleId :

Tot Ion:184 Resp: 869611




Ion	Ratio	Lower	Upper
184	100		
185	15.3	12.6	19.0
183	13.3	11.3	16.9



Abundance Ion 184.00 (183.70 to 184.70): T

Ion 185.00 (184.70 to 185.70): T

Ion 183.00 (182.70 to 183.70): T



#78

Pyrene

Concen: 44.738 ng

RT: 19.80 min Scan# 2844


Delta R.T. 0.00 min

Lab File: TEST7.D

Acq: 30 Apr 2020 1:38

Tgt Ion:202 Resp: 1411196



Ion	Ratio	Lower	Upper
202	100		
200	24.3	19.0	28.6
203	19.3	16.1	24.1




Abundance Ion 202.00 (201.70 to 202.70): T

Ion 200.00 (199.70 to 200.70): T

Ion 203.00 (202.70 to 203.70): T





#79

Terphenyl-d14

Concen: 100.384 ng

RT: 19.99 min Scan# 2878

Delta R.T. -0.00 min

Lab File: TEST7.D

Acq: 30 Apr 2020 1:38

Instrument :

BNA_G

ClientSampleId :


Tot Ion:244 Resp: 2107341

Ion Ratio Lower Upper

244 100

212 9.8 7.9 11.9




122 8.8 7.7 11.5



Abundance Ion 244.00 (243.70 to 244.70): T

Ion 212.00 (211.70 to 212.70): T

Ion 122.00 (121.70 to 122.70): T



#80

Butylbenzylphthalate

Concen: 44.927 ng

RT: 20.68 min Scan# 2995

Delta R.T. -0.00 min

Lab File: TEST7.D

Acq: 30 Apr 2020 1:38


Tgt Ion:149 Resp: 587435

Ion Ratio Lower Upper

149 100

91 71.4 58.5 87.7



206 24.2 19.2 28.8




Abundance Ion 149.00 (148.70 to 149.70): T

Ion 91.00 (90.70 to 91.70): TES

Ion 206.00 (205.70 to 206.70): T





#81

Benzo(a)anthracene

Concen: 46.375 ng

RT: 21.63 min Scan# 3156

Delta R.T. 0.00 min

Lab File: TEST7.D

Acq: 30 Apr 2020 1:38


Instrument :

BNA_G

ClientSampled :

Tot Ion:228 Resp: 1375297

Ion	Ratio	Lower	Upper
228	100		
226	29.1	24.0	36.0
229	21.9	17.5	26.3






Abundance

Ion 228.00 (227.70 to 228.70): T

Ion 226.00 (225.70 to 226.70): T

Ion 229.00 (228.70 to 229.70): T



#82

3,3'-Dichlorobenzidine

Concen: 31.545 ng

RT: 21.54 min Scan# 3141


Delta R.T. -0.00 min

Lab File: TEST7.D

Acq: 30 Apr 2020 1:38

Tgt Ion:252 Resp: 372347

Ion	Ratio	Lower	Upper
252	100		
254	64.7	52.6	79.0
126	8.3	7.5	11.3






Abundance

Ion 252.00 (251.70 to 252.70): T

Ion 254.00 (253.70 to 254.70): T

Ion 126.00 (125.70 to 126.70): T





#83

Chrysene

Concen: 46.648 ng

RT: 21.69 min Scan# 3167

Delta R.T. -0.00 min

Lab File: TEST7.D

Acq: 30 Apr 2020 1:38





Instrument :

BNA_G

ClientSampled :

Tot Ion:228 Resp: 1329182

Ion	Ratio	Lower	Upper
228	100		
226	32.8	26.2	39.4
229	21.1	17.0	25.6



#84

Bis(2-ethylhexyl)phthalate

Concen: 45.861 ng

RT: 21.54 min Scan# 3141




Delta R.T. -0.00 min


Lab File: TEST7.D

Acq: 30 Apr 2020 1:38

Tgt Ion:149 Resp: 824728

Ion	Ratio	Lower	Upper
149	100		
167	29.0	23.4	35.0
279	5.6	4.3	6.5









#85
Di-n-octyl phthalate
Concen: 45.265 ng
RT: 22.74 min Scan# 3346
Delta R.T. -0.00 min
Lab File: TEST7.D
Acq: 30 Apr 2020 1:38

Instrument :
BNA_G
ClientSampled :

Tot Ion:149 Resp: 1407609
Ion Ratio Lower Upper
149 100
167 1.7 1.4 2.2
43 7.2 5.7 8.5




Abundance Ion 149.00 (148.70 to 149.70): T
Ion 167.00 (166.70 to 167.70): T
Ion 43.00 (42.70 to 43.70): TES






#86
Indeno(1,2,3-cd)pyrene
Concen: 46.849 ng
RT: 28.42 min Scan# 4313
Delta R.T. -0.01 min
Lab File: TEST7.D
Acq: 30 Apr 2020 1:38

Tgt Ion:276 Resp: 1772329
Ion Ratio Lower Upper
276 100
138 10.7 9.6 14.4
277 25.8 20.6 31.0



Abundance Ion 276.00 (275.70 to 276.70): T
Ion 138.00 (137.70 to 138.70): T
Ion 277.00 (276.70 to 277.70): T









#87
Pervlene-d12
Concen: 20.000 ng
RT: 24.82 min Scan# 3699
Delta R.T. -0.00 min
Lab File: TEST7.D
Acq: 30 Apr 2020 1:38

Instrument :
BNA_G
ClientSampleId :

Tot Ion:264 Resp: 510556




Ion	Ratio	Lower	Upper
264	100		
260	22.8	19.2	28.8
265	21.2	19.0	28.4




#88
Benzo(b)fluoranthene
Concen: 46.350 ng
RT: 23.82 min Scan# 3529
Delta R.T. 0.00 min
Lab File: TEST7.D
Acq: 30 Apr 2020 1:38

Tgt Ion:252 Resp: 1482329

Ion	Ratio	Lower	Upper
252	100		
253	22.8	18.8	28.2
125	7.1	6.2	9.4









#89
Benzo(k)fluoranthene
Concen: 47.839 ng
RT: 23.88 min Scan# 3540
Delta R.T. -0.00 min
Lab File: TEST7.D
Acq: 30 Apr 2020 1:38




Instrument :
BNA_G
ClientSampleId :


Tot Ion:252 Resp: 1454968
Ion Ratio Lower Upper
252 100
253 22.9 19.2 28.8
125 6.8 5.9 8.9



#90
Benzo(a)pyrene
Concen: 45.390 ng
RT: 24.67 min Scan# 3674
Delta R.T. -0.00 min
Lab File: TEST7.D
Acq: 30 Apr 2020 1:38

Tgt Ion:252 Resp: 1370544
Ion Ratio Lower Upper
252 100
253 21.9 18.3 27.5
125 7.7 7.0 10.6





#91

Dibenzo(a,h)anthracene

Concen: 47.087 ng

RT: 28.49 min Scan# 4324

Delta R.T. 0.00 min

Lab File: TEST7.D

Acq: 30 Apr 2020 1:38

Instrument :

BNA_G

ClientSampleId :


Tot Ion:278 Resp: 1432670

Ion Ratio Lower Upper

278 100

139 10.5 9.0 13.6




279 25.0 18.4 27.6



Abundance Ion 278.00 (277.70 to 278.70): T

Ion 139.00 (138.70 to 139.70): T

Ion 279.00 (278.70 to 279.70): T



#92

Benzo(g,h,i)perylene

Concen: 47.412 ng

RT: 29.55 min Scan# 4505

Delta R.T. -0.00 min

Lab File: TEST7.D

Acq: 30 Apr 2020 1:38


Tgt Ion:276 Resp: 1446254

Ion Ratio Lower Upper

276 100

277 24.4 19.3 28.9

138 13.5 11.4 17.2



Abundance Ion 276.00 (275.70 to 276.70): T

Ion 277.00 (276.70 to 277.70): T

Ion 138.00 (137.70 to 138.70): T

