


Data Path : Z:\svoasrv\HPCHEM1\BNA_G\Data\BG050123\
Data File : BG057245.D
Acq On : 1 May 2023 9:18
Operator : CG/JU
Sample : DFTPP503
Misc :
ALS Vial : 1 Sample Multiplier: 1

Instrument :
BNA_G
ClientSampleId :
DFTPP503

Integration File: LSCINT.P

Method : Z:\svoasrv\HPCHEM1\BNA_G\Methods\SFAM-EPA-BG042823.MA.M
Title : SVOA CALIBRATION
Last Update : Mon May 01 12:45:38 2023



AutoFind: Scans 2482, 2483, 2484; Background Corrected with Scan 2476

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
51	198	10	80	31.7	14201	PASS
68	69	0.00	2	0.6	89	PASS
69	198	0.00	100	35.4	15864	PASS
70	69	0.00	2	0.0	0	PASS
127	198	10	80	40.8	18291	PASS
197	198	0.00	2	0.3	123	PASS
198	198	100	100	100.0	44827	PASS
199	198	5	9	7.0	3133	PASS
275	198	10	60	30.9	13845	PASS
365	198	1	100	5.5	2448	PASS
441	198	0.01	100	14.9	6668	PASS
442	198	50	100	93.2	41795	PASS
443	442	15	24	19.3	8070	PASS

m/z	Abundance
44.15	1009.0
206.95	253.0

Instrument :

BNA_G

ClientSampleId :

DFTPP503

m/z	Abundance
38.05	206.0
39.15	1271.0
44.15	1043.0
49.05	236.0
50.15	4432.0
51.15	17352.0
52.15	851.0
56.05	392.0
57.05	1247.0
60.95	280.0
62.05	288.0
63.05	674.0
64.05	152.0
65.05	424.0
69.05	19360.0
74.15	2170.0
75.05	3430.0
76.25	1197.0
77.15	21216.0
78.15	1433.0
79.15	1751.0
80.05	1068.0
81.15	1814.0
82.05	401.0
83.05	369.0
85.05	241.0
86.05	585.0
87.05	240.0
88.15	169.0
91.05	441.0
91.95	469.0
93.05	2972.0
96.25	175.0
98.05	2300.0
99.05	1298.0
101.05	1155.0
103.05	285.0
104.05	622.0
105.15	578.0
106.05	204.0
107.05	7153.0
108.05	1288.0
110.05	11011.0
111.05	2072.0
111.95	308.0
116.15	508.0
117.15	5887.0
118.05	375.0
122.15	647.0
122.95	821.0
124.05	179.0
124.95	247.0
127.05	20704.0
128.15	1897.0
129.05	9257.0
130.05	767.0
133.95	291.0
135.15	809.0
136.05	300.0
137.15	507.0
140.95	1462.0
142.05	472.0
143.15	351.0
145.95	320.0
147.15	633.0
148.05	1537.0
149.15	334.0
151.05	208.0
153.15	394.0
154.05	419.0
155.05	810.0
156.05	1162.0
157.15	192.0
158.05	262.0
160.05	434.0
161.15	811.0
161.95	189.0
164.05	161.0
165.05	584.0
165.95	505.0
167.05	2409.0
168.15	1031.0
169.05	198.0
172.05	286.0
173.05	418.0
174.15	577.0
175.05	875.0
176.15	207.0
177.05	424.0
178.05	158.0
178.95	2203.0
180.05	1572.0

Instrument :
BNA_G
ClientSampleId :
DFTPP503

181.05	512.0
184.25	206.0
185.15	1172.0
186.05	8005.0
187.15	2290.0
188.05	301.0
188.95	740.0
191.05	483.0
191.95	720.0
193.05	661.0
193.95	161.0
195.95	1252.0
198.05	49824.0
198.95	3610.0
200.05	428.0
201.45	183.0
203.05	641.0
204.05	2259.0
205.05	3658.0
206.05	14434.0
207.05	2589.0
208.05	628.0
209.05	208.0
210.05	268.0
210.25	268.0
211.05	625.0
217.05	4499.0
218.05	496.0
221.05	2168.0
221.95	745.0
223.05	955.0
224.05	8985.0
225.05	2128.0
227.05	4243.0
228.05	621.0
229.05	962.0
230.05	161.0
231.05	208.0
233.95	222.0
235.05	371.0
235.95	173.0
237.05	222.0
238.95	284.0
241.05	224.0
242.05	397.0
243.25	605.0
244.05	6369.0
245.05	669.0
246.15	1624.0
247.05	289.0
249.05	263.0
252.85	180.0
255.05	36664.0
256.05	5348.0
257.05	460.0
258.05	2247.0
259.05	409.0
265.05	952.0
270.95	190.0
272.05	173.0
272.85	919.0
274.05	2781.0
275.05	13854.0
276.05	1977.0
277.05	1396.0
277.85	242.0
283.15	193.0
285.15	280.0
292.95	382.0
296.05	4634.0
297.05	757.0
303.15	577.0
304.05	172.0
313.95	169.0
314.95	589.0
316.15	227.0
323.05	1333.0
324.05	374.0
327.05	342.0
328.15	191.0
332.95	191.0
333.95	1027.0
335.15	237.0
345.95	177.0
352.05	348.0
352.95	202.0
354.05	342.0
355.05	180.0
364.95	2143.0
366.05	368.0
371.95	486.0
373.05	226.0
382.95	199.0
401.85	302.0
402.95	295.0
421.05	322.0

Instrument :
BNA_G
ClientSampleId :
DFTPP503

421.95	346.0
422.95	1830.0
424.15	376.0
441.05	5677.0
442.05	36408.0
443.05	6800.0
444.05	694.0

Instrument :
BNA_G
ClientSampleId :
DFTPP503

m/z	Abundance
38.15	235.0
39.05	1324.0
44.05	1024.0
49.15	163.0
50.15	4246.0
51.15	16736.0
52.05	844.0
56.05	505.0
57.05	986.0
61.05	268.0
62.05	287.0
63.15	745.0
65.05	439.0
69.05	17712.0
73.25	214.0
74.15	2164.0
75.05	2804.0
76.05	910.0
77.15	20752.0
78.15	1242.0
79.05	1559.0
80.05	1404.0
81.15	1683.0
82.05	396.0
83.05	326.0
84.95	235.0
86.05	512.0
87.05	296.0
91.15	305.0
92.15	278.0
93.05	2801.0
98.05	2253.0
99.05	1468.0
101.05	792.0
103.05	244.0
104.05	560.0
105.05	535.0
106.05	166.0
107.05	6991.0
108.05	959.0
108.95	214.0
110.05	10873.0
111.05	1576.0
112.15	237.0
117.05	6207.0
118.05	438.0
122.15	665.0
123.15	694.0
124.05	371.0
125.05	334.0
127.05	21288.0
128.05	1984.0
129.05	9079.0
130.05	776.0
134.15	315.0
135.05	609.0
136.05	469.0
136.95	309.0
139.95	188.0
141.05	1355.0
142.05	397.0
143.05	199.0
146.15	171.0
147.05	760.0
148.05	1655.0
149.05	426.0
151.05	292.0
151.75	154.0
153.05	446.0
154.05	243.0
155.05	832.0
156.05	1126.0
157.15	190.0
157.95	249.0
158.85	245.0
160.15	460.0
161.05	623.0
162.05	332.0
165.05	571.0
166.15	401.0
167.05	2666.0
168.05	1230.0
169.15	273.0
170.95	169.0
172.05	308.0
173.05	324.0
174.05	549.0
175.05	978.0
176.05	319.0
177.15	771.0
178.15	190.0
179.05	2223.0

Instrument :
BNA_G
ClientSampleId :
DFTPP503

180.05	1499.0
181.05	777.0
183.95	162.0
185.05	890.0
186.15	8086.0
187.15	2656.0
188.05	253.0
189.15	527.0
191.05	234.0
192.05	619.0
193.05	911.0
196.05	1665.0
198.05	52024.0
198.95	3696.0
200.05	295.0
201.55	308.0
203.05	547.0
204.15	2567.0
205.05	3940.0
206.05	15789.0
207.05	2527.0
208.15	643.0
209.15	215.0
210.15	372.0
211.05	764.0
215.95	410.0
217.05	4764.0
217.95	627.0
221.05	2534.0
221.75	633.0
223.05	915.0
224.05	9337.0
225.05	2422.0
226.05	318.0
227.05	4035.0
228.05	665.0
229.05	910.0
230.15	153.0
231.15	527.0
234.05	223.0
235.05	259.0
236.05	166.0
237.05	324.0
238.85	161.0
240.05	188.0
241.15	245.0
242.15	560.0
243.15	619.0
244.05	7428.0
245.05	897.0
246.05	1650.0
246.95	253.0
249.15	283.0
251.15	171.0
252.95	168.0
255.05	39936.0
256.05	5514.0
257.05	339.0
258.05	2490.0
259.05	376.0
263.95	152.0
265.05	958.0
265.85	164.0
271.95	231.0
273.05	1415.0
274.05	3126.0
275.05	16600.0
276.05	2289.0
277.05	1464.0
278.05	223.0
283.05	306.0
285.15	272.0
293.05	297.0
295.05	187.0
296.05	5584.0
297.05	778.0
300.95	172.0
303.05	596.0
304.15	169.0
313.95	257.0
315.05	775.0
315.95	447.0
320.85	180.0
323.05	1635.0
324.05	398.0
326.95	401.0
332.85	258.0
334.05	1107.0
334.85	344.0
341.15	219.0
345.95	399.0
352.05	565.0
353.15	381.0
353.95	488.0
364.95	3021.0
365.95	376.0

Instrument :
BNA_G
ClientSampleId :
DFTPP503

372.05	712.0
373.05	306.0
383.05	245.0
402.15	305.0
403.05	475.0
403.95	247.0
421.15	339.0
423.05	2460.0
424.05	407.0
441.05	7864.0
442.05	50776.0
443.05	9814.0
443.95	801.0

Instrument :
BNA_G
ClientSampleId :
DFTPP503

m/z	Abundance
39.15	617.0
44.15	1010.0
50.15	2142.0
51.15	8515.0
52.05	380.0
56.05	289.0
56.95	603.0
63.05	402.0
65.05	210.0
68.15	266.0
69.05	10519.0
73.15	213.0
74.05	1024.0
75.05	1703.0
76.05	580.0
77.15	11644.0
78.05	690.0
79.05	885.0
80.05	530.0
81.05	840.0
82.05	336.0
83.25	165.0
85.05	192.0
86.05	191.0
91.05	229.0
92.15	217.0
93.05	1604.0
98.05	1149.0
98.95	690.0
101.05	559.0
103.95	328.0
105.05	316.0
107.05	4050.0
108.05	495.0
109.05	195.0
110.05	6427.0
111.05	1243.0
117.05	3683.0
117.95	204.0
121.95	345.0
123.15	441.0
127.05	12881.0
128.05	853.0
129.05	5029.0
129.95	437.0
133.95	186.0
135.05	529.0
137.05	239.0
141.05	784.0
142.95	180.0
147.05	601.0
148.05	928.0
148.95	178.0
152.95	342.0
154.15	236.0
154.95	461.0
156.05	707.0
159.95	294.0
161.25	369.0
165.05	422.0
166.05	333.0
167.05	1639.0
168.15	663.0
171.95	210.0
173.05	297.0
174.05	381.0
175.05	689.0
176.15	198.0
177.05	547.0
179.05	1255.0
180.05	996.0
181.05	374.0
185.15	696.0
186.05	5022.0
187.05	1339.0
188.95	342.0
191.25	164.0
192.05	568.0
193.15	421.0
196.05	931.0
196.85	370.0
198.05	32632.0
198.95	2093.0
199.85	194.0
201.45	307.0
203.05	376.0
204.05	1402.0
205.05	2254.0
206.05	9626.0
207.05	1497.0
208.15	264.0
211.05	368.0

Instrument :
BNA_G
ClientSampleId :
DFTPP503

215.95	185.0
216.95	3433.0
217.95	409.0
220.95	1069.0
222.05	312.0
222.95	549.0
224.05	6131.0
225.05	1337.0
227.05	2895.0
228.05	362.0
229.05	564.0
230.95	193.0
234.05	239.0
239.95	173.0
240.95	173.0
242.05	281.0
244.05	4556.0
245.05	411.0
246.05	1095.0
249.05	194.0
255.05	27616.0
256.05	3621.0
256.95	152.0
258.05	1797.0
259.15	259.0
264.95	774.0
270.85	261.0
273.05	970.0
273.95	1941.0
275.05	11082.0
276.05	1732.0
277.05	1077.0
277.95	194.0
284.95	184.0
293.05	258.0
296.05	4000.0
296.95	564.0
302.05	151.0
303.15	473.0
304.15	179.0
314.15	263.0
314.95	610.0
316.15	166.0
321.15	197.0
323.05	1333.0
323.95	266.0
327.05	272.0
328.05	180.0
334.05	616.0
335.15	200.0
345.95	237.0
352.05	567.0
352.95	229.0
354.05	363.0
365.05	2179.0
365.95	317.0
372.05	553.0
373.15	187.0
382.95	220.0
402.15	187.0
402.95	335.0
421.05	314.0
422.05	247.0
423.05	1941.0
424.05	523.0
441.05	6462.0
442.05	38200.0
443.05	7596.0
443.85	630.0

Instrument :
BNA_G
ClientSampleId :
DFTPP503