

Data Path : Z:\SVOASRV\HPCHEM1\BNA G\DATA\BG050420\
 Data File : BG045240.D
 Acq On : 4 May 2020 19:13
 Operator : CG/JU
 Sample : L2476-01
 Misc :
 ALS Vial : 11 Sample Multiplier: 1

Instrument :
 BNA_G
 ClientSampleId :
 FK-GF-03-008-A

Quant Time: May 05 01:27:05 2020
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA G\METHODS\8270-BG041520.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Mon May 04 13:03:14 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	7.99	152	101948	20.00	ng	0.00
21) Naphthalene-d8	10.80	136	404638	20.00	ng	0.00
39) Acenaphthene-d10	14.63	164	282148	20.00	ng	0.00
64) Phenanthrene-d10	17.38	188	584307	20.00	ng	0.00
76) Chrysene-d12	21.65	240	515885	20.00	ng	0.00
87) Perylene-d12	24.82	264	552839	20.00	ng	0.00

System Monitoring Compounds	R.T.	QIon	Response	Conc	Units	Dev(Min)
5) 2-Fluorophenol	5.56	112	408468	66.15	ng	0.00
7) Phenol-d6	7.15	99	630665	68.74	ng	0.00
23) Nitrobenzene-d5	9.15	82	417477	47.08	ng	0.00
42) 2,4,6-Tribromophenol	16.12	330	287063	65.86	ng	0.00
45) 2-Fluorobiphenyl	13.25	172	982031	51.04	ng	0.00
79) Terphenyl-d14	19.99	244	1364696	57.66	ng	0.00


Target Compounds	R.T.	QIon	Response	Conc	Units	Qvalue
9) Benzaldehyde	7.12	77	78	Below Cal	#	4
19) n-Nitroso-di-n-propylamine	9.15	70	55651	8.577 ng	#	81
32) Benzoic acid	10.02	122	68	6.386 ng	#	1
50) Dimethylphthalate	14.08	163	77759	3.385 ng	#	95
51) 2,6-Dinitrotoluene	14.63	165	37041	7.208 ng	#	25
57) 2,4-Dinitrotoluene	14.63	165	37041	4.933 ng	#	23
67) 4-Bromophenyl-phenylether	16.12	248	19996	2.653 ng	#	1
69) Atrazine	16.82	200	54	Below Cal	#	1
71) Phenanthrene	17.42	178	196962	6.560 ng		99
72) Anthracene	17.42	178	196962	6.539 ng		98
74) Di-n-butylphthalate	18.34	149	5969	Below Cal	#	89
75) Fluoranthene	19.43	202	348242	4.285 ng		96
77) Benzidine	19.70	184	971	Below Cal	#	1
78) Pyrene	19.79	202	313261	8.808 ng		98
81) Benzo(a)anthracene	21.62	228	165026	4.935 ng		97
83) Chrysene	21.69	228	154314	4.803 ng		95
84) Bis(2-ethylhexyl)phthalate	21.54	149	55658	2.745 ng		94
86) Indeno(1,2,3-cd)pyrene	28.43	276	108089	2.534 ng	#	95
88) Benzo(b)fluoranthene	23.82	252	212982	6.150 ng		97
89) Benzo(k)fluoranthene	23.82	252	212982	6.467 ng		98
90) Benzo(a)pyrene	24.67	252	157287	4.811 ng		98
92) Benzo(g,h,i)perylene	29.55	276	106879	3.236 ng		95


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\SVOASRV\HPCHEM1\BNA G\DATA\BG050420\
Data File : BG045240.D
Acq On : 4 May 2020 19:13
Operator : CG/JU
Sample : L2476-01
Misc :
ALS Vial : 11 Sample Multiplier: 1

Instrument :
BNA_G
ClientSampleId :
FK-GF-03-008-A

Quant Time: May 05 01:27:05 2020
Quant Method : Z:\SVOASRV\HPCHEM1\BNA G\METHODS\8270-BG041520.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Mon May 04 13:03:14 2020
Response via : Initial Calibration






#1



1,4-Dichlorobenzene-d4
Concen: 20.000 ng
RT: 7.99 min Scan# 834
Delta R.T. 0.00 min
Lab File: BG045240.D
Acq: 4 May 2020 19:13

Instrument :
BNA_G
ClientSampleId :
FK-GF-03-008-A

Tot Ion:152 Resp: 101948
Ion Ratio Lower Upper
152 100
150 164.2 123.6 185.4
115 57.8 49.1 73.7




Abundance Ion 152.00 (151.70 to 152.70): E
Ion 150.00 (149.70 to 150.70): E
Ion 115.00 (114.70 to 115.70): E





#5




2-Fluorophenol
Concen: 66.147 ng
RT: 5.56 min Scan# 421
Delta R.T. 0.00 min
Lab File: BG045240.D
Acq: 4 May 2020 19:13

Tgt Ion:112 Resp: 408468
Ion Ratio Lower Upper
112 100
64 54.9 42.4 63.6
63 33.0 23.1 34.7



Abundance Ion 112.00 (111.70 to 112.70): E
Ion 64.00 (63.70 to 64.70): BGC
Ion 63.00 (62.70 to 63.70): BGC





#7

Phenol-d6

Concen: 68.740 ng

RT: 7.15 min Scan# 692

Delta R.T. 0.00 min

Lab File: BG045240.D

Acq: 4 May 2020 19:13

Instrument :

BNA_G

ClientSampleId :

FK-GF-03-008-A





Tot Ion: 99 Resp: 630665

Ion Ratio Lower Upper

99 100

42 13.6 11.3 16.9

71 37.2 28.3 42.5



#19

n-Nitroso-di-n-propylamine

Concen: 8.577 ng

RT: 9.15 min Scan# 1032

Delta R.T. 0.35 min

Lab File: BG045240.D

Acq: 4 May 2020 19:13

Tgt Ion: 70 Resp: 55651


Ion Ratio Lower Upper


70 100

42 36.9 33.3 49.9

101 0.0 7.8 11.8#

130 2.3 15.8 23.6#









#21
Naphthalene-d8
Concen: 20.000 ng
RT: 10.80 min Scan# 1313
Delta R.T. 0.00 min
Lab File: BG045240.D
Acq: 4 May 2020 19:13

Instrument :
BNA_G
ClientSampleId :
FK-GF-03-008-A

Tot Ion:136	Resp:	404638
Ion Ratio	Lower	Upper
136	100	
137	10.3	8.6 13.0
54	6.7	4.8 7.2
68	5.0	5.2 7.8#




Abundance Ion 136.00 (135.70 to 136.70): E
Ion 137.00 (136.70 to 137.70): E
Ion 54.00 (53.70 to 54.70): BGC
Ion 68.00 (67.70 to 68.70): BGC






#23
Nitrobenzene-d5
Concen: 47.077 ng
RT: 9.15 min Scan# 1032
Delta R.T. 0.00 min
Lab File: BG045240.D
Acq: 4 May 2020 19:13

Tgt Ion: 82	Resp:	417477
Ion Ratio	Lower	Upper
82	100	
128	42.0	33.0 49.4
54	38.5	30.4 45.6



Abundance Ion 82.00 (81.70 to 82.70): BGC
Ion 128.00 (127.70 to 128.70): E
Ion 54.00 (53.70 to 54.70): BGC









#32
Benzoic acid
Concen: 6.386 ng
RT: 10.02 min Scan# 1180
Delta R.T. -0.09 min
Lab File: BG045240.D
Acq: 4 May 2020 19:13




Instrument :
BNA_G
ClientSampleId :
FK-GF-03-008-A


Tot Ion:122 Resp: 68
Ion Ratio Lower Upper
122 100
105 0.0 108.4 148.4#
77 0.0 78.8 118.8#



#39
Acenaphthene-d10
Concen: 20.000 ng
RT: 14.63 min Scan# 1965
Delta R.T. 0.00 min
Lab File: BG045240.D
Acq: 4 May 2020 19:13

Tgt Ion:164 Resp: 282148
Ion Ratio Lower Upper
164 100
162 98.9 78.7 118.1
160 42.7 32.8 49.2









#42
 2,4,6-Tribromophenol
 Concen: 65.858 ng
 RT: 16.12 min Scan# 2219
 Delta R.T. 0.00 min
 Lab File: BG045240.D
 Acq: 4 May 2020 19:13




Instrument :
 BNA_G
 ClientSampleId :
 FK-GF-03-008-A


Tot Ion	Ratio	Lower	Upper
330	100		
332	96.7	79.1	118.7
141	36.1	27.1	40.7



#45
 2-Fluorobiphenyl
 Concen: 51.036 ng
 RT: 13.25 min Scan# 1730
 Delta R.T. -0.00 min
 Lab File: BG045240.D
 Acq: 4 May 2020 19:13

Tgt Ion	Ratio	Lower	Upper
172	100		
171	36.5	30.6	46.0
170	24.1	20.5	30.7









#50
Dimethylphthalate
Concen: 3.385 ng
RT: 14.08 min Scan# 1871
Delta R.T. -0.00 min
Lab File: BG045240.D
Acq: 4 May 2020 19:13




Instrument :
BNA_G
ClientSampleId :
FK-GF-03-008-A


Tot Ion:163 Resp: 77759
Ion Ratio Lower Upper
163 100
194 4.2 3.5 5.3
164 13.0 8.2 12.4#



#51
2,6-Dinitrotoluene
Concen: 7.208 ng
RT: 14.63 min Scan# 1965
Delta R.T. 0.43 min
Lab File: BG045240.D
Acq: 4 May 2020 19:13

Tgt Ion:165 Resp: 37041
Ion Ratio Lower Upper
165 100
63 1.5 40.4 60.6#
89 2.0 49.1 73.7#









#57
 2,4-Dinitrotoluene
 Concen: 4.933 ng
 RT: 14.63 min Scan# 1965
 Delta R.T. -0.36 min
 Lab File: BG045240.D
 Acq: 4 May 2020 19:13

Instrument :
 BNA_G
 ClientSampleId :
 FK-GF-03-008-A

Tot Ion:165 Resp: 37041
 Ion Ratio Lower Upper
 165 100
 63 1.5 32.2 48.2#
 89 2.0 61.7 92.5#
 182 0.0 2.0 3.0#




Abundance Ion 165.00 (164.70 to 165.70): E






#64
 Phenanthrene-d10
 Concen: 20.000 ng
 RT: 17.38 min Scan# 2433
 Delta R.T. 0.00 min
 Lab File: BG045240.D
 Acq: 4 May 2020 19:13

Tgt Ion:188 Resp: 584307
 Ion Ratio Lower Upper
 188 100
 94 7.2 6.2 9.4
 80 7.1 7.0 10.6



Abundance Ion 188.00 (187.70 to 188.70): E









#67
 4-Bromophenyl-phenylether
 Concen: 2.653 ng
 RT: 16.12 min Scan# 2219
 Delta R.T. -0.45 min
 Lab File: BG045240.D
 Acq: 4 May 2020 19:13

Instrument :
 BNA_G
 ClientSampleId :
 FK-GF-03-008-A

Tot Ion:248 Resp: 19996
 Ion Ratio Lower Upper
 248 100
 250 167.8 75.4 113.2#
 141 430.6 54.4 81.6#




Abundance Ion 248.00 (247.70 to 248.70): E
 Ion 250.00 (249.70 to 250.70): E
 Ion 141.00 (140.70 to 141.70): E






#69
 Atrazine
 Concen: Below Cal
 RT: 16.82 min Scan# 2337
 Delta R.T. -0.02 min
 Lab File: BG045240.D
 Acq: 4 May 2020 19:13

Tgt Ion:200 Resp: 54
 Ion Ratio Lower Upper
 200 100
 173 0.0 4.5 44.5#
 215 149.3 30.0 70.0#



Abundance Ion 200.00 (199.70 to 200.70): E
 Ion 173.00 (172.70 to 173.70): E
 Ion 215.00 (214.70 to 215.70): E





#71

Phenanthrene

Concen: 6.560 ng

RT: 17.42 min Scan# 2440

Delta R.T. 0.00 min

Lab File: BG045240.D

Acq: 4 May 2020 19:13

Instrument :

BNA_G

ClientSampleId :

FK-GF-03-008-A





Tot Ion:178 Resp: 196962

Ion Ratio Lower Upper

178 100

176 20.4 16.8 25.2

179 15.8 13.2 19.8



#72

Anthracene

Concen: 6.539 ng

RT: 17.42 min Scan# 2440

Delta R.T. -0.09 min

Lab File: BG045240.D

Acq: 4 May 2020 19:13




Tgt Ion:178 Resp: 196962


Ion Ratio Lower Upper

178 100

176 20.4 16.0 24.0

179 15.8 13.6 20.4









#74
Di-n-butylphthalate
Concen: Below Cal
RT: 18.34 min Scan# 2597
Delta R.T. 0.00 min
Lab File: BG045240.D
Acq: 4 May 2020 19:13




Instrument :
BNA_G
ClientSampleId :
FK-GF-03-008-A


Tot Ion	Ratio	Lower	Upper
149	100		
150	12.8	8.2	12.2#
104	12.1	4.8	7.2#



#75
Fluoranthene
Concen: 4.285 ng
RT: 19.43 min Scan# 2782
Delta R.T. -0.00 min
Lab File: BG045240.D
Acq: 4 May 2020 19:13

Tgt Ion	Ratio	Lower	Upper
202	100		
101	7.9	0.0	29.6
203	18.2	0.0	39.7





#76

Chrysene-d12

Concen: 20.000 ng

RT: 21.65 min Scan# 3159

Delta R.T. 0.00 min

Lab File: BG045240.D

Acq: 4 May 2020 19:13

Instrument :

BNA_G

ClientSampleId :

FK-GF-03-008-A


Tot Ion:240 Resp: 515885

Ion Ratio Lower Upper

240 100

120 7.3 6.6 10.0




236 26.4 22.2 33.2



Abundance Ion 240.00 (239.70 to 240.70): E

Ion 120.00 (119.70 to 120.70): E

Ion 236.00 (235.70 to 236.70): E



#77

Benzidine

Concen: Below Cal

RT: 19.70 min Scan# 2827

Delta R.T. 0.09 min

Lab File: BG045240.D

Acq: 4 May 2020 19:13


Tgt Ion:184 Resp: 971

Ion Ratio Lower Upper

184 100

185 98.7 12.6 19.0#



183 167.3 11.3 16.9#




Abundance Ion 184.00 (183.70 to 184.70): E

Ion 185.00 (184.70 to 185.70): E

Ion 183.00 (182.70 to 183.70): E





#78

Pvrene

Concen: 8.808 ng

RT: 19.79 min Scan# 2843

Delta R.T. -0.00 min

Lab File: BG045240.D

Acq: 4 May 2020 19:13

Instrument :

BNA_G

ClientSampleId :

FK-GF-03-008-A


Tot Ion:202 Resp: 313261

Ion Ratio Lower Upper

202 100

200 22.4 19.0 28.6




203 19.1 16.1 24.1



Abundance Ion 202.00 (201.70 to 202.70): E

Ion 200.00 (199.70 to 200.70): E

Ion 203.00 (202.70 to 203.70): E



#79

Terphenyl-d14

Concen: 57.657 ng

RT: 19.99 min Scan# 2878

Delta R.T. 0.00 min

Lab File: BG045240.D

Acq: 4 May 2020 19:13


Tgt Ion:244 Resp: 1364696

Ion Ratio Lower Upper

244 100

212 8.8 7.9 11.9



122 7.6 7.7 11.5#




Abundance Ion 244.00 (243.70 to 244.70): E

Ion 212.00 (211.70 to 212.70): E

Ion 122.00 (121.70 to 122.70): E





#81

Benzo(a)anthracene

Concen: 4.935 ng

RT: 21.62 min Scan# 3155

Delta R.T. -0.00 min

Lab File: BG045240.D

Acq: 4 May 2020 19:13

Instrument :

BNA_G

ClientSampleId :

FK-GF-03-008-A


Tot Ion:228 Resp: 165026

Ion Ratio Lower Upper

228 100

226 27.4 24.0 36.0




229 21.1 17.5 26.3



Abundance Ion 228.00 (227.70 to 228.70): E

Ion 226.00 (225.70 to 226.70): E

Ion 229.00 (228.70 to 229.70): E



#83

Chrysene

Concen: 4.803 ng

RT: 21.69 min Scan# 3166

Delta R.T. -0.00 min

Lab File: BG045240.D

Acq: 4 May 2020 19:13


Tgt Ion:228 Resp: 154314

Ion Ratio Lower Upper

228 100

226 28.8 26.2 39.4



229 21.0 17.0 25.6




Abundance Ion 228.00 (227.70 to 228.70): E

Ion 226.00 (225.70 to 226.70): E

Ion 229.00 (228.70 to 229.70): E





#84

Bis(2-ethylhexyl)phthalate

Concen: 2.745 ng

RT: 21.54 min Scan# 3141

Delta R.T. -0.00 min

Lab File: BG045240.D

Acq: 4 May 2020 19:13

Instrument :

BNA_G

ClientSampleId :

FK-GF-03-008-A


Tot Ion:149 Resp: 55658

Ion Ratio Lower Upper

149 100

167 25.3 23.4 35.0




279 6.2 4.3 6.5



Abundance Ion 149.00 (148.70 to 149.70): E

Ion 167.00 (166.70 to 167.70): E

Ion 279.00 (278.70 to 279.70): E



#86

Indeno(1,2,3-cd)pyrene

Concen: 2.534 ng

RT: 28.43 min Scan# 4314

Delta R.T. 0.01 min

Lab File: BG045240.D

Acq: 4 May 2020 19:13


Tgt Ion:276 Resp: 108089

Ion Ratio Lower Upper

276 100

138 6.5 9.6 14.4#



277 25.2 20.6 31.0




Abundance Ion 276.00 (275.70 to 276.70): E

Ion 138.00 (137.70 to 138.70): E

Ion 277.00 (276.70 to 277.70): E





#87

Pervlene-d12

Concen: 20.000 ng

RT: 24.82 min Scan# 3700

Delta R.T. 0.00 min

Lab File: BG045240.D

Acq: 4 May 2020 19:13

Instrument :


BNA_G

ClientSampleId :

FK-GF-03-008-A

Tot Ion:264 Resp: 552839




Ion	Ratio	Lower	Upper
264	100		
260	25.1	19.2	28.8
265	22.2	19.0	28.4



Abundance Ion 264.00 (263.70 to 264.70): E

Ion 260.00 (259.70 to 260.70): E

Ion 265.00 (264.70 to 265.70): E



#88

Benzo(b)fluoranthene

Concen: 6.150 ng

RT: 23.82 min Scan# 3529


Delta R.T. 0.00 min

Lab File: BG045240.D

Acq: 4 May 2020 19:13

Tgt Ion:252 Resp: 212982



Ion	Ratio	Lower	Upper
252	100		
253	25.4	18.8	28.2
125	7.1	6.2	9.4




Abundance Ion 252.00 (251.70 to 252.70): E

Ion 253.00 (252.70 to 253.70): E

Ion 125.00 (124.70 to 125.70): E





#89

Benzo(k)fluoranthene

Concen: 6.467 ng

RT: 23.82 min Scan# 3529

Delta R.T. -0.06 min

Lab File: BG045240.D

Acq: 4 May 2020 19:13

Instrument :

BNA_G

ClientSampleId :

FK-GF-03-008-A


Tot Ion:252 Resp: 212982

Ion Ratio Lower Upper

252 100

253 25.4 19.2 28.8




125 7.1 5.9 8.9



Abundance Ion 252.00 (251.70 to 252.70): E

Ion 253.00 (252.70 to 253.70): E

Ion 125.00 (124.70 to 125.70): E



#90

Benzo(a)pyrene

Concen: 4.811 ng

RT: 24.67 min Scan# 3674

Delta R.T. -0.00 min

Lab File: BG045240.D

Acq: 4 May 2020 19:13


Tgt Ion:252 Resp: 157287

Ion Ratio Lower Upper

252 100

253 22.2 18.3 27.5



125 7.7 7.0 10.6




Abundance Ion 252.00 (251.70 to 252.70): E

Ion 253.00 (252.70 to 253.70): E

Ion 125.00 (124.70 to 125.70): E





#92

Benzo(a,h,i)perylene

Concen: 3.236 ng

RT: 29.55 min Scan# 4505

Delta R.T. 0.00 min

Lab File: BG045240.D

Acq: 4 May 2020 19:13

Instrument :

BNA_G

ClientSampleId :

FK-GF-03-008-A


Tot Ion:276 Resp: 106879

Ion Ratio Lower Upper

276 100

277 26.2 19.3 28.9

138 17.0 11.4 17.2



Abundance

Ion 276.00 (275.70 to 276.70): E

Ion 277.00 (276.70 to 277.70): E

Ion 138.00 (137.70 to 138.70): E

