


Data Path : Z:\SVOASRV\HPCHEM1\BNA\_G\DATA\BG050719\  
 Data File : BG040779.D  
 Acq On : 7 May 2019 22:57  
 Operator : HP/JU  
 Sample : DFTPP  
 Misc :  
 ALS Vial : 1 Sample Multiplier: 1

Instrument :  
 BNA\_G  
 ClientSampleId :  
 DFTPP

Integration File: rteint.p

Method : Z:\SVOASRV\HPCHEM1\BNA\_G\METHODS\8270-BG042319.M  
 Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION  
 Last Update : Wed Apr 24 05:35:33 2019



AutoFind: Scans 2522, 2523, 2524; Background Corrected with Scan 2515

| Target Mass | Rel. to Mass | Lower Limit% | Upper Limit% | Rel. Abn% | Raw Abn | Result Pass/Fail |
|-------------|--------------|--------------|--------------|-----------|---------|------------------|
| 51          | 198          | 10           | 80           | 35.4      | 51026   | PASS             |
| 68          | 69           | 0.00         | 2            | 0.4       | 204     | PASS             |
| 69          | 198          | 0.00         | 100          | 36.5      | 52696   | PASS             |
| 70          | 69           | 0.00         | 2            | 0.3       | 163     | PASS             |
| 127         | 198          | 10           | 80           | 40.8      | 58816   | PASS             |
| 197         | 198          | 0.00         | 2            | 0.0       | 0       | PASS             |
| 198         | 198          | 100          | 100          | 100.0     | 144197  | PASS             |
| 199         | 198          | 5            | 9            | 6.8       | 9831    | PASS             |
| 275         | 198          | 10           | 60           | 29.2      | 42064   | PASS             |
| 365         | 198          | 1            | 100          | 3.9       | 5598    | PASS             |
| 441         | 198          | 0.01         | 100          | 9.1       | 13190   | PASS             |
| 442         | 198          | 50           | 100          | 57.4      | 82784   | PASS             |
| 443         | 442          | 15           | 24           | 19.7      | 16280   | PASS             |