


Data Path : Z:\SVOASRV\HPCHEM1\BNA G\DATA\BG050720\  
 Data File : BG045275.D  
 Acq On : 7 May 2020 18:40  
 Operator : CG/JU  
 Sample : DFTPP  
 Misc :  
 ALS Vial : 1 Sample Multiplier: 1

Instrument :  
 BNA\_G  
 ClientSampleId :  
 DFTPP

Integration File: rteint.p

Method : Z:\SVOASRV\HPCHEM1\BNA\_G\METHODS\8270-BG041520.M  
 Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION  
 Last Update : Thu May 07 12:51:10 2020



AutoFind: Scans 2507, 2508, 2509; Background Corrected with Scan 2500

| Target Mass | Rel. to Mass | Lower Limit% | Upper Limit% | Rel. Abn% | Raw Abn | Result Pass/Fail |
|-------------|--------------|--------------|--------------|-----------|---------|------------------|
| 51          | 198          | 10           | 80           | 27.7      | 22752   | PASS             |
| 68          | 69           | 0.00         | 2            | 0.0       | 0       | PASS             |
| 69          | 198          | 0.00         | 100          | 35.0      | 28770   | PASS             |
| 70          | 69           | 0.00         | 2            | 0.3       | 90      | PASS             |
| 127         | 198          | 10           | 80           | 44.7      | 36693   | PASS             |
| 197         | 198          | 0.00         | 2            | 0.4       | 312     | PASS             |
| 198         | 198          | 100          | 100          | 100.0     | 82133   | PASS             |
| 199         | 198          | 5            | 9            | 7.2       | 5903    | PASS             |
| 275         | 198          | 10           | 60           | 31.4      | 25754   | PASS             |
| 365         | 198          | 1            | 100          | 5.0       | 4136    | PASS             |
| 441         | 198          | 0.01         | 100          | 14.9      | 12265   | PASS             |
| 442         | 198          | 50           | 100          | 90.9      | 74637   | PASS             |
| 443         | 442          | 15           | 24           | 20.0      | 14952   | PASS             |