

Data Path : Z:\SVOASRV\HPCHEM1\BNA_G\DATA\BG051119\
 Data File : BG040888.D
 Acq On : 12 May 2019 00:17
 Operator : HP/JU
 Sample : K2769-06
 Misc :
 ALS Vial : 17 Sample Multiplier: 1

Instrument :
 BNA_G
 ClientSampleId :
 RB-190506

Quant Time: May 13 04:25:17 2019
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_G\METHODS\8270-BG042319.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed Apr 24 05:35:33 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	8.13	152	50756	20.00	ng	-0.02
21) Naphthalene-d8	10.96	136	203805	20.00	ng	-0.03
39) Acenaphthene-d10	14.76	164	147848	20.00	ng	-0.03
64) Phenanthrene-d10	17.51	188	421949	20.00	ng	-0.03
76) Chrysene-d12	21.79	240	437008	20.00	ng	-0.04
87) Perylene-d12	25.14	264	461296	20.00	ng	-0.06
System Monitoring Compounds						
5) 2-Fluorophenol	5.67	112	190329	66.10	ng	-0.02
7) Phenol-d6	7.27	99	172210	38.84	ng	-0.03
23) Nitrobenzene-d5	9.31	82	421761	95.62	ng	-0.03
42) 2,4,6-Tribromophenol	16.25	330	355113	174.74	ng	-0.02
45) 2-Fluorobiphenyl	13.39	172	992321	96.93	ng	-0.03
79) Terphenyl-d14	20.10	244	1905590	96.60	ng	-0.03


Target Compounds	Qvalue
------------------	--------

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\SVOASRV\HPCHEM1\BNA_G\DATA\BG051119\
Data File : BG040888.D
Acq On : 12 May 2019 00:17
Operator : HP/JU
Sample : K2769-06
Misc :
ALS Vial : 17 Sample Multiplier: 1

Instrument :
BNA_G
ClientSampleId :
RB-190506

Quant Time: May 13 04:25:17 2019
Quant Method : Z:\SVOASRV\HPCHEM1\BNA_G\METHODS\8270-BG042319.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Wed Apr 24 05:35:33 2019
Response via : Initial Calibration








#1

1,4-Dichlorobenzene-d4
Concen: 20.000 ng
RT: 8.13 min Scan# 859
Delta R.T. -0.02 min
Lab File: BG040888.D
Acq: 12 May 2019 00:17

Instrument :
BNA_G
ClientSampleId :
RB-190506




Tgt Ion:152 Resp: 50756
Ion Ratio Lower Upper
152 100
150 147.6 120.2 180.2
115 60.9 46.2 69.2




#5

2-Fluorophenol
Concen: 66.102 ng
RT: 5.67 min Scan# 439
Delta R.T. -0.02 min
Lab File: BG040888.D
Acq: 12 May 2019 00:17

Tgt Ion:112 Resp: 190329
Ion Ratio Lower Upper
112 100
64 67.5 61.4 92.0
63 40.9 37.8 56.6





#7

Phenol-d6

Concen: 38.844 ng

RT: 7.27 min Scan# 711

Delta R.T. -0.03 min

Lab File: BG040888.D

Acq: 12 May 2019 00:17

Instrument :

BNA_G

ClientSampleId :

RB-190506


Tgt Ion: 99 Resp: 172210

Ion Ratio Lower Upper




99 100

42 22.7 21.8 32.8

71 45.1 38.6 57.8



Abundance Ion 99.00 (98.70 to 99.70): BGC
Ion 42.00 (41.70 to 42.70): BGC
Ion 71.00 (70.70 to 71.70): BGC



#21

Naphthalene-d8

Concen: 20.000 ng

RT: 10.96 min Scan# 1339

Delta R.T. -0.03 min

Lab File: BG040888.D

Acq: 12 May 2019 00:17

Tgt Ion: 136 Resp: 203805


Ion Ratio Lower Upper

136 100



137 10.4 9.7 14.5

54 9.1 8.6 12.8

68 5.4 4.8 7.2



Abundance Ion 136.00 (135.70 to 136.70): BGC
Ion 137.00 (136.70 to 137.70): BGC
Ion 54.00 (53.70 to 54.70): BGC
Ion 68.00 (67.70 to 68.70): BGC









#23
Nitrobenzene-d5
Concen: 95.621 ng
RT: 9.31 min Scan# 1059
Delta R.T. -0.03 min
Lab File: BG040888.D
Acq: 12 May 2019 00:17




Instrument :
BNA_G
ClientSampleId :
RB-190506


Tgt Ion	Ratio	Lower	Upper
82	100		
128	36.2	28.9	43.3
54	54.5	47.2	70.8



#39
Acenaphthene-d10
Concen: 20.000 ng
RT: 14.76 min Scan# 1987
Delta R.T. -0.03 min
Lab File: BG040888.D
Acq: 12 May 2019 00:17

Tgt Ion	Ratio	Lower	Upper
164	100		
162	105.3	81.9	122.9
160	46.1	35.4	53.2









#42
 2,4,6-Tribromophenol
 Concen: 174.743 ng
 RT: 16.25 min Scan# 2240
 Delta R.T. -0.02 min
 Lab File: BG040888.D
 Acq: 12 May 2019 00:17




Instrument :
 BNA_G
 ClientSampleId :
 RB-190506


Tgt Ion	Ratio	Lower	Upper
330	100		
332	96.5	75.9	113.9
141	35.6	32.4	48.6



#45
 2-Fluorobiphenyl
 Concen: 96.931 ng
 RT: 13.39 min Scan# 1753
 Delta R.T. -0.03 min
 Lab File: BG040888.D
 Acq: 12 May 2019 00:17

Tgt Ion	Ratio	Lower	Upper
172	100		
171	37.7	29.4	44.2
170	24.9	19.9	29.9









#64
Phenanthrene-d10
Concen: 20.000 ng
RT: 17.51 min Scan# 2454
Delta R.T. -0.03 min
Lab File: BG040888.D
Acq: 12 May 2019 00:17




Instrument :
BNA_G
ClientSampleId :
RB-190506


Tgt Ion	Ratio	Lower	Upper
188	100		
94	5.6	5.6	8.4
80	6.5	6.4	9.6



#76
Chrysene-d12
Concen: 20.000 ng
RT: 21.79 min Scan# 3183
Delta R.T. -0.04 min
Lab File: BG040888.D
Acq: 12 May 2019 00:17

Tgt Ion	Ratio	Lower	Upper
240	100		
120	4.9	5.0	7.6#
236	27.4	20.3	30.5





#79

Terphenyl-d14

Concen: 96.604 ng

RT: 20.10 min Scan# 2896

Delta R.T. -0.03 min

Lab File: BG040888.D

Acq: 12 May 2019 00:17

Instrument :

BNA_G

ClientSampleId :

RB-190506


Tgt Ion:244 Resp: 1905590

Ion Ratio Lower Upper

244 100

212 9.8 7.2 10.8




122 7.5 6.6 10.0



Abundance Ion 244.00 (243.70 to 244.70): E

Ion 212.00 (211.70 to 212.70): E

Ion 122.00 (121.70 to 122.70): E



#87

Perylene-d12

Concen: 20.000 ng

RT: 25.14 min Scan# 3753

Delta R.T. -0.06 min

Lab File: BG040888.D

Acq: 12 May 2019 00:17


Tgt Ion:264 Resp: 461296

Ion Ratio Lower Upper

264 100

260 24.2 19.2 28.8

265 23.5 16.6 25.0



Abundance Ion 264.00 (263.70 to 264.70): E

Ion 260.00 (259.70 to 260.70): E

Ion 265.00 (264.70 to 265.70): E

