


Data Path : Z:\SVOASRV\HPCHEM1\BNA G\DATA\BG051220\  
 Data File : BG045354.D  
 Acq On : 13 May 2020 9:02  
 Operator : CG/JU  
 Sample : DFTPP  
 Misc :  
 ALS Vial : 26 Sample Multiplier: 1

Instrument :  
 BNA\_G  
 ClientSampleId :  
 DFTPP

Integration File: rteint.p

Method : Z:\SVOASRV\HPCHEM1\BNA\_G\METHODS\8270-BG041520.M  
 Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION  
 Last Update : Wed May 13 11:40:01 2020



AutoFind: Scans 2505, 2506, 2507; Background Corrected with Scan 2498

| Target Mass | Rel. to Mass | Lower Limit% | Upper Limit% | Rel. Abn% | Raw Abn | Result Pass/Fail |
|-------------|--------------|--------------|--------------|-----------|---------|------------------|
| 51          | 198          | 10           | 80           | 27.2      | 14592   | PASS             |
| 68          | 69           | 0.00         | 2            | 0.8       | 148     | PASS             |
| 69          | 198          | 0.00         | 100          | 34.3      | 18385   | PASS             |
| 70          | 69           | 0.00         | 2            | 0.4       | 75      | PASS             |
| 127         | 198          | 10           | 80           | 42.4      | 22717   | PASS             |
| 197         | 198          | 0.00         | 2            | 0.0       | 0       | PASS             |
| 198         | 198          | 100          | 100          | 100.0     | 53592   | PASS             |
| 199         | 198          | 5            | 9            | 6.4       | 3412    | PASS             |
| 275         | 198          | 10           | 60           | 29.6      | 15873   | PASS             |
| 365         | 198          | 1            | 100          | 5.4       | 2915    | PASS             |
| 441         | 198          | 0.01         | 100          | 15.7      | 8419    | PASS             |
| 442         | 198          | 50           | 100          | 91.6      | 49072   | PASS             |
| 443         | 442          | 15           | 24           | 18.3      | 8988    | PASS             |