

Data Path : Z:\SVOASRV\HPCHEM1\BNA G\DATA\BG051520\
 Data File : BG045408.D
 Acq On : 15 May 2020 22:36
 Operator : CG/JU
 Sample : L2663-08
 Misc :
 ALS Vial : 9 Sample Multiplier: 1

Instrument :
 BNA_G
 ClientSampleId :
 1366

Quant Time: May 16 07:23:01 2020
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA G\METHODS\8270-BG051520.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Fri May 15 15:45:48 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	7.97	152	66148	20.00	ng	0.00
21) Naphthalene-d8	10.78	136	268466	20.00	ng	0.00
39) Acenaphthene-d10	14.61	164	185740	20.00	ng	0.00
64) Phenanthrene-d10	17.36	188	405729	20.00	ng	0.00
76) Chrysene-d12	21.62	240	390098	20.00	ng	0.00
87) Perylene-d12	24.78	264	423318	20.00	ng	0.00
System Monitoring Compounds						
5) 2-Fluorophenol	5.54	112	226073	66.79	ng	0.00
7) Phenol-d6	7.15	99	198972	41.30	ng	0.01
23) Nitrobenzene-d5	9.13	82	472212	95.92	ng	0.00
42) 2,4,6-Tribromophenol	16.11	330	348999	146.92	ng	0.01
45) 2-Fluorobiphenyl	13.23	172	1068343	97.56	ng	0.00
79) Terphenyl-d14	19.98	244	1602064	95.78	ng	0.00
Target Compounds						
3) Pyridine	3.83	79	14167	3.031	ng	Qvalue # 89
10) Phenol	7.18	94	28676	5.776	ng	92
15) Benzyl Alcohol	8.21	79	21702	5.453	ng	86
50) Dimethylphthalate	14.06	163	138245	10.850	ng	# 98
60) Diethylphthalate	15.43	149	45126	3.403	ng	98
74) Di-n-butylphthalate	18.33	149	187858	9.027	ng	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\SVOASRV\HPCHEM1\BNA G\DATA\BG051520\
Data File : BG045408.D
Acq On : 15 May 2020 22:36
Operator : CG/JU
Sample : L2663-08
Misc :
ALS Vial : 9 Sample Multiplier: 1

Instrument :
BNA_G
ClientSampleId :
1366

Quant Time: May 16 07:23:01 2020
Quant Method : Z:\SVOASRV\HPCHEM1\BNA G\METHODS\8270-BG051520.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Fri May 15 15:45:48 2020
Response via : Initial Calibration

