

Data Path : Z:\SVOASRV\HPCHEM1\BNA G\DATA\BG051520\
 Data File : BG045410.D
 Acq On : 15 May 2020 23:58
 Operator : CG/JU
 Sample : L2663-07
 Misc :
 ALS Vial : 11 Sample Multiplier: 1

Instrument :
 BNA_G
 ClientSampleId :
 1369

Manual Integrations
 APPROVED

mohammad
 5/18/2020 1:31:13 PM

Quant Time: May 16 01:30:33 2020
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA G\METHODS\8270-BG051520.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Fri May 15 15:45:48 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	7.97	152	72320	20.00	ng	0.00
21) Naphthalene-d8	10.78	136	290925	20.00	ng	0.00
39) Acenaphthene-d10	14.62	164	182734	20.00	ng	0.01
64) Phenanthrene-d10	17.39	188	358212	20.00	ng	0.02
76) Chrysene-d12	21.64	240	360038	20.00	ng	0.02
87) Perylene-d12	24.79	264	403483	20.00	ng	0.00
System Monitoring Compounds						
5) 2-Fluorophenol	5.54	112	255648	69.08	ng	0.00
7) Phenol-d6	7.15	99	329729	62.60	ng	0.01
23) Nitrobenzene-d5	9.13	82	385730	72.30	ng	0.00
42) 2,4,6-Tribromophenol	16.12	330	190696	81.60	ng	0.02
45) 2-Fluorobiphenyl	13.23	172	829913	77.04	ng	0.00
79) Terphenyl-d14	20.01	244	1253001	81.17	ng	0.03
Target Compounds						
15) Benzyl Alcohol	8.20	79	56270	12.933	ng	95
32) Benzoic acid	10.07	122	39918m	22.827	ng	
46) 1,1'-Biphenyl	13.44	154	27862	2.481	ng	94
50) Dimethylphthalate	14.06	163	148202	11.823	ng	# 92
84) Bis(2-ethylhexyl)phthalate	21.53	149	369517	30.345	ng	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\SVOASRV\HPCHEM1\BNA G\DATA\BG051520\
Data File : BG045410.D
Acq On : 15 May 2020 23:58
Operator : CG/JU
Sample : L2663-07
Misc :
ALS Vial : 11 Sample Multiplier: 1

Instrument :
BNA_G
ClientSampleId :
1369

Manual Integrations
APPROVED

mohammad
5/18/2020 1:31:13 PM

Quant Time: May 16 01:30:33 2020
Quant Method : Z:\SVOASRV\HPCHEM1\BNA G\METHODS\8270-BG051520.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Fri May 15 15:45:48 2020
Response via : Initial Calibration

