

Data Path : Z:\svoasrv\HPCHEM1\BNA_G\Data\BG051524\
 Data File : BG061202.D
 Acq On : 15 May 2024 12:02
 Operator : MA/JU
 Sample : PB160868TB
 Misc :
 ALS Vial : 5 Sample Multiplier: 1

Instrument :
 BNA_G
 ClientSampleId :
 PB160868TB

Quant Time: May 15 12:56:04 2024
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_G\Methods\8270-BG043024.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed May 01 05:11:25 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.929	152	18143	20.000	ng	-0.03
21) Naphthalene-d8	10.726	136	63378	20.000	ng	#-0.03
39) Acenaphthene-d10	14.557	164	55660	20.000	ng	-0.03
64) Phenanthrene-d10	17.307	188	164170	20.000	ng	-0.03
76) Chrysene-d12	21.566	240	177184	20.000	ng	-0.03
86) Perylene-d12	24.692	264	204204	20.000	ng	-0.04
System Monitoring Compounds						
5) 2-Fluorophenol	5.526	112	143126	160.437	ng	-0.02
7) Phenol-d6	7.107	99	196297	148.744	ng	-0.03
23) Nitrobenzene-d5	9.087	82	131573	102.912	ng	-0.03
42) 2,4,6-Tribromophenol	16.055	330	158990	142.906	ng	-0.02
45) 2-Fluorobiphenyl	13.182	172	346051	91.935	ng	-0.03
79) Terphenyl-d14	19.927	244	1000401	93.903	ng	-0.03

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_G\Data\BG051524\
Data File : BG061202.D
Acq On : 15 May 2024 12:02
Operator : MA/JU
Sample : PB160868TB
Misc :
ALS Vial : 5 Sample Multiplier: 1

Instrument :
BNA_G
ClientSampleId :
PB160868TB

Quant Time: May 15 12:56:04 2024
Quant Method : Z:\svoasrv\HPCHEM1\BNA_G\Methods\8270-BG043024.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Wed May 01 05:11:25 2024
Response via : Initial Calibration

