

Data Path : Z:\svoasrv\HPCHEM1\BNA_G\Data\BG051722\
 Data File : BG053583.D
 Acq On : 17 May 2022 14:30
 Operator : CG/JU
 Sample : N2806-02RE
 Misc :
 ALS Vial : 6 Sample Multiplier: 1

Instrument :
 BNA_G
 ClientSampleId :
 WC-8/9-1RE

Quant Time: May 17 15:28:37 2022
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_G\Methods\8270-BG050522.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed May 11 06:00:46 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	8.078	152	28183	20.000	ng	-0.01
21) Naphthalene-d8	10.892	136	106470	20.000	ng	0.00
39) Acenaphthene-d10	14.710	164	21275	20.000	ng	0.00
64) Phenanthrene-d10	17.453	188	99976	20.000	ng	0.00
76) Chrysene-d12	21.736	240	9458	20.000	ng	# 0.00
86) Perylene-d12	25.020	264	61	20.000	ng	# 0.00
System Monitoring Compounds						
5) 2-Fluorophenol	5.652	112	203362	122.461	ng	0.00
7) Phenol-d6	7.244	99	111105	44.085	ng	0.00
23) Nitrobenzene-d5	9.241	82	245386	98.109	ng	-0.01
42) 2,4,6-Tribromophenol	16.196	330	150999	481.223	ng	0.00
45) 2-Fluorobiphenyl	13.330	172	514729	353.543	ng	0.00
79) Terphenyl-d14	20.056	244	387492	763.049	ng	0.00

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_G\Data\BG051722\
Data File : BG053583.D
Acq On : 17 May 2022 14:30
Operator : CG/JU
Sample : N2806-02RE
Misc :
ALS Vial : 6 Sample Multiplier: 1

Instrument :
BNA_G
ClientSampleId :
WC-8/9-1RE

Quant Time: May 17 15:28:37 2022
Quant Method : Z:\svoasrv\HPCHEM1\BNA_G\Methods\8270-BG050522.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Wed May 11 06:00:46 2022
Response via : Initial Calibration

